

SLOVENSKI STANDARD
SIST EN ISO 11990-2:2015**01-januar-2015****Nadomešča:****SIST EN ISO 11990-2:2010**

Laserji in laserska oprema - Ugotavljanje odpornosti sapničnih (endotrahealnih) tubusov proti laserskemu žarku - 2. del: Sapnični (endotrahealni) tubusi (ISO 11990-2:2010)

Lasers and laser-related equipment - Determination of laser resistance of tracheal tubes - Part 2: Tracheal tube cuffs (ISO 11990-2:2010)

iTeh STANDARD PREVIEW

Laser und Laseranlagen - Bestimmung der Laserresistenz von Trachealtuben - Teil 2: Trachealtubusmanschetten (ISO 11990-2:2010)

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/SIST/EN/ISO/11990-2:2015)

Lasers et équipements associés aux lasers - Détermination de la résistance au laser des tubes trachéaux - Partie 2: Ballonnet de tubes trachéaux (ISO 11990-2:2010)

Ta slovenski standard je istoveten z: EN ISO 11990-2:2014**ICS:**

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
31.260	Optoelektronika, laserska oprema	Optoelectronics. Laser equipment

SIST EN ISO 11990-2:2015**en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015)

<https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015>

EUROPEAN STANDARD

EN ISO 11990-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2014

ICS 11.040.10; 31.260

Supersedes EN ISO 11990-2:2010

English Version

Lasers and laser-related equipment - Determination of laser resistance of tracheal tubes - Part 2: Tracheal tube cuffs (ISO 11990-2:2010)

Lasers et équipements associés aux lasers - Détermination de la résistance au laser des tubes trachéaux - Partie 2: Ballonnet de tubes trachéaux (ISO 11990-2:2010)

Laser und Laseranlagen - Bestimmung der Laserresistenz von Trachealtuben - Teil 2: Trachealtubumanschetten (ISO 11990-2:2010)

This European Standard was approved by CEN on 22 October 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword	3
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC (Medical Devices)	4

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015)
<https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015>

Foreword

The text of ISO 11990-2:2010 has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 11990-2:2014 by Technical Committee CEN/TC 123 "Lasers and photonics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2015, and conflicting national standards shall be withdrawn at the latest by April 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11990-2:2010.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/catalog/standards/sist/52fe9c46-4966-4040-9ac9-c9cf972b900/sist-en-iso-11990-2-2015)

The text of ISO 11990-2:2010 has been approved by CEN as EN ISO 11990-2:2014 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC (Medical Devices)

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on Medical Devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

NOTE When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC on Medical Devices

Clause(s)/sub-clause(s) of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
This entire standard	7.1 (first indent only)	This standard is intended to provide a test method that will allow an evaluation of the risk of ignition associated with the use of a tracheal tube and lasers during ear, nose and throat surgery as part of the risk assessment as set out in these essential requirements.
This entire standard	7.3	
This entire standard	9.3	

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

INTERNATIONAL STANDARD

ISO
11990-2

First edition
2010-07-15

Lasers and laser-related equipment — Determination of laser resistance of tracheal tubes —

Part 2: Tracheal tube cuffs

iTeh STANDARD PREVIEW
*Lasers et équipements associés aux lasers — Détermination de la
résistance au laser des tubes trachéaux —
(standards.iteh.ai)
Partie 2: Ballonnet de tubes trachéaux*

SIST EN ISO 11990-2:2015

<https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015>



Reference number
ISO 11990-2:2010(E)

© ISO 2010

ISO 11990-2:2010(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015)

<https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle.....	2
5 Significance and use of the test	2
6 Apparatus	3
6.1 Gas supply system	3
6.2 Containment box	4
6.3 Smoke evacuation device.....	7
6.4 Lasers and delivery systems.....	7
6.5 Oxygen analyser	7
7 Reagents and materials	8
8 Preparation of test specimens	8
9 Preparation of apparatus.....	8
10 Test procedure.....	9
11 Interpretation of results.....	10
12 Test report.....	10
Bibliography.....	11

ISO 11990-2:2010(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11990-2 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 9, *Electro-optical systems*.

ISO 11990 consists of the following parts, under the general title *Lasers and laser-related equipment — Determination of laser resistance of tracheal tubes*:

— *Part 1: Tracheal tube shafts*

— *Part 2: Tracheal tube cuffs*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015)

<https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015>

Introduction

A fire in the airway is always a serious matter. In addition to local damage in the larynx, injury can occur to the lower airway and the parenchymal tissue in the lung. The products of combustion may be blown into the lungs.

Procedures performed in the airway, where a tracheal tube and a laser are used, bring together an oxygen-enriched atmosphere, a fuel and high power, the three ingredients necessary to create a fire. The likelihood that a laser beam will contact the tracheal tube during airway procedures is high. This led to the development of a test method, described in ISO 11990-1, to assist the clinician in determining which tracheal tube shaft was the most laser-resistant under a defined set of conditions.

Unfortunately, fires with tracheal tubes, whose shafts were laser-resistant according to ISO 11990-1 have continued to occur. Investigations have shown that the cuff, and not the shaft, of the tracheal tube is the area of lowest laser resistance and most likely to be contacted by the laser beam, even when used according to the manufacturer's instructions. Clinical experience has shown that not only perforation of the part of the shaft below the cuff has happened, but also ignition of the outer surface of the cuff. This could then ignite other parts of the tracheal tube, such as the tip, which is normally unprotected.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 11990-2:2015](https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015)

<https://standards.iteh.ai/catalog/standards/sist/52fe96d6-4966-4049-9aca-c9cef972b900/sist-en-iso-11990-2-2015>