

SLOVENSKI STANDARD

SIST EN ISO 3405:2019

01-julij-2019

Nadomešča:

SIST EN ISO 3405:2011

Naftni in sorodni proizvodi iz naravnih ali sintetičnih virov - Določevanje destilacijskih značilnosti pri atmosferskem tlaku (ISO 3405:2019)

Petroleum and related products from natural or synthetic sources - Determination of distillation characteristics at atmospheric pressure (ISO 3405:2019)

Mineralölerzeugnisse und verwandte Produkte mit natürlichem oder synthetischem Ursprung - Bestimmung des Destillationsverlaufes bei Atmosphärendruck (ISO 3405:2019)

Produits pétroliers et connexes d'origine naturelle ou synthétique - Détermination des caractéristiques de distillation à pression atmosphérique (ISO 3405:2019)

Ta slovenski standard je istoveten z: EN ISO 3405:2019

ICS:

75.080	Naftni proizvodi na splošno	Petroleum products in general
--------	-----------------------------	-------------------------------

SIST EN ISO 3405:2019

en,fr,de

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

SIST EN ISO 3405:2019

<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 3405

May 2019

ICS 75.080

Supersedes EN ISO 3405:2011

English Version

Petroleum and related products from natural or synthetic sources - Determination of distillation characteristics at atmospheric pressure (ISO 3405:2019)

Produits pétroliers et connexes d'origine naturelle ou synthétique - Détermination des caractéristiques de distillation à pression atmosphérique (ISO 3405:2019)

Mineralölerzeugnisse und verwandte Produkte mit natürlichem oder synthetischem Ursprung - Bestimmung des Destillationsverlaufes bei Atmosphärendruck (ISO 3405:2019)

This European Standard was approved by CEN on 7 March 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 3405:2019
<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>

European foreword

This document (EN ISO 3405:2019) has been prepared by Technical Committee ISO/TC 28 "Petroleum and related products, fuels and lubricants from natural or synthetic sources" in collaboration with Technical Committee CEN/TC 19 "Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin," the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2019, and conflicting national standards shall be withdrawn at the latest by November 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 3405:2011.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

STANDARD PREVIEW
(standards.iteh.ai)

Endorsement notice

The text of ISO 3405:2019 has been approved by CEN as EN ISO 3405:2019 without any modification.

SIST EN ISO 3405:2019
<https://standards.iteh.ai/catalog/standards/sist/cf5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

SIST EN ISO 3405:2019

<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>

INTERNATIONAL STANDARD

**ISO
3405**

Fifth edition
2019-03

Petroleum and related products from natural or synthetic sources — Determination of distillation characteristics at atmospheric pressure

*Produits pétroliers et connexes d'origine naturelle ou synthétique —
Détermination des caractéristiques de distillation à pression
atmosphérique*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 3405:2019

<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>



Reference number
ISO 3405:2019(E)

© ISO 2019

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 3405:2019

<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Apparatus	3
5.1 General	3
5.2 Distillation flasks	4
5.3 Condenser tube and cooling bath	4
5.4 Metal shield or enclosure for flask (manual apparatus only)	4
5.5 Heat source(s)	8
5.6 Flask-support	8
5.7 Graduated cylinders	9
5.8 Temperature measurement system	10
5.9 Centring device	10
5.10 Barometer	10
6 Samples and sampling	12
6.1 Sample grouping	12
6.2 Sample maintenance prior to testing	12
6.2.1 General	12
6.2.2 Groups 1 and 2	13
6.2.3 Groups 3 and 4	13
6.3 Removing water from sample	13
6.3.1 General	13
6.3.2 Groups 1 and 2	14
6.3.3 Groups 3 and 4	14
7 Preparation of apparatus	14
8 Apparatus verification	16
8.1 Level follower	16
8.2 Electronic temperature-measurement devices	16
8.3 Electronic pressure measuring device	17
9 Procedure — Manual apparatus	17
10 Procedure — Automated apparatus	20
11 Calculations	21
12 Expression of results	24
13 Precision (Manual Apparatus)	25
13.1 General	25
13.2 Repeatability	26
13.3 Reproducibility	26
14 Precision (automated apparatus)	27
14.1 General	27
14.2 Repeatability	27
14.3 Reproducibility	27
14.4 Bias	28
14.4.1 Bias	28
14.4.2 Relative bias	29
15 Test report	29
Annex A (normative) Thermometer specifications	30

ISO 3405:2019(E)

Annex B (normative) Determination of temperature-sensor lag times	31
Annex C (normative) Determination of specified distillation data	32
Annex D (informative) Examples of data calculations	34
Annex E (informative) Emulation of emergent-stem errors	37
Annex F (informative) Examples of a test report	38
Bibliography	40

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 3405:2019

<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 28, *Petroleum and related products, fuels and lubricants from natural or synthetic sources*.

This fifth edition cancels and replaces the fourth edition (ISO 3405:2011), which has been technically revised. The main changes compared to the previous edition are as follows:

- extension of the scope to include synthetic and biological products in general and automotive petrol-ethanol blends and to diesel with up to 30 % (V/V) FAME specifically;
- the procedure has been aligned with ASTM D86^[1] and ASTM International has granted usage of its precision data on 5 July 2017;
- update of the precision (for automated apparatus) for groups 1, 2, and 3, with the slope-based precision obtained from a 2010 Interlaboratory Study^[2];
- for T95, group 4 now has a valid range of 260 °C to 360 °C and an updated precision, as a review of a 2006 Interlaboratory Study revealed the absence of some group 4 samples having a final boiling point near 360 °C, as well final boiling points above;
- the test report example in Annex F has been updated as group 0 is not addressed since the fourth edition of this document;
- introduction of a solution for the replacement of mercury-in-glass thermometers.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 3405:2019

<https://standards.iteh.ai/catalog/standards/sist/ef5677d0-6be1-480e-aa83-50fce2027069/sist-en-iso-3405-2019>

Petroleum and related products from natural or synthetic sources — Determination of distillation characteristics at atmospheric pressure

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of users of this document to take appropriate measures to ensure the safety and health of personnel prior to the application of the standard, and to determine the applicability of any other restrictions for this purpose.

1 Scope

This document specifies a laboratory method for the determination of the distillation characteristics of light and middle distillates derived from petroleum and related products of synthetic or biological origin with initial boiling points above 0 °C and end-points below approximately 400 °C, utilizing either manual or automated equipment. Light distillates are typically automotive engine petrol, automotive engine ethanol fuel blends with up to 85 % (V/V) ethanol, and aviation petrol. Middle distillates are typically aviation turbine fuel, kerosene, diesel, diesel with up to 30 % (V/V) FAME, burner fuel, and marine fuels that have no appreciable quantities of residua.

NOTE For the purposes of this document, the term “% (V/V)” is used to represent the volume fraction of a material.

The distillation (volatility) characteristics of hydrocarbons and related products of synthetic or biological origin have an important effect on their safety and performance, especially in the case of fuels and solvents. The boiling range gives important information on composition and behaviour during storage and use, and the rate of evaporation is an important factor in the application of many solvents. Limiting values to specified distillation characteristics are applied to most distillate petroleum product and liquid fuel specifications in order to control end-use performance and to regulate the formation of vapours which may form explosive mixtures with air, or otherwise escape into the atmosphere as emissions (VOC).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 918, *Volatile organic liquids for industrial use — Determination of distillation characteristics*

ISO 3170, *Petroleum liquids — Manual sampling*

ISO 3171, *Petroleum liquids — Automatic pipeline sampling*

ISO 4788, *Laboratory glassware — Graduated measuring cylinders*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

ISO 3405:2019(E)

— IEC Electropedia: available at <https://www.electropedia.org/>

3.1 decomposition point

thermometer reading (corrected) which coincides with the first indications of thermal decomposition of the liquid in the flask

Note 1 to entry: Characteristic indications of thermal decomposition are an evolution of fumes and erratic thermometer readings which usually show a decided decrease after any attempt has been made to adjust the heat.

3.2 dry point

thermometer reading (corrected) that is observed at the instant the last drop of liquid evaporates from the lowest point in the flask, any drops or film of liquid on the side of the flask or on the thermometer being disregarded

Note 1 to entry: The end-point (final boiling point), rather than the dry point is intended for general use. The dry point can be reported in connection with special purpose naphthas, such as those used in the paint industry. Also, it is substituted for the end-point (final boiling point) whenever the sample is of such a nature that the precision of the end-point cannot consistently meet the precision requirements given in [Clauses 13](#) or [14](#).

3.3 final boiling point

end-point maximum thermometer reading (corrected) obtained during the test

Note 1 to entry: This usually occurs after evaporation of all liquid from the bottom of the flask.

3.4 initial boiling point

thermometer reading (corrected) that is observed at the instant that the first drop of condensate falls from the lower end of the condenser tube

3.5 percent evaporated

sum of the percent recovered and the percent loss

3.6 percent loss

calculated amount of uncondensed material

Note 1 to entry: Sometimes called “front-end loss”; this is the amount of uncondensed material lost in the initial stages of the distillation.

3.7 corrected loss

percent loss corrected for barometric pressure

3.8 percent recovered

volume of condensate observed in the receiving cylinder at any point in the distillation, expressed as a percentage of the charge volume, in connection with a simultaneous temperature reading

3.9 percent recovery

maximum percent recovered, as observed in accordance with [9.10](#) or [10.10](#)

3.10 percent residue

volume of residue measured in accordance with [9.11](#) or [10.11](#), and expressed as a percentage of the charge volume

3.11**percent total recovery**

combined percent recovery and residue in the flask, as determined in accordance with [11.1](#)

3.12**thermometer reading**

temperature recorded by the sensor of the saturated vapour measured in the neck of the flask below the vapour tube, under the specified conditions of this test

3.13**temperature reading**

thermometer or temperature-measurement device reading ([3.12](#)) which is corrected to 101,3 kPa standard pressure

3.14**emergent stem effect**

offset in temperature reading caused by the use of a total immersion mercury-in-glass thermometer in the partial immersion mode

Note 1 to entry: The emergent part of the mercury column is at a lower temperature than the immersed portion, resulting in a lower temperature reading than that obtained when the thermometer was completely immersed for calibration.

3.15**temperature lag**

offset in temperature reading between a mercury-in-glass thermometer and an electronic temperature-measurement device, caused by the different response time of the systems involved

4 Principle

The sample is assigned into one of four groups based on its composition and expected volatility characteristics, each group defining the apparatus arrangement, condenser temperature and operational variables. A 100 ml test portion is distilled under the specified conditions appropriate to the group into which the sample falls, and systematic observations of thermometer readings and volumes of condensate recovered are made. The volume of the residue in the flask is measured, and the loss on distillation recorded. The thermometer readings are corrected for barometric pressure, and the data are then used for calculations appropriate to the nature of the sample and the specification requirements.

5 Apparatus**5.1 General**

Typical assemblies of the manual apparatus are shown in [Figures 1](#) and [2](#). In addition to the basic components described in [Clause 5](#), automated apparatus are equipped with a system to measure and automatically record the vapour temperature and the associated recovered volume in the receiving cylinder.

Automated equipment manufactured from 1999 onwards shall be equipped with a device to automatically shut down power to the unit and to spray an inert gas or vapour in the chamber where the distillation flask is mounted in the event of fire.

NOTE Some causes of fires are breakage of the distillation flask, electrical shorts, and foaming and spilling of liquid sample through the top opening of the flask.