

SLOVENSKI STANDARD
oSIST prEN ISO 80601-2-61:2017
01-januar-2017

Medicinska električna oprema - 2-61. del: Posebne zahteve za osnovno varnost in bistvene lastnosti pulznega oksimetra (ISO/DIS 80601-2-61:2016)

Medical electrical equipment - Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment (ISO/DIS 80601-2-61:2016)

Medizinische elektrische Geräte - Teil 2-61: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Pulsoximetriegeräten (ISO/DIS 80601-2-61:2016)

Appareils électromédicaux - Partie 2-61: Exigences particulières pour la sécurité de base et les performances essentielles pour les oxymètres de pouls (ISO/DIS 80601-2-61:2016)

Ta slovenski standard je istoveten z: prEN ISO 80601-2-61

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

oSIST prEN ISO 80601-2-61:2017

en

DRAFT INTERNATIONAL STANDARD

ISO/DIS 80601-2-61

ISO/TC 121/SC 3

Secretariat: ANSI

Voting begins on:
2016-10-24Voting terminates on:
2017-01-15

Medical electrical equipment —

Part 2-61:

Particular requirements for basic safety and essential performance of pulse oximeter equipment

Appareils électromédicaux —

Partie 2-61: Exigences particulières pour la sécurité de base et les performances essentielles pour les oxymètres de pouls

ICS: 11.040.10

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

SIST EN ISO 80601-2-61:2019

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

This document is circulated as received from the committee secretariat.

This draft is submitted to a parallel vote in ISO and in IEC.

ISO/CEN PARALLEL PROCESSING

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.



Reference number
ISO/DIS 80601-2-61:2016(E)

© ISO 2016

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

SIST EN ISO 80601-2-61:2019

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

CONTENTS

1			
2	CONTENTS.....		3
3	FOREWORD.....		5
4	INTRODUCTION.....		9
5	201.1 Scope, object and related standards.....		11
6	201.2 Normative references.....		13
7	201.3 Terminology and definitions.....		15
8	201.5 General requirements for testing of ME EQUIPMENT		21
9	201.6 Classification of ME EQUIPMENT and ME SYSTEMS		22
10	201.7 ME EQUIPMENT identification, marking and documents.....		22
11	201.8 Protection against electrical HAZARDS from ME EQUIPMENT		26
12	201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS		27
13	201.10 Protection against unwanted and excessive radiation HAZARDS.....		27
14	201.11 Protection against excessive temperatures and other HAZARDS.....		27
15	201.12 Accuracy of controls and instruments and protection against hazardous outputs.....		29
16	201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT		33
17	201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)		34
18	201.15 Construction of ME EQUIPMENT.....		34
19	201.16 ME SYSTEMS		35
20	201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS		35
21	201.101 PULSE OXIMETER PROBES and PROBE CABLE EXTENDERS		35
22	201.102 Saturation pulse INFORMATION SIGNAL		36
23	201.103 FUNCTIONAL CONNECTION.....		36
24	202 Electromagnetic disturbances – Requirements and tests.....		37
25	202.4.3.1 Configurations.....		37
26	202.5.2.2.1 Requirements applicable to all ME EQUIPMENT and ME SYSTEMS		37
27	202.8.1.101 Additional general requirements.....		37
28	202.8.2 PATIENT physiological simulation		37
29	206 Usability.....		38
30	208 General requirements, tests and guidance for alarm systems		
31	in medical electrical equipment and medical electrical systems.....		38
32	208.6.1.2.101 Additional requirements for ALARM CONDITION priority.....		38
33	208.6.5.4.101 Additional requirements for DEFAULT ALARM PRESET		39
34	208.6.8.5.101 Additional requirements for ALARM SIGNAL inactivation states,		
35	indication and access.....		39
36	211 Requirements for medical electrical equipment and medical electrical systems used in		
37	the home healthcare environment.....		39

38	212	General requirements, tests and guidance for EMS systems in medical electrical equipment and medical electrical systems.....	39
40	Annex C (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	40
42	Annex D (informative)	Symbols on marking.....	43
43	Annex AA (informative)	Particular guidance and rationale	44
44	Annex BB (informative)	Skin temperature at the PULSE OXIMETER PROBE	54
45	Annex CC (informative)	Determination of ACCURACY	59
46	Annex DD (informative)	Calibration standards.....	69
47	Annex EE (informative)	Guideline for evaluating and documenting SpO ₂ ACCURACY in human subjects	71
49	Annex FF (informative)	Simulators, calibrators and FUNCTIONAL TESTERS for PULSE OXIMETER EQUIPMENT	79
51	Annex GG (informative)	Concepts of ME EQUIPMENT response time.....	89
52	Annex HH (normative)	Data interface requirements.....	93
53	Annex II (informative)	Reference to the Essential Principles.....	97
54	Annex JJ (informative)	Terminology — alphabetized index of defined terms.....	101
55	Bibliography.....		105
56	Annex ZA (informative)	Relationship between this Document and the Essential Requirements of EU Directive 93/42/EEC.....	111

59

SIST EN ISO 80601-2-61:2019

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

INTERNATIONAL ORGANIZATION for STANDARDISATION

MEDICAL ELECTRICAL EQUIPMENT – Part 2-61: Particular requirements for the basic safety and essential performance of pulse oximetry equipment

FOREWORD

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

ISO 80601-2-61 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

This second edition of ISO 80601-2-61 cancels and replaces the first edition of ISO 80601-2-61^[1] (2011). This edition of ISO 80601-2-61 constitutes a technical revision and includes an alignment with Amendment 1 of both the third edition of IEC 60601-1 and the second edition of IEC 60601-1-8, as well as the fourth edition of IEC 60601-1-2, the third edition of IEC 60601-1-6, the second edition of IEC 60601-1-11 and IEC 60601-1-12.

¹ Numbers in square brackets refer to the Bibliography.

95 The most significant changes are the following modifications:

- 96 – Updated Rationale (Annex AA) and references related to advances in our understanding of hypoxia,
97 electronic health records and ALARM SYSTEMS;

98 and the following additions:

- 99 – Clause 211, requirements for use in the HOME HEALTHCARE ENVIRONMENT

- 100 – Clause 212, requirements for use in the emergency medical services (EMS) environment

- 101 – Annex HH, Data interface requirements

102

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

SIST EN ISO 80601-2-61:2019

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

European Foreword

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard “within the meaning of Annex ZA”, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When the ISO or IEC standard is referred to in the ISO text standard, this must be understood as a normative reference to the parallel EN standard or dated ISO standard, as outlined below, including the foreword and the Annexes ZZ.

NOTE The way in which these references documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table – Correlations between normative references and dated EN and ISO/IEC standards

Normative references as listed in Clause 2	Equivalent dated standard	
	EN	ISO/IEC
ISO 7000 (database)	—	ISO 7000 (database)
ISO 14155:2011	EN ISO 14155:2011+AC:2011	ISO 14155:2011
ISO 14937:2009	—	ISO 14937:2009
ISO 15223-1:— ²	prEN ISO 15223-1:—	ISO 15223-1:—
IEC 60068-2-27:2008+A1:2013	EN 60068-2-27:2009+A1:—	IEC 60068-2-27:2008+A1:2013
IEC 60068-2-31:2008	EN 60068-2-31:2008	IEC 60068-2-31:2008
IEC 60068-2-64:2008	EN 60068-2-64:2008	IEC 60068-2-64:2008
IEC 60601-1:2005+A1:2012	EN 60601-1:2006+A1:2013 +AC:2014+A12:2014	IEC 60601-1:2005+A1:2012
IEC 60601-1-2:2014	EN 60601-1-2:—1	IEC 60601-1-2:2014
IEC 60601-1-6:2010 +A1:2013	EN 60601-1-6:2010+A1:— 11	IEC 60601-1-6:2010 +A1:2013 1
IEC 60601-1-8:2006+A1:2012	EN 60601-1-8:2007 +A1:2013 +AC:2014	IEC 60601-1-8:2006 +A1:2012
IEC 60601-1-11:2015	EN 60601-1-11:2015	IEC 60601-1-11:2015
IEC 60601-1-12:2014	EN 60601-1-12:2015	IEC 60601-1-12:2014
IEC 60417 (database)	—	IEC 60417 (database)
IEC 60529:2013	EN 60529:—	IEC 60529:2013
IEC 60825-1:2014	EN 60825-1:2014	IEC 60825-1:2014

² Under preparation. Stage at the time of publication: ISO FDIS 15223-1:2016.

Normative references as listed in Clause 2	Equivalent dated standard	
	EN	ISO/IEC
IEC 60825-2:2004+A1:2006+A2:2010	EN 60825-2:2004+A1:2007 +A2:2010	IEC 60825- 2:2004+A1:2006+A2:2010

115

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[SIST EN ISO 80601-2-61:2019](https://standards.itih.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)

<https://standards.itih.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

INTRODUCTION

The approximation of arterial haemoglobin saturation and pulse rate using pulse oximetry is common practice in many areas of medicine. This document covers BASIC SAFETY and ESSENTIAL PERFORMANCE requirements achievable within the limits of existing technology.

The committees recognized the need to revise the first edition of this document due to the publication of the first edition of IEC 60601-1-12, as well as the fourth edition of IEC 60601-1-2, the second edition of IEC 60601-1-11 and the first Amendments to both the third edition of IEC 60601-1, the third edition of IEC 60601-1-6 and the second edition of IEC 60601-1-8.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committee that led to a requirement and identifying the HAZARDS that the requirement addresses.

Annex BB is a literature survey relevant to the determination of the maximum safe temperature of the interface between a PULSE OXIMETER PROBE and a PATIENT'S tissue.

Annex CC discusses both the formulae used to evaluate the SpO_2 ACCURACY of PULSE OXIMETER EQUIPMENT measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on when *in vitro* blood calibration of PULSE OXIMETER EQUIPMENT is needed.

Annex EE presents a guideline for a CONTROLLED DESATURATION STUDY for the calibration of PULSE OXIMETER EQUIPMENT.

Annex FF is a tutorial introduction to several kinds of testers used in pulse oximetry.

Annex GG describes concepts of PULSE OXIMETER EQUIPMENT response time.

Annex HH describes data interface requirements.

Annex II contains Reference to the Essential Principles formerly found in Annex HH.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- Requirements and definitions: roman type
- *Test specifications: italic type*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type
Normative text of tables is also in a smaller type
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD³ IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS

In referring to the structure of this document, the term

³ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance

146 – “clause” means one of the six numbered divisions within the table of contents, inclusive of all
147 subdivisions (e.g. Clause 201 includes subclauses 201.7.1, 201.7.2, etc.) and

148 – “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 7.2 and 201.7.2.1 are all
149 subclauses of Clause 201.7).

150 References to clauses within this document are preceded by the term “Clause” followed by the clause
151 number. References to subclauses within this document are by number only.

152 In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination
153 of the conditions is true.

154 The verbal forms used in this document conform to usage described in Annex H of the ISO/IEC Directives,
155 Part 2. For the purposes of this document, the auxiliary verb:

156 – “shall” means that compliance with a requirement or a test is mandatory for compliance with this
157 document;

158 – “should” means that compliance with a requirement or a test is recommended but is not mandatory
159 for compliance with this document; and

160 – “may” is used to describe a permissible way to achieve compliance with a requirement or test.

161 An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates
162 that there is guidance or rationale related to that item in Annex AA.

163 The attention of Member Bodies and National Committees is drawn to the fact that equipment
164 manufacturers and testing organizations may need a transitional period following publication of a new,
165 amended or revised ISO or IEC publication in which to make products in accordance with the new
166 requirements and to equip themselves for conducting new or revised tests. It is the recommendation of
167 the committees that the content of this document not be adopted for mandatory implementation
168 nationally earlier than 3 years from the date of publication for equipment newly designed, and not earlier
169 than 5 years from the date of publication for equipment already in production.

Medical Electrical Equipment — Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment

201.1 Scope, object and related standards

IEC 60601-1:2005+AMD1:2012, Clause 1 applies, except as follows:

201.1.1 * Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of PULSE OXIMETER EQUIPMENT intended for use on humans, hereafter referred to as ME EQUIPMENT. This includes any part necessary for NORMAL USE, including the PULSE OXIMETER MONITOR, PULSE OXIMETER PROBE, and PROBE CABLE EXTENDER.

These requirements also apply to PULSE OXIMETER EQUIPMENT, including PULSE OXIMETER MONITORS, PULSE OXIMETER PROBES and PROBE CABLE EXTENDERS, which have been REPROCESSED.

The intended use of PULSE OXIMETER EQUIPMENT includes, but is not limited to, the estimation of arterial oxygen haemoglobin saturation and pulse rate of PATIENTS in professional healthcare institutions as well as PATIENTS in the HOME HEALTHCARE ENVIRONMENT and the EMERGENCY SERVICES ENVIRONMENT.

This document is not applicable to PULSE OXIMETER EQUIPMENT intended for use in laboratory research applications nor to oximeters that require a blood sample from the PATIENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 201.11 and in 7.2.13 and 8.4.1 of the general standard.

NOTE 1 See also 4.2 of the general standard.

This document can also be applied to ME EQUIPMENT and their ACCESSORIES used for compensation or alleviation of disease, injury or disability.

This document is not applicable to PULSE OXIMETER EQUIPMENT intended solely for foetal use.

This document is not applicable to remote or slave (secondary) equipment that displays SpO_2 values that are located outside of the PATIENT ENVIRONMENT.

This document is not applicable to pulse haemoglobin monitors.

NOTE 2 ME EQUIPMENT that provides selection between diagnostic and monitoring functions is expected to meet the requirements of the appropriate document when configured for that function.

This document is applicable to PULSE OXIMETER EQUIPMENT intended for use under extreme or uncontrolled environmental conditions outside the hospital environment or physician's office, such as in ambulances and air transport. Additional standards can apply PULSE OXIMETER EQUIPMENT for those environments of use.

This document is a particular standard in the IEC 60601-1 and ISO/IEC 80601 series of standards.

201.1.2 Object

Subclause 1.2 of the general standard is replaced by:

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for PULSE OXIMETER EQUIPMENT [as defined in 201.3.203] and its ACCESSORIES.

NOTE ACCESSORIES are included because the combination of the PULSE OXIMETER MONITOR and the ACCESSORIES needs to be adequately safe. ACCESSORIES can have a significant impact on the BASIC SAFETY or ESSENTIAL PERFORMANCE of PULSE OXIMETER EQUIPMENT.

201.1.3 Collateral standards

IEC 60601-1:2005+AMD1:2012, subclause 1.3 applies with the following addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of the general standard⁴ and Clause 201.2 of this document.

IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-11 and IEC 60601-1-12 apply as modified in Clauses 202, 206, 208, 211 and 212 respectively. IEC 60601-1-3² does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

IEC 60601-1:2005+AMD1:2012, subclause 1.4 is replaced by:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard, including the collateral standards, as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY or ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard or the collateral standards.

For brevity, IEC 60601-1:2005+AMD1:2012 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

⁴ The general standard is IEC 60601-1:2005+AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*