

SLOVENSKI STANDARD

SIST EN ISO 80601-2-61:2019

01-april-2019

Nadomešča:

SIST EN ISO 80601-2-61:2011

Medicinska električna oprema - 2-61. del: Posebne zahteve za osnovno varnost in bistvene lastnosti pulznega oksimetra (ISO 80601-2-61:2017, popravljena verzija 2018-02)

Medical electrical equipment - Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment (ISO 80601-2-61:2017, Corrected version 2018-02)

iTeh STANDARD PREVIEW

Medizinische elektrische Geräte - Teil 2-61: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Pulsoximetriegeräten (ISO 80601-2-61:2017)

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

Appareils électromédicaux - Partie 2-61: Exigences particulières pour la sécurité de base et les performances essentielles pour les oxymètres de pouls (ISO 80601-2-61:2017, Version corrigée 2018-02)

Ta slovenski standard je istoveten z: EN ISO 80601-2-61:2019

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

SIST EN ISO 80601-2-61:2019

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 80601-2-61:2019](https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

EUROPEAN STANDARD

EN ISO 80601-2-61

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2019

ICS 11.040.55

Supersedes EN ISO 80601-2-61:2011

English Version

Medical electrical equipment - Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment (ISO 80601-2-61:2017, Corrected version 2018-02)

Appareils électromédicaux - Partie 2-61: Exigences particulières pour la sécurité de base et les performances essentielles pour les oxymètres de pouls (ISO 80601-2-61:2017, Version corrigée 2018-02)

Medizinische elektrische Geräte - Teil 2-61: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Pulsoximetriegeräten (ISO 80601-2-61:2017)

This European Standard was approved by CEN on 13 December 2018.

This European Standard was corrected and reissued by the CEN-CENELEC Management Centre on 06 March 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

[SIST EN ISO 80601-2-61:2019](#)

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 80601-2-61:2019](https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)
<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

European foreword

This document (EN ISO 80601-2-61:2019) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2019, and conflicting national standards shall be withdrawn at the latest by January 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 80601-2-61:2011.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of ISO 80601-2-61:2017, Corrected version 2018-02 has been approved by CEN as EN ISO 80601-2-61:2019 without any modification.

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 80601-2-61:2019](https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

INTERNATIONAL
STANDARD

ISO
80601-2-61

Second edition
2017-12

Corrected version
2018-02

Medical electrical equipment —
Part 2-61:
Particular requirements for basic
safety and essential performance of
pulse oximeter equipment

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Appareils électromédicaux —

Partie 2-61: Exigences particulières pour la sécurité de base et les performances essentielles pour les oxymètres de pouls

SIST EN ISO 80601-2-61:2019

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>



Reference number
ISO 80601-2-61:2017(E)

© ISO 2017

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 80601-2-61:2019](https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
201. 1 Scope, object and related standards	1
201. 2 Normative references	3
201. 3 Terms and definitions	4
201. 4 General requirements	9
201. 5 General requirements for testing of ME EQUIPMENT	10
201. 6 Classification of ME EQUIPMENT and ME SYSTEMS	10
201. 7 ME EQUIPMENT identification, marking and documents.....	11
201. 8 Protection against electrical HAZARDS from ME EQUIPMENT.....	15
201. 9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	15
201. 10 Protection against unwanted and excessive radiation HAZARDS.....	15
201. 11 Protection against excessive temperatures and other HAZARDS.....	15
201. 12 Accuracy of controls and instruments and protection against hazardous outputs.....	17
201. 13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	21
201. 14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	21
201. 15 Construction of ME EQUIPMENT.....	21
201. 16 ME SYSTEMS.....	23
201. 17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	23
201. 101 * PULSE OXIMETER PROBES and PROBE CABLE EXTENDERS	23
201. 102 Saturation pulse INFORMATION SIGNAL	23
201. 103 FUNCTIONAL CONNECTION	24
202 Electromagnetic disturbances – Requirements and tests.....	24
206 Usability	25
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	26
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.....	26
212 Requirements for medical electrical equipment and medical electrical systems used in the emergency medical services environment.....	27
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	28
Annex D (informative) Symbols on marking	31
Annex AA (informative) Particular guidance and rationale.....	32
Annex BB (informative) Skin temperature at the PULSE OXIMETER PROBE.....	40
ANNEX CC (informative) Determination of ACCURACY.....	44
Annex DD (informative) Calibration standards.....	53
Annex EE (informative) Guideline for evaluating and documenting SpO ₂ ACCURACY in human subjects	54
Annex FF (informative) Simulators, calibrators and FUNCTIONAL TESTERS for PULSE OXIMETER EQUIPMENT	61

ISO 80601-2-61:2017(E)

Annex GG (informative) Concepts of ME EQUIPMENT response time	70
Annex HH (normative) Data interface requirements.....	74
Annex II (informative) Reference to the ESSENTIAL PRINCIPLES	78
Annex JJ (informative) Terminology — alphabetized index of defined terms.....	82
Bibliography	87

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 80601-2-61:2019](https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared jointly by ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

This second edition of ISO 80601-2-61 cancels and replaces the first edition (ISO 80601-2-61:2011), which has been technically revised. It includes an alignment with Amendment 1 of both the third edition of IEC 60601-1 and the second edition of IEC 60601-1-8, as well as the fourth edition of IEC 60601-1-2, the third edition of IEC 60601-1-6, the second edition of IEC 60601-1-11 and IEC 60601-1-12.

The most significant changes are the following modifications:

- updated rationale (Annex AA) and references related to advances in the understanding of hypoxaemia, electronic health records and ALARM SYSTEMS;
- ingress protection changed from IPX1 to IPX2;

and the following additions:

- Clause 211, requirements for use in the HOME HEALTHCARE ENVIRONMENT;
- Clause 212, requirements for use in the emergency medical services (EMS) environment;
- Annex HH, Data interface requirements.

This corrected version of ISO 80601-2-61:2017 incorporates the following correction:

- headers have been corrected.

A list of all the parts of the ISO/IEC 80601 series is available on the ISO website.

ISO 80601-2-61:2017(E)

Introduction

The approximation of arterial haemoglobin saturation and pulse rate using pulse oximetry is common practice in many areas of medicine. This document covers BASIC SAFETY and ESSENTIAL PERFORMANCE requirements achievable within the limits of existing technology.

The committees recognized the need to revise the first edition of this document because of the publication of the first edition of IEC 60601-1-12, as well as the fourth edition of IEC 60601-1-2, the second edition of IEC 60601-1-11 and the first Amendments to both the third edition of IEC 60601-1, the third edition of IEC 60601-1-6 and the second edition of IEC 60601-1-8.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committees that led to a requirement and identifying the HAZARDS that the requirement addresses.

Annex BB is a literature survey relevant to the determination of the maximum safe temperature of the interface between a PULSE OXIMETER PROBE and a PATIENT'S tissue.

Annex CC discusses both the formulae used to evaluate the SpO_2 ACCURACY of PULSE OXIMETER EQUIPMENT measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on when in vitro blood calibration of PULSE OXIMETER EQUIPMENT is needed.

Annex EE presents a guideline for a CONTROLLED DESATURATION STUDY for the calibration of PULSE OXIMETER EQUIPMENT.

Annex FF is a tutorial introduction to several kinds of testers used in pulse oximetry.

Annex GG describes concepts of PULSE OXIMETER EQUIPMENT response time.

Annex HH describes data interface requirements.

Annex II contains Reference to the ESSENTIAL PRINCIPLES formerly found in Annex HH.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type; normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD¹ IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the six numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7.1, 201.7.2) and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document; and
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 80601-2-61:2019](https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019)

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 80601-2-61:2019

<https://standards.iteh.ai/catalog/standards/sist/e5fbd20b-b925-4c2e-982b-fadd9e833ce8/sist-en-iso-80601-2-61-2019>

Medical electrical equipment —

Part 2-61:

Particular requirements for basic safety and essential performance of pulse oximeter equipment

201.1 Scope, object and related standards

IEC 60601-1:2005+AMD1:2012, Clause 1 applies, except as follows:

201.1.1 * Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of PULSE OXIMETER EQUIPMENT intended for use on humans, hereafter referred to as ME EQUIPMENT. This includes any part necessary for NORMAL USE, including the PULSE OXIMETER MONITOR, PULSE OXIMETER PROBE, and PROBE CABLE EXTENDER.

These requirements also apply to PULSE OXIMETER EQUIPMENT, including PULSE OXIMETER MONITORS, PULSE OXIMETER PROBES and PROBE CABLE EXTENDERS, which have been REPROCESSED.

The intended use of PULSE OXIMETER EQUIPMENT includes, but is not limited to, the estimation of arterial oxygen haemoglobin saturation and pulse rate of PATIENTS in professional healthcare institutions as well as PATIENTS in the HOME HEALTHCARE ENVIRONMENT and the EMERGENCY MEDICAL SERVICES ENVIRONMENT.

This document is not applicable to PULSE OXIMETER EQUIPMENT intended for use in laboratory research applications nor to oximeters that require a blood sample from the PATIENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 201.11 and in 7.2.13 and 8.4.1 of the general standard.

NOTE 1 See also 4.2 of the general standard. "The general standard" is IEC 60601-1:2005+AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

This document can also be applied to ME EQUIPMENT and their ACCESSORIES used for compensation or alleviation of disease, injury or disability.

This document is not applicable to PULSE OXIMETER EQUIPMENT intended solely for foetal use.

This document is not applicable to remote or slave (secondary) equipment that displays SpO_2 values that are located outside of the PATIENT ENVIRONMENT.

NOTE 2 ME EQUIPMENT that provides selection between diagnostic and monitoring functions is expected to meet the requirements of the appropriate document when configured for that function.