

SLOVENSKI STANDARD

SIST EN 1865-2:2010+A1:2015

01-maj-2015

Nadomešča:
SIST EN 1865-2:2010

Oprema za ravnanje s pacienti v reševalnih vozilih - 2. del: Nosila s pomožnim pogonom

Patient handling equipment used in road ambulances - Part 2: Power assisted stretcher

Krankentransportmittel im Krankenkraftwagen - Teil 2: Kraftunterstützte Krankentrage

Spécifications d'équipements pour le transport de patient dans les ambulances routières
- Partie 2: Brancard motorisé

[SIST EN 1865-2:2010+A1:2015](https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-109758/sist-en-1865-2:2010+A1:2015)

<https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-109758/sist-en-1865-2:2010+A1:2015>

Ta slovenski standard je istoveten z: EN 1865-2:2010+A1:2015

ICS:

11.160	Prva pomoč	First aid
43.160	Vozila za posebne namene	Special purpose vehicles

SIST EN 1865-2:2010+A1:2015 **en,de**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 1865-2:2010+A1:2015

<https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-4bba89839908/sist-en-1865-2-2010a1-2015>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 1865-2:2010+A1

March 2015

ICS 11.160

Supersedes EN 1865-2:2010

English Version

**Patient handling equipment used in road ambulances - Part 2:
Power assisted stretcher**

Spécifications d'équipements pour le transport de patient
dans les ambulances routières - Partie 2: Brancard
motorisé

Krankentransportmittel im Krankenkraftwagen - Teil 2:
Kraftunterstützte Krankentrage

This European Standard was approved by CEN on 2 July 2010 and includes Amendment 1 approved by CEN on 20 December 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST EN 1865-2:2010+A1:2015](https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-4bba89839908/sist-en-1865-2-2010a1-2015)

<https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-4bba89839908/sist-en-1865-2-2010a1-2015>





EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definition	5
4 Requirements	6
4.1 General.....	6
4.2 Main stretcher	6
4.2.1 General.....	6
4.2.2 Dimensions.....	6
4.2.3 Mass	6
4.2.4 Loading capacity.....	7
4.2.5 Frame	7
4.2.6 Power source	8
4.2.7 Lying part of the stretcher	8
4.2.8 Restraint system	8
4.2.9 Flammability – Toxicity burning gases	8
4.2.10 Deformation of the frame	9
4.2.11 Fixation	9
4.2.12 Deformation of the lying area	9
4.2.13 Resistance to torsion	9
4.2.14 Splaying of the wheels	9
5 Marking	9
Annex ZA  (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices and of EU Directive 2006/42/EC on Essential Health and Safety Requirements 	10
Bibliography	12

Foreword

This document (EN 1865-2:2010+A1:2015) has been prepared by Technical Committee CEN/TC 239 "Rescue systems", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2015, and conflicting national standards shall be withdrawn at the latest by September 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document includes Amendment 1 approved by CEN on 20 December 2014.

■_{A1} This document supersedes EN 1865-2:2010. ■_{A1}

The start and finish of text introduced or altered by amendment is indicated in the text by tags ■_{A1} ■_{A1}.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

■_{A1} This European Standard is a part of EN 1865, *Patient handling equipment used in road ambulances*, which consists of the following parts:

- *Part 1: General stretcher systems and patient handling equipment;*
- *Part 2: Power assisted stretcher [the present document];*
- *Part 3: Heavy duty stretcher;*
- *Part 4: Foldable patient transfer chair;*
- *Part 5: Stretcher support.* ■_{A1}

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

In this standard reference is made to EN 1789 which specifies design requirements and test methods for road ambulances, which are relevant for checking requirements for such handling equipment.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 1865-2:2010+A1:2015](https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-4bba89839908/sist-en-1865-2-2010a1-2015)
<https://standards.iteh.ai/catalog/standards/sist/d62ee382-1697-44e8-b754-4bba89839908/sist-en-1865-2-2010a1-2015>

1 Scope

This European Standard defines minimum requirements for the design and performance of power assisted stretchers used in road ambulances for the treatment and transportation of patients. It aims to ensure patient safety and minimize the physical effort required by staff operating the equipment.

2 Normative references

^[A1] The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies. ^[A1]

^[A1] EN 597-1:1994 ^[A1], *Furniture - Assessment of the ignitability of mattresses and upholstered bed bases - Part 1: Ignition source: Smouldering cigarette*

^[A1] *deleted text* ^[A1]

^[A1] EN 1041:2008+A1:2013 ^[A1], *Information supplied by the manufacturer of medical devices*

^[A1] EN 1789:2007+A2:2014 ^[A1], *Medical vehicles and their equipment — Road ambulances*

^[A1] EN 1865-1:2010 ^[A1], *Patient handling equipment used in road ambulances — Part 1: General stretcher systems and patient handling equipment*

^[A1] EN 60601-1-2:2014, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic disturbances — Requirements and tests (IEC 60601-1-2:2014)* ^[A1]

^[A1] EN 62366:2008 ^[A1], *Medical devices — Application of usability engineering to medical devices (IEC 62366:2007)*

^[A1] EN ISO 14971:2012 ^[A1], *Medical devices - Application of risk management to medical devices (ISO 14971:2007, Corrected version 2007-10-01)*

^[A1] EN ISO 15223-1:2012, *Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements (ISO 15223-1:2012)* ^[A1]

3 Terms and definition

For the purposes of this document, the following terms and definitions apply.

3.1

power assisted stretcher

device designed for the conveyance of sick and/or injured persons or those in labour in transit in a recumbent position in safety and in comfort whilst facilitating treatment and providing powered movements in the operation of the stretcher to reduce the physical effort required by operatives

EN 1865-2:2010+A1:2015 (E)

4 Requirements

4.1 General

When operated and maintained in accordance with the manufacturer's instructions, power assisted stretchers shall not present any high level of risk. Any identified risk shall be reduced to an acceptable level by using risk management principles in accordance with A1 EN ISO 14971:2012 A1 .

Power assisted stretchers shall also:

- allow patients to be secured so that any movement during ambulance transport is minimized;
- be free of sharp edges that could cause injury to persons or damage to other equipment on board. The minimum radius should be 0,5 mm.

For all patients transported in the patient compartment, patient restraint-systems shall be available. All patient restraint-systems shall have quick release systems.

Patient restraint-systems for all patient handling equipment shall secure the patient, but at the same time shall permit treatment of the patient.

The lying-sitting part shall be made of a strong material which is bacterial resistant, fungal resistant, stain resistant, putrid resistant, easy to clean, washable, waterproof and petrol-oil resistant.

4.2 Main stretcher

4.2.1 General

The main stretcher shall consist of a stretcher part that can be used either alone or in combination with an integrated or detachable undercarriage.

The power assisted stretcher and ambulance loading system shall be compatible.

4.2.2 Dimensions

Dimensions shall be measured from the outermost edges:

- length: A1 (1 950 $^{+120}_{-50}$) mm A1 ;
- width: A1 (550 \pm $^{+60}_{-20}$) mm A1 ;
- height: maximum 300 mm from loading holding assembly to unloaded lying part. This height dimension does not apply to stretchers with monoblock undercarriages. If a monoblock is not available, the stretcher shall be constructed such that it is detachable from the undercarriage. Where a stretcher support is used the measurement shall be taken from the top surface of the stretcher support to the lying part of the stretcher.
- Undercarriage: length and width of the frame of the undercarriage when located in the ambulance shall not exceed length and width of the stretcher.

4.2.3 Mass

The total mass excluding mattress and patient restraints shall be not more than A1 75 kg A1 .

NOTE In all cases the mass should be as low as possible.

4.2.4 Loading capacity

The loading capacity shall be a minimum of 150 kg.

4.2.5 Frame

4.2.5.1 General

The frame shall be in sturdy lightweight non twisting construction enabling use of cardiopulmonary resuscitation. All corners of the frame shall be radiused for greater safety.

It shall be possible to lock and secure the stretcher and undercarriage against lateral, longitudinal, vertical movements.

All mechanisms shall be constructed to prevent damage to the user and the patient.

4.2.5.2 Stretcher parts

- a) If side rails are fitted, they shall have a minimum length of 500 mm and a minimum height between 150 mm and 200 mm measured from the top of the stretcher frame to the top of the side rail.
- b) If longitudinal handles are incorporated they shall be fitted to the ends of the longitudinal frame such that they lock and do not twist when they are stowed or in use. They shall be designed to minimize the risk of injuries to the hands and wrists when being operated or the stretcher is carried at angles. It shall allow the fixation and use of a carrying harness.
- c) The stretcher shall have a water and scratch resistant paint finish or be manufactured of corrosion resistant material. Both shall be unaffected by disinfectants.
- d) If intended to be used without undercarriage there shall be four wheels with a minimum diameter of 100 mm suitably placed to ensure stability.
- e) If intended to be used with undercarriage the stretcher shall be able to be fixed to the undercarriage without using supplementary means. A safe handling and lowering of the undercarriage shall be ensured.
- f) The fixed stretcher shall be easy to release from the stretcher fastener.

NOTE There should be a facility to attach a support for infusion.

4.2.5.3 Undercarriage

- a) The undercarriage shall be fitted with four wheels with a diameter of at least 100 mm. There shall be a minimum of two 360° swivel wheels at the foot end and at least two wheels shall be fitted with a brake.
- b) The undercarriage shall have variable height between the lowest and upper position and be able to hold the rated load capacity in any position.
- c) All the functions of the stretcher shall remain completely unimpaired when it is connected to the undercarriage.
- d) The undercarriage shall either be provided with a waterproof and scratch resistant finish or be made of corrosive resistant material or similar surface.

If the undercarriage is used with a separable stretcher it shall be possible both to connect the undercarriage to the stretcher and to separate them easily. The stretcher shall be secured to the undercarriage in such a manner that unintentional separation of the undercarriage and stretcher cannot occur. It shall be possible to