

# SLOVENSKI STANDARD

## SIST EN ISO 5360:2016

01-maj-2016

Nadomešča:  
SIST EN ISO 5360:2012

---

**Anestezijski hlapilniki (vaporizatorji) - Sistemi za nalivanje posebnih hlapnih anestetikov (ISO 5360:2016)**

Anesthetic vaporizers - Agent-specific filling systems (ISO 5360:2016)

Anästhesimittelverdampfer - Substanzspezifische Füllsysteme (ISO 5360:2016)

**iTeh STANDARD PREVIEW**

Évaporateurs d'anesthésie - Systèmes de remplissage spécifiques à l'agent (ISO 5360:2016)

[SIST EN ISO 5360:2016](#)

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-62b16f8is>

**Ta slovenski standard je istoveten: EN ISO 5360:2016**

---

**ICS:**

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

**SIST EN ISO 5360:2016**

**en**

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 5360:2016

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>

**EUROPEAN STANDARD**  
**NORME EUROPÉENNE**  
**EUROPÄISCHE NORM**

**EN ISO 5360**

March 2016

ICS 11.040.10

Supersedes EN ISO 5360:2012

English Version

**Anaesthetic vaporizers - Agent-specific filling systems (ISO  
5360:2016)**

Évaporateurs d'anesthésie - Systèmes de remplissage  
spécifiques à l'agent (ISO 5360:2016)

Anästhesimittelverdampfer - Substanzspezifische  
Füllsysteme (ISO 5360:2016)

This European Standard was approved by CEN on 7 November 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
 COMITÉ EUROPÉEN DE NORMALISATION  
 EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

European foreword.....	3
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices .....	4

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

SIST EN ISO 5360:2016

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>

## European foreword

This document (EN ISO 5360:2016) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2016, and conflicting national standards shall be withdrawn at the latest by March 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 5360:2012.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

## iTeh STANDARD PREVIEW (standards.itechhai)

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom. <http://standards.itechhai.org/standards/sist-en-iso-5360-2016?d2eed60cb16f/sist-en-iso-5360-2016>

## Endorsement notice

The text of ISO 5360:2016 has been approved by CEN as EN ISO 5360:2016 without any modification.

## Annex ZA (informative)

### **Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices**

This European standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

**Table ZA.1 — Correspondence between this European Standard and EU Directive 93/42/EEC**

Clause(s)/sub-clause(s) of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/notes
4, 5, 6, 7, 9, 10	7.5, first paragraph, first sentence	
14.3 f), 14.2.1 last paragraph	7.5, second paragraph  iTeh STANDARD PREVIEW  (standards.iteh.ai)	Only the presence of phthalates is addressed; Presumption of conformity to labelling requirement only provided if the symbol defined in EN 15896 is used
4, 5, 6, 7, 9 and 11	9.1  <a href="https://standards.iteh.ai/catalog/standards/sist/1cd2eed60cb16f/sist-en-iso-5360-2016">SIST EN ISO 5360:2016</a>	Clauses 4 to 7 of this standard specify the design of the filling system to ensure specificity for anaesthetic agent and avoid the anaesthetic agent escaping into environment.  Standard specifies colour coding of the anaesthetic agents including their generic names for a safe connection to anaesthetic systems  Information on restrictions on use is addressed in the clauses on labelling and instructions for use, see 14.1 c), 14.2.1 c), d), e), f), 14.2.2, 14.3 a) - d) and f).
14	13.1	
11	13.2	Standard specifies colour coding of the anaesthetic agents including their generic names.
14.1 a), 14.2.1 a)	13.3 a)	
11, 14.1 c), 14.2.1 b), 14.2.1 c)	13.3 b)	packaging is not addressed
14.1 b)	13.3 d)	Presumption of conformity to ER 13.3 d) only provided if the word "LOT" is used
14.2.1.d)	13.3 e)	
14.1 a), 14.2.1 a)	13.3 a)	

14.2.1 e)	13.3 f)	Consistency across the Community is not addressed
14.2.1 f)	13.3 i)	
14.1 c), 14.2.1 c), 14.2.2,	13.3 j)	
14.2.2), 14.3 b)	13.3 k)	
14.1 b)	13.5	
14.3 a), 14.3 b)	13.6 a)	
14.3 c), 14.3 d)	13.6 d)	
14.3 g	13.6 h)	
14.3 h)	13.6 q)	

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 5360:2016](#)

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 5360:2016

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>

# INTERNATIONAL STANDARD

ISO  
5360

Fourth edition  
2016-02-15

---

---

## Anaesthetic vaporizers — Agent-specific filling systems

Évaporateurs d'anesthésie — Systèmes de remplissage spécifiques à l'agent

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 5360:2016](#)

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>



Reference number  
ISO 5360:2016(E)

© ISO 2016

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 5360:2016

<https://standards.iteh.ai/catalog/standards/sist/1c155080-7d2d-4576-815e-d2eed60cb16f/sist-en-iso-5360-2016>



### COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

## Contents

	Page
<b>Foreword</b>	<b>iv</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Bottle</b>	<b>2</b>
<b>5 Bottle collar</b>	<b>3</b>
<b>6 Bottle adaptor</b>	<b>6</b>
<b>7 Filler receptacle</b>	<b>10</b>
<b>8 Filling rate</b>	<b>12</b>
<b>9 Leakage</b>	<b>14</b>
<b>10 Overfilling protection</b>	<b>14</b>
<b>11 Colour coding</b>	<b>14</b>
<b>12 Usability</b>	<b>14</b>
<b>13 Clinical evaluation</b>	<b>14</b>
<b>14 Information provided by the manufacturer</b>	<b>15</b>
14.1 Marking	15
14.2 Labelling	15
14.3 Instructions for use	15
<b>Annex A (informative) Recommendations on materials</b>	<b>17</b>
<b>Annex B (informative) Types of agent-specific filling systems</b>	<b>18</b>
<b>Annex C (normative) Determination of total leakage into atmosphere of anaesthetic agent during filling</b>	<b>19</b>
<b>Bibliography</b>	<b>21</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](http://Foreword-Supplementary-information.standards.iteh.ai).

The committee responsible for this document is ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

SIST EN ISO 5360:2016

This fourth edition cancels and replaces the third edition (ISO 5360:2012), of which it constitutes a minor revision with the following changes:

- [Figure 5](#) has been technically revised;
- minor editorial modifications have been incorporated into the text.