



# SLOVENSKI STANDARD SIST EN ISO 22442-1:2016

01-februar-2016

Nadomešča:

SIST EN ISO 22442-1:2008

---

**Medicinski pripomočki, ki uporabljajo živalska tkiva in njihove derivate - 1. del:  
Uporaba obvladovanja tveganja (ISO 22442-1:2015)**

Medical devices utilizing animal tissues and their derivatives - Part 1: Application of risk management (ISO 22442-1:2015)

Tierische Gewebe und deren Derivate, die zur Herstellung von Medizinprodukten eingesetzt werden - Teil 1: Anwendung des Risikomanagements (ISO 22442-1:2015)

Dispositifs médicaux utilisant des tissus animaux et leurs dérivés - Partie 1: Application de la gestion des risques (ISO 22442-1:2015)

**Ta slovenski standard je istoveten z: EN ISO 22442-1:2015**

---

**ICS:**

11.120.01 Farmacija na splošno      Pharmaceutics in general

**SIST EN ISO 22442-1:2016**      en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN ISO 22442-1:2016

<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 22442-1**

November 2015

ICS 11.100.99

Supersedes EN ISO 22442-1:2007

English Version

**Medical devices utilizing animal tissues and their derivatives - Part 1: Application of risk management (ISO 22442-1:2015)**

Dispositifs médicaux utilisant des tissus animaux et leurs dérivés - Partie 1: Application de la gestion des risques (ISO 22442-1:2015)

Tierische Gewebe und deren Derivate, die zur Herstellung von Medizinprodukten eingesetzt werden - Teil 1: Anwendung des Risikomanagements (ISO 22442-1:2015)

This European Standard was approved by CEN on 1 November 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>
<b>Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC.....</b>	<b>5</b>

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

SIST EN ISO 22442-1:2016  
<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>

## European foreword

This document (EN ISO 22442-1:2015) has been prepared by Technical Committee ISO/TC 194 "Biological evaluation of medical devices" in collaboration with Technical Committee CEN/TC 316 "Medical devices utilizing tissues" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2016 and conflicting national standards shall be withdrawn at the latest by May 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 22442-1:2007.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard 'within the meaning of Annex ZA', the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

## EN ISO 22442-1:2015 (E)

Table 1 – Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO
ISO 10993-1	EN ISO 10993-1:2009	ISO 10993-1:2009
ISO 14971	ISO 14971:2012	ISO 14971:2007
ISO 22442-2	EN ISO 22442-2:2016	ISO 22442-2:2016
ISO 22442-3	EN ISO 22442-3:2007	ISO 22442-3:2007

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

#### Endorsement notice

The text of ISO 22442-1:2015 has been approved by CEN as EN ISO 22442-1:2015 without any modification.

(standards.iteh.ai)

[SIST EN ISO 22442-1:2016](https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016)

<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>

**Annex ZA**  
(informative)  
**Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC**

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to the Essential Requirements of Directive 93/42/EEC, concerning medical devices, as amended by Commission Regulation (EU) No722/2012 in relation to detailed specifications regarding requirements for medical devices utilizing tissues of animal origin.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA Regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC, as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

**iTeh STANDARD PREVIEW**

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with essential requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

## EN ISO 22442-1:2015 (E)

**Table ZA.1 — Correspondence between this International Standard and Directive 93/42/EEC as amended by Commission Regulation (EU) No 722/2012**

Clause(s)/subclause(s) of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC as amended by Commission Regulation (EU) No 722/2012	Qualifying remarks/notes
4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	7.1	
4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	7.2	
4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	8.1	
4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	8.2	
Annexes C and D	Annex I of Commission Regulation (EU) No 722/2012	Annexes C and D are dedicated to TSE risk, but 4.1, 4.2, 4.3, 4.4 are also relevant.

**WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.**

SIST EN ISO 22442-1:2016  
<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>



INTERNATIONAL  
STANDARD

ISO  
22442-1

Second edition  
2015-11-01

---

---

**Medical devices utilizing animal  
tissues and their derivatives —**

**Part 1:  
Application of risk management**

*Dispositifs médicaux utilisant des tissus animaux et leurs dérivés —*

*Partie 1: Application de la gestion des risques*

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 22442-1:2016](https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016)

<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>



Reference number  
ISO 22442-1:2015(E)

© ISO 2015

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 22442-1:2016](https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016)

<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
[copyright@iso.org](mailto:copyright@iso.org)  
[www.iso.org](http://www.iso.org)

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Risk management process</b> .....	<b>3</b>
4.1 General.....	3
4.2 Risk analysis.....	3
4.2.1 Identification of qualitative and quantitative characteristics related to the safety of medical devices.....	3
4.2.2 Identification of hazards and hazardous situations.....	4
4.3 Risk evaluation.....	5
4.4 Risk control.....	5
4.4.1 General.....	5
4.4.2 Risk control for viruses and TSE agents.....	5
4.4.3 Risk control of other hazards.....	5
4.4.4 Residual risk evaluation.....	6
4.5 Evaluation of overall residual risk acceptability.....	6
4.5.1 General.....	6
4.5.2 Documentation.....	6
4.6 Production and post-production information system.....	7
<b>Annex A (informative) Guidance on the application of this part of ISO 22442</b> .....	<b>8</b>
<b>Annex B (informative) Graphical representation of part of the risk management process for medical devices utilizing animal material</b> .....	<b>9</b>
<b>Annex C (normative) Special requirements for some animal materials considering the risk management for TSE agents</b> .....	<b>11</b>
<b>Annex D (informative) Information relevant to the management of TSE risk</b> .....	<b>16</b>
<b>Bibliography</b> .....	<b>25</b>

## ISO 22442-1:2015(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 194, *Biological and clinical evaluation of medical devices*, SC 1, *Tissue product safety*.

This second edition cancels and replaces the first edition (ISO 22442-1:2007), of which it constitutes a minor revision.

ISO 22442 consists of the following parts, under the general title *Medical devices utilizing animal tissues and their derivatives*:

- *Part 1: Application of risk management*
- *Part 2: Controls on sourcing, collection and handling*
- *Part 3: Validation of the elimination and/or inactivation of viruses and transmissible spongiform encephalopathy (TSE) agents*
- *Part 4: Principles for elimination and/or inactivation of transmissible spongiform encephalopathy (TSE) agents and validation assays for those processes* [Technical Report]

## Introduction

Certain medical devices utilize materials of animal origin.

Animal tissues and their derivatives are used in the design and manufacture of medical devices to provide performance characteristics that have been chosen for advantages over non-animal based materials. The range and quantities of materials of animal origin in medical devices vary. These materials can comprise a major part of the device (e.g. bovine/porcine heart valves, bone substitutes for use in dental or orthopaedic applications, haemostatic devices), can be a product coating or impregnation (e.g. collagen, gelatine, heparin), or can be used in the device manufacturing process (e.g. tallow derivatives such as oleates and stearates, foetal calf serum, enzymes, culture media).

ISO 14971 is a general standard which specifies a process for a manufacturer by identifying hazards and hazardous situations associated with medical devices, including *in vitro* medical devices, to estimate and evaluate the risks associated with those hazards, to control these risks and to monitor the effectiveness of the control throughout the life cycle. This part of ISO 22442 provides additional requirements and guidance for the evaluation of medical devices manufactured utilizing animal tissues or derivatives which are non-viable or rendered non-viable.

This part of ISO 22442 is intended to cover medical devices including active implantable medical devices such as implantable infusion pumps.

This part of ISO 22442 does not apply to *in vitro* diagnostic devices.

This part of ISO 22442 can only be used in combination with ISO 14971 and is not a “standalone” standard.

To show compliance with this part of ISO 22442, its specified requirements should be fulfilled. The guidance given in the Notes and informative annexes is not normative and is not provided as a checklist for auditors.

[SIST EN ISO 22442-1:2016](https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016)

<https://standards.iteh.ai/catalog/standards/sist/e20d70db-54cd-42db-9fb1-30c3413f15e4/sist-en-iso-22442-1-2016>