

**SLOVENSKI STANDARD**  
**oSIST prEN ISO 13897:2017**  
**01-september-2017**

---

**Zobozdravstvo - Kapsule za ponovno uporabo zobnega amalgama (ISO/DIS 13897:2017)**

Dentistry - Dental amalgam reusable mixing-capsules (ISO/DIS 13897:2017)

Zahnheilkunde - Wiederverwendbare Mischkapseln für zahnärztliches Amalgam (ISO/DIS 13897:2017)

Art dentaire - Capsules réutilisables pour mélanger amalgame (ISO/DIS 13897:2017)

**Ta slovenski standard je istoveten z: prEN ISO 13897**

---

<https://standards.iteh.ai/catalog/standards/sist/4adebe40-9c80-423c-9807-6f51d9ddb164/sist-en-iso-13897-2018>

**ICS:**

11.060.10      Zobotehnični materiali      Dental materials

**oSIST prEN ISO 13897:2017**

**en**



# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 13897

ISO/TC 106/SC 6

Secretariat: DIN

Voting begins on:  
2017-05-11Voting terminates on:  
2017-08-02

---

---

## Dentistry — Dental amalgam reusable mixing-capsules

*Art dentaire — Capsules réutilisables pour mélanger amalgame*

ICS: 11.060.10

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

SIST EN ISO 13897:2018<https://standards.iteh.ai/catalog/standards/sist/4adebe40-9c80-423c-9807-6f51d9ddb164/sist-en-iso-13897-2018>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

**ISO/CEN PARALLEL PROCESSING**



Reference number  
ISO/DIS 13897:2017(E)

© ISO 2017

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

SIST EN ISO 13897:2018

<https://standards.iteh.ai/catalog/standards/sist/4adebe40-9c80-423c-9807-6f51d9ddb164/sist-en-iso-13897-2018>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
[copyright@iso.org](mailto:copyright@iso.org)  
[www.iso.org](http://www.iso.org)

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Requirements</b> .....	<b>1</b>
4.1 Dimensions.....	1
4.1.1 Length tolerance.....	1
4.1.2 Diameter tolerance.....	2
4.2 Loss of mass from the mixing-capsule during mixing.....	2
4.3 Retention of dental amalgam in the mixing-capsule.....	2
<b>5 Sampling</b> .....	<b>2</b>
<b>6 Measurements and test methods</b> .....	<b>2</b>
6.1 Dimensions.....	2
6.1.1 Apparatus.....	2
6.1.2 Test sample.....	2
6.1.3 Procedure.....	2
6.1.4 Report.....	3
6.2 Loss of mass from the mixing-capsule during mixing.....	3
6.2.1 Principle.....	3
6.2.2 Test sample.....	3
6.2.3 Apparatus.....	3
6.2.4 Test procedure.....	3
6.2.5 Expression of the results.....	5
6.2.6 Report.....	5
6.3 Retention of dental amalgam in the mixing-capsule.....	5
6.3.1 Principle.....	5
6.3.2 Test sample.....	5
6.3.3 Apparatus.....	6
6.3.4 Test procedure.....	6
6.3.5 Expression of the results.....	6
6.3.6 Additional testing.....	6
<b>7 Labelling</b> .....	<b>7</b>
<b>Bibliography</b> .....	<b>8</b>

## ISO/DIS 13897:2017(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental Equipment*.

This standard, of which this is first edition, cancels and replaces that part of ISO 13897 which is concerned with the requirements for reusable dental amalgam mixing-capsules.

The main change compared to ISO 13897 is as follows:

- The Scope has been revised and is now restricted to dental amalgam reusable mixing-capsules. Single-use capsules are now within the Scope of ISO 20749, (2017) *Dentistry – Pre-capsulated dental amalgam* and the requirements for such capsules may be found in that standard.