



# SLOVENSKI STANDARD

## SIST EN ISO 20570:2018

01-december-2018

---

Zobozdravstvo - Držalo skalpela za oralno kirurgijo (ISO 20570:2018)

Dentistry - Oral surgical scalpel handle (ISO 20570:2018)

Zahnheilkunde - Oralchirurgische Skalpellgriffe (ISO 20570:2018)

Médecine bucco-dentaire - Manche de bistouri pour chirurgie buccale (ISO 20570:2018)

**Ta slovenski standard je istoveten z:** **EN ISO 20570:2018**

[SIST EN ISO 20570:2018](#)

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bf0/sist-en-iso-20570-2018>

**ICS:**

11.060.20 Zobotehnična oprema Dental equipment

**SIST EN ISO 20570:2018** **en**

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

SIST EN ISO 20570:2018

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bf0/sist-en-iso-20570-2018>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

EN ISO 20570

October 2018

ICS 11.060.20

English Version

**Dentistry - Oral surgical scalpel handle (ISO 20570:2018)**

Médecine bucco-dentaire - Manche de bistouri pour  
chirurgie buccale (ISO 20570:2018)

Zahnheilkunde - Oralchirurgische Skalpellgriffe (ISO  
20570:2018)

This European Standard was approved by CEN on 10 August 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

**iTeh STANDARD PREVIEW**  
CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST EN ISO 20570:2018](#)

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bfe0/sist-en-iso-20570-2018>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

Contents	Page
European foreword.....	3

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 20570:2018  
<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bf0/sist-en-iso-20570-2018>

## European foreword

This document (EN ISO 20570:2018) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2019, and conflicting national standards shall be withdrawn at the latest by April 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Endorsement notice iTeh STANDARD PREVIEW

The text of ISO 20570:2018 has been approved by CEN as EN ISO 20570:2018 without any modification.  
<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bfe0/sist-en-iso-20570-2018>

[SIST EN ISO 20570:2018](#)

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bfe0/sist-en-iso-20570-2018>

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

SIST EN ISO 20570:2018

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bf0/sist-en-iso-20570-2018>

# INTERNATIONAL STANDARD

ISO  
**20570**

First edition  
2018-09

---

---

## Dentistry — Oral surgical scalpel handle

*Médecine bucco-dentaire — Manche de bistouri pour chirurgie buccale*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 20570:2018](#)

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bf0/sist-en-iso-20570-2018>



Reference number  
ISO 20570:2018(E)

© ISO 2018

# iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 20570:2018](#)

<https://standards.iteh.ai/catalog/standards/sist/b120b17c-42ee-4004-990f-1db9fd1bf0/sist-en-iso-20570-2018>



## COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

	Page
<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>v</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Classification</b>	<b>2</b>
<b>5 Requirements</b>	<b>2</b>
5.1 Shape	2
5.2 Dimensions	4
5.2.1 Maximum overall length	4
5.2.2 Dimensions of fitting feature	4
5.2.3 Dimensions of the shank	4
5.2.4 Angle between the handle and the working tip	4
5.3 Materials	4
5.4 Surface finish	4
5.5 Resistance to reprocessing	5
5.6 Hardness	5
5.7 Security of attachment	5
<b>6 Measurement and test methods</b>	<b>5</b>
6.1 Measurement of dimensions	5
6.2 Test on surface finish	5
6.3 Test on resistance to reprocessing	5
6.4 Hardness test	5
6.5 Security of attachment	5
<b>7 Marking, labelling and instructions for use</b>	<b>5</b>
7.1 Marking on the scalpel handle	5
7.2 Labelling on the package	6
7.3 Instructions for use	6