



SLOVENSKI STANDARD
kSIST-TS FprCEN/TS 17638:2021

01-marec-2021

**Emisije nepremičnih virov - Določevanje masne koncentracije formaldehida -
Ročna metoda**

Stationary source emissions - Determination of the mass concentration of formaldehyde
- Manual method

Emissionen aus stationären Quellen - Manuelles Verfahren zur Bestimmung der
Massenkonzentration von Formaldehyd - Referenzverfahren

Emissions de sources fixes - Détermination de la concentration massique en
formaldéhyde - Methode manuelle

[kSIST-TS FprCEN/TS 17638:2021](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a46448c261/ksist-ts-fprcen-ts-17638-2021)

[https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a46448c261/ksist-ts-fprcen-ts-17638-2021)

Ta slovenski standard je istoveten z: FprCEN/TS 17638

ICS:

13.040.40 Emisije nepremičnih virov Stationary source emissions

kSIST-TS FprCEN/TS 17638:2021 **en,fr,de**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[kSIST-TS FprCEN/TS 17638:2021](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021)

<https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021>

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

FINAL DRAFT
FprCEN/TS 17638

January 2021

ICS 13.040.40

English Version

**Stationary source emissions - Determination of the mass
concentration of formaldehyde - Manual method**

Emissions de sources fixes - Détermination de la
concentration massique en formaldéhyde - Methode
manuelle

Emissionen aus stationären Quellen - Manuelles
Verfahren zur Bestimmung der Massenkonzentration
von Formaldehyd - Referenzverfahren

This draft Technical Specification is submitted to CEN members for Vote. It has been drawn up by the Technical Committee CEN/TC 264.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a Technical Specification. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a Technical Specification.

<https://standards.iteh.ai/catalog/standards/sist/2077f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents		Page
European foreword.....		5
Introduction		6
1	Scope.....	7
2	Normative references.....	7
3	Terms and definitions	7
4	Symbols and abbreviations	12
5	Principle	12
6	Sampling strategy.....	12
6.1	General.....	12
6.2	Non-isokinetic sampling.....	13
6.3	Isokinetic sampling with a side stream.....	13
7	Sampling equipment.....	14
7.1	Sampling probe.....	14
7.1.1	Non-isokinetic sampling.....	14
7.1.2	Isokinetic sampling.....	14
7.1.3	Particle filter.....	14
7.2	Absorbers.....	15
7.3	Sampling pump	17
7.4	Gas volume meter.....	17
7.5	Filter housing.....	18
7.6	Temperature controller.....	18
7.7	Suction and volume flow meter	18
7.8	Additional apparatus.....	18
7.9	Additional apparatus for isokinetic sampling: Connection to the main line.....	18
7.10	Materials.....	18
7.11	Absorption solution.....	18
8	Performance characteristics of the method.....	19
8.1	General.....	19
8.2	Performance characteristics and performance criteria of the sampling system	19
8.3	Performance characteristics of the analysis	20
8.3.1	Sources of uncertainty.....	20
8.3.2	Performance criterion of analysis	20
8.4	Establishment of the uncertainty budget.....	21
9	Sampling procedure.....	22
9.1	Preparation and installation of equipment.....	22
9.1.1	Sampling location.....	22
9.1.2	Measurement point(s).....	22
9.1.3	Preparation	22
9.1.4	Checks.....	22
9.1.5	Sampling.....	22
9.1.6	Transport and storage of samples	23
9.1.7	Other parameters to be recorded	23
9.2	Validation of results.....	24

9.2.1	Parameters depending on the stationary source.....	24
9.2.2	Leak tests.....	24
9.2.3	Field blank.....	24
9.2.4	Other elements	25
10	Analysis	25
11	Calibration.....	25
11.1	General	25
11.2	Formaldehyde standard solution	25
11.3	Formaldehyde calibration solution	26
11.4	Determination of the calibration curve.....	26
12	Expression of results	27
13	Measurement report	27
Annex A (normative) Analytical methods.....		29
A.1	Method A (AHMT method).....	29
A.1.1	Principle.....	29
A.1.2	Apparatus, reagents and solutions.....	29
A.1.2.1	Apparatus.....	29
A.1.2.2	Reagents and solutions.....	29
A.1.3	Procedure.....	29
A.1.4	Calculation of the mass of absorbed formaldehyde in the wash bottle	30
A.2	Method B (Pararosaniline method).....	30
A.2.1	Principle.....	30
A.2.2	Apparatus, reagents and solutions.....	30
A.2.2.1	Apparatus.....	30
A.2.2.2	Reagents.....	31
A.2.3	Procedure.....	31
A.2.4	Calculation of the mass of absorbed formaldehyde in the wash bottle	31
A.3	Method C (acetylacetone method).....	31
A.3.1	Principle.....	31
A.3.2	Apparatus, reagents and solutions.....	32
A.3.2.1	Apparatus.....	32
A.3.2.2	Reagents and solutions.....	32
A.3.3	Procedure.....	32
A.3.4	Calculation of the mass of formaldehyde absorbed in the wash bottles	32
A.4	Method D (DNPH method).....	33
A.4.1	Principle.....	33
A.4.2	Apparatus, reagents and solutions.....	33
A.4.2.1	Apparatus.....	33

FprCEN/TS 17638:2021 (E)

A.4.2.2 Reagents and solutions	33
A.4.3 Procedure.....	33
Annex B (informative) Measurement uncertainty	37
B.1 General.....	37
B.2 Elements required for the uncertainty determinations	37
B.3 Example of an uncertainty calculation	37
B.3.1 Specific conditions in the field.....	37
B.3.2 Performance characteristics of the method.....	38
B.3.3 Model equation and application of rule of uncertainty propagation	40
B.3.3.1 Concentration of formaldehyde.....	40
B.3.3.2 Calculation of the combined uncertainty of C_m and V_{std}	40
B.3.3.3 Calculation of sensitivity coefficients	41
B.3.3.4 Results of the standard uncertainties calculations.....	41
B.3.4 Estimation of the combined uncertainty	42
B.3.5 Calculation of the expanded uncertainty.....	42
B.3.6 Uncertainty associated to the mass concentration of formaldehyde at O ₂ reference concentration	43
Annex C (informative) Validation tests.....	45
Bibliography.....	46

[kSIST-TS FprCEN/TS 17638:2021](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021)

<https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021>

European foreword

This document (FprCEN/TS 17638:2021) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

This document is currently submitted to the Vote on TS.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[kSIST-TS FprCEN/TS 17638:2021](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021)

<https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021>

FprCEN/TS 17638:2021 (E)**Introduction**

Formaldehyde is a carcinogenic pollutant that is generated in different industrial sectors, like energy industries (combustion plants (e.g. for wood and gas), combustion engines (gas engines and turbines), chemical industry (e.g. formaldehyde production), food industry (e.g. smoking plants), wood industry (e.g. production of wood-based panels or wood pellets) and thus contained in emissions of these processes.

Currently, no European (EN) or International (ISO) Standard exists for the continuous or periodic measurement of formaldehyde emissions, which are being addressed, e.g. by the European Commission in its implementing decision 2015/2119 [1] establishing best available techniques (BAT) conclusions, under Directive 2010/75/EU [2], for the production of wood-based panels.

Instead, different national methods for formaldehyde measurements are currently applied, e.g. US EPA M316 [3], VDI 3862 Part 4 [4], VDI 3862 Part 6 [5], and FD X43-319 [6], all of them based on sampling in aqueous absorption solutions. Several comparison studies have shown that the equivalence of these methods is not ensured.

This measurement method is specified as a Technical Specification because currently no sufficient validation data are available. A comprehensive validation programme has been developed and will be carried out as soon as the funding is ensured.

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[kSIST-TS FprCEN/TS 17638:2021](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021)

<https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021>

1 Scope

This document specifies the reference method for the determination of the concentration of formaldehyde in emissions from stationary sources. Waste gas samples are taken by absorption in water and subsequently analysed by spectrophotometry or HPLC. The method applies to waste gases in which the formaldehyde concentration is 2 mg/m³ to 60 mg/m³, on dry basis, at the reference conditions of 273 K and 101,3 kPa.

The specific components and the requirements for the measuring system are described. A number of performance characteristics with associated minimum performance criteria are specified for the measuring system.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13284-1:2017, *Stationary source emissions - Determination of low range mass concentration of dust - Part 1: Manual gravimetric method*

EN 15259:2007, *Air quality - Measurement of stationary source emissions - Requirements for measurement sections and sites and for the measurement objective, plan and report*

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

absorber

device in which formaldehyde is absorbed into an absorption solution

Note 1 to entry: For formaldehyde absorption wash bottles are used as absorbers.

3.2

limit of quantification

lowest amount of an analyte that is quantifiable with a given confidence level

Note 1 to entry: For a manual method the limit of quantification is usually calculated as ten times the standard deviation of blank measurements provided that the blank value is negligible. This corresponds to a confidence level of 95 %.

3.3

analytical repeatability

closeness of the agreement between the results of successive measurements of the same measurand carried out under the same conditions of measurement

Note 1 to entry: Analytical repeatability conditions include:

FprCEN/TS 17638:2021 (E)

- the same measurement procedure;
- the same laboratory;
- the same sampling equipment, used under the same conditions and at the same location;
- repetition over a short period of time.

Note 2 to entry: Analytical repeatability may be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 3 to entry: In this document the analytical repeatability is expressed as a value with a level of confidence of 95 %.

3.4 calibration

set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring method or measuring system, and the corresponding values given by the applicable reference

3.5 certified reference material CRM

reference material, accompanied by documentation issued by an authoritative body and providing one or more specified property values with associated uncertainties and traceabilities, using valid procedures

[SOURCE: ISO/IEC Guide 99:2007 [7]]

iTech STANDARD PREVIEW
(standards.iteh.ai)

3.6 chemical blank

content of an unexposed sample of the absorption solution, plus reagents that are added to the solution before analysis if necessary

kSIST-TS FprCEN/TS 17638:2021
http://standards.iteh.ai/catalog/standards/sist/17638-2021/ksist-ts-17638-2021
63a464d8c261/ksist-ts-fprcen-ts-17638-2021

3.7 combined uncertainty

standard uncertainty attached to the measurement result calculated by combination of several standard uncertainties

Note 1 to entry: According to the principles laid down in ISO/IEC Guide 98-3

3.8 emission limit value ELV

limit value given in regulations such as EU Directives, ordinances, administrative regulations, permits, licences, authorisations or consents

Note 1 to entry: ELV can be stated as concentration limits expressed as half-hourly, hourly and daily averaged values, or mass flow limits expressed as hourly, daily, weekly, monthly or annually aggregated values.

3.9 expanded uncertainty

quantity defining a level of confidence about the result of a measurement that may be expected to encompass a specific fraction of the distribution of values that could reasonably be attributed to a measurand

$$U = k \times uc$$

Note 1 to entry: In this Technical Specification, the expanded uncertainty is calculated with a coverage factor of $k = 2$, and with a level of confidence of 95 %.

Note 2 to entry: The expression overall uncertainty is sometimes used to express the expanded uncertainty.

3.10 field blank

value determined by a specific procedure used to ensure that no significant contamination has occurred during all steps of the measurement and to check that the operator can achieve a quantification level adapted to the task

3.11 isokinetic sampling

sampling at a rate such that the velocity and direction of the gas entering the sampling nozzle are the same as the velocity and direction of the gas in the duct at the measurement point

[SOURCE: EN 13284-1:2017]

3.12 measurand

particular quantity subject to measurement

Note 1 to entry: The measurand is a quantifiable property of the stack gas under test, for example mass concentration of a measured component, temperature, velocity, mass flow, oxygen content and water vapour content.

[SOURCE: EN 15259:2007]

3.13 measurement method

method described in a written procedure containing all the means and procedures required to sample and analyse, namely field of application, principle and/or reactions, definitions, equipment, procedures, presentation of results, other requirements and measurement report

[SOURCE: EN 14793:2017 [8]]

3.14 measurement plane

plane normal to the centreline of the duct at the sampling position

Note 1 to entry: Measurement plane is also known as sampling plane.

[SOURCE: EN 15259:2007]

3.15 measurement point

position in the measurement plane at which the sample stream is extracted or the measurement data are obtained directly

Note 1 to entry: Measurement point is also known as sampling point.

[SOURCE: EN 15259:2007]

FprCEN/TS 17638:2021 (E)**3.16****measurement port**

opening in the waste gas duct along the measurement line, through which access to the waste gas is gained

Note 1 to entry: Measurement port is also known as sampling port or access port.

[SOURCE: EN 15259:2007]

3.17**measurement series**

successive measurements carried out at the same measurement plane and at the same operating conditions of the industrial process

[SOURCE: EN 13284-1:2017]

3.18**performance characteristic**

one of the quantities (described by values, tolerances, range, etc.) assigned to equipment in order to define its performance

3.19**reference method****RM**

measurement method taken as a reference by convention, which gives the accepted reference value of the measurand

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Note 1 to entry: A reference method is fully described.

Note 2 to entry: A reference method can be a manual or an automated method.

Note 3 to entry: Alternative methods can be used if equivalence to the reference method has been demonstrated.

[SOURCE: EN 15259:2007]

3.20**repeatability in the field**

closeness of the agreement between the results of simultaneous measurements of the same measurand carried out with two sets of equipment under the same conditions of measurement

Note 1 to entry: These conditions include:

- the same measurement procedure;
- two sets of equipment, the performances of which are fulfilling the requirements of the reference method, used under the same conditions;
- the same location;
- implemented by the same laboratory;
- typically calculated over short periods of time in order to avoid the effect of changes of influence parameters (e.g. 30 min).

Note 2 to entry: Repeatability may be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 3 to entry: In this European Standard the repeatability under field conditions is expressed as a value with a level of confidence of 95 %.

3.21

reproducibility in the field

closeness of the agreement between the results of simultaneous measurements of the same measurand carried out with several sets of equipment under the same conditions of measurement

Note 1 to entry: These conditions include:

- the same measurement procedure;
- several sets of equipment, the performance of which fulfils the requirements of the reference method, used under the same conditions;
- the same location;
- implemented by several laboratories.

Note 2 to entry: Reproducibility may be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 3 to entry: In this European Standard the reproducibility under field conditions is expressed as a value with a level of confidence of 95 %.

3.22

standard uncertainty uncertainty of the result of a measurement expressed as a standard deviation
(standards.iteh.ai)

[SOURCE: ISO/IEC Guide 98-3:2008 [9]]

[kSIST-TS FprCEN/TS 17638:2021](https://standards.iteh.ai/catalog/standards/sist/2677f3c3-7588-4a91-ace1-63a464d8c261/ksist-ts-fprcen-ts-17638-2021)

3.23

uncertainty

parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand

[SOURCE: ISO/IEC Guide 98-3:2008 [9]]

3.24

uncertainty budget

statement of a measurement uncertainty, of the components of that measurement uncertainty, and of their calculation and combination

Note 1 to entry: An uncertainty budget should include the measurement model, estimates, and measurement uncertainties associated with the quantities in the measurement model, covariances, type of applied probability density functions, degrees of freedom, type of evaluation of measurement uncertainty, and any coverage factor.

[SOURCE: ISO/IEC Guide 99:2007 [7]]

3.25

waste gas

any gas leaving a process which is not a product (includes exhaust gas, off-gas and flue-gas)