
Merjenje radioaktivnosti v okolju - Tla - 7. del: Meritve radionuklidov, ki sevajo žarke gama, na kraju samem (ISO 18589-7:2013)

Measurement of radioactivity in the environment - Soil - Part 7: In situ measurement of gamma-emitting radionuclides (ISO 18589-7:2013)

Ermittlung der Radioaktivität in der Umwelt - Erdboden - Teil 7: In-situ-Messung von Gammastrahlung emittierenden Radionukliden (ISO 18589-7:2013)

Mesurage de la radioactivité dans l'environnement - Sol - Partie 7: Mesurage in situ des radionucléides émetteurs gamma (ISO 18589-7:2013)

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

Ta slovenski standard je istoveten z: EN ISO 18589-7:2016

ICS:

13.080.99	Drugi standardi v zvezi s kakovostjo tal	Other standards related to soil quality
17.240	Merjenje sevanja	Radiation measurements

SIST EN ISO 18589-7:2016**en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 18589-7:2016

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

EUROPEAN STANDARD

EN ISO 18589-7

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2016

ICS 17.240; 13.080.01

English Version

Measurement of radioactivity in the environment - Soil - Part 7: In situ measurement of gamma-emitting radionuclides (ISO 18589-7:2013)

Mesurage de la radioactivité dans l'environnement -
Sol - Partie 7: Mesurage in situ des radionucléides
émetteurs gamma (ISO 18589-7:2013)

Ermittlung der Radioaktivität in der Umwelt -
Erdboden - Teil 7: In-situ-Messung von
Gammastrahlung emittierenden Radionukliden (ISO
18589-7:2013)

This European Standard was approved by CEN on 21 February 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 18589-7:2016
<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

European foreword

The text of ISO 18589-7:2013 has been prepared by Technical Committee ISO/TC 85 “Nuclear energy, nuclear technologies, and radiological protection” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 18589-7:2016 by Technical Committee CEN/TC 430 “Nuclear energy, nuclear technologies, and radiological protection” the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2016, and conflicting national standards shall be withdrawn at the latest by October 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of ISO 18589-7:2013 has been approved by CEN as EN ISO 18589-7:2016 without any modification.

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 18589-7:2016](https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016)

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

INTERNATIONAL
STANDARD

ISO
18589-7

First edition
2013-10-01

**Measurement of radioactivity in the
environment — Soil —**

Part 7:
**In situ measurement of gamma-
emitting radionuclides**

iTeh STANDARD PREVIEW
Mesurage de la radioactivité dans l'environnement — Sol —
(standards.iteh.ai) Partie 7: Mesurage in situ des radionucléides émetteurs gamma

[SIST EN ISO 18589-7:2016](https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016)

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>



Reference number
ISO 18589-7:2013(E)

© ISO 2013

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 18589-7:2016

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms, definitions, symbols, and units.....	2
3.1 Terms and definitions.....	2
3.2 Symbols and units.....	3
4 Principles.....	6
4.1 Measurement method.....	6
4.2 Uncertainties of the measurement method.....	6
5 Equipment.....	6
5.1 Portable <i>in situ</i> spectrometry system.....	6
5.2 Detector System.....	7
5.3 Pulse processing electronics.....	8
5.4 Assembly jig for a detector system.....	9
5.5 Collimated detector.....	9
6 Procedure.....	12
6.1 Calibration.....	12
6.2 Method of combined calibrations.....	12
7 Quality assurance and quality control program.....	17
7.1 General.....	17
7.2 Influencing variables.....	17
7.3 Instrument verification.....	17
7.4 Method verification.....	17
7.5 Quality control program.....	17
7.6 Standard operating procedure.....	19
8 Expression of results.....	19
8.1 Calculation of activity per unit of surface area or unit of mass.....	19
8.2 Calculation of the characteristic limits and the best estimate of the measurand as well as its standard uncertainty.....	19
8.3 Calculation of the radionuclide specific ambient dose rate.....	21
9 Test report.....	22
Annex A (informative) Influence of radionuclides in air on the result of surface or mass activity measured by <i>in situ</i> gamma spectrometry.....	23
Annex B (informative) Influence quantities.....	24
Annex C (informative) Characteristics of germanium detectors.....	27
Annex D (informative) Field-of-view of an <i>in situ</i> gamma spectrometer as a function of the photon energy for different radionuclide distributions in soil.....	29
Annex E (informative) Methods for calculating geometry factors and angular correction factors.....	33
Annex F (informative) Example for calculation of the characteristic limits as well as the best estimate of the measurand and its standard uncertainty.....	41
Annex G (informative) Conversion factors for surface or mass activity to air kerma rate and ambient dose equivalent rate for different radionuclide distribution in soil.....	45
Annex H (informative) Mass attenuation factors for soil and attenuation factors for air as a function of photon energy and deviation of $G(E,V)$ for different soil compositions.....	52
Bibliography.....	54

ISO 18589-7:2013(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

The committee responsible for this document is ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

ISO 18589 consists of the following parts, under the general title *Measurement of the radioactivity in the environment — Soil*:

- *Part 1: General guidelines and definitions*
- *Part 2: Guidance for the selection of the sampling strategy, sampling and pre-treatment of samples*
- *Part 3: Measurements of gamma-emitting radionuclides*
- *Part 4: Measurement of plutonium isotopes (plutonium 238 and plutonium 239 + 240) by alpha spectrometry*
- *Part 5: Measurement of strontium 90*
- *Part 6: Measurement of gross alpha and gross beta activities*
- *Part 7: In situ measurement of gamma-emitting radionuclides*

Introduction

In situ gamma spectrometry is a rapid and accurate technique to assess the activity concentration of gamma-emitting radionuclides present in the top soil layer or deposited onto the soil surface. This method is also used to assess the dose rates of individual radionuclides.

In situ gamma spectrometry is a direct physical measurement of radioactivity that does not need any soil samples, thus reducing the time and cost of laboratory analysis of large number of soil samples.

The quantitative analysis of the recorded line spectra requires a suitable area for the measurement. Furthermore, it is required to know the physicochemical properties of the soil and the vertical distribution in the soil to assess the activity of the radionuclides.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 18589-7:2016](https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016)

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 18589-7:2016](https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016)

<https://standards.iteh.ai/catalog/standards/sist/8124c85c-d2a0-433b-86bb-49a2ad1706b9/sist-en-iso-18589-7-2016>

Measurement of radioactivity in the environment — Soil —

Part 7:

In situ measurement of gamma-emitting radionuclides

1 Scope

This part of 18589 specifies the identification of radionuclides and the measurement of their activity in soil using *in situ* gamma spectrometry with portable systems equipped with germanium or scintillation detectors.

This part of ISO 18589 is suitable to rapidly assess the activity of artificial and natural radionuclides deposited on or present in soil layers of large areas of a site under investigation.

This part of ISO 18589 can be used in connection with radionuclide measurements of soil samples in the laboratory (ISO 18589-3) in the following cases:

- routine surveillance of the impact of radioactivity released from nuclear installations or of the evolution of radioactivity in the region;
- investigations of accident and incident situations;
- planning and surveillance of remedial action;
- decommissioning of installations or the clearance of materials.

It can also be used for the identification of airborne artificial radionuclides, when assessing the exposure levels inside buildings or during waste disposal operations.

Following a nuclear accident, *in situ* gamma spectrometry is a powerful method for rapid evaluation of the gamma activity deposited onto the soil surface as well as the surficial contamination of flat objects.

NOTE The method described in this part of ISO 18589 is not suitable when the spatial distribution of the radionuclides in the environment is not precisely known (influence quantities, unknown distribution in soil) or in situations with very high photon flux. However, the use of small volume detectors with suitable electronics allows measurements to be performed under high photon flux.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

IEC 61275, *Radiation protection instrumentation — Measurement of discrete radionuclides in the environment — In situ photon spectrometry system using a germanium detector*

ISO 11929, *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application*

ISO 18589-7:2013(E)

3 Terms, definitions, symbols, and units

3.1 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1.1

intrinsic efficiency η_0

cross section of a detector for photons from the direction of the crystal symmetry axis

Note 1 to entry: The intrinsic efficiency depends on the energy of the photon.

3.1.2

detector efficiency $\eta_0(E)$

detector efficiency in the direction of the crystal symmetry axis as a function of the photon energy E

3.1.3

detector height d

distance between the geometrical centre of the crystal and the soil surface

3.1.4

efficiency per unit of surface area or unit of mass ϵ

ratio between the net count rate of an absorption line with energy E and the photon emission rate per unit area or mass

3.1.5

relative detection efficiency

ratio, expressed in percentage, of the count rate in the ^{60}Co 1.333 keV total absorption peak to the one obtained with a 3 x 3 inch NaI(Tl) scintillator for normal incidence and at 0,25 m from the source

3.1.6

geometry factor G

ratio between the flux density without scattered photons measured at the detector location and the photon emission rate per unit area or mass

3.1.7

aperture angle of collimator θ_{col}

characteristic angle for an *in situ* gamma spectrometer with collimator

3.1.8

relaxation mass per unit area^[Z] β

mathematical parameter describing radionuclide distribution as a function of soil depth

Note 1 to entry: It indicates the soil mass per unit of surface area at which gamma activity decreases to 1/e (37 %).

3.1.9

field-of-view of a detector

soil surface area, from which 90 % of the unscattered detected photons originate

3.1.10

distribution model V

entity of all physical and geometrical parameters to describe the distribution of the radionuclide in the environment as well as the interaction of an emitted photon with soil and air

3.1.11 angular coefficient

k_m

factor taking into account the angular response of the detector and the angular distribution of the incident flux

3.1.12 measurement area

area in the soil and/or on the soil surface having radionuclide activity per unit of surface area or unit of mass

3.1.13 mass per unit area (collimator)[Z]

ζ_{col}

product of material density and wall thickness of a collimator

Note 1 to entry: The mass per unit area is reported for a polar angle, ϑ , of 90° in relation to the crystal centre.

3.1.14 cross section of the detector

ratio of the net rate of the total absorption line at energy E and the flux density of unscattered photons of the energy E in the detector

3.1.15 calibration factor per unit of surface area or unit of mass

w

ratio of the activity of surface area or unit of mass of the radionuclide to the net count rate of the total absorption line

(standards.iteh.ai)

3.2 Symbols and units

For the purposes of this part of ISO 18589, the symbols and units defined in ISO 11929 and given in Table 1 apply.

Table 1 — Symbols

Symbols	Designation	Unit
a	Activity of a given radionuclide at the time of measurement	
	a) per unit of surface area	Bq · m ⁻²
	b) per unit of mass	Bq · kg ⁻¹
\hat{a}	Best estimate of the measurand of the activity of the radionuclide in question	
	a) per unit of surface area	Bq · m ⁻²
	b) per unit of mass	Bq · kg ⁻¹
a_K	Activity of the calibration standard at the time of measurement	Bq
a_0	Activity of the radionuclide in question at the soil surface	Bq · m ⁻²
$a(\zeta)$	Projected surface activity as a function of mass per unit at the surface of the soil	Bq · m ⁻²
a^*	Decision threshold of the measurand of the radionuclide in question at the time of measurement	
	a) per unit of surface area	Bq · m ⁻²
	b) per unit of mass	Bq · kg ⁻¹
$a^\#$	Detection limit of the measurand of the radionuclide in question at the time of measurement	
	a) per unit of surface area	Bq · m ⁻²