SIST

SLOVENSKI STANDARD
SIST EN IEC 60601-2-46:2020

01-januar-2020

Nadomešča:

SIST EN 60601-2-46:2011

Medicinska električna oprema - 2-46. del: Posebne zahteve za osnovno varnost in bistvene lastnosti operacijskih miz (IEC 60601-2-46:2016)

Medical electrical equipment - Part 2-46: Particular requirements for the basic safety and essential performance of operating tables (IEC 60601-2-46:2016)

Medizinische elektrische Geräte - Teil 2-46: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Operationstischen (IEC 60601-2-46:2016)

Appareils électromédicaux - Partie 2-46: Exigences particulières pour la sécurité de base et les performances essentielles des tables d'opération (IEC 60601-2-46:2016)

Ta slovenski standard je istoveten z: **EN IEC 60601-2-46:2019**

ICS:

11.140	Oprema bolnišnic	Hospital equipment

SIST EN IEC 60601-2-46:2020 **en**

EUROPEAN STANDARD

NORME EUROPÉENNE

EUROPÄISCHE NORM

EN IEC 60601-2-46

November 2019

ICS 11.140

Supersedes EN 60601-2-46:2011 and all of its amendments and corrigenda (if any)

English Version

Medical electrical equipment - Part 2-46: Particular requirements for the basic safety and essential performance of operating tables (IEC 60601-2-46:2016)

Appareils électromédicaux - Partie 2-46: Exigences particulières pour la sécurité de base et les performances essentielles des tables d'opération (IEC 60601-2-46:2016)

Medizinische elektrische Geräte - Teil 2-46: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Operationstischen (IEC 60601-2-46:2016)

This European Standard was approved by CENELEC on 2016-09-14. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62D/1365/FDIS, future edition 3.0 of IEC 60601-2-46, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-46:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-05-15
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-11-15

This document supersedes EN 60601-2-46:2011 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 60601-2-46:2016 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-1-8	NOTE	Harmonized as EN 60601-1-8
IEC 60601-1-10:2007	NOTE	Harmonized as EN 60601-1-10:2008 (not modified)
IEC 60601-1-11:2015	NOTE	Harmonized as EN 60601-1-11:2015 (not modified)
IEC 60601-1-12:2014	NOTE	Harmonized as EN 60601-1-12:2015 (not modified)
IEC 80601-2-35	NOTE	Harmonized as EN 80601-2-35

Annex ZA
(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

The Annex ZA of EN 60601-1:2006 is applicable, except as follows:

Publication	Year	Title	EN/HD	Year
Replacement				
IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests	EN 60601-1-2	2015
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3	2008
			+EN 60601-1-3:2008/corrigendum Mar. 2010	2010
			+A11	2016
Addition				
IEC 60601-2-2	-	Medical electrical equipment - Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories	EN IEC 60601-2-2	-
IEC 60601-2-43	-	Medical electrical equipment - Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures	EN 60601-2-43	-
			+AC	2014

IEC 60601-2-46

Edition 3.0 2016-08

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Part 2-46: Particular requirements for the basic safety and essential performance of operating tables

Appareils électromédicaux – Partie 2-46: Exigences particulières pour la sécurité de base et les performances essentielles des tables d'opération

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.140

ISBN 978-2-8322-3565-2

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

INTERNATIONAL ELECTROTECHNICAL COMMISSION

____

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-46: Particular requirements for the basic safety and essential performance of operating tables

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-46 has been prepared by IEC subcommittee 62D Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2010 and constitutes a technical revision. This edition of IEC 60601-2-46 was revised to align structurally with the 2005 edition of IEC 60601-1 and with IEC 60601-1:2005/AMD1:2012.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/1365/FDIS	62D/1371/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

– Requirements and definitions: roman type.

– *Test specifications: italic type*.

– Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

– TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

– “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);

– “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

– “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;

– “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

– “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of OPERATING TABLES. It amends and supplements IEC 60601-1 (third edition, 2005) and its Amendment 1 (IEC 60601-1:2005/AMD1:2012), hereinafter referred to as the general standard.

The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this Standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-46: Particular requirements for the basic safety and essential performance of operating tables

201.1 Scope, object and related standards

Clause 1 of the general standard[1)] applies, except as follows:

201.1.1 Scope

Replacement:

This particular standard specifies safety requirements for OPERATING TABLES, whether or not having electrical parts, including TRANSPORTERS, used for the transportation of the OPERATING TABLE top to or from the base or pedestal of an OPERATING TABLE with detachable OPERATING TABLE top.

NOTE See also 4.2 of the General Standard.

This particular standard does not apply to

– dental PATIENT chairs;
– examination chairs and couches;
– PATIENT-supporting systems of diagnostic and therapeutic devices; (see IEC 60601-2-43)
– OPERATING TABLE heating blankets; (see IEC 80601-2-35)
– PATIENT transfer equipment;
– delivery tables and beds;
– medical beds; (see IEC 60601-2-52)
– field tables.

If OPERATING TABLES will be used in combination with diagnostic and/or therapeutic devices the relevant requirements of each related particular standard are also applicable.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for OPERATING TABLES as defined in 201.3.201.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

---

1) The general standard is IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.