

SLOVENSKI STANDARD

SIST EN 12014-2:2018

01-februar-2018

Nadomešča:
SIST EN 12014-2:1999

Živila - Določevanje vsebnosti nitratov in/ali nitritov - 2. del: Metoda HPLC/IC za določevanje vsebnosti nitratov v zelenjavi in zelenjavnih proizvodih

Foodstuffs - Determination of nitrate and/or nitrite content - Part 2: HPLC/IC method for the determination of nitrate content of vegetables and vegetable products

Lebensmittel - Bestimmung des Nitrat- und/oder Nitritgehaltes - Teil 2: HPLC/IC-Verfahren für die Bestimmung des Nitratgehaltes in Gemüse und Gemüseerzeugnissen

Produits alimentaires - Détermination de la teneur en nitrates et/ou en nitrites - Partie 2: Méthode de détermination par CLHP/CI de la teneur en nitrates des légumes et des produits à base de légumes

Ta slovenski standard je istoveten z: EN 12014-2:2017

ICS:

67.050	Splošne preskusne in analizne metode za živilske proizvode	General methods of tests and analysis for food products
67.080.20	Zelenjava in zelenjavni proizvodi	Vegetables and derived products

SIST EN 12014-2:2018

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 12014-2:2018

<https://standards.iteh.ai/catalog/standards/sist/98ae6054-e144-47e6-8649-d6acb68da855/sist-en-12014-2-2018>

EUROPEAN STANDARD
 NORME EUROPÉENNE
 EUROPÄISCHE NORM

EN 12014-2

December 2017

ICS 67.080.20

Supersedes EN 12014-2:1997

English Version

**Foodstuffs - Determination of nitrate and/or nitrite
 content - Part 2: HPLC/IC method for the determination of
 nitrate content of vegetables and vegetable products**

Produits alimentaires - Détermination de la teneur en
 nitrates et/ou en nitrites - Partie 2: Méthode de
 détermination par CLHP/CI de la teneur en nitrates des
 légumes et des produits à base de légumes

Lebensmittel - Bestimmung des Nitrat- und/oder
 Nitritgehaltes - Teil 2: HPLC/IC-Verfahren für die
 Bestimmung des Nitratgehaltes in Gemüse und
 Gemüseerzeugnissen

This European Standard was approved by CEN on 15 October 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
 COMITÉ EUROPÉEN DE NORMALISATION
 EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3
Introduction	4
1 Scope.....	5
2 Normative references.....	5
3 Principle	5
4 Reagents	5
5 Apparatus and equipment	6
6 Procedure.....	7
6.1 Sample preparation.....	7
6.1.1 Liquid samples, e.g. vegetable juice.....	7
6.1.2 Solid samples, e.g. leaf vegetable.....	7
6.1.3 Pasty samples, e.g. mashed vegetables.....	7
6.2 Hot water extraction of samples (solid and pasty samples).....	7
6.3 Preparation of sample test solutions.....	7
6.4 Preparation of the calibration curve.....	7
6.5 Determination	8
6.5.1 High-performance liquid chromatography.....	8
6.5.2 Ion chromatography	8
7 Evaluation	8
7.1 Identification and quantitative determination.....	8
7.2 Calculation	8
7.3 Precision.....	8
7.3.1 General.....	8
7.3.2 Repeatability.....	9
7.3.3 Reproducibility.....	9
8 Test report.....	9
Annex A (informative) Examples of HPLC/IC chromatograms	10
Annex B (informative) Precision data.....	13
Annex C (informative) Examples of suitable HPLC conditions	15
Annex D (informative) Examples of suitable IC conditions.....	16
Bibliography.....	18

European foreword

This document (EN 12014-2:2017) has been prepared by Technical Committee CEN/TC 275 "Food analysis - Horizontal methods", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2018, and conflicting national standards shall be withdrawn at the latest by June 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 12014-2:1997.

The following changes have been made to the former edition:

- a) the scope has been reduced from 50 mg/kg nitrate to 25 mg/kg nitrate and the upper limit has been deleted;
- b) the purification method 1 and 2 used in the preparation of sample test solutions has been deleted;
- c) a new matrix (iceberg lettuce) has been verified in an interlaboratory test;
- d) update of the HPLC/IC-conditions and chromatograms in Annex A;
- e) the procedure has been extensively revalidated and precision data in Annex B have been revised;
- f) editorially revised.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

EN 12014-2:2017 (E)**Introduction**

The method described in this standard has been developed and validated for investigations regarding the European legislation for nitrate in vegetables and vegetable products. Laboratory experience has shown that this analytical method is also suitable for the determination of nitrite in other matrices; however, this has not been validated in the interlaboratory test scheme cited here.

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

SIST EN 12014-2:2018

<https://standards.iteh.ai/catalog/standards/sist/98ae6054-e144-47e6-8649-d6acb68da855/sist-en-12014-2-2018>

1 Scope

This European Standard specifies a high-performance liquid chromatographic (HPLC) and an ion chromatographic (IC) method for determination of the nitrate level in vegetables and vegetable products. This method is applicable for samples with a content of 25 mg/kg or greater.

It has been validated on naturally contaminated and spiked samples as beetroot juice with nitrate mass fractions of 194 mg/kg and 691 mg/kg, pureed carrots with nitrate mass fractions of 26 mg/kg and 222 mg/kg and with iceberg lettuce with nitrate mass fractions of 623 mg/kg and 3 542 mg/kg.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Principle

The nitrate is extracted from the foodstuff with hot water. The determination is performed either by reverse-phase HPLC and UV detection, or by IC and conductivity or UV detection.

4 Reagents

iTeh STANDARD PREVIEW

Use only reagents of recognized analytical grade, unless otherwise specified. Water shall be nitrate-free and shall comply with grade 1 of ISO 3696.

4.1 Potassium nitrate.

SIST EN 12014-2:2018

[https://standards.iteh.ai/catalog/standards/sist/98ae6054-e144-47e6-8649-](https://standards.iteh.ai/catalog/standards/sist/98ae6054-e144-47e6-8649-d6acb68da855/sist-en-12014-2-2018)

4.2 Sodium nitrate.

d6acb68da855/sist-en-12014-2-2018

4.3 Sulfuric acid, 96 % ($\rho = 1,84 \text{ g/ml}^1$).

4.4 Regenerator for suppressor, e.g. sulfuric acid ($c = 0,0125 \text{ mol/l}^2$).

Carefully pipette 20 ml of sulfuric acid (4.3) into a 1 000 ml volumetric flask containing 800 ml of water. Mix the solution, dilute to the mark with water and mix again. Transfer 33 ml of this solution to a 1 000 ml volumetric flask containing 500 ml of water. Dilute to the mark with water and mix.

Alternatively, use a method with electrolytic suppressor.

4.5 Nitrate stock solution, $\rho = 1\ 000 \text{ mg/l}$ (expressed as ion NO_3^-).

Weigh 1,630 7 g of potassium nitrate (4.1) to the nearest 0,1 mg and dissolve in water in a 1 000 ml volumetric flask.

Alternatively, 1,370 9 g of sodium nitrate (4.2) can be used. If required, 15 ml of sulfuric acid (4.3) can be added as a preservative. Mix the solution, dilute to the mark with water and mix again. This solution is stable for at least two months if stored at +4 °C in the dark (refrigerator).

Alternatively, commercially available standard solutions can also be used.

1) ρ = mass concentration

2) c = substance concentration

EN 12014-2:2017 (E)**4.6 Nitrate standard solutions.**

Prepare suitable dilutions from the nitrate stock solution (4.5), preferably in the range of 1 mg/l to 50 mg/l. Pipette an appropriate volume of nitrate stock solution into a 100 ml volumetric flask, dilute to the mark with water, mix and filter through a membrane filter (5.4), if required.

4.7 Mobile phase, for the HPLC.

Examples of suitable eluents are given in Annex C.

4.8 Mobile phase, for the IC.

Examples of suitable eluents are given in Annex D.

5 Apparatus and equipment

Usual laboratory equipment and, in particular, the following:

5.1 Commercially available laboratory mixer, e.g. a laboratory cutter.

5.2 Homogenizer.

5.3 Fluted filter paper (nitrogen-deficient).

5.4 Membrane filter, for aqueous solutions, with a pore size of 0,45 µm (e.g. regenerated cellulose).

5.5 Filter holder for membrane filter, with suitable syringe.

5.6 HPLC equipment.

5.6.1 High-performance liquid chromatograph, consisting of an eluent reservoir, a pump, a sample applicator, a UV detector adjustable from 190 nm to 800 nm and an evaluation unit.

5.6.2 Analytical (reversed-phase) separating column.

Examples of suitable separating columns are given in Annex C.

Use a pre-column with filling material of the same type to protect the analytical separating column.

5.7 IC apparatus.

5.7.1 Ion-exchange chromatography unit, consisting of an eluent reservoir, a pump, a pulsation damper, a dialysis unit (if required), a suppressor (if required), a sample applicator with a sample loop, a conductivity detector or UV detector adjustable from 190 nm to 800 nm and an evaluation unit.

5.7.2 Analytical separating column.

Examples of suitable separating columns are given in Annex D.

Use a pre-column with filling material of the same type to protect the analytical separating column.

6 Procedure

6.1 Sample preparation

6.1.1 Liquid samples, e.g. vegetable juice

Vigorously shake a representative sample until it is thoroughly mixed. If required, filter the sample through a fluted filter paper (5.3) or, if necessary, predilute it with water to achieve a nitrate concentration of approximately 25 mg/l in the solution.

6.1.2 Solid samples, e.g. leaf vegetable

Take a representative sample and remove adhering animals, soil or dirty particles. If required, remove any inedible or damaged external leaves. Shred the sample prepared in this way, e.g. in a laboratory cutter (5.1) and thoroughly homogenize it using a homogenizer (5.2). Analyse the sample immediately after homogenization.

The requirements of the European legislation shall be taken into account, e.g. [1], [2] ³⁾. If immediate testing of the sample homogenate is not possible, immediately freeze the sample. For analysis, thaw the sample as gently as possible (e.g. overnight in the refrigerator).

If it is not possible to homogenize a fresh produced sample within 24 h, the whole sample should be frozen and analysed within a maximum of 6 weeks.

6.1.3 Pasty samples, e.g. mashed vegetables

Repeatedly stir a representative sample, e.g. by using a homogenizer until it is thoroughly mixed.

6.2 Hot water extraction of samples (solid and pasty samples)

From the sample prepared in accordance with 6.1.2 or 6.1.3, weigh approximately 10 g of material to the nearest of 1 mg into a 500 ml conical flask. Add approximately 400 ml of hot water (approximately 80 °C) and place the conical flask into a boiling water bath. When the water bath has reached boiling point temperature again, keep the flask there for 15 min. Cool to room temperature, transfer the solution into a 500 ml volumetric flask, dilute to the mark, shake thoroughly and filter the contents through a fluted filter paper (5.3).

For sample material with low nitrate content, make appropriate adjustments, if necessary, to the initial test portions and volumetric ratios. When the test portion is increased or decreased (e.g. dried products) it is recommended to check the effect of this modification on the method performance.

6.3 Preparation of sample test solutions

If necessary, filter the sample solutions (6.1.1 and 6.2) through a suitable membrane filter (5.4). At least a minimum of 15 ml of test solution (e.g. starch containing samples) is required. Discard the first few millilitres of sample (to avoid possible contamination). The filtrate is dialysed or directly injected. Obtain at least one reagent blank value in parallel.

6.4 Preparation of the calibration curve

To plot a calibration curve, prepare a series of standard solutions (4.6) of suitable concentrations (generally five different concentration levels and one blank value). Inject equal volumes of the sample solution and standard solutions as described in 7.1. Check the calibration curve using freshly prepared standard solutions.

3) For example, the provisions of Commission Regulation (EC) No. 1882/2006 regarding sample preparation (in the version of 2015) stipulate that samples shall not be washed.

EN 12014-2:2017 (E)**6.5 Determination****6.5.1 High-performance liquid chromatography**

Examples of HPLC conditions which have proven useful in the interlaboratory test are given in Annex C.

6.5.2 Ion chromatography

Examples of IC conditions which have proven useful in the interlaboratory test are given in Annex D.

7 Evaluation**7.1 Identification and quantitative determination**

Inject a suitable volume (see Table C.1 and D.1) of the sample test solution (6.3) and standard solution (4.6) into the HPLC/IC-system. Examples of HPLC/IC chromatograms are given in Annex A. Identify the nitrate peak by comparing the retention times for standard solution and sample test solution.

Perform the quantitative determination in mg/kg or mg/l using a calibration function produced on the basis of the peak area or the peak height.

If the peak area/peak height of the measured sample test solution falls outside the range of the calibration curve, dilute the sample test solution with water and repeat the analysis.

The injected volumes of sample test solutions and standard solutions shall be identical.

7.2 Calculation

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Calculate the mass fraction of nitrate w in mg/kg or the mass concentration ρ in mg/l of the sample using Formula (1):

$$w \text{ or } \rho = \frac{x \cdot F \cdot 1000 \cdot V}{m} \quad \text{SIST EN 12014-2:2018} \quad \text{https://standards.iteh.ai/catalog/standards/sist/98ae6054-e144-47e6-8649-d6acb68da855/sist-en-12014-2-2018} \quad (1)$$

where

x is the nitrate concentration of the sample test solution read from the calibration curve minus the reagent blank value, in mg/l;

F is the dilution factor resulting from any further dilution steps;

m is the used sample amount, in g or ml;

V is the volume of the extraction solution (6.2), in l, here 0,5 l.

Report the result in mg/kg or mg/l with no decimal places, according to current legislation.

7.3 Precision**7.3.1 General**

Details of an interlaboratory test for the precision of the method for beetroot juice, carrot purée and iceberg lettuce are given in Table B.1. The values derived from the interlaboratory test are possibly not applicable to concentration ranges and/or matrices other than those given in Annex B.

7.3.2 Repeatability

The absolute difference between two single test results, found on identical test material by one operator using the same apparatus within the shortest feasible time interval will exceed the repeatability limit r in not more than 5 % of the cases.

The values are:

Beetroot juice, naturally contaminated	$\bar{x} = 194 \text{ mg/kg}$	$r = 14 \text{ mg/kg}$
Beetroot juice, spiked	$\bar{x} = 691 \text{ mg/kg}$	$r = 17 \text{ mg/kg}$
Pur�ed carrots, naturally contaminated	$\bar{x} = 26 \text{ mg/kg}$	$r = 3,3 \text{ mg/kg}$
Pur�ed carrots, spiked	$\bar{x} = 222 \text{ mg/kg}$	$r = 12 \text{ mg/kg}$
Iceberg lettuce, naturally contaminated	$\bar{x} = 623 \text{ mg/kg}$	$r = 39 \text{ mg/kg}$
Iceberg lettuce, spiked	$\bar{x} = 3\,542 \text{ mg/kg}$	$r = 105 \text{ mg/kg}$

7.3.3 Reproducibility

The absolute difference between two single test results found on identical test material reported by two laboratories will exceed the reproducibility limit R not more than 5 % of the cases.

The values are:

Beetroot juice, naturally contaminated	$\bar{x} = 194 \text{ mg/kg}$	$R = 51 \text{ mg/kg}$
Beetroot juice, spiked	$\bar{x} = 691 \text{ mg/kg}$	$R = 32 \text{ mg/kg}$
Pur�ed carrots, naturally contaminated	$\bar{x} = 26 \text{ mg/kg}$	$R = 12 \text{ mg/kg}$
Pur�ed carrots, spiked	$\bar{x} = 222 \text{ mg/kg}$	$R = 71 \text{ mg/kg}$
Iceberg lettuce, naturally contaminated	$\bar{x} = 623 \text{ mg/kg}$	$R = 122 \text{ mg/kg}$
Iceberg lettuce, spiked	$\bar{x} = 3\,542 \text{ mg/kg}$	$R = 876 \text{ mg/kg}$

8 Test report

The test report should consider the requirements as in EN ISO/IEC 17025 and shall contain as a minimum the following data:

- all information necessary for the identification of the sample (kind of sample, origin of sample, designation);
- a reference to this European Standard;
- date and type of sampling procedure (if known);
- date of the sample receipt;
- date of testing;
- test results and units in which they have been expressed;
- any particular points observed in the course of the test;
- any operations not specified in the method or regarded as optional, which might have affected the results.