



SLOVENSKI STANDARD

SIST EN 15984:2017

01-oktober-2017

Nadomešča:
SIST EN 15984:2011

Naftna industrija in proizvodi - Določevanje sestave rafinerijskega plina za ogrevanje in izračunavanje vsebnosti ogljika in kalorične vrednosti - Plinska kromatografska metoda

Petroleum industry and products - Determination of composition of refinery heating gas and calculation of carbon content and calorific value - Gas chromatography method

iTeh STANDARD PREVIEW

Mineralölindustrie und -produkte - Bestimmung der Zusammensetzung von Heizgas für Raffinerien und Berechnung des Kohlenstoffgehaltes und des Heizwertes - Gaschromatographisches Verfahren

[SIST EN 15984:2017](#)

[https://standards.iteh.ai/catalog/standards/sist/402d2ad7-44bf-44a0-9780-](https://standards.iteh.ai/catalog/standards/sist/402d2ad7-44bf-44a0-9780-4040731e20/sist-en-15984-2017)

Industries et produits pétroliers - Détermination de la composition des gaz combustibles de raffinerie, de leur pouvoir calorifique et de leur teneur en carbone - Méthode par chromatographie en phase gazeuse

Ta slovenski standard je istoveten z: EN 15984:2017

ICS:

71.040.50	Fizikalnokemijske analitske metode	Physicochemical methods of analysis
75.160.30	Plinska goriva	Gaseous fuels

SIST EN 15984:2017 **en,fr,de**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 15984:2017](#)

<https://standards.iteh.ai/catalog/standards/sist/402d2ad7-44bf-44a0-9780-4b0040736a20/sist-en-15984-2017>

EUROPEAN STANDARD

EN 15984

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2017

ICS 75.160.30

Supersedes EN 15984:2011

English Version

Petroleum industry and products - Determination of composition of refinery heating gas and calculation of carbon content and calorific value - Gas chromatography method

Industries et produits pétroliers - Détermination de la composition des gaz combustibles de raffinerie, de leur pouvoir calorifique et de leur teneur en carbone - Méthode par chromatographie en phase gazeuse

Mineralölindustrie und -produkte - Bestimmung der Zusammensetzung von Heizgas für Raffinerien und Berechnung des Kohlenstoffgehaltes und des Heizwertes - Gaschromatographisches Verfahren

This European Standard was approved by CEN on 19 June 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3
1 Scope	4
2 Principle	4
3 Reagents and materials	4
4 Apparatus	6
5 Gas chromatographic analysis	6
5.1 Analysis systems.....	6
5.2 System configuration.....	6
5.3 Columns.....	7
6 Calibration	7
6.1 General.....	7
6.2 Absolute response factors.....	7
6.3 Relative response factors.....	7
7 Calculation	8
7.1 General.....	8
7.2 Calculation of the non normalized mole fractions.....	8
7.3 Validation of normalized composition (reference components).....	8
7.4 Calculation of carbon content.....	9
7.5 Calculation of lower calorific value, on mass basis.....	10
8 Expression of results	10
9 Precision.....	10
9.1 General.....	10
9.2 Repeatability.....	10
9.3 Reproducibility.....	11
10 Test report	11
Annex A (informative) Possible column combination	12
Annex B (normative) Detector linearity check	18
Annex C (informative) Calculation test data	20
Annex D (normative) Data for calculation	21
Bibliography	22

European foreword

This document (EN 15984:2017) has been prepared by Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2018, and conflicting national standards shall be withdrawn at the latest by February 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 15984:2011.

In this revision only the precision values have been updated to reflect the realistic values found in five consecutive interlaboratory studies performed by the DIN/FAM over the years 2009 to 2014. These pooled precision statements were calculated and approved by CEN/TC 19/WG 36.

This document is based on a DIN Standard [3] with the same scope.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/402d2ad7-44bf-44a0-9780-4b0040736a20/sist-en-15984-2017>

1 Scope

This European Standard specifies a gas chromatographic analysis for the determination of the composition of fuel gases, as used in refinery heating gas. These results are used to calculate the carbon content and the lower calorific value.

With this gas chromatographic analysis, an overall of 23 refinery heating gas components are determined in concentrations as typically found in refineries (see Table 1 for further details).

Water is not analysed. The results represent dry gases.

NOTE 1 Depending on the equipment used, there is a possibility to determine higher hydrocarbons as well.

NOTE 2 For the purposes of this European Standard, the terms “% (V/V)” is used to represent the volume fraction (φ).

WARNING — The use of this standard can involve hazardous materials, operations and equipment. This document does not purport to address all of the safety problems associated with its use. It is the responsibility of users of this standard to take appropriate measures to ensure the safety and health of personnel prior to application of the document, and fulfil statutory and regulatory requirements for this purpose.

2 Principle

This European Standard defines a procedure that is used to determine all components that are present in a typical refinery heating gas, as indicated in Table 1.

The composition range in which each component can be analysed does depend on the actual sample composition as higher amounts of a certain component may affect the detection range of other components eluting close by. The general ranges which apply to all the individual components are:

- hydrocarbons from 0,01 (mol/100 mol) up to 100 (mol/100 mol);
- non-condensable gases from 0,02 (mol/100 mol) up to 100 (mol/100 mol);
- for hydrogen sulfide a range between 0,1 (mol/100 mol) up to 10 (mol/100 mol) has been found applicable.

Three different analysis systems are necessary; they may be built in three separate gas chromatographs, or be integrated into one.

Depending on the configuration, hydrocarbons with more than five carbon atoms are reported as a sum parameter. The composition of the refinery heating gas is used to calculate the carbon content and the calorific value. A typical procedure is described hereafter. A configuration is acceptable when the determination gives the precision as described in Clause 9.

3 Reagents and materials

3.1 Gases.

3.1.1 Hydrogen, with a minimum purity of 99,995 % (V/V).

3.1.2 Helium, with a minimum purity of 99,995 % (V/V).

3.1.3 Nitrogen, with a minimum purity of 99,995 % (V/V).

3.1.4 Air, free of oil and water.

3.1.5 Argon, as alternative for analysis system 2, with a minimum purity of 99,995 % (V/V).

3.2 Calibration sample.

A certified reference gas mixture in concentrations that allow the determination of the necessary response factors (see 6.1) and retention times is required.

Table 1 — Refinery heating gas components

Number	Description	Chemical formula
1	Hydrogen	H ₂
2	Oxygen/Argon	O ₂ /Ar
3	Nitrogen	N ₂
4	Carbon monoxide	CO
5	Carbon dioxide	CO ₂
6	Hydrogen sulfide	H ₂ S
7	Methane	CH ₄
8	Ethane	C ₂ H ₆
9	Ethene	C ₂ H ₄
10	Ethyne /Acetylene	C ₂ H ₂
11	Propane	C ₃ H ₈
12	Propene	C ₃ H ₆
13	Prop-1-yne (Propyne, Methylacetylene)	C ₃ H ₄
14	Allene (Propadiene)	C ₃ H ₄
15	Methylpropane (<i>iso</i> -Butane)	C ₄ H ₁₀
16	<i>n</i> -Butane	C ₄ H ₁₀
17	<i>But-2-ene Trans</i> (<i>trans</i> -2-Butene)	C ₄ H ₈
18	But-1-ene (1-Butene)	C ₄ H ₈
19	2-methylprop-1-ene (2-Methyl-Propene)	C ₄ H ₈
20	<i>Cis But-2-ene (cis</i> -2-Butene)	C ₄ H ₈
21	Buta-1,3-diene (1,3-Butadiene)	C ₄ H ₆
22	2-Methylbutane (<i>iso</i> -Pentane)	C ₅ H ₁₂
23	<i>n</i> -Pentane	C ₅ H ₁₂
24	Other components with 5 or more Carbon atoms, excluding 2-Methylbutane and Pentane (C ₅₊)	

4 Apparatus

4.1 Usual laboratory apparatus and glassware.

4.2 Gas chromatographic apparatus, consisting of at least three separation systems able to work simultaneously in one gas chromatograph, with a Thermal Conductivity Detector (TCD) and Flame Ionization Detector (FID) that should be available, and confirming to the requirements as given in Clause 5.

5 Gas chromatographic analysis

5.1 Analysis systems

The gas chromatographic system consists of the following three parts:

1) Analysis system 1

All components except hydrogen (see Table 1) are retained on a porous polymer and a molecular sieve column and back flushed.

Hydrogen is determined on a TCD with nitrogen (3.1.3) as the carrier gas (see Figure A.4).

2) Analysis system 2

The second analysis system separates non-condensable gases, carbon dioxide, and hydrocarbons with two carbon atoms and hydrogen sulfide with helium (3.1.2) as the carrier gas and a TCD as a detector. After a pre-separation on a porous polymer column (column 3) propane and higher hydrocarbons are back flushed and vented. See Figure A.5 for details.

When the inert gases (O₂/Ar, N₂, CH₄ and CO) are on the molecular sieve 13X column, this column is isolated. Carbon dioxide, the C₂-hydrocarbons and hydrogen sulfide are eluted from the porous polymer column (column 4) and are detected. The C₂-hydrocarbons from this fraction are not used for quantification. After this the molecular sieve 13X column (column 6) is eluted and the components are determined on the TCD. Methane is quantified on Analysis System 3.

3) Analysis system 3

The third analysis system separates and quantifies all hydrocarbons by an FID and hydrogen (3.1.1) or helium (3.1.2) as the carrier gas. Two columns are used in series. From the methyl silicone column (column 1), the components above a certain cut point e.g. *n*-pentane are back flushed and determined as a summed peak. The hydrocarbons from the alumina oxide column, column 2, are then separated and quantified. See Figure A.6 for details.

5.2 System configuration

The gas chromatographic system may consist of one or more gas chromatographs, with the possibility for isothermal or temperature programmed runs or both depending on the selected system configuration. A TCD and an FID shall be available.

A gas flow control system and gas sampling valves and switching valves are used.

Means for quantification shall be provided.

Optionally a vaporizer may be used.

5.3 Columns

A combination of packed and capillary columns is possible.

It shall be ascertained that a quantitative separation of all the components that are to be determined (see Table 1) is possible and that the summed total of C₅₊ is quantitative. Adequate separation is required between the components on all three separation systems.

If cyclopropane can be determined, it shall be summed to propene.

An example for a typical configuration is given in Annex A.

6 Calibration

6.1 General

Depending on the production procedure of the reference gas mixture, it might be necessary to convert the concentration units of the components in the calibration gas to (mol/mol) %.

Linearity of the detectors shall be checked according to Annex B.

If calibration is necessary, e.g. daily or for an analysis series, a reference gas mixture shall be analysed. For each analysis system (see 5.1) at least one component, the reference component, needs to be calibrated. For example, for analysis system 1 hydrogen, for analysis system 2 nitrogen and for analysis system 3 propane. With these three components, an external calibration is done. The absolute response factor is calculated.

For all the other components, relative response factors will be used (see 6.3).

The relative response factor of the reference component will be set at 1,0.

As alternative all absolute response factors can be determined for all the components that have to be analysed in a reference gas mixture. A check of this calibration can take place with a gas mixture with fewer components.

6.2 Absolute response factors

The determination of absolute response factors is possible when a repeatable injection is possible under constant analysis conditions. The ratio between concentration and peak area is determined and the absolute response factor is calculated using Formula (1):

$$RF_i = \frac{x_i}{A_i} \quad (1)$$

where

RF_i is the absolute response factor of component i ;

x_i is the concentration of component i in (mol/100 mol) in the calibration gas;

A_i is the peak area of component i in the calibration gas.

6.3 Relative response factors

For quantification relative response factors can be used too. In one analysis system, one component is used as a reference (reference component).

The relative response factor is calculated using Formula (2):

EN 15984:2017 (E)

$$RRF_i = \frac{RF_i}{RF_{St}} \quad (2)$$

where

- RRF_i is the relative response factor of component i ;
 RF_i is the absolute response factor of component i ;
 RF_{St} is the absolute response factor of the reference component St .

7 Calculation

7.1 General

For the calculation of parameters of gas mixtures, as described in this standard, the molar composition of the gas mixture is necessary.

A more accurate result may be obtained by analysing the C5 Alkenes (olefins) separately and not adding them to the sum of all heavier components.

7.2 Calculation of the non normalized mole fractions

The concentrations of the components, $x_{i, sample}$, mentioned in Table 1 are calculated as mole fractions. The calculation of these follows Formula (3):

$$x_{i, sample} = A_{i, sample} \times RRF_{i, St} \times RF_{StK} \quad (3)$$

where

- $A_{i, sample}$ is the area of component i in the sample;
 $RRF_{i, St}$ is the relative response factor of component i relative to the reference component St ;
 RF_{StK} is the absolute response factor of the reference component St in the reference gas mixture K .

NOTE If only absolute response factors are used, then reference component St and component i are identical. Therefore, in Formula (3) $RRF = 1$ and each RF represents the individual component i .

7.3 Validation of normalized composition (reference components)

The results can be normalized, when the sum of all the mole fractions of the components are not smaller than 0,98 or greater than 1,02. If not, the analysis should be repeated. If the duplicate analysis does not improve this result, the calibration and the apparatus shall be checked.

Normalization is done according to Formula (4):

$$x_i^* = \frac{x_{i, sample}}{\sum_{i=1}^k x_{i, sample}} * 100 \quad (4)$$

where

- x_i^* is the corrected concentration of component i in the sample in mol/100 mol;

$x_{i,sample}$ is the uncorrected concentration of component i in the sample;

$\sum_{i=1}^k x_{i,sample}$ is the sum of the non-normalized mol fractions of the sample;

k is the number of components.

7.4 Calculation of carbon content

The determined and normalized concentrations of each component, calculated according to Formulae (3) and (4) are independent of pressure and temperature. By multiplying the concentration of each component with their respective molar mass and normalizing the results, the (mol/100 mol) fractions are converted to (g/100 g) of gas, $w_{x_i}^*$ according to Formula (5):

$$w_{x_i}^* = \frac{x_i^* M_i}{\sum_{i=1}^k (x_i^* M_i)} * 100 \quad (5)$$

where

x_i^* is the corrected concentration of the component i in the sample in (mol/100 mol);

M_i is the molar mass of the component i in g/mol;

$\sum_{i=1}^k (x_i^* M_i)$ is the sum of the normalized molar fractions of the sample multiplied by the molar mass;

k is the number of components.

The carbon content of the individual components, w_{C_i} , are given in Table D.1.

The calculation is according to Formula (6):

$$w_{C_i} = \frac{n_{C_i} * M_C}{M_i} \quad (6)$$

where

n_{C_i} is the number of carbon atoms in the component i ;

M_C is the molar mass of carbon in g/mol;

M_i is the molar mass of the component i in g/mol.

The sum of the values found by multiplication of the concentration w_i in (g/100g gas) with the molar carbon concentration w_{C_i} of component i results in the carbon content of the mixture C_c (g C/100g gas), (Formula (7)).

$$C_c = \sum_{i=1}^k w_{x_i}^* * w_{C_i} \quad (7)$$