

---

---

**Ophthalmic optics — Contact lens  
care products — Guidelines for  
determination of shelf-life**

*Optique ophtalmique — Produits d'entretien pour lentilles de  
contact — Lignes directrices pour la détermination de la durée de  
conservation*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 13212:2014](https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014)

[https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-  
1956f0379795/iso-13212-2014](https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014)



**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 13212:2014

<https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword.....	iv
Introduction.....	v
<b>1 Scope.....</b>	<b>1</b>
<b>2 Normative references.....</b>	<b>1</b>
<b>3 Terms and definitions.....</b>	<b>1</b>
<b>4 General requirements.....</b>	<b>1</b>
<b>5 Determination of finished product stability.....</b>	<b>1</b>
5.1 Objective.....	1
5.2 Study methods.....	2
5.3 Description of the product under study.....	3
5.4 Characteristics.....	3
5.5 Evaluation methods.....	4
5.6 Presentation of results.....	4
5.7 Discussion, interpretation, and conclusions.....	4
5.8 Ongoing stability.....	5
<b>Annex A (informative) Example of a stability testing plan for the finished contact lens care product.....</b>	<b>6</b>
<b>Bibliography.....</b>	<b>7</b>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 13212:2014

<https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This third edition cancels and replaces the second edition (ISO 13212:2011), which has been technically revised.

## Introduction

The purpose of stability tests of contact lens care products is to obtain sufficient information to enable the manufacturer to establish an appropriate shelf-life and identify any unique storage conditions required to appear on the labelling of the product.

The quality of a contact lens care product is determined by its content of active ingredient(s), its purity, and its physicochemical and microbiological properties. It is important to take into account the possible interaction of the container/closure with the contents.

The stability studies are intended to ascertain how the quality of a product varies as a function of time and under the influence of a variety of environmental factors.

On the basis of the information obtained, storage conditions are recommended, which will guarantee the maintenance of the quality of the product in relation to its safety, performance, and acceptability throughout the proposed shelf-life.

The design of the finished product stability studies for a care product is based on the knowledge obtained from studies on the active ingredient(s) and from the development studies.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 13212:2014](https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014)

<https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 13212:2014

<https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014>

# Ophthalmic optics — Contact lens care products — Guidelines for determination of shelf-life

## 1 Scope

This International Standard provides guidance on the design of stability studies for use in gathering information to enable determination of the shelf-life of contact lens care products.

This International Standard does not address studies designed to obtain information to establish the in-use stability (i.e. notice of discard date) of contact lens care products.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14729, *Ophthalmic optics — Contact lens care products — Microbiological requirements and test methods for products and regimens for hygienic management of contact lenses*

ISO 14730, *Ophthalmic optics — Contact lens care products — Antimicrobial preservative efficacy testing and guidance on determining discard date*

ISO 18369-1, *Ophthalmic optics — Contact lenses — Part 1: Vocabulary, classification system and recommendations for labelling specifications*

ISO 13212:2014  
<https://standards.iteh.ai/catalog/standards/sist/4122160b-0066-45cc-8ad4-1956f0379795/iso-13212-2014>

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18369-1 apply.

## 4 General requirements

**4.1** The specified shelf-life of the contact lens care product shall be based on the evaluation of the results of stability studies.

**4.2** Analytical methods that have been validated and are stability indicating shall be used to assay for active ingredients. Validation includes, but is not limited to, being able to differentiate between the active ingredient and its degradation products. The test methods used shall be described in full.

## 5 Determination of finished product stability

### 5.1 Objective

The objective of stability testing on contact lens care products is to provide data for determining the time period during which the product performance characteristics are maintained and to define appropriate storage conditions.

The design of the stability tests is based on the known properties of the active ingredient(s), the properties of the chosen formulation, and the recommendations for use of the product.

The relevant assay methods shall be determined prior to the start of the stability testing.

The specifications proposed from the time of manufacture to the end of the proposed shelf-life shall reflect, as far as possible, the results of the stability studies, particularly in relation to any parameters which could have a bearing on performance and safety and on product acceptability.

## 5.2 Study methods

### 5.2.1 General

Before starting stability studies, a suitable testing plan should be set up, taking into consideration the properties of the active ingredient(s) as well as the proposed mode of action of the care product.

### 5.2.2 Real-time studies

These studies should be carried out under a range of controlled test conditions, when applicable, which will enable the shelf-life and the storage requirements which are to appear on the product container label/package insert to be defined. This will normally include studies which are intended to allow the properties of the product at temperatures between 20 °C and 30 °C to be evaluated. However, 25 °C ± 2 °C should be used as the mean kinetic testing temperature.

Relative humidity should be controlled for all products that are not in hermetically sealed containers. In the case of tablets, high humidity conditions should be considered. In the case of aqueous liquids, low humidity conditions should be considered (see [Annex A](#)).

For each study, the mean temperature, the ranges of temperature, and mean humidity, if applicable, shall be stated in the stability report.

These studies are intended to support the initial shelf-life request and, for shelf-life extensions, any changes that could significantly impact the safety and performance of the product (e.g. certain changes in formulation, packaging materials, or manufacturing methods).

NOTE Real-time studies are performed in conjunction with accelerated ageing studies to establish an initial shelf-life.

### 5.2.3 Studies under varying storage conditions

These studies shall be carried out to provide important additional information. They can fulfil a number of objectives, such as

- supporting the initial shelf-life request, by complementing the limited results of the early real-time studies because decomposition, if it occurs, is likely to be accelerated,
- producing useful data at an early stage of development, demonstrating the effects of adverse storage in the packaging and product, and enabling storage conditions and suitable labelling to be provided, and
- supporting a request to extend the shelf-life.

The various test conditions should be stated. Depending on the nature and objectives of the stability study, the following points might need to be considered:

- a) various test temperatures: three or more, particularly if long-term real-time data are unavailable. In addition, the effect of low temperatures might need to be considered, such as below -15 °C (freezer), 2 °C to 8 °C (refrigerator) and freeze-thaw cycling;
- b) high humidity: relative humidity up to 75 % ± 5 % (see [Annex A](#)). Storage under high humidity conditions applies particularly to solid dosage forms. For products such as solutions, suspensions, etc. contained in packs designed to provide a permanent barrier to water loss, storage under high humidity is not necessary. However, low humidity can have an adverse effect on products packaged in semi-permeable containers;



- c) elevated temperature and humidity in combination: e.g. temperature of 40 °C associated with a relative humidity of up to 75 %, possibly with the effects of cycling between different temperatures and humidities;
- d) since most contact lens care products are water-based, relative humidity of 40 % or less (for example, 25 °C and 40 % relative humidity) should be considered (see [Annex A](#));
- e) light: either natural daylight or defined artificial illumination (see [Annex A](#)).

### 5.3 Description of the product under study

#### 5.3.1 Number and nature of the batches tested

The number of batches tested shall be stated with the batch number, details of the composition, date of manufacture, size of the batch and the name of the manufacturer of the active ingredient(s) used. The material used for all packaging that can impact stability of the product shall be stated and supplier identified.

Normally, three batches of the finished product are studied. If the number of batches tested is less than three, it shall be justified.

Satisfactory performance of the product in the smallest size of container with the highest surface/volume ratio shall allow the extension of shelf-life to containers which are up to eight times larger in volume.

#### 5.3.2 Primary container

The product batches shall be packed in the primary containers proposed for marketing. The smallest primary container size should be tested. Satisfactory performance in the smallest size shall allow marketing of the product in containers with eight times the volume present in the smallest size.

Details of the packaging should be stated, including

- a) type(s) of container and closure and nature of the constituent material(s),
- b) nature of any desiccant if used, and
- c) the complete range of sizes of the product proposed for marketing.

### 5.4 Characteristics

#### 5.4.1 General

The following characteristics should be studied:

- a) those in the finished product specification that are likely to be affected by storage, and
- b) those not monitored routinely at the time of manufacture, but which can be indicative of the stability/instability of the particular product (e.g. dissolution of tablets).

#### 5.4.2 Physical characteristics of the finished product

The following physical characteristics should be tested:

- a) physical properties specific to the product, such as tablet hardness and hygroscopicity, or pH, colour, clarity, and viscosity for solutions;
- b) important quality parameters such as the *in vitro* dissolution, moisture content (e.g. in relation to any desiccant used in the packaging), and particle size;