

SLOVENSKI STANDARD

SIST EN ISO 8041-1:2017

01-julij-2017

Nadomešča:

SIST EN ISO 8041:2005

SIST EN ISO 8041:2005/AC:2008

Odzivanje človeka na vibracije - Merilni instrumentarij - 1. del: Splošna uporaba vibracijskih merilnikov (ISO 8041-1:2017)

Human response to vibration - Measuring instrumentation - Part 1: General purpose vibration meters (ISO 8041-1:2017)

iTeh STANDARD PREVIEW

Schwingungseinwirkung auf den Menschen - Messeinrichtung - Teil 1: Schwingungsmesser für allgemeine Anwendungen (ISO 8041-1:2017)

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standard/standard/sist/1c96c6a5-20b3-4757-a4b3-64dea58b5045/sist-en-iso-8041-1-2017)

Réponse des individus aux vibrations - Appareillage de mesure - Partie 1: Instrument de mesure à usage général (ISO 8041-1:2017)

Ta slovenski standard je istoveten z: EN ISO 8041-1:2017

ICS:

13.160	Vpliv vibracij in udarcev na ljudi	Vibration and shock with respect to human beings
--------	------------------------------------	--

SIST EN ISO 8041-1:2017

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)

<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

EUROPEAN STANDARD

EN ISO 8041-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2017

ICS 13.160

Supersedes EN ISO 8041:2005

English Version

Human response to vibration - Measuring instrumentation - Part 1: General purpose vibration meters (ISO 8041- 1:2017)

Réponse des individus aux vibrations - Appareillage de
mesure - Partie 1: Instrument de mesure à usage
général (ISO 8041-1:2017)

Schwingungseinwirkung auf den Menschen -
Messeinrichtung - Teil 1: Schwingungsmesser für
allgemeine Anwendungen (ISO 8041-1:2017)

This European Standard was approved by CEN on 22 April 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

iTeh STANDARD PREVIEW

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)

<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

European foreword

This document (EN ISO 8041-1:2017) has been prepared by Technical Committee ISO/TC 108 "Mechanical vibration, shock and condition monitoring" in collaboration with Technical Committee CEN/TC 231 "Mechanical vibration and shock" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2017 and conflicting national standards shall be withdrawn at the latest by November 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8041:2005.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1e06c6f5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)

<https://standards.iteh.ai/catalog/standards/sist/1e06c6f5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

Endorsement notice

The text of ISO 8041-1:2017 has been approved by CEN as EN ISO 8041-1:2017 without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)

<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

INTERNATIONAL
STANDARD

ISO
8041-1

First edition
2017-05

**Human response to vibration —
Measuring instrumentation —**

**Part 1:
General purpose vibration meters**

Réponse des individus aux vibrations — Appareillage de mesure —

Partie 1: Instrument de mesure à usage général

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)

<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>



Reference number
ISO 8041-1:2017(E)

© ISO 2017

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 8041-1:2017

<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	vi
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms, definitions and symbols	2
3.1 Terms and definitions.....	2
3.1.1 General.....	2
3.1.2 Frequency-weighted values.....	4
3.2 Symbols.....	6
4 Reference environmental conditions	8
5 Performance specifications	8
5.1 General characteristics.....	8
5.2 Display of signal magnitude.....	10
5.2.1 General.....	10
5.2.2 Resolution and refresh rate.....	10
5.2.3 Stabilization, measurement start and display times.....	11
5.3 Electrical output.....	11
5.4 Vibration sensitivity.....	11
5.5 Accuracy of indication at reference frequency under reference conditions.....	12
5.6 Frequency weightings and frequency responses.....	12
5.6.1 Parameters.....	12
5.6.2 Band-limiting filter.....	13
5.6.3 a-v transition filter.....	13
5.6.4 Upward-step filter.....	14
5.6.5 Overall frequency weighting.....	14
5.6.6 Tolerances.....	14
5.7 Amplitude linearity.....	15
5.8 Instrument noise.....	16
5.9 Signal-burst response.....	16
5.10 Overload indication.....	19
5.11 Under-range indication.....	20
5.12 Time averaging.....	20
5.13 Running r.m.s. acceleration.....	20
5.14 Reset.....	21
5.15 Timing facilities.....	21
5.16 Electrical cross-talk.....	21
5.17 Vibration transducer characteristics.....	21
5.18 Power supply.....	21
6 Mounting	22
7 Environmental and electromagnetic criteria	22
7.1 General.....	22
7.2 Air temperature.....	22
7.3 Surface temperature.....	23
7.4 Electrostatic discharge.....	23
7.5 Radio-frequency emissions and public-power-supply disturbances.....	23
7.6 Immunity to AC power-frequency fields and radio-frequency fields.....	24
7.7 Ingress of water and dust.....	24
8 Provision for use with auxiliary devices	25
9 Instrument marking	25
10 Instrument documentation	25

ISO 8041-1:2017(E)

11	Testing and calibration	25
12	Pattern evaluation	28
12.1	General	28
12.2	Testing requirements	28
12.3	Submission for testing	29
12.4	Marking of the vibration meter and information in the instrument documentation	29
12.5	Mandatory facilities and general requirements	29
12.6	Initial instrument preparation	29
12.7	Indication at the reference frequency under reference conditions	29
12.8	Electrical cross-talk	30
12.9	Vibration transducer	30
12.10	Amplitude linearity and under-range indication	31
12.10.1	Electrical tests of amplitude linearity	31
12.10.2	Mechanical tests of amplitude linearity	32
12.11	Frequency weightings and frequency responses	33
12.11.1	General	33
12.11.2	Mechanical tests of frequency response	33
12.11.3	Electrical tests of frequency response	34
12.11.4	Conformance	35
12.12	Instrument noise	35
12.13	Signal-burst response	36
12.14	Overload indication	36
12.15	Reset	36
12.16	Combined axis outputs	37
12.17	AC electrical output	37
12.18	Timing facilities	37
12.19	Power supply	37
12.20	Environmental, electrostatic and radio-frequency tests	37
12.20.1	General	37
12.20.2	Expanded uncertainties for measurements of environmental conditions	38
12.20.3	Acclimatization requirements for tests of the influence of air temperature and relative humidity	38
12.20.4	Test of the influence of air temperature and relative humidity combined	38
12.20.5	Influence of surface temperature	39
12.20.6	Influence of electrostatic discharges	39
12.20.7	Radio-frequency emissions and public-power-supply disturbances	39
12.20.8	Immunity to AC power-frequency fields and radio-frequency fields	40
12.21	Test report	41
13	Validation of one-off instruments	42
13.1	General	42
13.2	Testing requirements	42
13.3	Test object	42
13.4	Submission for testing	43
13.5	Marking of the one-off instrument and information in the instrument documentation	43
13.6	Mandatory facilities and general requirements	43
13.7	Initial instrument preparation	43
13.8	Test procedure	43
13.9	Indication at the reference frequency under reference conditions	44
13.10	Test parameters	45
13.10.1	Vibration measurement chain for hand-arm vibration	45
13.10.2	Vibration measurement chain for whole-body vibration	45
13.10.3	Vibration measurement chain low-frequency whole-body vibration	46
13.11	Conducting the test	46
13.12	Instrument noise	47
13.13	Electrical cross-talk	47
13.14	Overload indication	47
13.15	Timing facilities	47

13.16	Test report.....	48
14	Periodic verification.....	48
14.1	General.....	48
14.2	Testing requirements.....	48
14.3	Test object.....	48
14.4	Submission for testing.....	48
14.5	Preliminary inspection.....	49
14.6	Marking of the vibration meter and information in the instrument documentation.....	49
14.7	Test procedure.....	49
14.8	Test parameters.....	50
	14.8.1 Vibration measurement chain for hand-arm vibration.....	50
	14.8.2 Vibration measurement chain for whole-body vibration.....	50
	14.8.3 Vibration measurement chain low-frequency whole-body vibration.....	50
14.9	Conducting the test.....	51
14.10	Test report.....	51
15	<i>In situ</i> checks.....	51
15.1	General.....	51
15.2	Preliminary inspection.....	52
15.3	Vibration sensitivity (field calibration).....	52
	Annex A (normative) Specification for field vibration calibrator.....	53
	Annex B (informative) Frequency weightings.....	55
	Annex C (informative) Realization of frequency weighting filters.....	74
	Annex D (informative) Running r.m.s. time averaging.....	78
	Annex E (informative) Vibration transducer characteristics.....	81
	Annex F (informative) Tests for mounting systems.....	84
	Annex G (normative) Instrument documentation.....	87
	Annex H (normative) Phase response requirements for measurement of non-r.m.s. quantities.....	92
	Annex I (informative) Guidelines for the estimation of the instrumental measurement uncertainty.....	100
	Bibliography.....	106

ISO 8041-1:2017(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html. (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 108, *Mechanical vibration, shock and condition monitoring*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.
SIST EN ISO 8041-1:2017
<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

This first edition cancels and replaces ISO 8041:2005, which has been technically revised. It also incorporates the Technical Corrigendum ISO 8041:2005/Cor. 1:2007. The following main changes have been made:

- addition of an Introduction explaining the reasons for this revision;
- addition of a validation test for one-off instruments;
- revision and simplification of the verification test;
- addition of Annex I, which gives example estimates of the instrumental measurement uncertainty;
- correction of errors in formulae, numbers and figures.

A list of parts in the ISO 8041 series can be found on the ISO website.

Introduction

Until 2005, when the previous edition of this document was published, measuring instrumentation for human response to vibration (vibration meters) normally consisted of a signal processing unit and a detachable vibration transducer. According to recent developments, however, part of the signal processing steps can be integrated in the transducer unit, so that the signal coming out of the transducer's sensing element and going into the signal conditioning unit is not accessible any more. These transducer units include, for example, IEPE and MEMS transducers.

Some of the test procedures specified in this document, however, presume that this point in the signal path is accessible (electrical input). Since such an input is not mandatory these tests can only be performed on a vibration meter having an electrical input or after some technical modifications to the instrumentation, e.g. internal access to signal paths. Or those tests can only be performed mechanically, which in certain cases requires modifications to some test procedures. Such modifications to test procedures, however, are beyond the present scope of this document.

Some of the test procedures specified in this document can only be performed if an electrical output is available, see for example 5.13. Since such an output is not mandatory these tests can only be performed on a vibration meter having an electrical output or after some technical modifications to the instrumentation, e.g. internal access to signal paths.

The verification test now specified in this document is practicable and has the objective of identifying an instrument which is adequately calibrated for the intended applications and is suitable for its purpose, at a cost reasonable for the calibration laboratory and affordable for the end user. Therefore, the verification test is strongly reduced in its extent compared to the full pattern evaluation, or validation, and only tests the most relevant characteristics of a vibration meter.

ITeH STANDARD PREVIEW
(standards.iteh.ai)
[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)
<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 8041-1:2017](https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017)

<https://standards.iteh.ai/catalog/standards/sist/1c96c6a5-20b3-4757-a4b3-64eea58bc045/sist-en-iso-8041-1-2017>

Human response to vibration — Measuring instrumentation —

Part 1: General purpose vibration meters

1 Scope

This document specifies the performance specifications and tolerance limits for instruments designed to measure vibration values, for the purpose of assessing human response to vibration. It includes requirements for pattern evaluation, or validation, periodic verification and *in situ* checks, and the specification of vibration calibrators for *in situ* checks.

Vibration instruments specified in this document can be single instruments, combinations of instrumentation or computer-based acquisition and analysis systems.

Vibration instruments specified in this document are intended to measure vibration for one or more applications, such as the following:

- hand-transmitted vibration (see ISO 5349-1);
- whole-body vibration (see ISO 2631-1, ISO 2631-2 and ISO 2631-4);
- low-frequency whole-body vibration in the frequency range from 0,1 Hz to 0,5 Hz (see ISO 2631-1).

Vibration instruments can be designed for measurement according to one or more of the frequency weightings defined within each of these applications.

Three levels of performance testing are defined in this document:

- a) pattern evaluation or validation:
 - 1) pattern evaluation, i.e. a full test of the instrument against the specifications defined in this document;
 - 2) validation of one-off instruments, i.e. a limited set of tests of an individual vibration measuring system against the relevant specifications defined in this document;
- b) periodic verification, i.e. an intermediate set of tests designed to ensure that an instrument remains within the required performance specification;
- c) *in situ* checks, i.e. a minimum level of testing required to indicate that an instrument is likely to be functioning within the required performance specification.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2631-1, *Mechanical vibration and shock — Evaluation of human exposure to whole-body vibration — Part 1: General requirements*

ISO 2631-2, *Mechanical vibration and shock — Evaluation of human exposure to whole-body vibration — Part 2: Vibration in buildings (1 Hz to 80 Hz)*