



Designation: D4765–98 Designation: D 4765 – 03 (Reapproved 2008)

Standard Test Method for Fluorides in Workplace Atmospheres¹

This standard is issued under the fixed designation D 4765; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the simultaneous collection and separate measurements of gaseous and particulate fluoride found in certain industrial workplaces. The gaseous inorganic fluorides collected are reported in terms of fluoride; the procedure is not applicable to the collection or analysis of other fluoride-bearing gases (for example, fluorocarbon or fluorosulfur compounds). This test method covers sample collection, preparation, and fluoride measurement.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

D 1193 Specification for Reagent Water

D 1356 Terminology Relating to Sampling and Analysis of Atmospheres

D 1357 Practice for Planning the Sampling of the Ambient Atmosphere

D 5337 Practice for Flow Rate for Calibration of Personal Sampling Pumps

3. Terminology

3.1 *Definitions*— For definitions of terms used in this test method, refer to Terminology D 1356.

4. Summary of Test Method

4.1 Particulate material from a measured volume of air is collected by means of a membrane filter. Gaseous fluoride, from the sample of air, is absorbed by an alkali-impregnated cellulose pad placed immediately behind the membrane filter. The membrane filter and collected solids are made alkaline, ashed, and the residue fused with additional alkali. Finally, the fluoride is determined in a solution of the melt by use of a fluoride ion-selective electrode. Gaseous fluoride is determined in an aqueous extract of the cellulose pad, also by means of the fluoride ion-selective electrode.

5. Significance and Use

5.1 The capability of this test method to collect and quantitate both particulate and gaseous fluorides over the ranges normally encountered in industrial atmospheres makes it applicable for industrial hygiene evaluation and control purposes. The recommended range of this test method is from 0.005 to 5 mg F⁻/m³ air.

6. Interferences

6.1 Because an ion-selective electrode responds to ionic activity, insoluble and complex forms of fluoride must be released by appropriate combinations of fusion, adjustment of pH, and addition of complexing agents.

6.2 Acidity (pH) and ionic strengths of fluoride standard solutions must be matched to those of samples.

6.3 Temperature of sample and standard solutions must be controlled within $\pm 2^{\circ}\text{C}$.

¹ This test method is under the jurisdiction of ASTM Committee D-22 on Sampling and Analysis of Atmospheres and is the direct responsibility of Subcommittee D22.04 on Workplace Atmospheres.

Current edition approved May 10, 1998. Published July 1998. Originally published as D4765–88. Last previous edition D4765–93.

¹ This test method is under the jurisdiction of ASTM Committee D22 on Air Quality and is the direct responsibility of Subcommittee D22.04 on Workplace Air Quality

² Current edition approved April 1, 2008. Published July 2008. Originally approved in 1988. Last previous edition approved in 2003 as D 4765 – 03.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

7. Apparatus

7.1 *Personal Sampling Pump*, Equipped with a flow-monitoring device (rotameter, critical orifice) or a constant-flow device capable of drawing 2 L/min of air through the 0.8- μm membrane filter and pad for a period of 8 h.

7.2 *Filter Holder*— Plastic holders of the preloaded personal monitor type, that accept filters of 37-mm diameter, are preferred. The holder is to be numbered for identification.

7.3 *Membrane Filter*, of mixed-cellulose esters, 0.8- μm pore size, and of diameter to fit the filter holder (see 7.2).

7.4 *Cellulose Pad*, of size to fit the filter holder (see 7.2). The pad is commercially available as a plain, unregrenated pad or as an alkali-impregnated pad.

7.5 *Crucibles*, 20-mL, nickel, Inconel,³ or platinum.

7.6 *Fluoride Ion-Selective Electrode* .

7.7 *Reference Electrode*, calomel type, preferably combined with the fluoride ion-selective electrode.

7.8 *Electrometer or Expanded Scale pH Meter*, with a millivolt scale for measurement of potentials.

7.9 *Magnetic Stirrer*.

7.10 *Plastic Beakers*, 50 and 100-mL capacities.

7.11 *Beakers*, 250-mL capacity.

7.12 *Volumetric Flasks*, 100-mL capacity.

8. Reagents

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁴

8.2 *Purity of Water*— Unless otherwise indicated, references to water shall be understood to mean Type I Reagent Water conforming to Specification D 1193.

8.3 *Alkaline Fixative Solution* —Dissolve 25 g of sodium carbonate (Na_2CO_3) in water, add 20 mL glycerol, and dilute to 1 L.

8.4 *Buffer Solution (ALCOA)*—Dissolve 60 g of citric acid monohydrate ($\text{C}_2\text{H}_8\text{O}_7 \cdot \text{H}_2\text{O}$), 210 g of sodium citrate ($\text{Na}_3\text{C}_6\text{H}_5\text{O}_7 \cdot 2\text{H}_2\text{O}$) and 53.5 g of ammonium chloride (NH_4Cl) in 500 mL water. Add 67 mL of ammonium hydroxide (NH_4OH) (sp gr = 0.90) and dilute to 1 L with water.

8.5 *Fluoride Solution, Standard (100 $\mu\text{g}/\text{mL}$)*—Dissolve 0.2211 g sodium fluoride (NaF , dried at 105°C for 2 h) in water and dilute to volume in a 1-L volumetric flask.

8.6 *Ethanol, Denatured*—Formula 30 denatured alcohol is satisfactory.

8.7 *Borate-Carbonate Fusion Mixture* —Intimately mix a 1+2 (w/w) combination of sodium tetraborate ($\text{Na}_2\text{B}_4\text{O}_7$) and sodium carbonate (Na_2CO_3).

8.8 *Hydrochloric Acid (1+1)*—Mix one part hydrochloric acid to one part distilled water as a solution.

9. Sampling

9.1 For general information on sampling, refer to Practice D 1357.

9.2 Equip the worker whose exposure is to be evaluated with a personal monitor connected by a 75-cm length of hose to a belt-supported sampling pump. Attach the monitor to the worker's collar and remove the plug for closed-face sampling. Air is drawn through the filter at the calibrated rate of approximately 2.0 L/min and maintained at that rate by occasional checking and adjustment. On termination of sampling, note the duration of sampling, reseal the monitor and return the monitor to the laboratory. Filter a minimum air sample of 250 L.

9.3 Total particulate loading may be determined, if required, by pre- and post-weighing of the membrane filter.

10. Preparation of Monitors

10.1 Disassemble the personal monitor (see 7.2), removing the membrane filter and cellulose pads. Moisten the pad with a measured volume of alkaline fixative solution (see 8.3); 0.8 mL is required for a pad of 37-mm diameter. Dry the pad at 105°C for 30 to 45 min.

NOTE 1—Preparation of alkali-impregnated pads must be carried out in a low-fluoride environment with minimum exposure.

10.2 Reassemble the filter monitor, inserting an impregnated pad and membrane filter, and closing with the filter retaining ring and front cover. Seal the assembly against air leakage by a wrap of masking tape or cellulose shrink bands, covering the crevice between the retaining ring and back cover. Close the inlet and outlet openings of the monitor with plastic plugs.

² Annual Book of ASTM Standards, Vol 11.05.

³ Inconel is a trademark for a group of corrosion-resistant alloys of nickel and chromium.

⁴ Inconel is a trademark for a group of corrosion-resistant alloys of nickel and chromium.

⁴ *Reagent Chemicals, American Chemical Society Specifications* , American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.