

# SLOVENSKI STANDARD

## SIST EN ISO 8536-4:2020

01-marec-2020

Nadomešča:

SIST EN ISO 8536-4:2013

SIST EN ISO 8536-4:2013/A1:2013

---

**Infuzijska oprema za uporabo v medicini - 4. del: Infuzijski seti za enkratno uporabo, delujoči na osnovi gravitacije (ISO 8536-4:2019)**

Infusion equipment for medical use - Part 4: Infusion sets for single use, gravity feed (ISO 8536-4:2019)

**iTeh STANDARD PREVIEW**

Infusionsgeräte zur medizinischen Verwendung - Teil 4: Infusionsgeräte für Schwerkraftinfusionen zur einmaligen Verwendung (ISO 8536-4:2019)

SIST EN ISO 8536-4:2020

Matériel de perfusion à usage médical - Partie 4: Appareils de perfusion non réutilisables, à alimentation par gravité (ISO 8536-4:2019)

**Ta slovenski standard je istoveten z: EN ISO 8536-4:2020**

---

**ICS:**

11.040.20	Transfuzijska, infuzijska in injekcijska oprema	Transfusion, infusion and injection equipment
-----------	---	---

**SIST EN ISO 8536-4:2020**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-4:2020](https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020)

<https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 8536-4**

January 2020

ICS 11.040.20

Supersedes EN ISO 8536-4:2013

English Version

## Infusion equipment for medical use - Part 4: Infusion sets for single use, gravity feed (ISO 8536-4:2019)

Matériel de perfusion à usage médical - Partie 4:  
Appareils de perfusion non réutilisables, à  
alimentation par gravité (ISO 8536-4:2019)

Infusionsgeräte zur medizinischen Verwendung - Teil  
4: Infusionsgeräte für Schwerkraftinfusionen zur  
einmaligen Verwendung (ISO 8536-4:2019)

This European Standard was approved by CEN on 17 October 2019.

This European Standard was corrected and reissued by the CEN-CENELEC Management Centre on 04 March 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

**iTeh STANDARD PREVIEW**

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

Contents	Page
European foreword.....	3

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-4:2020](https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020)

<https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020>

## European foreword

This document (EN ISO 8536-4:2020) has been prepared by Technical Committee ISO/TC 76 "Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2020, and conflicting national standards shall be withdrawn at the latest by July 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8536-4:2013.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN ISO 8536-4:2020](https://standards.iteh.ai/c/en/iso/8536-4/2020)

<https://standards.iteh.ai/c/en/iso/8536-4/2020>  
**Endorsement notice**  
[b4708056be5a/sist-en-iso-8536-4-2020](https://standards.iteh.ai/c/en/iso/8536-4/2020)

The text of ISO 8536-4:2019 has been approved by CEN as EN ISO 8536-4:2020 without any modification.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-4:2020](https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020)

<https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020>

INTERNATIONAL  
STANDARD

ISO  
8536-4

Sixth edition  
2019-09

---

---

**Infusion equipment for medical use —  
Part 4:  
Infusion sets for single use, gravity feed**

*Matériel de perfusion à usage médical —*

*Partie 4: Appareils de perfusion non réutilisables, à alimentation  
par gravité*

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 8536-4:2020](https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020)

[https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-  
b4708056be5a/sist-en-iso-8536-4-2020](https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020)



Reference number  
ISO 8536-4:2019(E)

© ISO 2019

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 8536-4:2020

<https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland



<b>Contents</b>	<b>Page</b>
<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 General requirements</b> .....	<b>1</b>
<b>5 Designation</b> .....	<b>4</b>
<b>6 Materials</b> .....	<b>4</b>
<b>7 Physical requirements</b> .....	<b>4</b>
7.1 Particulate contamination.....	4
7.2 Leakage.....	4
7.3 Tensile strength.....	4
7.4 Closure-piercing device.....	5
7.5 Air-inlet device.....	5
7.6 Tubing.....	5
7.7 Fluid filter.....	5
7.8 Drip chamber and drip tube.....	6
7.9 Flow regulator.....	6
7.10 Flow rate of infusion set.....	6
7.11 Injection site.....	6
7.12 Male conical fitting.....	6
7.13 Protective caps.....	6
<b>8 Chemical requirements</b> .....	<b>6</b>
8.1 Reducing (oxidizable) matter.....	6
8.2 Metal ions.....	6
8.3 Titration acidity or alkalinity.....	6
8.4 Residue on evaporation.....	7
8.5 UV absorption of extract solution.....	7
<b>9 Biological requirements</b> .....	<b>7</b>
9.1 General.....	7
9.2 Sterility.....	7
9.3 Pyrogenicity.....	7
9.4 Haemolysis.....	7
9.5 Toxicity.....	7
<b>10 Labelling</b> .....	<b>7</b>
10.1 General.....	7
10.2 Unit container.....	7
10.3 Shelf or multi-unit container.....	8
<b>11 Packaging</b> .....	<b>8</b>
<b>12 Disposal</b> .....	<b>9</b>
<b>Annex A (normative) Physical tests</b> .....	<b>10</b>
<b>Annex B (normative) Chemical tests</b> .....	<b>15</b>
<b>Annex C (normative) Biological tests</b> .....	<b>17</b>
<b>Bibliography</b> .....	<b>18</b>

## ISO 8536-4:2019(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

This sixth edition cancels and replaces the fifth edition (ISO 8536-4:2010), which has been technically revised. It also incorporates the Amendment ISO 8536-4:2010/Amd.1:2013.

The main changes compared to the previous edition are as follows:

- [Clause 5](#) 'Designation' now refers to [Clause 10](#) 'Labelling';
- the physical requirements – especially regarding stand-alone air-inlet devices – have been further clarified;
- [Clause 10](#) 'Labelling' has been updated;
- test for leakage in [A.3](#) has been updated;
- determination of flow rate in [A.5](#) has been totally reviewed;
- normative references in [Clause 2](#) and the Bibliography have been updated.

A list of all parts in the ISO 8536 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# Infusion equipment for medical use —

## Part 4: Infusion sets for single use, gravity feed

### 1 Scope

This document specifies requirements for single use, gravity feed infusion sets for medical use in order to ensure their compatibility with containers for infusion solutions and intravenous equipment.

Secondary aims of this document are to provide guidance on specifications relating to the quality and performance of materials used in infusion sets and to present designations for infusion set components.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 7864, *Sterile hypodermic needles for single use — Requirements and test methods*

ISO 8536-13, *Infusion equipment for medical use — Part 13: Graduated flow regulators for single use with fluid contact*

<https://standards.iteh.ai/catalog/standards/sist/37701ac7-368a-4680-bb8c-b4708056be5a/sist-en-iso-8536-4-2020>

ISO 8536-14, *Infusion equipment for medical use — Part 14: Clamps and flow regulators for transfusion and infusion equipment without fluid contact*

ISO 14644-1, *Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness by particle concentration*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

### 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 4 General requirements

**4.1** The nomenclature to be used for components of infusion sets and of a stand-alone air-inlet device is given in [Figures 1, 2](#) and [3](#). These figures illustrate examples of the configuration of infusion sets and air-inlet devices; other configurations may be used provided they lead to the same results. Infusion sets