



SLOVENSKI STANDARD
oSIST prEN ISO 10650:2017
01-junij-2017

Zobozdravstvo - Polimerizacijski aktivatorji (ISO/DIS 10650:2017)

Dentistry - Powered polymerization activators (ISO/DIS 10650:2017)

Zahnheilkunde - Polymerisationslampen (ISO/DIS 10650:2017)

Médecine bucco-dentaire - Activateurs électriques de polymérisation (ISO/DIS 10650:2017)

Ta slovenski standard je istoveten z: prEN ISO 10650

ICS:

11.060.20

Zobotehnična oprema

Dental equipment

oSIST prEN ISO 10650:2017

en

DRAFT INTERNATIONAL STANDARD

ISO/DIS 10650

ISO/TC 106/SC 6

Secretariat: DIN

Voting begins on:
2017-04-17Voting terminates on:
2017-07-09

Dentistry — Powered polymerization activators

Médecine bucco-dentaire — Activeurs électriques de polymérisation

ICS: 11.060.20

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[SIST EN ISO 10650:2018](https://standards.iteh.ai/catalog/standards/sist/1e090818-3feb-4cb0-a22a-fce23b4b46ff/sist-en-iso-10650-2018)<https://standards.iteh.ai/catalog/standards/sist/1e090818-3feb-4cb0-a22a-fce23b4b46ff/sist-en-iso-10650-2018>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 10650:2017(E)

© ISO 2017

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

SIST EN ISO 10650:2018

<https://standards.iteh.ai/catalog/standards/sist/1e090818-3feb-4cb0-a22a-fce23b4b46ff/sist-en-iso-10650-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification	2
5 Requirements	2
5.1 General	2
5.1.1 Design	2
5.1.2 Connection	2
5.1.3 Operating controls	2
5.1.4 Cleaning, disinfection and sterilization	2
5.1.5 Excessive temperatures	2
5.2 Irradiance	3
5.2.1 Irradiance in the 385 nm to 515 nm wavelength region	3
5.2.2 Irradiance in the wavelength region below 385 nm	3
5.2.3 Irradiance in the wavelength region above 515 nm	3
5.3 Radiant exposure (dose)	3
5.4 Electrical safety requirements	3
5.5 Usability	3
6 Sampling	4
7 Measurement and test methods	4
7.1 General	4
7.1.1 General provisions for tests	4
7.1.2 Atmospheric conditions	4
7.2 Measurement of the optical cross-sectional area of the light guide	4
7.2.1 Apparatus	4
7.2.2 Procedure	4
7.3 Measurement of the irradiance	4
7.3.1 Apparatus	4
7.3.2 Procedure	5
7.3.2 Treatment of results	7
8 Information to be supplied by the manufacturer	7
8.1 Instructions for use	7
8.2 Technical description	8
9 Marking	8
10 Packaging	9
Bibliography	10

ISO/DIS 10650:2017(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee 6, *Dental equipment*.

This second edition of ISO 10650 cancels and replaces the first edition ISO 10650:2015, which has been technically revised with the following changes:

- a) The test procedure using filters was deleted.
- b) A test procedure using spectrophotometer was included.
- c) The term "radiant exitance" was changed to "irradiance".
- d) An upper limit to the irradiance for the 385 nm to 515 nm wavelength region was added.
- e) A requirement for the setting delivering a radiant exposure (dose) was added.