

SLOVENSKI STANDARD
SIST EN IEC 60601-2-49:2019**01-december-2019****Nadomešča:****SIST EN 60601-2-49:2015**

Medicinska električna oprema - 2-49. del: Posebne zahteve za osnovno varnost in bistvene lastnosti večfunkcijske opreme za nadzor pacientov (IEC 80601-2-49:2018)

Medical electrical equipment - Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitoring equipment (IEC 80601-2-49:2018)

iTeh STANDARD PREVIEW

Medizinische elektrische Geräte (Teil 2-49: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von multifunktionalen Patientenüberwachungsgeräten (IEC 80601-2-49:2018))⁹

<https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019>

Appareils électromédicaux - Partie 2-49: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de surveillance multifonction des patients (IEC 80601-2-49:2018)

Ta slovenski standard je istoveten z: EN IEC 80601-2-49:2019

ICS:

11.040.55 Diagnostična oprema Diagnostic equipment

SIST EN IEC 60601-2-49:2019 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN IEC 60601-2-49:2019

<https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019>

EUROPEAN STANDARD

EN IEC 80601-2-49

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2019

ICS 11.040.55

Supersedes EN 60601-2-49:2015 and all of its
amendments and corrigenda (if any)

English Version

Medical electrical equipment - Part 2-49: Particular requirements
for the basic safety and essential performance of multifunction
patient monitoring equipment
(IEC 80601-2-49:2018)

Appareils électromédicaux - Partie 2-49: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils de surveillance multifonction des
patients
(IEC 80601-2-49:2018)

Medizinische elektrische Geräte - Teil 2-49: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von multifunktionalen
Patientenüberwachungsgeräten
(IEC 80601-2-49:2018)

This European Standard was approved by CENELEC on 2019-08-07. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 80601-2-49:2019 (E)**European foreword**

The text of document 62D/1547/FDIS, future edition 1.0 of IEC 80601-2-49, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 80601-2-49:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-05-07
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-08-07

This document supersedes EN 60601-2-49:2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

iTeh STANDARD PREVIEW (standards.iteh.ai)

Endorsement notice

[SIST EN IEC 60601-2-49:2019](https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019)

[https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-](https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019)

[5353f27693e9/sist-en-iec-60601-2-49-2019](https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019)

The text of the International Standard IEC 80601-2-49:2018 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-16	NOTE	Harmonized as EN IEC 60601-2-16
ISO 80601-2-13	NOTE	Harmonized as EN ISO 80601-2-13
ISO 80601-2-56	NOTE	Harmonized as EN ISO 80601-2-56
ISO 80601-2-72	NOTE	Harmonized as EN ISO 80601-2-72
IEC 62366-1:2015	NOTE	Harmonized as EN 62366-1:2015 (not modified)

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

The Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement</i> IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2:EN 60601-1-2 General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests		2015
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6:EN 60601-1-6 General requirements for basic safety and essential performance - Collateral standard: Usability		2010
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8:EN 60601-1-8 General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems		2007
			+EN 60601-1-2010 8:2007/corrigendum Mar. 2010	
IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	+A11 by EN 60529	2017 1991
			+EN 60529:1991/corrigendum May 1993	1993

EN IEC 80601-2-49:2019 (E)

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Addition</i> IEC 60601-1	2005	Medical electrical equipment - Part 1:EN 60601-1 General requirements for basic safety and essential performance		2006
			+A12	2014
			+EN 60601-2010 1:2006/corrigendum Mar. 2010	
			+AC	2014
			+A11	2011
IEC 60601-1-11	2015	Medical electrical equipment – Part 1-11:- General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment		-
IEC 60601-1-12	2014	Medical electrical equipment - Part 1-12:- General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment		-
IEC 60601-2-2	2017	Medical electrical equipment - Part 2-2:EN IEC 60601-2-2 Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories		2018
IEC 60601-2-27	2011	Medical electrical equipment - Part 2-27:EN 60601-2-27 Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment		2014
IEC 60601-2-34	2011	Medical electrical equipment - Part 2-34:EN 60601-2-34 Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment		2014



IEC 80601-2-49

Edition 1.0 2018-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
Part 2-49: Particular requirements for the basic safety and essential performance
of multifunction patient monitors

Appareils électromédicaux –
Partie 2-49: Exigences particulières pour la sécurité de base et les performances
essentiels des moniteurs multifonctions des patients

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.55

ISBN 978-2-8322-5359-5

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	11
201.4 General requirements	11
201.5 General requirements for testing ME EQUIPMENT	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7 ME EQUIPMENT identification, marking and documents	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	14
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	16
201.10 Protection against unwanted and excessive radiation HAZARDS	16
201.11 Protection against excessive temperatures and other HAZARDS	16
201.12 Accuracy of controls and instruments and protection against hazardous outputs	17
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	18
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	18
201.15 Construction of ME EQUIPMENT	18
201.16 ME SYSTEMS	19
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	19
202 Electromagnetic disturbances – Requirements and tests.....	19
206 USABILITY	24
208 General requirements, tests and guidance for ALARM SYSTEMS IN MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.....	25
Annexes	28
Annex AA (informative) Particular guidance and rationale.....	29
Bibliography.....	38
Index of defined terms used in this particular standard.....	39
Figure 201.101 – MULTIFUNCTION PATIENT MONITOR with single PATIENT circuit (6) with multiple PHYSIOLOGICAL MONITORING UNITS and multiple PATIENT circuits (7) each with a single PHYSIOLOGICAL MONITORING UNIT	15
Figure 202.101 – Test layout for conducted and radiated EMISSIONS and IMMUNITY test	20
Figure 202.102 – Test circuit for HF SURGICAL EQUIPMENT protection measurement according to 202.8.102 with PATIENT CONNECTIONS	22
Figure 202.103 – Test setup for HF SURGICAL EQUIPMENT protection measurement according to 202.8.102	23
Figure 202.104 – Test circuit for HF SURGICAL EQUIPMENT protection measurement according to 202.8.102 with non-conductive APPLIED PART.....	24
Figure AA.1 – Example of a pre-configured MULTIFUNCTION PATIENT MONITOR	30
Figure AA.2 – Example of a modular MULTIFUNCTION PATIENT MONITOR	30
Figure AA.3 – Example of a MULTIFUNCTION PATIENT MONITOR connected to a central station	30

Figure AA.4 – Example of a MULTIFUNCTION PATIENT MONITOR integrated into a ventilator 31

Figure AA.5 – Single PATIENT circuit with multiple PHYSIOLOGICAL MONITORING UNITS and PATIENT cables 33

Table 201.101 – ESSENTIAL PERFORMANCE requirements..... 12

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN IEC 60601-2-49:2019](https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019)

<https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitors**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-49 has been prepared by a Joint Working Group of IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and of ISO subcommittee SC3: Lung ventilators and related equipment, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This first edition cancels and replaces the second edition of IEC 60601-2-49, published in 2011. This edition constitutes a technical revision to align with the current edition and Amendment to IEC 60601-1, new versions of collateral standards and amendments thereto. Major changes are in Clause 208 because many of the former requirements are now addressed by IEC 60601-1-8.

It is published as a double logo standard.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/1547/FDIS	62D/1559/RVD

Full information on the voting for the approval of this document can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by XXX P members out of YYY having cast a vote.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

<https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019>

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the 80601 International Standard, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN IEC 60601-2-49:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019>

INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of MULTIFUNCTION PATIENT MONITORS. It amends and supplements IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

The aim of this edition is to bring this particular standard up to date with reference to the edition 3.1 of the general standard and new versions of collateral standards and amendments thereto through technical changes.

The requirements of this particular standard take priority over those of the general standard.

A "Particular guidance and rationale" for the requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this Annex AA does not form part of the requirements of this document.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN IEC 60601-2-49:2019](https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019)

<https://standards.iteh.ai/catalog/standards/sist/bc8e5538-30fd-4cb3-9758-5353f27693e9/sist-en-iec-60601-2-49-2019>