



**SLOVENSKI STANDARD**  
**oSIST prEN ISO 7405:2017**  
**01-maj-2017**

---

**Zobozdravstvo - Ovrednotenje biokompatibilnosti medicinskih pripomočkov v zobozdravstvu (ISO/DIS 7405:2017)**

Dentistry - Evaluation of biocompatibility of medical devices used in dentistry (ISO/DIS 7405:2017)

Zahnheilkunde - Beurteilung der Biokompatibilität von in der Zahnheilkunde verwendeten Medizinprodukten (ISO/DIS 7405:2017)

Médecine bucco-dentaire - Évaluation de la biocompatibilité des dispositifs médicaux utilisés en médecine bucco-dentaire (ISO/DIS 7405:2017)

**Ta slovenski standard je istoveten z: prEN ISO 7405**

<https://standards.iteh.ai/catalog/standards/sist/cf4b08e7-2ede-4a94-8d00-8a96dd3ea14e/sist-en-iso-7405-2017>

---

**ICS:**

11.060.01	Zobozdravstvo na splošno	Dentistry in general
11.100.20	Biološko ovrednotenje medicinskih pripomočkov	Biological evaluation of medical devices

**oSIST prEN ISO 7405:2017**

**en**



# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 7405

ISO/TC 106

Secretariat: SCC

Voting begins on:  
2017-03-07Voting terminates on:  
2017-05-29

---

---

## Dentistry — Evaluation of biocompatibility of medical devices used in dentistry

*Médecine bucco-dentaire — Évaluation de la biocompatibilité des dispositifs médicaux utilisés en médecine bucco-dentaire*

ICS: 11.060.10; 11.100.99

### iTeh Standards (<https://standards.iteh.ai>) Document Preview

SIST EN ISO 7405:2019<https://standards.iteh.ai/catalog/standards/sist/cf4b08e7-2ede-4a94-8d00-8a96dd3ea14e/sist-en-iso-7405-2019>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

**ISO/CEN PARALLEL PROCESSING**



Reference number  
ISO/DIS 7405:2017(E)

© ISO 2017

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

SIST EN ISO 7405:2019

<https://standards.iteh.ai/catalog/standards/sist/cf4b08e7-2ede-4a94-8d00-8a96dd3ea14e/sist-en-iso-7405-2019>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

	Page
<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Categorization of medical devices</b> .....	<b>3</b>
4.1 Categorization by nature of contact.....	3
4.1.1 General.....	3
4.1.2 Non-contact devices.....	3
4.1.3 Surface-contacting devices.....	3
4.1.4 External communicating devices.....	3
4.1.5 Implant devices used in dentistry.....	3
4.2 Categorization by duration of contact.....	4
4.2.1 General.....	4
4.2.2 Limited exposure devices.....	4
4.2.3 Prolonged exposure devices.....	4
4.2.4 Permanent contact devices.....	4
<b>5 Biological evaluation process</b> .....	<b>4</b>
5.1 General.....	4
5.2 Selection of tests and overall assessment.....	5
5.3 Selection of test methods.....	5
5.4 Types of test.....	5
5.5 Re-evaluation of biocompatibility.....	6
<b>6 Test procedures specific to dental materials</b> .....	<b>6</b>
6.1 Recommendations for sample preparation.....	6
6.1.1 General.....	6
6.1.2 General recommendations for sample preparation.....	7
6.1.3 Specific recommendations for light curing materials.....	7
6.1.4 Specific recommendations for chemically setting materials.....	8
6.1.5 Positive control material.....	8
6.2 Agar diffusion test.....	8
6.2.1 Objective.....	8
6.2.2 Cell line.....	8
6.2.3 Culture medium, reagents and equipment.....	8
6.2.4 Sample preparation.....	9
6.2.5 Controls.....	9
6.2.6 Test procedure.....	9
6.2.7 Parameters of assessment.....	10
6.2.8 Assessment of results.....	10
6.2.9 Test report.....	11
6.3 Filter diffusion test.....	11
6.3.1 Objective.....	11
6.3.2 Cell line.....	11
6.3.3 Culture medium, reagents and equipment.....	11
6.3.4 Sample preparation.....	12
6.3.5 Controls.....	12
6.3.6 Test procedure.....	12
6.3.7 Assessment of cell damage.....	13
6.3.8 Assessment of results.....	13
6.3.9 Test report.....	13
6.4 Pulp and dentine usage test.....	14
6.4.1 Objective.....	14

## ISO/DIS 7405:2017(E)

6.4.2	Animals and animal welfare.....	14
6.4.3	Test procedure.....	14
6.4.4	Assessment of results.....	19
6.4.5	Test report.....	19
6.5	Pulp capping test.....	19
6.5.1	Objective.....	19
6.5.2	Animals and animal welfare.....	19
6.5.3	Test procedure.....	20
6.5.4	Assessment of results.....	22
6.5.5	Test report.....	22
6.6	Endodontic usage test.....	22
6.6.1	Objective.....	22
6.6.2	Animals and animal welfare.....	22
6.6.3	Test procedure.....	22
6.6.4	Assessment of results.....	24
6.6.5	Test report.....	25
<b>Annex A (informative) Types of test to be considered for evaluation of biocompatibility of medical devices used in dentistry.....</b>		<b>26</b>
<b>Annex B (informative) Dentine barrier cytotoxicity test.....</b>		<b>29</b>
<b>Annex C (informative) Endosseous dental implant usage test.....</b>		<b>37</b>
<b>Bibliography.....</b>		<b>41</b>

iTeh Standards  
(<https://standards.itih.ai>)  
Document Preview

SIST EN ISO 7405:2019

<https://standards.itih.ai/catalog/standards/sist/cf4b08e7-2ede-4a94-8d00-8a96dd3ea14e/sist-en-iso-7405-2019>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

The committee responsible for this document is ISO/TC 106, *Dentistry*.

This third edition cancels and replaces the second edition (ISO 7405:2008) which has been technically revised. The following changes have been made:

- a) inclusion of Amendment 1:2013;
- b) deletion of [Annex C](#) (Acute toxicity testing);
- c) addition of ISO/TS 22911 as new [Annex C](#).

**ISO/DIS 7405:2017(E)****Introduction**

This document describes the evaluation of the biocompatibility of medical devices used in dentistry. It is to be used in conjunction with the ISO 10993- series of standards. This document contains special tests, for which ample experience exists in dentistry and which acknowledge the special needs of dentistry.

Only test methods for which the members of the committee considered there was sufficient published data have been included. In recommending test methods, the need to minimize the use of animals was given a high priority. It is essential that the decision to undertake tests involving animals be reached only after a full and careful review of the evidence indicating that a similar outcome cannot be achieved by other types of test. In order to keep the number of animals required for tests to an absolute minimum, consistent with achieving the objective indicated, it can be appropriate to conduct more than one type of test on the same animal at the same time, e.g. pulp and dentin usage test and pulp capping test. However, in accordance with ISO 10993-2 these tests are performed both in an efficient and humane way. On all occasions when animal testing is undertaken, such tests are conducted empathetically and according to standardized procedures as described for each test.

This document does not explicitly describe test methods for occupationally related risks.

[Annex B](#) is included to encourage the development of *in vitro* and *ex vivo* test methods which will further reduce the use of animals in the evaluation of the biocompatibility of medical devices used in dentistry. [Annex C](#) is based on and will replace the former ISO/TS 22911.

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[SIST EN ISO 7405:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/cf4b08e7-2ede-4a94-8d00-8a96dd3ea14e/sist-en-iso-7405-2019>



# Dentistry — Evaluation of biocompatibility of medical devices used in dentistry

## 1 Scope

This document specifies test methods for the evaluation of biological effects of medical devices used in dentistry. It includes testing of pharmacological agents that are an integral part of the device under test.

This document does not cover testing of materials and devices that do not come into direct or indirect contact with the patient's body.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6344-1, *Coated abrasives — Grain size analysis — Part 1: Grain size distribution test*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-2, *Biological evaluation of medical devices — Part 2: Animal welfare requirements*

ISO 10993-3, *Biological evaluation of medical devices — Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*

ISO 10993-5, *Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity*

ISO 10993-6, *Biological evaluation of medical devices — Part 6: Tests for local effects after implantation*

ISO 10993-10, *Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

ISO 10993-12:2012, *Biological evaluation of medical devices — Part 12: Sample preparation and reference materials*

ISO 10993-18, *Biological evaluation of medical devices — Part 18: Chemical characterization of materials*

ISO 10993-19, *Biological evaluation of medical devices — Part 19: Physico-chemical, morphological and topographical characterization of materials*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO/TR 15499, *Biological evaluation of medical devices — Guidance on the conduct of biological evaluation within a risk management process*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942, ISO 10993-1, ISO 10993-12, ISO 16443 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at <http://www.electropedia.org/>

**ISO/DIS 7405:2017(E)**

— ISO Online browsing platform: available at <http://www.iso.org/obp>

**3.1  
medical device**  
any instrument, apparatus, appliance, software, material or other article, whether used alone or in combination, together with any accessories, including the software necessary for its proper application intended by the manufacturer to be used for medical purposes for human beings for the purpose of:

- diagnosis, prevention, monitoring, treatment or alleviation of disease;
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury or handicap;
- investigation, replacement or modification of the anatomy or of a physiological process;
- control of conception;

and which does not achieve its principal intended action in or on the human body by pharmacological, immunological or metabolic means, but which may be assisted in its function by such means

**3.2  
dental material**  
material and/or substance or combination of materials and/or substances specially formulated and prepared for use in the practice of dentistry and/or associated procedures

**3.3  
final product**  
medical device in its “as-used” state, as defined by the manufacturer’s specification or labelling

Note 1 to entry: Many dental materials are used in a freshly mixed state, and evaluation of the materials in both freshly mixed and set conditions should be considered.

**3.4  
positive control  
positive control material**  
any well characterized material and/or substance that, when evaluated by a specific test method, demonstrates the suitability of the test system to yield a reproducible, appropriately positive or reactive response in the test system

**3.5  
negative control  
negative control material**  
any well characterized material and/or substance that, when evaluated by a specific test method, demonstrates the suitability of the test system to yield a reproducible, appropriately negative, non-reactive or minimal response in the test system

Note 1 to entry: In practice, negative controls include blanks, vehicles/solvents and reference materials.

**3.6  
reference material**  
material with one or more property values that are sufficiently reproducible and well established to enable use of the material or substance for the calibration of an apparatus, the assessment of a measurement method or for the assignment of values to materials

Note 1 to entry: For the purpose of this document, a reference material is any well characterized material and/or substance that, when tested by the procedure described, demonstrates the suitability of the procedure to yield a reproducible, predictable response. The response may be negative or positive.

**3.7  
in vitro pulp chamber**  
device that holds a thin slice of dentine between two chambers and allows fluid and molecules to filter or to diffuse across the “dentine barrier”

### 3.8

#### **hydraulic conductance**

active movement or filtration of fluid and its constituents through a barrier such as dentin by means of a pressure gradient

### 3.9

#### **diffusion**

establishment of passive movement of solutes (solubilized constituents) by means of a diffusion gradient through the “dentin barrier”

## 4 Categorization of medical devices

### 4.1 Categorization by nature of contact

#### 4.1.1 General

For the purposes of this document, the classification of medical devices used in dentistry is derived from ISO 10993-1. If a device or material can be placed in more than one category, the more rigorous testing requirements shall apply. With multiple exposures the decision into which category a device is placed shall take into account the potential cumulative effect, bearing in mind the period of time over which these exposures occur.

NOTE In this context the term dentistry includes the oromaxillofacial environment.

#### 4.1.2 Non-contact devices

These devices do not contact the patient's body directly or indirectly, and are not included in ISO 10993-1.

#### 4.1.3 Surface-contacting devices

These devices include those that contact the surface of intact or breached or otherwise compromised skin, the surface of intact or breached or otherwise compromised oral mucosa, and those that contact the external surfaces of dental hard tissue, including enamel, dentine and cementum.

NOTE In some circumstances, dentine and cementum are considered as surfaces, e.g. after gingival recession.

#### 4.1.4 External communicating devices

These devices include dental devices that penetrate and are in contact with oral mucosa, dental hard tissues, dental pulp tissue or bone, or any combination of these, and are exposed to the oral environment.

NOTE This group also includes any kind of lining or base material to be used under a restoration.

#### 4.1.5 Implant devices used in dentistry

These devices include dental implants and other dental devices that are partially or fully embedded in one or more of the following:

- a) soft tissue, e.g. subperiosteal implants and subdermal implants;
- b) bone, e.g. endosteal implants and bone substitutes;
- c) pulpodental system of the tooth, e.g. endodontic materials;
- d) any combination of these, e.g. transosteal implants.

## ISO/DIS 7405:2017(E)

### 4.2 Categorization by duration of contact

#### 4.2.1 General

For the purposes of this document, medical devices used in dentistry are classified by duration of contact as described in ISO 10993-1 and listed in [4.2.2](#) to [4.2.4](#).

#### 4.2.2 Limited exposure devices

Devices whose cumulative single or multiple use or contact is likely to be up to 24 h.

#### 4.2.3 Prolonged exposure devices

Devices whose cumulative single, multiple or long-term use or contact is likely to exceed 24 h but not 30 d.

#### 4.2.4 Permanent contact devices

Devices whose cumulative single, multiple or long-term use or contact exceeds 30 d.

NOTE 1 The definition of the term “permanent” is meant to be applied solely for the use of this document. It is consistent with the definition given in ISO 10993-1.

NOTE 2 With multiple exposures to the device, the decision into which category a device is placed should take into account the potential cumulative effect, bearing in mind the period of time over which these exposures occur.

## 5 Biological evaluation process

### 5.1 General

Each medical device used in dentistry shall be subjected to a structured biological evaluation programme within a risk management process (see ISO 10993-1). Guidance on the implementation of this programme is provided in ISO 14971 and ISO 10993-1. Further information can be obtained from ISO/TR 15499.

The biological evaluation programme shall include the review of data sets concerning the biological properties of each medical device used in dentistry. When this part of the biological evaluation programme indicates that one or more data sets are incomplete and that further testing is necessary, the tests should be selected from the methods described in the ISO 10993- series of standards or in this International Standard, or in both. If tests that are not included in these International Standards are selected, a statement shall be made that indicates that the tests described in these International Standards have been considered and shall include a justification for the selection of other tests.

For combination products the final product should be evaluated according to this document in conjunction with any applicable standards.

NOTE 1 In this context, combination products are dental devices of any kind that incorporate, or are intended to incorporate, as an integral part, a substance that:

- a) if used separately, would be a medicine or a biological product;
- b) is liable to affect the patient’s body by an ancillary action.

An example would be a bone filling/augmentation device containing a growth factor (i.e. a biological product).

NOTE 2 For combination products, where the device and pharmacological components are packaged separately, it may be informative to test the device components alone.

All tests shall be conducted according to recognized current/valid best laboratory/quality practices, where applicable.

NOTE 3 Examples of relevant guidance include GLP (Good Laboratory Practice) or ISO/IEC 17025[11].

## 5.2 Selection of tests and overall assessment

The selection of tests and the overall assessment of the results shall be carried out by an expert who has the appropriate chemical, physical and biological data concerning the device and who is aware of the intended conditions of use.

## 5.3 Selection of test methods

The selection of test methods shall be based upon consideration of:

- a) the intended use of the medical device;
- b) the tissue(s) which the medical device may contact;
- c) the duration of the contact.

If a test selected is not included in the International Standards, a justification for the choice of the methods shall be included in the test report for each device. If more than one test method in the same category is recommended by the standards, the selection of one test over the others should be justified.

## 5.4 Types of test

According to the categorization of the device, tests shall be considered for use as summarised in [Table A.1](#). This table indicates which types of test method shall be considered, but not that they are necessarily required to be carried out. A decision not to carry out a type of test identified in [Table A.1](#) shall be justified in the test report on each device. The types of test listed are regarded as a framework for the evaluation of the biocompatibility of medical devices used in dentistry. For most types of test, particular methods are identified, although for some devices it is recognized that alternative methods not included in the International Standards listed may be more appropriate.

For convenience, the types of test have been listed in three groups.

### a) Group I

This group comprises *in vitro* tests of cytotoxicity. General guidance for *in vitro* cytotoxicity tests is presented in ISO 10993-5 and shall be followed. Detailed test protocols for the agar or agarose diffusion and filter diffusion methods, appropriate to dental materials, are included in this document. The *in vitro* cytotoxicity methods include:

- 1) agar diffusion test (see [6.2](#));
- 2) filter diffusion test (see [6.3](#));
- 3) direct contact or extract tests in accordance with ISO 10993-5;
- 4) dentine barrier cytotoxicity test (see [Annex B](#));

NOTE 1 The order of listing does not indicate any preference for one method over another.

NOTE 2 This list does not indicate that all cytotoxicity tests mentioned have to be performed for each medical device under consideration.

NOTE 3 The use of the dentine barrier cytotoxicity test is encouraged and a description of the test is presented in [Annex B](#). References to this test are presented in the Bibliography.

### b) Group II

**ISO/DIS 7405:2017(E)**

This group comprises tests in accordance with the 10993 series of standards and particular tests, where appropriate, are identified:

- 1) acute systemic toxicity — oral application — in accordance with ISO 10993-11;
- 2) acute systemic toxicity — application by inhalation — in accordance with ISO 10993-11;
- 3) subacute and subchronic systemic toxicity — oral application — in accordance with ISO 10993-11;
- 4) skin irritation and intracutaneous reactivity in accordance with ISO 10993-10;
- 5) delayed-type hypersensitivity in accordance with ISO 10993-10;
- 6) genotoxicity in accordance with ISO 10993-3;
- 7) local effects after implantation in accordance with ISO 10993-6.

NOTE 1 In order to allow use of the latest edition of the referenced document only, an undated cross-reference is possible. An indication of the appropriate clause and subclause is only possible for dated references. Therefore, the user of this International Standard is requested to check the referenced documents for the appropriate clause numbers.

NOTE 2 In the evaluation of materials following local implantation involving mineralized tissues in accordance with ISO 10993-6, examination of undemineralized sections, in addition to routine demineralized sections, is recommended.

NOTE 3 If appropriate, the local effects after implantation may be evaluated in accordance with dental implant usage test instead of ISO 10993-6. (See 5.4 c) 4.)

**c) Group III**

This group comprises tests, specific for medical devices used in dentistry, not referred to in the 10993 series of standards:

- 1) pulp and dentine usage test (see 6.4);
- 2) pulp capping test (see 6.5);
- 3) endodontic usage test (see 6.6);
- 4) endosseous dental implant usage test (see Annex C).

NOTE Endosseous dental implant usage test is not required, but if applicable, is recommended.

**5.5 Re-evaluation of biocompatibility**

In accordance with ISO 10993-1, a device shall be considered for re-evaluation of its biocompatibility as described in 5.4 when revisions or modifications to the formula, quality and/or performance specifications are made.

NOTE See also ISO/TR 15499, which provides indications on when to commence a re-evaluation.

**6 Test procedures specific to dental materials****6.1 Recommendations for sample preparation****6.1.1 General**

These recommendations have been designed for *in vitro* testing, but can also be used for other purposes, if suitable.