



Designation: D7266 – 07^{ε1}

Standard Test Method for Analysis of Cyclohexane by Gas Chromatography (External Standard)¹

This standard is issued under the fixed designation D7266; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^{ε1} NOTE—Table 2 was corrected editorially in July 2008.

1. Scope

1.1 This test method covers the determination of the purity of cyclohexane by gas chromatography. Calibration of the gas chromatography system is done by the external standard calibration technique.

1.2 This test method has been found applicable to the measurement of impurities such as those found in Table 1, which are impurities that may be found in cyclohexane. The impurities can be analyzed over a range of 5 to 180 mg/kg by this method, but may be applicable to a wider range.

1.3 The limit of detection is 1 mg/kg.

1.4 In determining the conformance of the test results using this test method to applicable specifications, results shall be rounded off in accordance with the rounding-off method of Practice E29.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Section 7.

2. Referenced Documents

2.1 *ASTM Standards:*²

D3437 Practice for Sampling and Handling Liquid Cyclic Products

D4307 Practice for Preparation of Liquid Blends for Use as Analytical Standards

D4790 Terminology of Aromatic Hydrocarbons and Related Chemicals

D6809 Guide for Quality Control and Quality Assurance Procedures for Aromatic Hydrocarbons and Related Materials

E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

E355 Practice for Gas Chromatography Terms and Relationships

E1510 Practice for Installing Fused Silica Open Tubular Capillary Columns in Gas Chromatographs

2.2 *Other Document:*

OSHA Regulations, 29 CFR, paragraphs 1910.1000 and 1910.1200³

3. Terminology

3.1 See Terminology D4790 for definitions of terms used in this test method.

4. Summary of Test Method

4.1 Cyclohexane is analyzed using a gas chromatograph (GC) equipped with a flame ionization detector (FID). A precisely repeatable volume of the sample to be analyzed is injected onto the gas chromatograph. The peak areas of the impurities are measured and converted to concentrations via an external standard methodology. Purity by GC (the cyclohexane content) is calculated by subtracting the sum of the impurities from 100.00. Individual impurities are reported in mg/kg. The cyclohexane purity is reported in weight percent.

5. Significance and Use

5.1 This test method is suitable for setting specifications on the materials referenced in Table 1 and for use as an internal quality control tool where cyclohexane is produced or is used

¹ This test method is under the jurisdiction of ASTM Committee D16 on Aromatic Hydrocarbons and Related Chemicals and is the direct responsibility of Subcommittee D16.01 on Benzene, Toluene, Xylenes, Cyclohexane and Their Derivatives.

Current edition approved Jan. 1, 2007. Published January 2007. DOI: 10.1520/D7266-07E01.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, http://www.access.gpo.gov.

in a manufacturing process. It may also be used in development or research work involving cyclohexane.

5.2 This test method is useful in determining the purity of cyclohexane with normal impurities present. If extremely high boiling or unusual impurities are present in the cyclohexane, this test method would not necessarily detect them and the purity calculation would be erroneous.

6. Apparatus

6.1 *Gas Chromatograph*—Any instrument having a flame ionization detector that can be operated at the conditions given in **Table 2**. The system should have sufficient sensitivity to obtain a minimum peak height response for 1 mg/kg benzene of twice the height of the signal background noise.

6.2 *Columns*—The choice of column is based on resolution requirements. Any column may be used that is capable of resolving all significant impurities from cyclohexane. The column described in **Table 2** has been used successfully.

6.3 *Recorder*—Electronic integration is required.

TABLE 1 Impurities Known or Suggested to be Present in Commercial Cyclohexane

| | |
|----------------|---|
| C ₄ | |
| (1) | <i>n</i> -butane |
| (2) | isobutene |
| C ₅ | |
| (3) | <i>n</i> -pentane ^A |
| (4) | isopentane ^A |
| (5) | cyclopentane ^A |
| C ₆ | |
| (6) | <i>n</i> -hexane |
| (7) | 2-methylpentane ^A |
| (8) | 3-methylpentane ^A |
| (9) | methylcyclopentane ^A |
| (10) | benzene ^A |
| (11) | cyclohexene ^A |
| (12) | 2,2-dimethylbutane ^A |
| (13) | 2,3-dimethylbutane ^A |
| C ₇ | |
| (14) | 3,3-dimethylpentane |
| (15) | 2,2-dimethylpentane ^A |
| (16) | 2,3-dimethylpentane ^A |
| (17) | 2,4-dimethylpentane ^A |
| (18) | 1,1-dimethylcyclopentane ^A |
| (19) | <i>trans</i> -1,3-dimethylcyclopentane ^A |
| (20) | <i>trans</i> -1,2-dimethylcyclopentane ^A |
| (21) | <i>cis</i> -1,2-dimethylcyclopentane |
| (22) | 2,2-dimethylcyclopentane |
| (23) | 2,4-dimethylcyclopentane |
| (24) | <i>cis</i> -1,3-dimethylcyclopentane ^A |
| (25) | ethylcyclopentane ^A |
| (26) | methylcyclohexane ^A |
| (27) | 3-ethylpentane ^A |
| (28) | 3-methylhexane ^A |
| (29) | 2-methylhexane ^A |
| (30) | <i>n</i> -heptane ^A |
| (31) | toluene ^A |
| C ₈ | |
| (32) | <i>iso</i> -octane ^A |
| (33) | <i>p</i> -xylene ^A |
| C ₉ | |
| (34) | isopropylcyclohexane ^A |

^A These components were used to prepare the standard used in the repeatability program.

TABLE 2 Instrumental Parameters

| | |
|--------------------|---|
| Detector | flame ionization |
| Injection Port | capillary splitter |
| Column A: | |
| Tubing | fused silica |
| Stationary phase | bonded and crosslinked 100 % dimethylpolysiloxane† |
| Film thickness, μm | 0.5 |
| Length, m | 100 |
| Diameter, mm | 0.25 |
| Temperatures: | |
| Injector, °C | 230 |
| Detector, °C | 250 |
| Oven, °C | 32 hold for 12 min Ramp 1 = 8°C/min to 64°C, hold for 10 min Ramp 2 = 10°C/min to 200°C, hold for 5 min |
| Carrier gas | Hydrogen |
| Flow rate, mls/min | 3 |
| Split ratio | 100:1 |
| Sample size, μl | 1.0 |

† Corrected editorially.

6.4 *Injector*—The specimen must be precisely and repeatably injected into the gas chromatograph. An automatic sample injection device is highly recommended. Manual injection can be employed if the precision stated in **Table 3** can be reliably and consistently satisfied.

7. Reagents and Materials

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society,

TABLE 3 Summary of Precision Data (mg/kg)

| Impurity | Expected Value | Average | Repeatability |
|--|----------------|---------|---------------|
| Isopentane | 21.1 | 21.1 | 2.0 |
| <i>n</i> -pentane | 24.7 | 25.8 | 3.0 |
| 2,2-dimethylbutane | 9.9 | 9.9 | 1.0 |
| cyclopentane | 11.5 | 11.4 | 0.8 |
| 2,3-dimethylbutane | 10.0 | 10.2 | 1.0 |
| 2-methylpentane | 17.3 | 18.1 | 2.2 |
| 3-methylpentane | 23.9 | 24.8 | 2.0 |
| <i>n</i> -hexane | 46.7 | 48.4 | 5.2 |
| 2,2-dimethylpentane | 4.9 | 5.2 | 0.8 |
| methylcyclopentane | 36.1 | 36.8 | 2.1 |
| 2,4-dimethylpentane | 49.7 | 51.7 | 5.1 |
| benzene | 12.1 | 12.4 | 1.3 |
| 2,3-dimethylpentane | 57.3 | 58.3 | 4.2 |
| 1,1-dimethylcyclopentane | 23.5 | 23.0 | 1.2 |
| cyclohexene | 29.4 | 29.6 | 1.3 |
| 3-methylhexane | 9.9 | 10.5 | 0.9 |
| <i>cis</i> -1,3-dimethylcyclopentane | 22.6 | 23.6 | 1.6 |
| <i>trans</i> -1,3-dimethylcyclopentane | 10.8 | 11.1 | 0.8 |
| 3-ethylpentane | 29.9 | 31.0 | 2.5 |
| <i>trans</i> -1,2-dimethylcyclopentane | 41.2 | 40.4 | 2.3 |
| isooctane | 10.0 | 10.5 | 1.1 |
| <i>n</i> -heptane | 37.1 | 38.4 | 3.7 |
| methylcyclohexane | 178.5 | 181.0 | 10.0 |
| ethylcyclopentane | 19.0 | 19.6 | 1.7 |
| toluene | 19.9 | 20.9 | 1.9 |
| <i>para</i> -xylene | 19.9 | 20.9 | 2.0 |
| isopropylcyclohexane | 19.6 | 20.4 | 1.8 |