This document is not an ASTM standard and is intended only to provide the user of an ASTM standard an indication of what changes have been made to the previous version. Because it may not be technically possible to adequately depict all changes accurately, ASTM recommends that users consult prior editions as appropriate. In all cases only the current version of the standard as published by ASTM is to be considered the official document.

INTERNATIONAL

Designation: E2589-08 Designation: E 2589 - 08a

Standard Terminology Relating to Nonsieving Methods of Powder Characterization¹

This standard is issued under the fixed designation E 2589; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

Particle size distribution, surface area, and other forms of particle analysis have been commonly adopted methods of verifying compliance with desired particle specifications for some time. Greater emphasis is now being placed on inter- and intralaboratory correlation of all particle measurement systems.

To ensure a better understanding of the comparison of testing results from particle measurement systems, terminology relating to the measurements must be clearly defined and documented so that both the recipient and generator of the data are in full agreement as to the meaning of the data. Every effort has been made here to ensure accuracy, precision, and clarity for the terms included in this terminology document. For Committee E29, this is an ongoing process with new terms being developed and defined for future inclusion. Suggestions and comments for additions, corrections, and revisions are welcomed.

1. Scope

1.1 This terminology covers the definitions of terms used in the description and procedures of analysis of particulate materials not ordinarily analyzed using test sieves. The terms relate directly to the equipment used in analysis, the physical forms of the materials to be analyzed, and selected descriptive data reduction and analysis formats.

1.2 Committee E29 on Particle and Spray Characterization believes that it is essential to include terms and definitions explicit to the committee's scope, regardless of whether the terms appear in existing ASTM standards. Terms that are in common usage and appear in common-language dictionaries are generally not included, unless they have specific meanings in the context of particle characterization different from the common-language definitions.

2. Significance and Use

STM E2589-08a

2.1 Interpretation and use of data generated by particle characterization methods is highly dependent on the definitions of terms describing that data. It is extremely important that those terms be defined in precisely the same way both when comparing data from different characterization techniques and even when correlating data from the same technique.

2.2 It is likewise important that users of particle characterization methods and the data generated therefrom understand the principles of the methods, so that differences and similarities in the data can be interpreted in relation to those principles. That understanding can help to avoid disagreements when data from different characterization methods are compared.

2.3 The definitions contained in this terminology will aid in the interpretation of particle characterization data with respect to the method(s) used to produce that data.

3. Terminology

blind pore, *n*—open pore having only a single connection with an external surface.

dispersion, *n*—system consisting of particles distributed in a solid, liquid, or gas.

dynamic image analysis, *n*—particle size and shape analysis using computer image analysis techniques on instantaneouslycaptured still-frame projected images of particles in motion.

DISCUSSION-Some instruments use a moving measurement apparatus on static particles.

electrical sensing zone analysis, n— particle size analysis in which particles suspended in a conductive liquid medium pass

Copyright © ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States.

¹ This terminology is under the jurisdiction of ASTM Committee E29 on Particle and Spray Characterization and is the direct responsibility of Subcommittee E29.02 on Non-Sieving Methods.

Current edition approved May 15, 2008. Published June 2008.

Current edition approved July 1, 2008. Published July 2008. Originally approved in 2007. Last previous edition approved in 2008 as E 2589 - 08.