

INTERNATIONAL  
STANDARD

ISO/ASTM  
51607

Third edition  
2013-06-01

---

---

**Practice for use of an alanine-EPR  
dosimetry system**

*Pratique pour l'utilisation d'un système dosimétrique à l'alanine  
utilisant la résonance paramagnétique électronique*

(<https://standards.iteh.ai>)  
Document Preview

[ISO/ASTM 51607:2013](https://standards.iteh.ai/catalog/standards/iso/bce12010-303e-4376-a0f9-4b957265ba9c/iso-astm-51607-2013)

<https://standards.iteh.ai/catalog/standards/iso/bce12010-303e-4376-a0f9-4b957265ba9c/iso-astm-51607-2013>



Reference number  
ISO/ASTM 51607:2013(E)

© ISO/ASTM International 2013

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO/ASTM 51607:2013](https://standards.iteh.ai/catalog/standards/iso/bcef2010-303e-4376-a0f9-4b957265ba9c/iso-astm-51607-2013)

<https://standards.iteh.ai/catalog/standards/iso/bcef2010-303e-4376-a0f9-4b957265ba9c/iso-astm-51607-2013>

© ISO/ASTM International 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

ASTM International, 100 Barr Harbor Drive, PO Box C700,  
West Conshohocken, PA 19428-2959, USA  
Tel. +610 832 9634  
Fax +610 832 9635  
E-mail [khooper@astm.org](mailto:khooper@astm.org)  
Web [www.astm.org](http://www.astm.org)

Published in Switzerland

<b>Contents</b>	<b>Page</b>
1 Scope .....	1
2 Referenced documents .....	1
3 Terminology .....	1
4 Significance and use .....	2
5 Overview .....	2
6 Influence quantities .....	2
7 Dosimetry system and its verification .....	3
8 Incoming dosimeter stock assessment .....	4
9 Calibration procedures .....	4
10 Routine use .....	4
11 Minimum documentation requirements .....	5
12 Measurement uncertainty .....	5
13 Keywords .....	5
<b>Bibliography</b> .....	<b>5</b>

iTeh Standards  
 (https://standards.itih.ai)  
 Document Preview

[ISO/ASTM 51607:2013](https://standards.itih.ai/catalog/standards/iso/bcef2010-303e-4376-a0f9-4b957265ba9c/iso-astm-51607-2013)

<https://standards.itih.ai/catalog/standards/iso/bcef2010-303e-4376-a0f9-4b957265ba9c/iso-astm-51607-2013>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

A pilot project between ISO and ASTM International has been formed to develop and maintain a group of ISO/ASTM radiation processing dosimetry standards. Under this pilot project, ASTM Committee E61, Radiation Processing, is responsible for the development and maintenance of these dosimetry standards with unrestricted participation and input from appropriate ISO member bodies.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. Neither ISO nor ASTM International shall be held responsible for identifying any or all such patent rights.

[ISO/ASTM 51607:2013](https://standards.iteh.ai/catalog/standards/iso/bce2010-303e-4376-a09-4b957265ba9c/iso-astm-51607-2013)

International Standard ISO/ASTM 51607 was developed by ASTM Committee E61, Radiation Processing, through Subcommittee E61.02, Dosimetry Systems, and by Technical Committee ISO/TC 85, Nuclear energy, nuclear technologies and radiological protection.

This third edition of ISO/ASTM 51607 cancels and replaces ISO/ASTM 51607:2004(E)



## Standard Practice for Use of an Alanine-EPR Dosimetry System<sup>1</sup>

This standard is issued under the fixed designation ISO/ASTM 51607; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

### 1. Scope

1.1 This practice covers dosimeter materials, instrumentation, and procedures for using the alanine-EPR dosimetry system for measuring the absorbed dose in the photon and electron radiation processing of materials. The system is based on electron paramagnetic resonance (EPR) spectroscopy of free radicals derived from the amino acid alanine.<sup>2</sup>

1.2 The alanine dosimeter is classified as a type I dosimeter as it is affected by individual influence quantities in a well-defined way that can be expressed in terms of independent correction factors (see ASTM Practice E2628). The alanine dosimeter may be used in either a reference standard dosimetry system or in a routine dosimetry system.

1.3 This document is one of a set of standards that provides recommendations for properly implementing dosimetry in radiation processing, and describes a means of achieving compliance with the requirements of ASTM E2628 “Practice for Dosimetry in Radiation Processing” for alanine dosimetry system. It should be read in conjunction with ASTM E2628.

1.4 This practice covers alanine-EPR dosimetry systems for dose measurements under the following conditions:

1.4.1 The absorbed dose range is between 1 and  $1.5 \times 10^5$  Gy.

1.4.2 The absorbed dose rate is up to  $10^2$  Gy s<sup>-1</sup> for continuous radiation fields and up to  $3 \times 10^{10}$  Gy s<sup>-1</sup> for pulsed radiation fields **(1-4)**.<sup>3</sup>

1.4.3 The radiation energy for photons and electrons is between 0.1 and 30 MeV **(1, 2, 5-8)**.

1.4.4 The irradiation temperature is between -78 °C and +70 °C **(2, 9-12)**.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.02 on Dosimetry Systems, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved April 9, 2013. Published June 2013. Originally published as ASTM E 1607 – 94. Last previous ASTM edition E 1607 – 96<sup>1</sup>. ASTM E 1607 – 94 was adopted by ISO in 1998 with the intermediate designation ISO 15566:1998(E). The present International Standard ISO/ASTM 51607:2013(E) replaces ISO 15566 and is a major revision of the last previous edition ISO/ASTM 51607–2004(E).

<sup>2</sup> The term “electron spin resonance” (ESR) is used interchangeably with electron paramagnetic resonance (EPR).

<sup>3</sup> The boldface numbers in parentheses refer to the bibliography at the end of this standard.

### 2. Referenced documents

#### 2.1 ASTM Standards:<sup>4</sup>

E170 Terminology Relating to Radiation Measurements and Dosimetry

E2628 Practice for Dosimetry in Radiation Processing

E2701 Guide for Performance Characterization of Dosimeters and Dosimetry Systems for Use in Radiation Processing

#### 2.2 ISO/ASTM Standards:<sup>4</sup>

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51707 Guide for Estimating Uncertainties in Dosimetry for Radiation Processing

#### 2.3 ICRU Reports:<sup>5</sup>

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

ICRU Report 80 Dosimetry Systems for Use in Radiation Processing

#### 2.4 Joint Committee for Guides in Metrology (JCGM) Reports:

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement<sup>6</sup>

JCGM 100:2008, VIM , International vocabulary of metrology – Basis and general concepts and associated terms<sup>7</sup>

### 3. Terminology

#### 3.1 Definitions:

3.1.1 *alanine dosimeter*—specified quantity and physical form of the radiation-sensitive material alanine and any added inert substance such as a binder.

3.1.2 *alanine-EPR dosimetry system*—system used for determining absorbed dose, consisting of alanine dosimeters, an EPR spectrometer and its associated reference materials, and procedures for the system’s use.

<sup>4</sup> For referenced ASTM and ISO/ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.

<sup>5</sup> Available from International Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, U.S.A.

<sup>6</sup> Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1). Available free of charge at the BIPM website (<http://www.bipm.org>).

<sup>7</sup> Document produced by Working Group 2 of the Joint Committee for Guides in Metrology (JCGM/WG 2). Available free of charge at the BIPM website (<http://www.bipm.org>).



3.1.3 *alanine-EPR dosimeter response*—value resulting from applied adjustments to the EPR signal amplitude.

3.1.4 *check standard*—a standard prepared independently of the calibration standards that is measured to verify the performance of a dosimetry system.

3.1.5 *EPR intensity reference material*—a stable paramagnetic material whose measurement by EPR is applied to the dosimeter EPR signal amplitude as part of the dosimeter response determination.

3.1.6 *EPR signal amplitude*—peak-to-peak amplitude of the central signal of the EPR spectrum.

3.1.6.1 *Discussion*—This signal is proportional to the alanine-derived free radical concentration in the alanine dosimeter.

3.1.7 *EPR spectroscopy*—measurement of resonant absorption of electromagnetic energy resulting from the transition of unpaired electrons between different energy levels, upon application of radio frequencies to a paramagnetic substance in the presence of a magnetic field.

3.1.8 *EPR spectrum*—first derivative of the electron paramagnetic absorption spectrum measured as a function of the magnetic field.

3.1.9 *zero dose amplitude*—EPR signal amplitude of an unirradiated alanine dosimeter with the same EPR spectrometer parameters used for the lowest measurable absorbed dose value.

3.2 Definitions of other terms used in this standard that pertain to radiation measurement and dosimetry may be found in ASTM Terminology E170. Definitions in E170 are compatible with ICRU Report 85a; that document, therefore, may be used as an alternative reference.

## 4. Significance and use

4.1 The alanine-EPR dosimetry system provides a means for measuring absorbed dose. It is based on the measurement of specific stable free radicals in crystalline alanine generated by ionizing radiation.

4.2 Alanine-EPR dosimetry systems are used in reference- or transfer-standard or routine dosimetry systems in radiation applications that include: sterilization of medical devices and pharmaceuticals, food irradiation, polymer modifications, medical therapy and radiation damage studies in materials (1, 13-15).

## 5. Overview

5.1 The dosimeter is prepared using  $\alpha$ -alanine,  $\text{CH}_3\text{-CH}(\text{NH}_2)\text{-COOH}$ , in the form of polycrystalline powder.

5.2 All stereoisomers of  $\alpha$ -alanine are suitable for dosimetry; L-alanine is used most commonly.

5.3 Usual physical shapes are films or pellets (cylinders).

NOTE 1—Additives, capsules, or film support materials used in the preparation of dosimeters should not add any significant intrinsic or radiation-induced EPR signal. Examples of suitable binders are ethylene-propylene rubber, gelatin, paraffin, polyethylene, polyethylene vinyl acetate, polystyrene, polyvinylpyrrolidone, polyvinyl propylene, and stearin. Lubricants added in the dosimeter manufacturing process are optional. An example of a suitable lubricant is stearic acid (16-21).

5.4 The dosimeter contains crystalline alanine and registers the absorbed dose by the formation of alanine-derived free radicals (22). Identification and measurement of alanine-derived free radicals are performed by EPR spectroscopy. ICRU Report 80 provides information on the scientific basis and historical development of this dosimetry system.

5.5 The measurement of free radicals by EPR spectroscopy is nondestructive. This can be repeated and hence can be used for archival purposes (23-25).

## 6. Influence quantities

6.1 Factors other than absorbed dose which influence the dosimeter response are referred to as influence quantities, and are discussed in the following sections (see also ASTM Guide E2701). Examples of such influence quantities are temperature and dose rate.

### 6.2 Pre-Irradiation Conditions:

6.2.1 *Dosimeter Conditioning and Packaging*—Alanine dosimeter conditioning and packaging may be important under certain conditions (see 6.2.4).

NOTE 2—The sorting of alanine pellet dosimeters by mass into sub-lots will improve the measurement uncertainty.

6.2.2 *Time Since Manufacture*—There is no known influence of time since manufacture on alanine dosimeters when stored under recommended conditions.

6.2.3 *Temperature*—There is no known influence of pre-irradiation temperature. However, it is recommended that alanine dosimeters be stored at manufacturer recommended temperatures. Exposure to temperatures outside the manufacturer's recommended range should be avoided to reduce the potential for adverse effects on dosimeter response.

6.2.4 *Relative Humidity*—The humidity during pre-irradiation storage may influence the EPR signal amplitude of alanine dosimeters (24, 25). The effect of humidity may be reduced by sealing dosimeters in a material impervious to water.

6.2.5 *Exposure to Light*—There is no known influence of ambient light.

### 6.3 Conditions During Irradiation:

6.3.1 *Irradiation Temperature*—The irradiation temperature influences the EPR signal amplitude of alanine dosimeters.

NOTE 3—The effect of irradiation temperature on the dosimeter EPR signal amplitude may be dependent on the dosimeter type. The temperature coefficient,  $R_t$  ( $\text{K}^{-1}$ ) is described by the relationship,  $(\Delta m/m)/\Delta T$ , where  $m$  is the EPR signal amplitude (in arbitrary units) and  $T$  is the irradiation temperature (in K). For dosimeters with L-alanine, a positive temperature coefficient, expressed in percent, in the range of +0.1 to +0.2 %  $^\circ\text{C}^{-1}$  is typical for irradiation temperatures from  $-10^\circ\text{C}$  to  $+70^\circ\text{C}$  (10, 11, 26-28); refer to Ref (9, 12) for irradiation temperatures below  $-10^\circ\text{C}$ . The temperature coefficient for dosimeters prepared with the DL stereoisomer of alanine is more than 50 % higher than one prepared with L-alanine (29). A summary of published temperature coefficients is tabulated in Ref (26, 29).

6.3.2 *Absorbed-Dose Rate*—Under normal radiation processing conditions there is no measurable effect of absorbed dose rate; however, a dose dependent effect has been characterized for alanine dosimeters irradiated to high doses at low dose rates (30).