

DRAFT INTERNATIONAL STANDARD

ISO/DIS 1135-5

ISO/TC 76

Secretariat: DIN

Voting begins on:
2013-10-03

Voting terminates on:
2014-03-03

Transfusion equipment for medical use —

Part 5:

Transfusion sets for single use with pressure infusion apparatus

Matériel de transfusion à usage médical —

Partie 5: Appareils de transfusion non réutilisables avec des appareils de perfusion sous pression

[Revision of first edition (ISO 1135:1997) and of five edition ISO 1135-4:2012]

ICS: 11.040.20

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/1135-5-2013/3e93-4981-94ef-4b9aff19dbb9/iso-1135-5-2013>

ISO/CEN PARALLEL PROCESSING

This draft has been developed within the International Organization for Standardization (ISO), and processed under the **ISO lead** mode of collaboration as defined in the Vienna Agreement.

This draft is hereby submitted to the ISO member bodies and to the CEN member bodies for a parallel five month enquiry.

Should this draft be accepted, a final draft, established on the basis of comments received, will be submitted to a parallel two-month approval vote in ISO and formal vote in CEN.

To expedite distribution, this document is circulated as received from the committee secretariat. ISO Central Secretariat work of editing and text composition will be undertaken at publication stage.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.



Reference number
ISO/DIS 1135-5:2013(E)

© ISO 2013

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/937c7148-3e93-4981-94ef-4b9aff19dbb9/iso-1135-5-2015>

Copyright notice

This ISO document is a Draft International Standard and is copyright-protected by ISO. Except as permitted under the applicable laws of the user's country, neither this ISO draft nor any extract from it may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, photocopying, recording or otherwise, without prior written permission being secured.

Requests for permission to reproduce should be addressed to either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Reproduction may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

Contents

Page

Foreword	v
1 Scope	1
2 Normative references	1
3 General requirements	2
3.1 Nomenclature for components of the transfusion set	2
3.2 Maintenance of sterility.....	2
3.3 Designation	3
4 Materials	3
5 Physical requirements	3
5.1 Particulate contamination.....	3
5.2 Leakage	3
5.3 Tensile strength.....	3
5.4 Closure-piercing device.....	3
5.5 Tubing.....	4
5.6 Filter for blood and blood components	4
5.7 Drip chamber and drip tube.....	4
5.8 Flow regulator.....	4
5.9 Flow rate of blood and blood components.....	4
5.10 Injection site.....	5
5.11 Male conical fitting	5
5.12 Protective caps	5
6 Chemical requirements.....	5
6.1 Reducing (oxidizable) matter	5
6.2 Metal ions	5
6.3 Titration acidity or alkalinity.....	5
6.4 Residue on evaporation.....	5
6.5 UV absorption of extract solution.....	5
7 Biological requirements.....	6
7.1 General	6
7.2 Sterility.....	6
7.3 Pyrogenicity	6
7.4 Haemolysis.....	6
7.5 Toxicity	6
7.6 Assessment of blood component depletion.....	6
7.7 Assessment of damage to blood components.....	6
8 Labelling	7
8.1 General	7
8.2 Unit container	7
8.3 Shelf or multi-unit container	7
9 Packaging.....	8
10 Disposal.....	8
Annex A (normative) Physical tests.....	9
Annex B (normative) Chemical tests	13
Annex C (normative) Biological tests	15
Bibliography.....	16

Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EC Directive 93/42/EEC on medical devices 17

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/937c7148-3e93-4981-94ef-4b9aff19dbb9/iso-1135-5-2015>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 1135-5 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

This first edition of ISO 1135-5 together with ISO 1135-4, sixth edition, cancels and replaces the fifth edition (ISO 1135-4:2012), of which the scope has been restricted to gravity feed and the whole document aligned accordingly.

ISO 1135 consists of the following parts, under the general title *Transfusion equipment for medical use*:

- Part 3: *Blood-taking set*
- Part 4: *Transfusion sets for single use, gravity feed*
- Part 5: *Transfusion sets for single use with pressure infusion apparatus*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/937c7148-3e93-4981-94ef-4b9aff19dbb9/iso-1135-5-2015>

Transfusion equipment for medical use — Part 5: Transfusion sets for single use with pressure infusion apparatus

1 Scope

This part of ISO 1135 specifies requirements for single use transfusion sets for use with pressure infusion equipment capable of generating pressures up to 200 kPa (2 bar). The standard ensures compatibility with containers for blood and blood components as well as intravenous equipment.

Secondary aims of this part of ISO 1135 are to provide guidance on specifications relating to the quality and performance of materials used in transfusion sets, to present designations for transfusion set components and to ensure the compatibility of sets with red cell and plasma blood components.

Platelet components should not be transfused under pressure using these sets.

In some countries, the national pharmacopoeia or other national regulations are legally binding and take precedence over this part of ISO 1135.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-1:1986, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements*¹⁾

ISO 594-2:1998, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*¹⁾

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 3826-1:2013, *Plastics collapsible containers for human blood and blood components — Part 1: Conventional containers*

ISO 7000, *Graphical symbols for use on equipment — Index and synopsis*

ISO 7864, *Sterile hypodermic needles for single use*

ISO 14644-1:1999, *Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness*


ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

1) Will be replaced by ISO 80369-7.

3 General requirements

3.1 Nomenclature for components of the transfusion set

The nomenclature for components of transfusion sets is given in Figure 1.



Key

- | | |
|-------------------------------------------------|-----------------------------------------------|
| 1 protective cap of the closure-piercing device | 7 tubing |
| 2 closure-piercing device | 8 flow regulator |
| 3 fluid channel | 9 injection site |
| 4 drip tube | 10 male conical fitting |
| 5 drip chamber | 11 protective cap of the male conical fitting |
| 6 filter for blood and blood components | |

^a Indicates alternative locations of the filter for blood and blood components. Other designs are acceptable if the same safety aspects are ensured.

^b Injection site is optional.

Figure 1 — Example of a transfusion set

3.2 Maintenance of sterility

The transfusion set shall be provided with protective caps to maintain sterility of the internal parts of the set until the set is used.

3.3 Designation

A transfusion set complying with the requirements of this part of ISO 1135 is designated as follows:

Transfusion set ISO 1135-5

4 Materials

The materials from which the transfusion sets given in Clause 3 are manufactured shall comply with the requirements specified in Clause 5. If components of the transfusion set come into contact with blood and blood components, they shall additionally comply with the requirements specified in Clauses 6 and 7.

5 Physical requirements

5.1 Particulate contamination

The transfusion sets shall be manufactured under conditions that minimize particulate contamination. All parts shall be smooth and clean at the fluid pathway surfaces. When tested as specified in A.1, the number of particles detected shall not exceed the contamination index limit.

5.2 Leakage

The transfusion set, when tested in accordance with A.2, shall show no signs of air leakage.


5.3 Tensile strength

Any connections between the components of the transfusion set, excluding protective caps, shall withstand a static tensile force of not less than 15 N for 15 s.

5.4 Closure-piercing device

5.4.1 The dimensions of the closure-piercing device shall conform to the dimensions shown in Figure 2.

NOTE The dimension of 15 mm in Figure 2 is a reference measurement. The cross-section of the piercing device at this site is a circle.



Dimensions in millimetres

Figure 2 — Dimensions of the closure-piercing device

5.4.2 The closure-piercing device shall be capable of piercing and penetrating the closure of a container for blood and blood components without pre-piercing. No coring should occur during this procedure.

NOTE 1 A carefully controlled surface treatment of the closure-piercing device (e.g. siliconization) is recommended to facilitate its insertion into the blood bag port. The same effect may be achieved by a careful selection of material for the closure-piercing device. Typical results including test equipment for penetration forces between spikes and blood bag ports have been published. See References [11] and [12].

NOTE 2 A central closure-piercing device tip is preferred to an asymmetric design in order to aid its insertion.

5.4.3 When inserted into a blood bag port conforming to ISO 3826-1, the closure-piercing device shall resist a pull force of 15 N for 15 s.

5.4.4 When tested in accordance with ISO 3826-1:2013, 5.3, the connection between the closure-piercing device and the blood bag port shall show no evidence of leakage.

5.5 Tubing

5.5.1 The tubing, made of flexible material, shall be transparent or sufficiently translucent so that the interface of air and water during the passage of air bubbles can be observed with normal or corrected-to-normal vision.

5.5.2 The tubing from the distal end to the drip chamber shall be not less than 1 500 mm in length, including the injection site, when provided, and the male conical fitting.

5.5.3 The tubing from the distal end to the drip chamber shall be capable of resisting (without collapsing) negative pressures of at least -50 kPa.

5.6 Filter for blood and blood components

The transfusion set shall be provided with a filter for blood and blood components. The filter shall have uniform pores and shall cover a total area of not less than 10 cm^2 . When tested in accordance with A.3, the mass of solid material retained on the filter shall be not less than 80 % (mass fraction) of that retained on the reference filter.

If the filter has a confirmed thread diameter of $(100 \pm 10)\text{ }\mu\text{m}$ and a pore size of $(200 \pm 20)\text{ }\mu\text{m}$, with a single warp and a single weft, a filtration performance test can be exempted.

Pore size measurement can be performed by microscopic inspection.

5.7 Drip chamber and drip tube

The drip chamber shall permit continuous observation of the fall of drops. The liquid shall enter the drip chamber through a tube which projects into the chamber. There shall be a distance of not less than 40 mm between the end of the drip tube and the outlet of the chamber, or a distance of not less than 20 mm between the drip tube and the filter for blood and blood components. The wall of the drip chamber shall not be closer than 5 mm to the end of the drip tube. The drip tube shall be such that 20 drops of distilled water at $(23 \pm 2)\text{ }^\circ\text{C}$ and at a flow rate of (50 ± 10) drops/min deliver $(1 \pm 0,1)$ ml $[(1 \pm 0,1)\text{ g}]$.

The drip chamber should permit and facilitate the procedure of priming.

5.8 Flow regulator

The flow regulator shall adjust the flow of the blood and blood components between zero and maximum.

The flow regulator should be capable of continuous use throughout a transfusion without the tubing being damaged. There should be no deleterious reaction between the flow regulator and the tubing when stored in such a manner that there is contact.

5.9 Flow rate of blood and blood components

The transfusion set shall deliver not less than 1 000 ml of blood at $(23 \pm 2)\text{ }^\circ\text{C}$ in 30 min with a pressure difference of 10 kPa. The transfusion set shall also deliver not less than 500 ml of blood in 2 min under a pressure of 30 kPa above atmospheric pressure.