



SLOVENSKI STANDARD SIST EN IEC 61207-3:2019

01-november-2019

Nadomešča:
SIST EN 61207-3:2002

Analizatorji plina - Izražanje lastnosti - 3. del: Paramagnetni analizatorji kisika (IEC 61207-3:2019)

Gas Analyzers - Expression of performance - Part 3: Paramagnetic oxygen analysers (IEC 61207-3:2019)

Angabe zum Betriebsverhalten von Gasanalysatoren - Teil 3: Paramagnetische Sauerstoffanalysatoren (IEC 61207-3:2019)

Analyseurs de gaz - Expression des performances - Partie 3: Analyseurs d'oxygène paramagnétiques (IEC 61207-3:2019)

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>

Ta slovenski standard je istoveten z: EN IEC 61207-3:2019

ICS:

71.040.40 Kemijska analiza Chemical analysis

SIST EN IEC 61207-3:2019 en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN IEC 61207-3:2019

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>

EUROPEAN STANDARD

EN IEC 61207-3

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2019

ICS 19.040; 71.040.40

Supersedes EN 61207-3:2002 and all of its amendments
and corrigenda (if any)

English Version

Gas Analyzers - Expression of performance - Part 3: Paramagnetic oxygen analysers (IEC 61207-3:2019)

Analyseurs de gaz - Expression des performances - Partie
3: Analyseurs d'oxygène paramagnétiques
(IEC 61207-3:2019)

Angabe zum Betriebsverhalten von Gasanalysatoren - Teil
3: Paramagnetische Sauerstoffanalysatoren
(IEC 61207-3:2019)

This European Standard was approved by CENELEC on 2019-07-31. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

[SIST EN IEC 61207-3:2019](#)

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 61207-3:2019 (E)**European foreword**

The text of document 65B/1155/FDIS, future edition 3 of IEC 61207-3, prepared by SC 65B "Measurement and control devices" of IEC/TC 65 "Industrial-process measurement, control and automation" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 61207-3:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-04-30
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-07-31

This document supersedes EN 61207-3:2002 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Endorsement notice

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>

The text of the International Standard IEC 61207-3:2019 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 61115	NOTE	Harmonized as EN 61115
IEC 60654-1	NOTE	Harmonized as EN 60654-1
ISO 9001	NOTE	Harmonized as EN ISO 9001

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61207-1	-	Expression of performance of gas analyzers - Part 1: General	EN 61207-1	-

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN IEC 61207-3:2019

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN IEC 61207-3:2019

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>



IEC 61207-3

Edition 3.0 2019-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Gas analyzers – Expression of performance –
Part 3: Paramagnetic oxygen analyzers

Analyseurs de gaz – Expression des performances –
Partie 3: Analyseurs d'oxygène paramagnétiques

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 19.040; 71.040.40

ISBN 978-2-8322-7046-2

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms and definitions	7
4 Procedures for specification	15
4.1 General.....	15
4.2 Specification of essential ancillary units and services	15
4.2.1 Sampling system	15
4.2.2 Services	15
4.3 Additional characteristics related to specification of performance.....	16
4.4 Important aspects related to specification of performance	16
4.4.1 General	16
4.4.2 Rated range of ambient temperature.....	16
4.4.3 Rated range of sample gas temperature	16
4.4.4 Rated range of ambient pressure	17
4.4.5 Rated range of sample pressure	17
4.4.6 Rated range of sample flow	17
4.4.7 Rated range of sample dew point.....	17
4.4.8 Rated range of sample particulate content.....	17
4.4.9 Rated range of interference uncertainties	18
4.4.10 Rated range of linearity uncertainty.....	18
4.4.11 Rated ranges of influence quantities	18
5 Procedures for compliance testing.....	18
5.1 Analyzer testing	18
5.1.1 General	18
5.1.2 Test equipment.....	18
5.2 Testing procedures	19
5.2.1 General	19
5.2.2 Interference uncertainty.....	19
5.2.3 Wet samples.....	20
5.2.4 Delay times, rise time, fall time	20
Annex A (informative) Interfering gases	22
Annex B (informative) Methods of preparation of water vapour in test gases	26
Bibliography.....	28
Figure 1 – Magnetic auto-balance system with current feedback.....	9
Figure 2 – Thermomagnetic oxygen sensor.....	11
Figure 3 – Differential pressure oxygen sensor	12
Figure 4 – Typical sampling systems – Filtered and dried system with pump for wet samples	13
Figure 5 – Typical sampling system – Steam-aspirated system with water wash for wet samples	14
Figure 6 – General test arrangement – Dry gases	19
Figure 7 – Test apparatus to apply gases and water vapour to analysis systems	21

Table A.1 – Zero correction factors for current gases.....23

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN IEC 61207-3:2019

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**GAS ANALYZERS –
EXPRESSION OF PERFORMANCE –****Part 3: Paramagnetic oxygen analyzers**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a333f/sist-en-iec-61207-3-2019>
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61207-3 has been prepared by sub-committee 65B: Measurement and control devices, of IEC technical committee 65: Industrial-process measurement, control and automation.

This third edition cancels and replaces the second edition published in 2002. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) all references (normative and informative) have been updated, deleted or added to as appropriate;
- b) all the terms, descriptions and definitions relating to the document have been updated where appropriate;

- c) all references to “errors” have been replaced by “uncertainties” and appropriate updated definitions applied.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
65B/1155/FDIS	65B/1157/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

This International Standard is to be used in conjunction with IEC 61207-1:2010.

A list of all parts in the IEC 61207 series, published under the general title *Gas analyzers – Expression of performance*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN IEC 61207-3:2019](https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019)
<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>

INTRODUCTION

Paramagnetic oxygen analyzers respond to the partial pressure of oxygen in the measured gas, and the volumetric concentration is then determined by knowledge of the total pressure, as in many other gas analyzers. Due to this fact, many paramagnetic oxygen analyzers use pressure compensation (see 4.4.4 and 4.4.5). They are used in a wide range of industrial, laboratory, medical, and other applications where the rated measuring range of the analyzer is between 0 % to 1 % and 0 % to 100 %, at reference pressure (usually near atmospheric).

Only a few gases display significant paramagnetism (for example, oxygen, nitric oxide and nitrogen dioxide), and oxygen has the strongest paramagnetic susceptibility (see Annex A) among gases. By employing this particular property of oxygen, analyzers have been designed that can be highly specific to the measurement in most industrial and medical applications, where, for example, high background levels of hydrocarbons or moisture may be present.

There are several different techniques described for measuring oxygen by its paramagnetic property, but three main methods have evolved over many years of commercial application.

The three methods are:

- automatic null balance;
- thermomagnetic or magnetic wind;
- differential pressure or Quincke.

These methods all require the sample gas to be clean and non-condensing, though some versions work at elevated temperatures so that samples that are likely to condense at a lower temperature can be analyzed. Because of this requirement, analyzers often require a sample system to condition the sample prior to measurement.

<https://standards.iteh.ai/catalog/standards/sist/553f7335-7b7d-4abb-891e-0c39ed5a33fc/sist-en-iec-61207-3-2019>