



SLOVENSKI STANDARD SIST EN ISO 14971:2020

01-maj-2020

Nadomešča:
SIST EN ISO 14971:2012

Medicinski pripomočki - Uporaba obvladovanja tveganja pri medicinskih pripomočkih (ISO 14971:2019)

Medical devices - Application of risk management to medical devices (ISO 14971:2019)

Medizinprodukte - Anwendung des Risikomanagements auf Medizinprodukte (ISO 14971:2019)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Dispositifs médicaux - Application de la gestion des risques aux dispositifs médicaux (ISO 14971:2019)

[SIST EN ISO 14971:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

Ta slovenski standard je istoveten z: EN ISO 14971:2019

ICS:

11.040.01	Medicinska oprema na splošno	Medical equipment in general
-----------	------------------------------	------------------------------

SIST EN ISO 14971:2020	en
-------------------------------	-----------

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14971:2020](https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020)

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

EUROPEAN STANDARD

EN ISO 14971

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2019

ICS 11.040.01

Supersedes EN ISO 14971:2012

English version

Medical devices - Application of risk management to medical devices (ISO 14971:2019)

Dispositifs médicaux - Application de la gestion des
risques aux dispositifs médicaux (ISO 14971:2019)

Medizinprodukte - Anwendung des
Risikomanagements auf Medizinprodukte (ISO
14971:2019)

This European Standard was approved by CEN on 5 August 2019.

CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN and CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN and CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



**CEN-CENELEC Management Centre:
Rue de la Science 23, B-1040 Brussels**

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14971:2020](https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020)
<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

European foreword

This document (EN ISO 14971:2019) has been prepared by Technical Committee ISO/TC 210 "Quality management and corresponding general aspects for medical devices" in collaboration with Technical Committee CEN/CLC/JTC 3 "Quality management and corresponding general aspects for medical devices" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2020, and conflicting national standards shall be withdrawn at the latest by June 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 14971:2012.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW

(standards.iteh.ai)

Endorsement notice

The text of ISO 14971:2019 has been approved by CEN as EN ISO 14971:2019 without any modification.

SIST EN ISO 14971:2020
<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8b1e-a466d8f8598e/sist-en-iso-14971-2020>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 14971:2020

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

INTERNATIONAL
STANDARD

ISO
14971

Third edition
2019-12

**Medical devices — Application of risk
management to medical devices**

*Dispositifs médicaux — Application de la gestion des risques aux
dispositifs médicaux*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 14971:2020](https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020)

[https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-
a466d8f8598e/sist-en-iso-14971-2020](https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020)



Reference number
ISO 14971:2019(E)

© ISO 2019

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 14971:2020

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements for risk management system	7
4.1 Risk management process	7
4.2 Management responsibilities.....	8
4.3 Competence of personnel.....	9
4.4 Risk management plan	9
4.5 Risk management file	10
5 Risk analysis	10
5.1 Risk analysis process	10
5.2 Intended use and reasonably foreseeable misuse	10
5.3 Identification of characteristics related to safety	11
5.4 Identification of hazards and hazardous situations	11
5.5 Risk estimation	11
6 Risk evaluation	12
7 Risk control	12
7.1 Risk control option analysis.....	12
7.2 Implementation of risk control measures.....	13
7.3 Residual risk evaluation.....	13
7.4 Benefit-risk analysis.....	14
7.5 Risks arising from risk control measures.....	14
7.6 Completeness of risk control	14
8 Evaluation of overall residual risk	14
9 Risk management review	15
10 Production and post-production activities	15
10.1 General.....	15
10.2 Information collection.....	15
10.3 Information review.....	16
10.4 Actions.....	16
Annex A (informative) Rationale for requirements	17
Annex B (informative) Risk management process for medical devices	26
Annex C (informative) Fundamental risk concepts	30
Bibliography	36

ISO 14971:2019(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*, and IEC/SC 62A, *Common aspects of electrical equipment used in medical practice*.

This third edition cancels and replaces the second edition (ISO 14971:2007), which has been technically revised. The main changes compared to the previous edition are as follows:

- A clause on normative references has been included, in order to respect the requirements for fixed in Clause 15 of ISO/IEC Directives, Part 2:2018.
- The defined terms are updated and many are derived from ISO/IEC Guide 63:2019. Defined terms are printed in italic to assist the reader in identifying them in the body of the document.
- Definitions of *benefit*, *reasonably foreseeable misuse* and *state of the art* have been introduced.
- More attention is given to the *benefits* that are expected from the use of the *medical device*. The term *benefit-risk* analysis has been aligned with terminology used in some regulations.
- It is explained that the *process* described in ISO 14971 can be used for managing *risks* associated with *medical devices*, including those related to data and systems security.
- The method for the evaluation of the overall *residual risk* and the criteria for its acceptability are required to be defined in the *risk management* plan. The method can include gathering and reviewing data and literature for the *medical device* and for similar *medical devices* and similar other products on the market. The criteria for the acceptability of the overall *residual risk* can be different from the criteria for acceptability of individual *risks*.
- The requirements to disclose *residual risks* have been moved and merged into one requirement, after the overall *residual risk* has been evaluated and judged acceptable.
- The review before commercial distribution of the *medical device* concerns the execution of the *risk management* plan. The results of the review are documented as the *risk management* report.

- The requirements for production and *post-production* activities have been clarified and restructured. More detail is given on the information to be collected and the actions to be taken when the collected information has been reviewed and determined to be relevant to *safety*.
- Several informative annexes are moved to the guidance in ISO/TR 24971, which has been revised in parallel. More information and a rationale for the requirements in this third edition of ISO 14971 have been provided in [Annex A](#). The correspondence between the clauses of the second edition and those of this third edition is given in [Annex B](#).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 14971:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

ISO 14971:2019(E)

Introduction

The requirements contained in this document provide *manufacturers* with a framework within which experience, insight and judgment are applied systematically to manage the *risks* associated with the use of *medical devices*.

This document was developed specifically for *manufacturers* of *medical devices* on the basis of established principles of *risk management* that have evolved over many years. This document could be used as guidance in developing and maintaining a *risk management process* for other products that are not necessarily *medical devices* in some jurisdictions and for suppliers and other parties involved in the *medical device life cycle*.

This document deals with *processes* for managing *risks* associated with *medical devices*. *Risks* can be related to injury, not only to the patient, but also to the user and other persons. *Risks* can also be related to damage to property (for example objects, data, other equipment) or the environment.

Risk management is a complex subject because each stakeholder can place a different value on the acceptability of *risks* in relation to the anticipated *benefits*. The concepts of *risk management* are particularly important in relation to *medical devices* because of the variety of stakeholders including medical practitioners, the organizations providing health care, governments, industry, patients and members of the public.

It is generally accepted that the concept of *risk* has two key components:

- the probability of occurrence of *harm*; and
- the consequences of that *harm*, that is, how severe it might be.

All stakeholders need to understand that the use of a *medical device* involves an inherent degree of *risk*, even after the *risks* have been reduced to an acceptable level. It is well known that in the context of a clinical *procedure* some *residual risks* remain. The acceptability of a *risk* to a stakeholder is influenced by the key components listed above and by the stakeholder's perception of the *risk* and the *benefit*. Each stakeholder's perception can vary depending upon their cultural background, the socio-economic and educational background of the society concerned and the actual and perceived state of health of the patient. The way a *risk* is perceived also takes into account other factors, for example, whether exposure to the *hazard* or *hazardous situation* seems to be involuntary, avoidable, from a man-made source, due to negligence, arising from a poorly understood cause, or directed at a vulnerable group within society.

As one of the stakeholders, the *manufacturer* reduces *risks* and makes judgments relating to the *safety* of a *medical device*, including the acceptability of *residual risks*. The *manufacturer* takes into account the generally acknowledged *state of the art*, in order to determine the suitability of a *medical device* to be placed on the market for its *intended use*. This document specifies a *process* through which the *manufacturer* of a *medical device* can identify *hazards* associated with the *medical device*, estimate and evaluate the *risks* associated with these *hazards*, control these *risks*, and monitor the effectiveness of the controls throughout the *life cycle* of the *medical device*.

The decision to use a *medical device* in the context of a particular clinical *procedure* requires the *residual risks* to be balanced against the anticipated *benefits* of the *procedure*. Such decisions are beyond the scope of this document and take into account the *intended use*, the circumstances of use, the performance and *risks* associated with the *medical device*, as well as the *risks* and *benefits* associated with the clinical *procedure*. Some of these decisions can be made only by a qualified medical practitioner with knowledge of the state of health of an individual patient or the patient's own opinion.

For any particular *medical device*, other standards or regulations could require the application of specific methods for managing *risk*. In those cases, it is necessary to also follow the requirements outlined in those documents.

The verbal forms used in this document conform to the usage described in [Clause 7](#) of the ISO/IEC Directives, Part 2:2018. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe permission (e.g. a permissible way to achieve compliance with a requirement or test);
- “can” is used to express possibility and capability; and
- “must” is used to express an external constraint that is not a requirement of the document.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 14971:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 14971:2020

<https://standards.iteh.ai/catalog/standards/sist/99598415-139c-4897-8bfe-a466d8f8598e/sist-en-iso-14971-2020>

Medical devices — Application of risk management to medical devices

1 Scope

This document specifies terminology, principles and a *process* for *risk management* of *medical devices*, including software as a *medical device* and *in vitro diagnostic medical devices*. The *process* described in this document intends to assist *manufacturers* of *medical devices* to identify the *hazards* associated with the *medical device*, to estimate and evaluate the associated *risks*, to control these *risks*, and to monitor the effectiveness of the controls.

The requirements of this document are applicable to all phases of the *life cycle* of a *medical device*. The *process* described in this document applies to *risks* associated with a *medical device*, such as *risks* related to biocompatibility, data and systems security, electricity, moving parts, radiation, and usability.

The *process* described in this document can also be applied to products that are not necessarily *medical devices* in some jurisdictions and can also be used by others involved in the *medical device life cycle*.

This document does not apply to:

- decisions on the use of a *medical device* in the context of any particular clinical *procedure*; or
- business *risk management*.

This document requires *manufacturers* to establish objective criteria for *risk* acceptability but does not specify acceptable *risk* levels.

Risk management can be an integral part of a quality management system. However, this document does not require the *manufacturer* to have a quality management system in place.

NOTE Guidance on the application of this document can be found in ISO/TR 24971^[9].

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

accompanying documentation

materials accompanying a *medical device* (3.10) and containing information for the user or those accountable for the installation, use, maintenance, decommissioning and disposal of the *medical device* (3.10), particularly regarding safe use

Note 1 to entry: The *accompanying documentation* can consist of the instructions for use, technical description, installation manual, quick reference guide, etc.