



SLOVENSKI STANDARD SIST EN IEC 60601-2-66:2020

01-julij-2020

Nadomešča:
SIST EN 60601-2-66:2016

Medicinska električna oprema - 2-66. del: Posebne zahteve za osnovno varnost in bistvene lastnosti slušnih pripomočkov in sistemov (IEC 60601-2-66:2019)

Medical electrical equipment - Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems (IEC 60601-2-66:2019)

Medizinische elektrische Geräte - Teil 2-66: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Hörgeräten und Hörgerätesystemen (IEC 60601-2-66:2019)

Appareils électromédicaux - Partie 2-66: Exigences particulières pour la sécurité de base et les performances essentielles des instruments d'audition et systèmes d'audition (IEC 60601-2-66:2019)

Ta slovenski standard je istoveten z: EN IEC 60601-2-66:2020

ICS:

11.180.15 Pripomočki za gluhe osebe in Aids for deaf and hearing
osebe z okvaro sluha impaired people

SIST EN IEC 60601-2-66:2020 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN IEC 60601-2-66:2020](https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020)

<https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020>

EUROPEAN STANDARD

EN IEC 60601-2-66

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2020

ICS 11.180.15; 17.140.50

Supersedes EN 60601-2-66:2015 and all of its
amendments and corrigenda (if any)

English Version

**Medical electrical equipment - Part 2-66: Particular requirements
for the basic safety and essential performance of hearing aids
and hearing aid systems
(IEC 60601-2-66:2019)**

Appareils électromédicaux - Partie 2-66: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils de correction auditive et des
systèmes de correction auditive
(IEC 60601-2-66:2019)

Medizinische elektrische Geräte - Teil 2-66: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Hörgeräten und
Hörgerätesystemen
(IEC 60601-2-66:2019)

This European Standard was approved by CENELEC on 2020-02-19. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 60601-2-66:2020 (E)**European foreword**

The text of document 29/1023/FDIS, future edition 3 of IEC 60601-2-66, prepared by IEC/TC 29 "Electroacoustics" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-66:2020.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2020-11-19
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2023-02-19

This document supersedes EN 60601-2-66:2015 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZZ, which is an integral part of this document.

STANDARD PREVIEW
(standards.iteh.ai)
[SIST EN IEC 60601-2-66:2020](https://standards.iteh.ai/catalog/standards/sist/60601-2-66-2020)
Endorsement notice
<https://standards.iteh.ai/catalog/standards/sist/60601-2-66-2020>
a900903c0d7f/sist-en-iec-60601-2-66-2020

The text of the International Standard IEC 60601-2-66:2019 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60118-4:2014	NOTE	Harmonized as EN 60118-4:2015 (not modified)
IEC 60318-5:2006	NOTE	Harmonized as EN 60318-5:2006 (not modified)
IEC 60601-1-9	NOTE	Harmonized as EN 60601-1-9
IEC 60601-1-10	NOTE	Harmonized as EN 60601-1-10
IEC 60645-1:2017	NOTE	Harmonized as EN 60645-1:2017 (not modified)
IEC 61000-4-2	NOTE	Harmonized as EN 61000-4-2
IEC 61000-4-8	NOTE	Harmonized as EN 61000-4-8
IEC 62489-1:2010	NOTE	Harmonized as EN 62489-1:2010 (not modified)
IEC 62489-1:2010/A1:2014	NOTE	Harmonized as EN 62489-1:2010/A1:2015 (not modified)
IEC 62489-1:2010/A2:2017	NOTE	Harmonized as EN 62489-1:2010/A2:2018 (not modified)
IEC 62366-1:2015	NOTE	Harmonized as EN 62366-1:2015 (not modified)
CISPR 11	NOTE	Harmonized as EN 55011

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Clause 2 of the general standard applies except as follows:

Replacement:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60950-1 (mod)	2005	Information technology equipment - Safety - Part 1: General requirements	EN 60950-1	2006
-	-		+ A11	2009
+ A1 (mod)	2009		+ A1	2010
-	-		+ A12	2011
+ A2 (mod)	2013		+ A2	2013
IEC 62368-1 (mod)	2014	Audio/video, information and communication technology equipment - Part 1: Safety requirements	EN 62368-1	2014
-	-		+ A11	2017

Addition:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60118-0	2015	Electroacoustics - Hearing aids - Part 0: Measurement of the performance characteristics of hearing aids	EN 60118-0	2015
IEC 60118-13	-	Electroacoustics - Hearing aids - Part 13: Electromagnetic compatibility (EMC)	EN 60118-13	2019
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
+ A1	2012		+ A1	2013
-	-		+ A12	2014

EN IEC 60601-2-66:2020 (E)

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests	EN 60601-1-2	2015
IEC 60601-1-11	2015	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2015
IEC 62304	-	Medical device software - Software life cycle processes	EN 62304 + A1	2006 2015

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN IEC 60601-2-66:2020](https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020)

<https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020>

Annex ZZ (informative)

Coverage of Essential Requirements of EU Directives

This European standard has been prepared under a Commission's standardisation request M/023 and M432 concerning the development of European Standards related to medical devices given to CENELEC by the European Commission and the European Free Trade Association to provide a means of conforming to the Essential Requirements given in Annex I of the EU Directives 93/42/EEC as amended by 2007/47/EC.

General Guidance:

Once this standard will be cited in the Official Journal of the European Union under that Directive, compliance with the clauses of this standard given in Table ZZ.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements (ERs) of that Directive and associated EFTA regulations.

NOTE 1 This standard is intended to be applied in its entirety only. Selected clauses or subclauses may be not applicable due to the specific type of equipment under consideration. It is necessary to understand and apply Clauses 1 to 5. It is also recommended to understand and apply those clauses which contain general requirements related to a specific subclause. Elements of the standard that are not cited in Table ZZ.1 may be relevant for the appropriate fulfilment of certain essential requirements through indirect reference, and for safety and performance aspects of the device, that are not addressed through essential requirements.

NOTE 2 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance to the MDD (Directive 93/42/EEC amended by 2007/47/EC). This means that risks have to be reduced "as far as possible", "to a minimum", "to the lowest possible level", "minimized" or "removed", according to the wording of the corresponding essential requirement.

NOTE 3 With respect to Note 4 of 4.2.2 General requirement for risk management, the manufacturer's policy for determining **acceptable risk** must be in compliance with essential requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the directive.

NOTE 4 References in the Clauses 3 to 17 or in the Annexes of this standard specify whether the normative references listed in Clause 2 as cited in Annex ZA are to be applied in whole or in part.

NOTE 5 This Annex ZZ is based on Normative References according to Annex ZA, replacing the references in the core text.

NOTE 6 According to the scope of this standard the coverage in Table ZZ.1 only applies to the design and construction of HEARING AIDS and HEARING AID SYSTEMS. This European Standard lists in Table ZZ.1 only the essential requirements covered.

WARNING: Other requirements and other EU Directives and Regulations may be applicable to the product(s) falling within the scope of this standard.

EN IEC 60601-2-66:2020 (E)

Table ZZ.1 – Relationship between Essential Requirements of Directive 93/42/EEC amended by 2007/47/EC, and clauses and subclauses of this standard

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
1, second indent	201.7.9.1 201.7.9.2.1 201.7.9.2.2	This document (201.7.9.1, 201.7.9.2.1, 201.7.9.2.2) covers requirements related to instructions for use, including safety warnings and notices by considering the technical knowledge, experience, education and training and where applicable the medical and physical conditions of intended users (design for lay, professional, disabled or other users).
5	201.7.2.17 201.7.9.2.2 201.15.3 201.15.3.7	Covered by requirements to design and packaging (201.7.2.17) to withstand transport and storage with regards of mechanical strength (201.15.3), resistance to environmental conditions (201.15.3.7) and the necessary instructions (201.7.9.2.2).
6	201.7.9.2.1 201.9.6	The reduction of unintentional exposure to excessive acoustic noise is covered in 201.9.6 regarding the design. 201.7.9.2.1 covers information and instructions for the user related to side effects.
7.1, first indent	201.11.1.1 201.13.1.2	Covered for risks of fire and high temperatures.
7.3	201.11.6.6 201.15.3.7	This document covers (201.15.3.7, 201.11.6.6) the design of devices in such a way that they can be used safely with the materials, substances and gases with which they enter into contact during their normal use or during routine procedures. Requirements for HEARING AIDS that are intended to be used in explosive and oxygenenriched atmospheres are not contained in this standard (see 201.11.2).
7.5, first sentence of first paragraph only	201.7.9.2.4	Covered for the risk of leakage from the battery in situ.
7.6	201.11.6.5	Covered in 201.11.6.5.
8.1, first sentence only	201.7.9.2.12 201.11.6.6 201.12.2	Design covered in 201.12.2, 201.11.6.6 and instruction covered in 201.7.9.2.12.
9.1	201.5.5 c) 201.6.2 201.7.9.2.5 201.7.9.2.9 201.8.1 201.8.2.1 201.8.4.2	

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
9.2, first indent	201.9	Mechanical risks (e.g. 'entanglement') are covered by 201.9.
9.2, second indent	201.5.3 201.5.7 201.7.2.17 201.7.9.2.1 201.7.9.2.2 201.15.3.7	Covered in respect of environmental temperatures, humidity or variations in pressure: 201.5.3, 201.5.7, 201.7.2.17, 201.7.9.2.1, 201.7.9.2.2, 201.15.3.7 of this document provide design and test requirements with regards to climatic environmental conditions which are suitable to remove or minimize as far as possible these risks.
9.2, third indent	201.7.9.2.2 201.17	To fully cover this ER, risks must be removed or minimised as far as possible.
9.2, fourth indent	201.15.2	Covered for the serviceability of hearing aids subject to mechanical wear, electrical degradation or ageing.
9.3	201.11.1.1 201.13.1.2	Risks of fire and high temperatures covered in 201.11.1.1, 201.13.1.2. HEARING AIDS are normally not exposed to flammable substances or to substances which could cause combustion. The requirements for HEARING AIDS that are intended to be used in explosive and oxygenenriched atmospheres are not covered in this document.
12.5	201.17	To fully cover this ER, risks must be removed or minimised as far as possible.
12.6	201.8 201.13	Electrical risks covered in 201.8 for normal conditions and 201.13 in fault conditions.
12.7.1	201.15.3.1	Covered by the mechanical requirements in 201.15.3.1.
12.7.3	201.7 201.9.6 201.13.1.2	The reduction of unintentional exposure to excessive acoustic noise is covered in 201.9.6 regarding the design, in 201.7 regarding correct application 201.13.1.2 in case of faults.
12.7.4	201.8.1 201.8.2.1 201.8.7 201.16	Covered in respect of the following: Electrical Risks: 201.8.1 Fundamental rule of protection against electric shock 201.8.2.1 Connection to power sources 201.16 Limitation of voltage current or energy 201.8.7 Leakage current. Gas or hydraulic and pneumatic energy supplies not applicable to HEARING AIDS.
12.7.5	201.11.1	Covered by 201.11.1 Excessive temperatures

EN IEC 60601-2-66:2020 (E)

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
12.8.1	201.9.6 201.12.4.4	Requirements of sufficient accurate indication of output is covered in 201.9.6. 201.12.4.4 covers possible sources of incorrect output.
12.8.2	201.9.6	The reduction of unintentional exposure to excessive acoustic noise is covered in 201.9.6
12.9	201.7.9.1 201.7.9.2.1 201.7.9.2.9	Requirements to specification of function of controls and indicators on the device or in the instructions for use are covered in 201.7.9.2.1 and 201.7.9.2.9.
13.1 – first and second paragraph	201.7.9.2 201.7.9.2.16	Covered in respect of the following: 201.7.9.2 Instructions for use. 201.7.9.2.16 Technical description, provided that the technical description is included in the instructions for use.
13.1 - first and third paragraph	201.7.2.2 201.7.2.17 201.7.9.2	Covered in respect of information on the packaging in 201.7.2.17. Covered in respect of information in the instruction for use in 201.7.9.2.
13.1 - first and fourth paragraph	201.7.9.2	Requirements for instruction for use covered in 201.7.9.2.
13.2	201.7.8.1 201.7.9.2.9	Description of symbols in the documentation covered in 201.7.9.2.9. Description of colours in the documentation covered in 201.7.8.1.
13.3 a)	201.7.2.2	To fully cover this ER, the name and address of the authorised representative must be provided, if applicable.
13.3 d)	201.7.2.2	To fully cover this ER the batch number (if provided) must be preceded by the symbol LOT.
13.3 i)	201.7.2.17	
13.3 j)	201.7.2	
13.3 l)	201.7.2.2	
13.4	201.7.9.1	Covered for the Instructions for Use.
13.5	201.7.2.2	Serial number required in 201.7.2.2.

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
13.6 b)	201.7.9.3	Covered provided the technical description is included in the instructions for use.
13.6 d)	201.7.9.2.1 201.7.9.2.12	
13.6 k)	201.7.9.2.1	
13.6 l)	201.7.9.2.1	

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN IEC 60601-2-66:2020](https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020)

<https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN IEC 60601-2-66:2020

<https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020>



IEC 60601-2-66

Edition 3.0 2019-10

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
**Part 2-66: Particular requirements for the basic safety and essential
performance of hearing aids and hearing aid systems**

Appareils électromédicaux –
**Partie 2-66: Exigences particulières pour la sécurité de base et les
performances essentielles des appareils de correction auditive et des systèmes
de correction auditive**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.180.15; 17.140.50

ISBN 978-2-8322-7350-0

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD	4
INTRODUCTION	7
201.1 Scope, object and related standards	8
201.2 Normative references	10
201.3 Terms and definitions	11
201.4 General requirements	12
201.5 General requirements for testing ME EQUIPMENT	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	15
201.7 ME EQUIPMENT identification, marking and documents	15
201.8 * Protection against electrical HAZARDS from ME EQUIPMENT	21
201.9 * Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10 * Protection against unwanted and excessive radiation HAZARDS	25
201.11 * Protection against excessive temperatures and other HAZARDS	25
201.12 * Accuracy of controls and instruments and protection against hazardous outputs	27
201.13 * HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	28
201.14 * PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	30
201.15 * Construction of ME EQUIPMENT	31
201.16 * ME SYSTEMS	33
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	33
Annexes	35
Annex E (informative) Examples of the connection of the measuring device (MD) for measurement of the PATIENT LEAKAGE CURRENT and PATIENT AUXILIARY CURRENT	36
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures	37
Annex H (informative) PEMS structure, PEMS DEVELOPMENT LIFE-CYCLE and documentation	38
Annex I (informative) ME SYSTEMS aspects	39
Annex J (informative) Survey of insulation paths	40
Annex K (informative) Simplified PATIENT LEAKAGE CURRENT diagrams	41
Annex L (normative) Insulated winding wires for use without interleaved insulation	42
Annex AA (informative) Particular guidance and rationale	43
Annex BB (informative) Abbreviations	48
Annex CC (informative) Essential performance	49
Annex DD (informative) Electromagnetic compatibility of ME EQUIPMENT	50
Bibliography	52
Index of defined terms used in this particular standard	54
Figure 201.101 – Measuring circuit for LEAKAGE CURRENT (see 201.8.7.4.7)	22
Table 201.101 – MECHANICAL HAZARDS to be considered	24
Table AA.101 – Summary of the approach of this document	44

Table DD.101 – Example of applicable standards for HEARING AID specific setting for RF radiated emissions	50
Table DD.102 – Example of applicable tests	51

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN IEC 60601-2-66:2020](https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020)

<https://standards.iteh.ai/catalog/standards/sist/0892194b-e752-4c37-ad12-a900903c0d7f/sist-en-iec-60601-2-66-2020>