



Designation: ~~D3559-03~~ Designation: D 3559 – 08

Standard Test Methods for Lead in Water¹

This standard is issued under the fixed designation D 3559; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope *

1.1 These test methods cover the determination of dissolved and total recoverable lead in water and waste water by atomic-absorption spectrophotometry² and differential pulse anodic stripping voltammetry. Four test methods are included as follows:

	Concentration Range	Sections
Test Method A—Atomic Absorption, Direct	1.0 to 10 mg/L	7 to 15
Test Method B—Atomic Absorption, Chelation-Extraction	100 to 1000 $\mu\text{g/L}$	16 to 24
Test Method C—Differential Pulse Anodic Stripping Voltammetry	1 to 100 $\mu\text{g/L}$	25 to 35
Test Method D—Atomic Absorption, Graphite Furnace	5 to 100 $\mu\text{g/L}$	36 to 44

1.2 Test Method B can be used to determine lead in brines. Test Method D has been used successfully with reagent water, lake water, well water, filtered tap water, condensate from a medium Btu coal gasification process, waste treatment plant effluent, and a production plant process water.

1.3 It is the user's responsibility to ensure the validity of these test methods for waters of untested matrices.

1.4

1.4 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazards statements, see 10.4.1, Note 2, 11.2, 11.3, 21.7, 21.8, 21.11, 23.7, 23.10, 32.2.1, and 33.1.

2. Referenced Documents

2.1 *ASTM Standards:*³ <http://www.astm.org/catalog/standards/sist/ecaa/c47-530d-4d2d-b798-84d12cf6f62f/astm-d3559-08>

D 858 Test Methods for Manganese in Water

D 1066 Practice for Sampling Steam

D 1068 Test Methods for Iron in Water

D 1129 Terminology Relating to Water³

D 1192 Specification for Equipment for Sampling Water and Steam in Closed Conduits³ Terminology Relating to Water

D 1193 Specification for Reagent Water

D 1687 Test Methods for Chromium in Water

D 1688 Test Methods for Copper in Water

D 1691 Test Methods for Zinc in Water

D 1886 Test Methods for Nickel in Water

D 2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D-19 D19 on Water

D 3370 Practices for Sampling Water from Closed Conduits

¹ These test methods are under the jurisdiction of ASTM Committee D19 on Water and are the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved Jan. 10, 2003. Published January 2003. Originally approved in 1977. Last previous edition approved in 1996 as D3559-96.

Current edition approved Oct. 1, 2008. Published October 2008. Originally approved in 1977. Last previous edition approved in 2003 as D 3559 – 03.

² Platte, J. A., and Marcy, V. M., "A New Tool for the Water Chemist," *Industrial Water Engineering*, May 1965 .

Brown, E., Skougstad, M. W., and Fishman, M. J., "Methods for Collection and Analysis of Water Samples for Dissolved Minerals and Gases," *Techniques of Water-Resources Investigations of the U. S. Geological Survey*, Book 5, Chapter 7, 1970, p. 115.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards*, Vol 11.01, volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard.

- D 3557 Test Methods for Cadmium in Water
- D 3558 Test Methods for Cobalt in Water
- D 3919 Practice for Measuring Trace Elements in Water by Graphite Furnace Atomic Absorption Spectrophotometry
- D 4841 Practice for Estimation of Holding Time for Water Samples Containing Organic and Inorganic Constituents
- D 5810 Guide for Spiking into Aqueous Samples
- D 5847 Practice for the Writing Quality Control Specifications for Standard Test Methods for Water Analysis
- ~~E60 Practice for Photometric and Spectrophotometric Methods for Chemical Analysis of Metals~~ 60 Practice for Analysis of Metals, Ores, and Related Materials by Molecular Absorption Spectrometry
- E 275 Practice for Describing and Measuring Performance of Ultraviolet, Visible, and Near-Infrared Spectrophotometers

3. Terminology

3.1 *Definitions*—For definition of terms used in these test methods, refer to Terminology D 1129.

3.2 *total recoverable lead*—an arbitrary analytical term relating to the recoverable forms of lead that are determined by the digestion method which are included in the procedure.

4. Significance and Use

4.1 The test for lead is necessary because it is a toxicant and because there is a limit specified for lead in potable water in the National Interim Primary Drinking Water Regulations. This test serves to determine whether the lead content of potable water is above or below the acceptable limit.

5. Purity of Reagents

5.1 Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society.⁴ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D 1193, Type I. Other reagent water types may be used provided it is first ascertained that the water is of sufficiently high purity to permit its use without adversely affecting the bias and precision of the test method. Type II water was specified at the time of round-robin testing of this test method.

6. Sampling

6.1 Collect the samples in accordance with Practice D 1066, ~~Specification D 1192, and Practices D 3370 and Practices D 3370,~~ as applicable.

6.2 Samples shall be preserved with HNO₃ (sp gr 1.42) to a pH of 2 or less immediately at the time of collection, normally about 2 mL/L of HNO₃. If only dissolved lead is to be determined, the sample shall be filtered through a 0.45- μ m membrane filter before acidification.

TEST METHOD A—ATOMIC ABSORPTION, DIRECT

7. Scope

7.1 This test method covers the determination of dissolved and total recoverable lead in most waters and wastewaters.

7.2 The test method is applicable in the range from 1.0 to 10 mg/L of lead. The upper limits of detectability can be increased to concentrations greater than 10 mg/L by dilution of the sample.

8. Summary of Test Method

8.1 Lead is determined by atomic absorption spectrophotometry. Dissolved lead is determined by aspirating the filtered and preserved sample directly with no pretreatment. Total recoverable lead is determined by aspirating the sample following hydrochloric-nitric acid digestion and filtration. The same digestion procedure may be used to determine total recoverable cadmium (Test Methods D 3557), chromium (Test Methods D 1687), cobalt (Test Methods D 3558), copper (Test Methods D 1688), iron (Test Methods D 1068), manganese (Test Methods D 858), nickel (Test Methods D 1886), and zinc (Test Methods D 1691).

9. Interferences

9.1 Other metals usually do not interfere in the determination of lead by increasing or decreasing the amount of absorbed radiation. The most common interference is caused by a chemical reaction in the flame that prevents conversion of the lead to the atomic state.

⁴ *Annual Book of ASTM Standards*, Vol 03.05-Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

9.2 High concentrations of calcium, such as those connected with the coal industry, will give lead concentrations higher than which actually exist. This can be corrected by using a background correction technique, or by the chelation-extraction procedure (Test Method B).

9.2.1 The equipment manufacturer's instructions for use of specific correction technique shall be followed.

10. Apparatus

10.1 *Atomic Absorption Spectrophotometer*, for use at 283.3 nm.

NOTE 1—The manufacturer's instructions shall be followed for all instrumental parameters. Wavelengths other than 283.3 nm may be used if they have been determined to be equally suitable.

10.2 *Lead Light Source*, hollow-cathode lamps or electrodeless-discharge lamps have been found satisfactory.

10.3 *Oxidant*:

10.3.1 *Air*, which has been passed through a suitable filter to remove oil, water, and other foreign substances, is the usual oxidant.

10.4 *Fuel*:

10.4.1 *Acetylene*—Standard, commercially available acetylene is the usual fuel. Acetone, always present in acetylene cylinders, can affect analytical results. The cylinder should be replaced at 50 psig (345 kPa).

10.4.1.1 **Warning:**—“Purified” grade acetylene containing a special proprietary solvent rather than acetone should not be used with poly(vinyl chloride) tubing as weakening of the walls can cause a potential hazardous situation.

10.5 *Pressure-Reducing Valves*—The supplies of fuel and oxidant shall be maintained at pressures somewhat higher than the controlled operating pressure of the instrument by suitable valves.

11. Reagents

11.1 *Hydrochloric Acid (sp gr 1.19)*—Concentrated hydrochloric acid (HCl).

NOTE 2—If the reagent blank concentration is greater than the method detection limit, distill the HCl or use a spectrograde acid. **Caution/Warning:**—~~When~~ ~~When~~ HCl is distilled, an azeotropic mixture is formed (approximately 6 N HCl is formed). Therefore, whenever concentrated HCl is used in the preparation of a reagent or in the procedure, use double the volume specified if distilled HCl is used.

11.2 *Lead Solution, Stock (1 mL = 1 mg lead)*—~~Dissolve 1.5999 g of lead nitrate (Pb(NO₃)₂)~~—Commercially purchase or dissolve 1.5999 g of lead nitrate (Pb(NO₃)₂) in a mixture of 10 mL of HNO₃ (sp gr 1.42) and 100 mL of water. Dilute to 1 L with water. **Warning:**—~~Lead~~ ~~Lead~~ salts are toxic. Handle with care and avoid personal contamination.

11.3 *Lead Solution, Standard (1 mL = 0.1 mg lead)*—Dilute 100.0 mL of stock lead solution to 1 L with HNO₃(1 + 499). Store all solutions in polyethylene bottles. **Warning:**—~~Lead~~ ~~Lead~~ salts are toxic. Never pipette by mouth. Pipette with the end of a suction device or employ other conventional means of quantitative measurement.

11.4 *Nitric Acid (sp gr 1.42)*—Concentrated nitric acid (HNO₃).

NOTE 3—If the reagent blank concentration is greater than the method detection limit, distill the HNO₃ or use a trace metal grade acid.

11.5 *Nitric Acid (1 + 499)*—Add 1 volume of HNO₃(sp gr 1.42) to 499 volumes of water.

12. Standardization

12.1 Prepare 100 mL each of a blank and at least four standard solutions to bracket the expected lead concentration range to be analyzed by diluting the lead standard solution (11.3) with HNO₃(1 + 499). Prepare the standards each time the test is to be performed.

12.2 When determining total recoverable lead, add 0.5 mL of HNO₃(sp gr 1.42) to each blank and standard solution and proceed as directed in 13.2 through 13.4. After the digestion of the blank and standard solutions has been completed in 13.4, return to 12.3 to complete the standardization for total recoverable determinations. When determining dissolved lead, proceed with 12.3.

12.3 Aspirate the blank and standards and record the instrument readings. Aspirate HNO₃(1 + 499) between standards.

12.4 Prepare an analytical curve by plotting the absorbance versus the concentration for each standard on linear graph paper. Alternatively, read directly in concentration if this capability is provided with an instrument.

13. Procedure

13.1 Measure 100.0 mL of a well-mixed acidified sample into a 125-mL beaker or flask.

NOTE 4—If only dissolved lead is to be determined, start with 13.5.

13.2 Add 5 mL of HCl (sp gr 1.19) to each sample.

13.3 Heat the samples on a steam bath or hot plate in a well-ventilated hood until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil.

NOTE 5—For samples having appreciable amounts of suspended matter or dissolved matter, the amount of reduction in volume is left to the discretion of the analyst.

13.4 Cool and filter the samples through a suitable filter such as fine-textured, acid washed, ashless paper, into 100-mL

volumetric flasks. Wash the filter paper two or three times with water and adjust to volume.

13.5 Aspirate each filtered and acidified sample and determine its absorbance or concentration at 283.3 nm. Aspirate HNO₃ (1 + 499) between samples.

14. Calculation

14.1 Calculate the concentration of lead in each sample, in milligrams per litre, using the calibration curve established in 12.4.

15. Precision and Bias ⁵

15.1 Fourteen operators from nine laboratories participated in this study. One operator performed the analysis in quadruplicate, twelve in triplicate and one in duplicate at each concentration level.

15.2 The bias of this test method for lead is listed in Table 1.

15.3 These data may not apply to waters of other matrices.

15.4 This section on precision and bias conforms to Practice D 2777 – 77 which was in place at the time of collaborative testing. Under the allowances made in 1.4 of D 2777–98, –06, these precision and bias data do meet existing requirements of interlaboratory studies of Committee D19 test methods.

16. Quality Control

16.1 In order to be certain that analytical values obtained using these test methods are valid and accurate within the confidence limits of the test, the following QC procedures must be followed when analyzing lead.

16.2 *Calibration and Calibration Verification:*

16.2.1 Analyze at least ~~three~~four working standards containing concentrations of lead that bracket the expected sample concentration, prior to analysis of samples, to calibrate the instrument. The calibration correlation coefficient shall be equal to or greater than 0.990. In addition to the initial calibration blank, a calibration blank shall be analyzed at the end of the batch run to ensure contamination was not a problem during the batch analysis.

16.2.2 Verify instrument calibration after standardization by analyzing a standard at the concentration of one of the calibration standards. The concentration of a mid-range standard should fall within $\pm 15\%$ of the known concentration.

16.2.3 If calibration cannot be verified, recalibrate the instrument.

16.3 *Initial Demonstration of Laboratory Capability:*

16.3.1 If a laboratory has not performed the test before, or if there has been a major change in the measurement system, for example, new analyst, new instrument, etc., a precision and bias study must be performed to demonstrate laboratory capability.

16.3.2 Analyze seven replicates of a standard solution prepared from an Independent Reference Material containing a mid-range concentration of lead. The matrix and chemistry of the solution should be equivalent to the solution used in the collaborative study. Each replicate must be taken through the complete analytical test method including any sample preservation and pretreatment steps. ~~The replicates may be interspersed with samples.~~

16.3.3 Calculate the mean and standard deviation of the seven values and compare to the acceptable ranges of bias in Table 1. This study should be repeated until the recoveries are within the limits given in Table 1. If a concentration other than the recommended concentration is used, refer to Practice D 5847 for information on applying the *F* test and *t* test in evaluating the acceptability of the mean and standard deviation.

16.4 *Laboratory Control Sample (LCS) :*

16.4.1 To ensure that the test method is in control, analyze a LCS containing a known concentration of lead with each batch or 10 samples. If large numbers of samples are analyzed in the batch, analyze the LCS after every 10 samples. The laboratory

⁶ Annual Book of ASTM Standards, Vol 03.06.

⁵ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR: D19-1030.

TABLE 1 Determination of Bias, Direct

Amount Added, mg/L	Amount Found, mg/L	S _r , mg/L	S _o , mg/L	Bias, %	Statistically Significant (95 % Confidence Level)
Reagent Water, Type II					
1	1.01	0.08	0.04	+1.00	no
6	6.01	0.28	0.14	+0.17	no
8	8.02	0.34	0.14	+0.25	no
Selected Water Matrices					
1	1.00	0.00	0.06	0.00	no
6	6.11	0.25	0.16	+1.83	yes
8	7.99	0.36	0.23	-0.13	no

control samples for a large batch should cover the analytical range when possible. The LCS must be taken through all of the steps of the analytical method including sample preservation and pretreatment. The result obtained for a mid-range LCS shall fall within $\pm 15\%$ of the known concentration.

16.4.2 If the result is not within these limits, analysis of samples is halted until the problem is corrected, and either all the samples in the batch must be reanalyzed, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

16.5 *Method Blank:*

16.5.1 Analyze a reagent water test blank with each batch. The concentration of lead found in the blank should be less than 0.5 times the lowest calibration standard. If the concentration of lead is found above this level, analysis of samples is halted until the contamination is eliminated, and a blank shows no contamination at or above this level, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

16.6 *Matrix Spike (MS):*

16.6.1 To check for interferences in the specific matrix being tested, perform a MS on at least one sample from each batch by spiking an aliquot of the sample with a known concentration of lead and taking it through the analytical method.

16.6.2 The spike concentration plus the background concentration of lead must not exceed the high calibration standard. The spike must produce a concentration in the spiked sample that is 2 to 5 times the analyte concentration in the unspiked sample, or 10 to 50 times the detection limit of the test method, whichever is greater.

16.6.3 Calculate the percent recovery of the spike (P) using the following formula:

$$P = 100 [A(V_s + V) - B V_s] / C V \quad (1)$$

where:

A = analyte concentration (mg/L) in spiked sample,

B = analyte concentration (mg/L) in unspiked sample,

C = concentration (mg/L) of analyte in spiking solution,

V_s = volume (mL) of sample used, and

V = ~~volume (mL) added with spike.~~ volume (mL) of spiking solution added.

16.6.4 The percent recovery of the spike shall fall within the limits, based on the analyte concentration, listed in Guide D 5810, Table 1. If the percent recovery is not within these limits, a matrix interference may be present in the sample selected for spiking. Under these circumstances, one of the following remedies must be employed: the matrix interference must be removed, all samples in the batch must be analyzed by a test method not affected by the matrix interference, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

NOTE 6—Acceptable spike recoveries are dependent on the concentration of the component of interest. See Guide D 5810 for additional information.

16.7 *Duplicate:*

16.7.1 To check the precision of sample analyses, analyze a sample in duplicate with each batch. If the concentration of the analyte is less than five times the detection limit for the analyte, a matrix spike duplicate (MSD) should be used.

16.7.2 Calculate the standard deviation of the duplicate values and compare to the precision in the collaborative study using an F test. Refer to 6.4.4 of Practice D 5847 for information on applying the F test.

16.7.3 If the result exceeds the precision limit, the batch must be reanalyzed or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

16.8 *Independent Reference Material (IRM):*

16.8.1 In order to verify the quantitative value produced by the test method, analyze an Independent Reference Material (IRM) submitted as a regular sample (if practical) to the laboratory at least once per quarter. The concentration of the IRM should be in the concentration mid-range for the method chosen. The value obtained must fall within the control limits established by the laboratory.

TEST METHOD B—ATOMIC ABSORPTION, CHELATION-EXTRACTION

17. Scope

17.1 This test method covers the determination of dissolved and total recoverable lead in most waters and brines.

17.2 This test method is applicable in the range from 100 to 1000 $\mu\text{g/L}$ of lead. The range may be extended upward by dilution of the samples.

18. Summary of Test Method

18.1 Lead is determined by atomic absorption spectrophotometry. The element, either dissolved or total recoverable, is chelated with pyrrolidine dithiocarbamic acid and extracted with chloroform. The extract is evaporated to dryness, treated with hot HCl and diluted to a specified volume with water. The resulting solution is then aspirated into the air-acetylene flame of the spectrophotometer. The digestion procedure summarized in 8.1 is used for total recoverable lead. The same chelation-extraction procedure may be used to determine total recoverable cadmium (Test Methods D 3557), cobalt (Test Methods D 3558), copper (Test Methods D 1688), iron (Test Methods D 1068), nickel (Test Methods D 1886), and zinc (Test Methods D 1691).

19. Interferences

19.1 See Section 9.

20. Apparatus

20.1 All apparatus described in Section 10 are required.

21. Reagents

21.1 *Bromphenol Blue Indicator Solution* (1 g/L)—Dissolve 0.1 g of bromphenol blue in 100 mL of 50 % ethanol or 2-propanol.

21.2 *Carbon Disulfide* (CS₂).

21.3 *Chloroform* (CHCl₃).

21.4 *Hydrochloric Acid* (sp gr 1.19)—Concentrated hydrochloric acid (HCl).

21.5 *Hydrochloric Acid* (1 + 2)—Add 1 volume of HCl (sp gr 1.19) to 2 volumes of water.

21.6 *Hydrochloric Acid* (1 + 49)—Add 1 volume of HCl (sp gr 1.19) to 49 volumes of water.

21.7 *Lead Solution, Stock* (1.0 mL = 200 µg lead)—~~Dissolve~~ Commercially purchase or dissolve 0.3198 g of lead nitrate (Pb(NO₃)₂) in water containing 1 mL of HNO₃ (sp gr 1.42) and dilute to 1 L with water. **Warning:**—~~Lead~~ Lead salts are toxic. Handle with care and avoid personal contamination.

21.8 *Lead Solution, Intermediate* (1.0 mL = 2.0 µg lead)—Dilute 10 mL of lead stock solution and 1 mL of HNO₃ (sp gr 1.42) to 1 L with water. **Warning:**—~~Lead~~ Lead salts are toxic. Never pipette by mouth. Pipette with the end of a suction device or employ other convenient means of quantitative measurement.

21.9 *Lead Solution, Standard* (1.0 mL = 0.2 µg lead)—Immediately before use, dilute 10.0 mL of lead intermediate solution and 1 mL of HNO₃ (sp gr 1.42) to 100 mL with water. This standard is used to prepare working standards at the time of analysis.

21.10 *Nitric Acid* (sp gr 1.42)—Concentrated nitric acid (HNO₃).

21.11 *Pyrrolidine Dithiocarbamic Acid-Chloroform Reagent*—Add 36 mL of pyrrolidine to 1 L of CHCl₃. Cool the solution and add 30 mL of CS₂ in small portions, swirling between additions. Dilute to 2 L with CHCl₃. The reagent can be used for several months if stored in a cool, dark place. **Warning:** All components of this reagent are highly toxic. Carbon disulfide is also highly flammable, prepare and use in a well-ventilated hood. Avoid inhalation and direct contact.

21.12 *Sodium Hydroxide Solution* (100 g/L)—~~Dissolve 100 g of sodium hydroxide (NaOH) in water and dilute to 1 L.~~ Dissolve 100 g of sodium hydroxide (NaOH) in water and dilute to 1 L. **Warning:** This is a very exothermic reaction.

22. Standardization

22.1 Prepare a blank and sufficient standards from 0.0 to 1000 µg/L lead from the lead standard solution (21.9) by making appropriate dilutions in water. Prepare standards immediately prior to use.

22.2 When determining total recoverable lead use 125-mL beakers or flasks, add 0.5 mL HNO₃ (sp gr 1.42) and proceed as directed in 23.2 thru 23.15. When determining dissolved lead use 250-mL separatory funnels and proceed as directed in 23.5 thru 23.15.

22.3 Construct an analytical curve by plotting the absorbances of standards versus micrograms of lead. Alternatively, read directly in concentration if this capability is provided with the instrument.

23. Procedure

23.1 Measure a volume of a well-mixed acidified sample containing less than 100 µg lead (100-mL maximum) into a 125-mL beaker or flask and adjust the volume to 100 mL with water.

NOTE 7—If only dissolved lead is to be determined, measure a volume of filtered and acidified sample containing less than 100 µg of lead (100-mL maximum) into a 250-mL separatory funnel, and start with 23.5.

23.2 Add 5 mL of HCl (sp gr 1.19) to each sample.

23.3 Heat the samples on a steam bath or hot plate in a well-ventilated hood until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil.

NOTE 8—When analyzing brines and samples containing appreciable amounts of suspended matter, the amount of reduction in volume is left to the discretion of the analyst.

23.4 Cool and filter the samples through a suitable filter such as fine-textured, acid-washed, ashless paper, into 250-mL separatory funnels. Wash the filter paper two or three times with water and bring to approximately a 100-mL volume.

23.5 Add 2 drops of bromphenol blue indicator solution and mix.

23.6 Adjust the pH by addition of NaOH (100 g/L) solution until a blue color persists. Add HCl (1 + 49) by drops until the blue color just disappears; then add 2.5 mL of HCl (1 + 49) in excess. The pH at this point should be 2.3.

NOTE 9—The pH adjustment in 23.6 may be made with a pH meter instead of using an indicator.

23.7 Add 10 mL of pyrrolidine dithiocarbamic acid-chloroform reagent and shake vigorously for 20 min (**Warning**—See Warning: See 21.11).

23.8 Plug the tip of the separatory funnel with cotton, allow the phases to separate, and drain the chloroform phase into a 100-mL beaker.

23.9 Repeat the extraction with 10 mL of chloroform and drain the chloroform layer into the same beaker.

NOTE 10—If color still remains in the CHCl_3 extract, reextract the aqueous phase until the chloroform layer is colorless.

23.10 Place the beaker on a hot plate at low heat and evaporate just to near dryness. Remove beaker from heat and allow residual solvent to evaporate without further heating. ~~Warning—Perform in a well-ventilated hood.~~ **Warning:** Perform in a well-ventilated hood.

23.11 Hold the beaker at a 45° angle and slowly add dropwise 2 mL of HNO_3 (sp gr 1.42), rotating the beaker to effect thorough contact of the acid with the residue.

23.11.1 If acid is added to the beaker in a vertical position, a violent reaction will occur accompanied by high heat and spattering.

23.12 Place the beaker on a hot plate at low heat and evaporate just to dryness.

23.13 Add 2 mL of HCl (1 + 2) to the beaker and heat while swirling for 1 min.

NOTE 11—If a precipitate appears when the HCl (1 + 2) is added to the dried residue, obtain a fresh supply of pyrrolidine which has a different lot number or redistill the pyrrolidine just before preparing the pyrrolidine dithiocarbamic acid-chloroform reagent.

23.14 Cool and quantitatively transfer the solution to a 10-mL volumetric flask and bring to volume with water.

23.15 Aspirate each sample and record the scale reading or concentration.

24. Calculation

24.1 Determine the weight of lead in each sample by referring to 22.3. Calculate the concentration of lead in micrograms per litre as follows:

$$\text{Lead, } \mu\text{g/L} = (1000/A) \times B \quad (2)$$

where:

$1000 = 1000 \text{ mL / Liter}$

$A = \text{volume of original sample, mL, and}$

$B = \text{weight of lead in sample, } \mu\text{g.}$

25. Precision and Bias

25.1 Seven operators from six laboratories participated in this study. Five operators performed the analysis in triplicate and two in duplicate at each concentration level.

25.2 The bias of this test method for lead is listed in Table 2.

25.3 These data may not apply to waters of other matrices.

25.4 This section on precision and bias conforms to Practice D 2777 – 77, which was in place at the time of collaborative testing. Under the allowances made in 1.4 of ~~D 2777 – 98~~ D 2777 – 06, these precision and bias data do meet existing requirements of interlaboratory studies of Committee D19 test methods.

26. Quality Control

26.1 In order to be certain that analytical values obtained using these test methods are valid and accurate within the confidence limits of the test, the following QC procedures must be followed when analyzing lead.

26.2 *Calibration and Calibration Verification:*

26.2.1 Analyze at least three working standards containing concentrations of lead that bracket the expected sample concentration, prior to analysis of samples, to calibrate the instrument. The calibration correlation coefficient shall be equal to or greater than 0.990. In addition to the initial calibration blank, a calibration blank shall be analyzed at the end of the batch run to ensure contamination was not a problem during the batch analysis.

TABLE 2 Determination of Bias, Chelation-Extraction

Amount Added, $\mu\text{g/L}$	Amount Found, $\mu\text{g/L}$	S_T , $\mu\text{g/L}$	S_o , $\mu\text{g/L}$	Bias, %	Statistically Significant (95 % Confidence Level)
Reagent Water Type II					
100	86.1	17.7	8	-13.9	yes
400	364	55	27	-9.0	yes
800	674	124	24	-15.8	yes
Selected Water Matrices					
100	83	20	6.5	-17	yes
400	352	51	21	-12	yes
800	669	78	50	-16	yes

26.2.2 Verify instrument calibration after standardization by analyzing a standard at the concentration of one of the calibration standards. The concentration of a mid-range standard should fall within $\pm 15\%$ of the known concentration.

26.2.3 If calibration cannot be verified, recalibrate the instrument.

26.3 *Initial Demonstration of Laboratory Capability:*

26.3.1 If a laboratory has not performed the test before, or if there has been a major change in the measurement system, for example, new analyst, new instrument, etc., a precision and bias study must be performed to demonstrate laboratory capability.

26.3.2 Analyze seven replicates of a standard solution prepared from an Independent Reference Material containing a mid-range concentration of lead. The matrix and chemistry of the solution should be equivalent to the solution used in the collaborative study. Each replicate must be taken through the complete analytical test method including any sample preservation and pretreatment steps. ~~The replicates may be interspersed with samples.~~

26.3.3 Calculate the mean and standard deviation of the seven values and compare to the acceptable ranges of bias in Table 2. This study should be repeated until the recoveries are within the limits given in Table 2. If a concentration other than the recommended concentration is used, refer to Practice D 5847 for information on applying the F test and t test in evaluating the acceptability of the mean and standard deviation.

26.4 *Laboratory Control Sample (LCS) :*

26.4.1 To ensure that the test method is in control, analyze a LCS containing a known concentration of lead with each batch or 10 samples. If large numbers of samples are analyzed in the batch, analyze the LCS after every 10 samples. The laboratory control samples for a large batch should cover the analytical range when possible. The LCS must be taken through all of the steps of the analytical method including sample preservation and pretreatment. The result obtained for a mid-range LCS shall fall within $\pm 15\%$ of the known concentration.

26.4.2 If the result is not within these limits, analysis of samples is halted until the problem is corrected, and either all the samples in the batch must be reanalyzed, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

26.5 *Method Blank:*

26.5.1 Analyze a reagent water test blank with each batch. The concentration of lead found in the blank should be less than 0.5 times the lowest calibration standard. If the concentration of lead is found above this level, analysis of samples is halted until the contamination is eliminated, and a blank shows no contamination at or above this level, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

26.6 *Matrix Spike (MS):*

26.6.1 To check for interferences in the specific matrix being tested, perform a MS on at least one sample from each batch by spiking an aliquot of the sample with a known concentration of lead and taking it through the analytical method.

26.6.2 The spike concentration plus the background concentration of lead must not exceed the high calibration standard. The spike must produce a concentration in the spiked sample that is 2 to 5 times the analyte concentration in the unspiked sample, or 10 to 50 times the detection limit of the test method, whichever is greater.

26.6.3 Calculate the percent recovery of the spike (P) using the following formula: 8-84d12cf6f62fastm-d3559-08

$$P = 100 [A(V_s + V) - B V_s] / C V \quad (3)$$

where:

A = analyte concentration (mg/L) in spiked sample,

B = analyte concentration (mg/L) in unspiked sample,

C = concentration (mg/L) of analyte in spiking solution,

V_s = volume (mL) of sample used, and

V = ~~volume (mL) added with spike.~~ volume (mL) of spiking solution added.

26.6.4 The percent recovery of the spike shall fall within the limits, based on the analyte concentration, listed in Guide D 5810, Table 2. If the percent recovery is not within these limits, a matrix interference may be present in the sample selected for spiking. Under these circumstances, one of the following remedies must be employed: the matrix interference must be removed, all samples in the batch must be analyzed by a test method not affected by the matrix interference, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

NOTE 12—Acceptable spike recoveries are dependent on the concentration of the component of interest. See Guide D 5810 for additional information.

26.7 *Duplicate:*

26.7.1 To check the precision of sample analyses, analyze a sample in duplicate with each batch. If the concentration of the analyte is less than five times the detection limit for the analyte, a matrix spike duplicate (MSD) should be used.

26.7.2 Calculate the standard deviation of the duplicate values and compare to the precision in the collaborative study using an F test. Refer to 6.4.4 of Practice D 5847 for information on applying the F test.

26.7.3 If the result exceeds the precision limit, the batch must be reanalyzed or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

26.8 *Independent Reference Material (IRM):*

26.8.1 In order to verify the quantitative value produced by the test method, analyze an Independent Reference Material (IRM)

submitted as a regular sample (if practical) to the laboratory at least once per quarter. The concentration of the IRM should be in the concentration mid-range for the method chosen. The value obtained must fall within the control limits established by the laboratory.

TEST METHOD C—DIFFERENTIAL PULSE ANODIC STRIPPING VOLTAMMETRY

27. Scope

27.1 This test method describes the determination of lead in water and waste waters using differential pulse anodic stripping voltammetry.

27.2 This test method is applicable up to a concentration of 100 µg/L lead. Higher concentrations can be determined by dilution.

27.3 The lower limit of detection for lead is 1.0 µg/L.

NOTE 13—The lower limit of detection for differential pulse anodic stripping voltammetry is not absolute and can easily be lowered by changing the experimental parameters as described in Appendix X1. However, these variations have not been interlaboratory tested.

28. Terminology

28.1 *Definitions*—See 3.1.

28.2 *Definitions of Terms Specific to This Standard:*

28.2.1 *spiking solution*—the standard solution added to the polarographic cell that is used to quantitate the sample.

28.2.2 *stripping peak potential*—the applied potential versus SCE at which the stripping peak current is a maximum.

28.2.2.1 *SCE*—saturated calomel electrode.

28.2.3 *stripping peak signal*—the current measured at the stripping peak maximum for a metal.

29. Summary of Test Method

29.1 This test method determines the total recoverable concentration of lead in water and waste water. The same digestion, sample preparation, and analysis procedure may be used to determine total recoverable cadmium (Test Methods D 3557) simultaneously with lead.

29.2 The sample is digested with nitric acid in a polarographic cell: 0.2 M ammonium citrate buffer (pH 3.0) and 10 % hydroxylamine solution are added. The solution is warmed to dissolve the lead. Warming with hydroxylamine eliminates interference from ferric iron by reducing it to ferrous.

29.3 After cooling, this sample is deaerated, and the lead is deposited into a hanging mercury drop electrode with surface area of 1.5 to 3.0 mm² at a constant potential of – 0.80 V versus saturated calomel electrode (SCE). The lead is then stripped back into solution using the differential pulse scanning mode, and the current is measured during the stripping step.

29.4 The stripping peak height is proportional to the concentration of the lead, and the stripping peak potential is a qualitative measure of the lead in solution.

30. Interferences

30.1 Selenium does not interfere up to 50 µg/L. Interference from selenium concentration up to 1000 µg/L may be overcome by adding ascorbic acid which reduces selenium (IV) to selenium metal and eliminates the interference.

30.2 When ferric ions are present at levels greater than cadmium, interference may occur from oxidizing the deposited metal out of the amalgam. Interference by ferric iron at concentrations as high as 20 mg/L is eliminated by warming with hydroxylamine. Ferric ions are reduced to ferrous ions by the hydroxylamine, and the interference caused by the presence of iron is eliminated.

30.3 The presence of a neighboring stripping peak which is <100 mV from that of lead will interfere.

31. Apparatus

31.1 *Polarographic Instrumentation*, capable of performing differential pulse work.⁶

31.2 *Hanging Mercury Drop Electrode*.⁷

⁶ Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

⁶ Two suitable instruments are the Princeton Applied Research, Princeton, NJ, Model 174A polarographic analyzer with mechanical drop timer, and Houston Omnigraphic X-Y Recorder Model 2200-3-3. An equally suitable instrument is the Environmental Sciences Associates (ESA), Bedford, MA, Model 3040 Charge Transfer Analyzer. For settings on ESA Model 3040 equivalent to those in paragraph 33.10, see ESA Application Note CTA-AN-1. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend.

⁷ Supporting data are available from ASTM Headquarters. Request RR:D 19-1030.

⁷ The Model 9323 hanging, mercury drop electrode or the Model 314 automated hanging mercury drop electrode manufactured by Princeton Applied Research has been found satisfactory. The Metrohm E-410 hanging mercury drop electrode is equally satisfactory. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend.

31.3 *Reagent Purifier System.*⁸

31.4 *Counter Electrode*, platinum.

31.5 *Salt Bridge*, with slow leakage fritted glass tip, tip,⁹ to isolate saturated calomel electrode from the test solution.

31.6 *Magnetic Stirrer*—The magnetic stirrer used must have a separate on/off switch, so that uniform rotational speed can be maintained. ~~mm~~13-mm (0.5-in.) magnetic stirring bar is also required.

31.7 *pH Meter*.

31.8 *Hot Plate*.

31.9 *Micropipets* Micropipettes incorporating disposable plastic tips are used. The sizes required are 10, 20, 50, and 100 μL .

32. Reagents

32.1 *Citrate Buffer Solution*—Dissolve 42 g of citric acid in 800 mL of water and add enough ammonium hydroxide to bring the pH to 3.0 ± 0.2 . Dilute to 1000 mL with water and place in the cell of the reagent purifier system. Purify for a minimum of 36 h at a potential of -1.3 V versus SCE at a mercury pool working electrode. Deaerate the supporting electrolyte during the purification process. If the buffer contains less than $1 \mu\text{g/L}$ of lead, the purification step may be omitted, providing new buffer is prepared every 2 weeks. The electrolyzed buffer is stable against bacterial growth for at least 1 month.

NOTE 14—To prevent bacterial growth in the unpurified buffer for a month, sterilize by autoclaving for 15 min at 121°C and 1.03×10^5 Pa (15 psi).

32.2 *Aqua Regia (1 + 1)*—Add 1 volume of nitric acid (sp gr 1.42), reagent grade, to 4 volumes of water. Then add 3 volumes of hydrochloric acid.

32.2.1 **Warning:**—~~Toxic~~ Toxic fumes may be released. Prepare and use in a ventilated hood.

32.3 *Ascorbic Acid Solution (100 g/L)*—Dissolve 10.0 g of L-ascorbic acid in reagent water and dilute to 100 mL.

32.4 *Hydrochloric Acid (sp gr 1.19)*—Concentrated hydrochloric acid (HCl).

32.5 *Hydroxylamine Solution (100 g/L)*—Dissolve 5.00 g of hydroxylamine hydrochloride ($\text{NH}_2\text{OH}\cdot\text{HCl}$) in reagent water and dilute to 50 mL.

32.6 *Nitric Acid (sp gr 1.42)*¹⁰—Redistilled concentrated nitric acid (HNO_3).

32.7 *Nitric Acid (sp gr 1.42)*—Concentrated nitric acid (HNO_3).

32.8 *Nitric Acid (1 + 4)*—Add 1 volume of nitric acid (sp gr 1.42) to 4 volumes of water.

32.9 *Nitric Acid (2 + 3)*—Add 2 volumes of nitric acid (sp gr 1.42) reagent grade,⁶ to 3 volumes of water.

32.10 *Purified Nitrogen*—Nitrogen employed for deoxygenation must be sufficiently oxygen-free so that a differential pulse polarographic scan from -0.20 to -0.80 V versus SCE of the citrate buffer solution, after 10 min deaeration at 10^5 mm^3/min , gives a signal no more than $0.1 \mu\text{A}$. See Appendix X2 to learn methods of gas purification.

32.11 *Lead Solution, Stock (1 mL = 0.1 mg Pb)*—~~Clean oxide from lead metal with HNO_3~~ Commercially purchase a stock solution. Alternately, clean oxide from lead metal with HNO_3 (1 + 4). Wash the cleaned metal with water and dry. Dissolve 0.1000 g of the lead in 25 mL of HNO_3 (1 + 4). Dilute to 1 L with water.

33. Caution

33.1 The liquid mercury used for the hanging mercury drop electrode⁷ forms a toxic vapor, and the liquid itself is toxic. Handle with gloves in a ventilated hood.

34. Calibration

34.1 After a differential pulse anodic stripping curve is run on the sample solution, the anodic stripping curve is quantitated using the technique of standard addition.

34.2 Prepare spiking solution as directed in 32.11. Alternatively, if cadmium is to be quantified too, both metals may be added to a single spiking solution. The best procedure here is to prepare the spiking solution with each metal in the ratio expected in the sample. (Example: If lead is expected to be 5 times the cadmium, prepare a spiking solution with lead and cadmium in a 5 to 1 ratio).

⁸ Two suitable instrument are the Princeton Applied Research, Princeton, NJ, Model 174A polarographic analyzer with mechanical drop timer, and Houston Omnigraphic X-Y Recorder Model 2200-3-3. An equally suitable instrument is the Environmental Sciences Associates (ESA), Bedford, MA, Model 3040 Charge Transfer Analyzer. For settings on ESA Model 3040 equivalent to those in paragraph 33.10, see ESA Application Note CTA-AN-1.

⁸ Both the Model 9500 Electrolyte Purification System (Princeton Applied Research, Princeton, NJ) and the Model 2014 PM Reagent Cleaning System (Environmental Sciences Associates, Bedford, MA) are equally suitable. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend.

⁹ The Model 9323 hanging, mercury drop electrode or the Model 314 automated hanging mercury drop electrode manufactured by Princeton Applied Research has been found satisfactory. The Metrohm E-410 hanging mercury drop electrode is equally satisfactory.

⁹ A Vycor (Corning Glass Works, Corning, NY) tip has been found suitable. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend.

¹⁰ Both the Model 9500 Electrolyte Purification System (Princeton Applied Research, Princeton, NJ) and the Model 2014 PM Reagent Cleaning System (Environmental Sciences Associates, Bedford, MA) are equally suitable.

¹⁰ Acids that may contain suitably low levels of lead (and cadmium) are the redistilled reagents or equivalent from G. Frederick Smith Chemical Co., 867 McKinley Ave., Columbus, OH 43223.