



SLOVENSKI STANDARD SIST EN ISO 81060-2:2020

01-februar-2020

Nadomešča:

SIST EN ISO 81060-2:2014

Neinvazivni sfigmomanometri - 2. del: Klinične raziskave avtomatiziranih vrst merjenja s prekinitvami (ISO 81060-2:2018)

Non-invasive sphygmomanometers - Part 2: Clinical investigation of intermittent automated measurement type (ISO 81060-2:2018)

Nichtinvasive Blutdruckmessgeräte Teil 2: Klinische Prüfung der intermittierenden automatisierten Bauart (ISO 81060-2:2018)

Sphygmomanomètres non invasifs - Partie 2: Investigation clinique pour type ponctuel à mesurage automatique (ISO 81060-2:2018)

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>

Ta slovenski standard je istoveten z: EN ISO 81060-2:2019

ICS:

11.040.55 Diagnostična oprema Diagnostic equipment

SIST EN ISO 81060-2:2020

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 81060-2:2020

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>

EUROPEAN STANDARD

EN ISO 81060-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2019

ICS 11.040.10

Supersedes EN ISO 81060-2:2014

English Version

Non-invasive sphygmomanometers - Part 2: Clinical investigation of intermittent automated measurement type (ISO 81060-2:2018)

Sphygmomanomètres non invasifs - Partie 2:
Investigation clinique pour type ponctuel à mesurage
automatique (ISO 81060-2:2018)

Nichtinvasive Blutdruckmessgeräte - Teil 2: Klinische
Prüfung der intermittierenden automatisierten Bauart
(ISO 81060-2:2018)

This European Standard was approved by CEN on 20 November 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

iTeh STANDARD PREVIEW

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Annex ZA (informative) Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [O] L 169] aimed to be covered	5

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 81060-2:2020
<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>

European foreword

This document (EN ISO 81060-2:2019) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2020, and conflicting national standards shall be withdrawn at the latest by November 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 81060-2:2014.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZA, which is an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard "within the meaning of Annex ZA", the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>

When the ISO or IEC standard is referred to in the ISO text standard, this must be understood as a normative reference to the parallel EN standard or dated ISO standard, as outlined below, including the foreword and the Annexes ZZ.

NOTE The way in which these references documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table — Correlations between normative references and dated EN and ISO/IEC standards

Normative references as listed in Clause 2	Equivalent dated standard	
	EN	ISO/IEC
ISO 14155:2011	EN ISO 14155:2011	ISO 14155:2011
ISO 14971:2007	EN ISO 14971:2012	ISO 14971:2007
ISO 16142-1:2016	—	ISO 16142-1:2016
IEC 60601-1:2005+AMD1:2012	EN 60601-1:2006 +AMD1:2013 +AMD12:2014	IEC 60601-1:2005 +AMD1:2012
IEC 60601-1-11:2015	EN 60601-1-11:2015	IEC 60601-1-11:2015
IEC 60601-2-34:2011	EN 60601-2-34:2014	IEC 60601-2-34:2011
IEC 80601-2-30:2018	EN 80601-2-30:2019	IEC 80601-2-30:2018
ISO 81060-1:2007	EN ISO 81060-1:2012	ISO 81060-1:2007

EN ISO 81060-2:2019 (E)

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 81060-2:2018 has been approved by CEN as EN ISO 81060-2:2019 without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 81060-2:2020](https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020)

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>

Annex ZA (informative)

Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered

This European Standard has been prepared under a Commission's joint standardization request M/295 concerning the development of European standards relating to medical devices to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Directive 93/42/EEC [OJ L 169]

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
10.1	4, 5 and 6	Only the characteristics of the measurement performance (accuracy), as well as the corresponding tests methods, are addressed.

EN ISO 81060-2:2019 (E)

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
13.6 a) j)	5.1.6 e), 6.2.1 d) 2), 6.2.2 a), 6.2.7, 7 h) and 7 i)	<p>Covered only in respect of certain additional warnings contained in the indicated subclauses related to the following:</p> <ul style="list-style-type: none"> – definition of special patient population; – that effectiveness has not been established in the presence of any dysrhythmias included in the exclusion criteria, where applicable; – specifying the arterial reference site; – disclosure of the method used to determine and verify the mean arterial pressure; – suitability for use with pregnant (including pre-eclamptic) patients, where applicable; – that effectiveness has not been established in pregnant (including pre-eclamptic) patients, where applicable.
Annex X, 2.3.1 to 2.3.3	5, 6 and 7	<p>Covered only in respect of carrying our clinical investigations with</p> <ul style="list-style-type: none"> – reference auscultatory sphygmomanometers; – reference invasive blood pressure monitoring equipment; and – pregnant patient populations, where applicable.

WARNING 1 — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

WARNING 2 — Other Union legislation may be applicable to the products falling within the scope of this standard.

INTERNATIONAL
STANDARD

ISO
81060-2

Third edition
2018-11

**Non-invasive sphygmomanometers —
Part 2:
Clinical investigation of intermittent
automated measurement type**

Sphygmomanomètres non invasifs —

*Partie 2: Investigation clinique pour type ponctuel à mesurage
automatique*

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

SIST EN ISO 81060-2:2020

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>



Reference number
ISO 81060-2:2018(E)

© ISO 2018

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 81060-2:2020

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 General requirements for CLINICAL INVESTIGATIONS	2
4.1 CLINICAL INVESTIGATION methods.....	2
4.2 Good clinical practice.....	3
4.3 Status of previous CLINICAL INVESTIGATIONS.....	3
4.4 Disclosure of summary of CLINICAL INVESTIGATION.....	3
5 CLINICAL INVESTIGATION with an auscultatory REFERENCE SPHYGMOMANOMETER	3
5.1 Subject requirements.....	3
5.1.1 * Number.....	3
5.1.2 * Gender distribution.....	3
5.1.3 * Age distribution.....	4
5.1.4 * Limb size distribution.....	4
5.1.5 Blood pressure distribution.....	4
5.1.6 * Special PATIENT populations.....	5
5.2 CLINICAL INVESTIGATION method with a REFERENCE SPHYGMOMANOMETER.....	5
5.2.1 * Subject preparation.....	5
5.2.2 * Observer preparation.....	6
5.2.3 * REFERENCE readings.....	6
5.2.4 CLINICAL INVESTIGATION methods.....	7
5.2.5 * Additional requirements for a SPHYGMOMANOMETER intended for use in exercise stress testing environments.....	15
5.2.6 * Additional requirements for a SPHYGMOMANOMETER intended for use in ambulatory monitoring.....	16
6 CLINICAL INVESTIGATION with REFERENCE INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT	17
6.1 PATIENT requirements.....	17
6.1.1 Number.....	17
6.1.2 * Gender distribution.....	17
6.1.3 * Age distribution.....	17
6.1.4 Limb size distribution.....	18
6.1.5 BLOOD PRESSURE distribution.....	18
6.1.6 Special PATIENT populations.....	19
6.2 CLINICAL INVESTIGATION methods with REFERENCE INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT.....	19
6.2.1 * REFERENCE measurement.....	19
6.2.2 * Arterial REFERENCE site.....	20
6.2.3 PROCEDURE.....	20
6.2.4 * Determining the REFERENCE BLOOD PRESSURE.....	21
6.2.5 Determining the error of the BLOOD PRESSURE measurement.....	22
6.2.6 Data analysis.....	22
6.2.7 MEAN ARTERIAL PRESSURE (MAP).....	23
7 * Pregnant PATIENT populations	23
Annex A (informative) Rationale and guidance	25
Annex B (informative) Reference to the ESSENTIAL PRINCIPLES	33
Annex C (informative) Terminology — alphabetized index of defined terms	34
Bibliography	35

ISO 81060-2:2018(E)

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see <http://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electromedical equipment*.

This third edition cancels and replaces the second edition (ISO 81060-2:2013), which has been technically revised.

The main changes compared to the previous edition are as follows:

- same arm simultaneous method has been deleted;
- numerous clarifications have been added and kPa equivalent values for the mmHg values have been included.

A list of all parts in the ISO/IEC 81060 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Determining BLOOD PRESSURE is an important PROCEDURE that is clinically used to assess the status of a PATIENT.

BLOOD PRESSURE serves as aid to control the drug titration and fluid management and to provide warning about the changes in PATIENT'S state of health.

Frequently determining BLOOD PRESSURE is routine during anaesthesia. BLOOD PRESSURE serves to aid to control drug titration and fluid management and to provide warning about the changes in the PATIENT'S state of health.

In this document, the following print types are used:

- requirements, compliance with which can be verified, and definitions: roman type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller roman type. Normative text of tables is also in a smaller roman type;
- *test methods: italic type*; and
- TERMS DEFINED IN [CLAUSE 3](#) OF THE GENERAL STANDARD, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS TYPE.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document; <https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in [Annex A](#).

[Annex B](#) maps the clauses and subclauses of this document with the ESSENTIAL PRINCIPLES of ISO 16142-1:2016.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 81060-2:2020

<https://standards.iteh.ai/catalog/standards/sist/d398ea49-45bc-4020-9471-64ae2bd643c9/sist-en-iso-81060-2-2020>

Non-invasive sphygmomanometers —

Part 2: Clinical investigation of intermittent automated measurement type

1 Scope

This document specifies the requirements and methods for the CLINICAL INVESTIGATION OF ME EQUIPMENT used for the INTERMITTENT non-invasive automated estimation of the arterial BLOOD PRESSURE by utilizing a CUFF.

This document is applicable to all SPHYGMOMANOMETERS that sense or display pulsations, flow or sounds for the estimation, display or recording of BLOOD PRESSURE. These SPHYGMOMANOMETERS need not have automatic CUFF inflation.

This document covers SPHYGMOMANOMETERS intended for use in all PATIENT populations (e.g. all age and weight ranges), and all conditions of use (e.g. ambulatory BLOOD PRESSURE monitoring, stress testing BLOOD PRESSURE monitoring and BLOOD PRESSURE monitors for the HOME HEALTHCARE ENVIRONMENT for self-measurement as well as use in a professional healthcare facility).

EXAMPLE AUTOMATED SPHYGMOMANOMETER as given in IEC 60601-2-30 undergoing CLINICAL INVESTIGATION according to this document.

This document specifies additional disclosure requirements for the ACCOMPANYING DOCUMENTS of SPHYGMOMANOMETERS that have passed a CLINICAL INVESTIGATION according to this document.

This document is not applicable to CLINICAL INVESTIGATIONS of NON-AUTOMATED SPHYGMOMANOMETERS as given in ISO 81060-1 or INVASIVE BLOOD PRESSURE MONITORING EQUIPMENT as given in IEC 60601-2-34.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the Bibliography.

ISO 14155:2011, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14971:2007, *Medical devices — Application of risk management to medical devices*

ISO 16142-1:2016, *Medical devices — Recognized essential principles of safety and performance of medical devices — Part 1: General essential principles and additional specific essential principles for all non-IVD medical devices and guidance on the selection of standards*

ISO 81060-1:2007, *Non-invasive sphygmomanometers — Part 1: Requirements and test methods for non-automated measurement type*

IEC 60601-1:2005+Amendment 1:2012, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*