

INTERNATIONAL  
STANDARD

ISO  
19344

IDF  
232

First edition  
2015-12-15

---

---

**Milk and milk products — Starter  
cultures, probiotics and fermented  
products — Quantification of lactic  
acid bacteria by flow cytometry**

*Lait et produits laitiers — Cultures, probiotiques et produits fermentés  
— Quantification de bactéries lactiques par cytométrie en flux*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 19344:2015

<https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015>



Reference numbers  
ISO 19344:2015(E)  
IDF 232:2015(E)

© ISO and IDF 2015

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 19344:2015](https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015)

<https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO and IDF 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

International Dairy Federation  
Silver Building • Bd Auguste Reyers 70/B • B-1030 Brussels  
Tel. + 32 2 325 67 40  
Fax + 32 2 325 67 41  
info@fil-idf.org  
www.fil-idf.org

# Contents

Page

<b>Forewords</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>2</b>
<b>5 Diluents and reagents</b> .....	<b>3</b>
5.1 General .....	3
5.2 Peptone-salt solution .....	3
5.3 Diluents and reagents for staining protocols .....	3
5.3.1 Protocol A .....	4
5.3.2 Protocol B .....	4
5.3.3 Protocol C .....	5
<b>6 Apparatus</b> .....	<b>6</b>
<b>7 Sampling</b> .....	<b>7</b>
<b>8 Preparation of test sample</b> .....	<b>7</b>
8.1 General .....	7
8.2 Freeze-dried cultures .....	7
8.3 Frozen cultures .....	8
8.4 Fermented milk products .....	8
<b>9 Procedure</b> .....	<b>8</b>
9.1 General .....	8
9.2 Staining .....	8
9.2.1 Protocol A .....	9
9.2.2 Protocol B .....	9
9.2.3 Protocol C .....	9
9.3 Flow cytometry analysis .....	10
9.3.1 General .....	10
9.3.2 Instruments and settings .....	10
9.4 Gating .....	11
9.4.1 General .....	11
9.4.2 Protocol A .....	11
9.4.3 Protocol B .....	13
9.4.4 Protocol C .....	14
<b>10 Calculation and expression of results</b> .....	<b>15</b>
<b>11 Critical factors affecting results</b> .....	<b>16</b>
<b>12 Precision</b> .....	<b>17</b>
12.1 Interlaboratory test .....	17
12.2 Repeatability .....	18
12.3 Reproducibility .....	18
<b>13 Test report</b> .....	<b>18</b>
<b>Annex A (informative) Diagram of staining protocols</b> .....	<b>19</b>
<b>Annex B (informative) Calculation example of the appropriate sample dilution</b> .....	<b>21</b>
<b>Annex C (informative) Interlaboratory test</b> .....	<b>22</b>
<b>Bibliography</b> .....	<b>25</b>

## Forewords

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products* and the International Dairy Federation (IDF). This document is being published jointly by ISO and IDF.

<https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015>

**IDF (the International Dairy Federation)** is a non-profit private sector organization representing the interests of various stakeholders in dairying at the global level. IDF members are organized in National Committees, which are national associations composed of representatives of dairy-related national interest groups including dairy farmers, dairy processing industry, dairy suppliers, academics and governments/food control authorities.

ISO and IDF collaborate closely on all matters of standardization relating to methods of analysis and sampling for milk and milk products. Since 2001, ISO and IDF jointly publish their International Standards using the logos and reference numbers of both organizations.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. IDF shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

ISO 19344 | IDF 232 was prepared by the IDF Standing Committee on *Analytical Methods for Dairy Microorganisms* and the ISO Technical Committee ISO/TC 34 on *Food products*, Subcommittee SC 5 on *Milk and milk products*.

The work was carried out by the IDF/ISO Project Group on *Quantification of Lactic Acid Bacteria by Flow Cytometry* of the Standing Committee on *Analytical Methods for Dairy Microorganisms* under the aegis of its project leader, Sandra Casani (DK), Ph.D.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 19344:2015](https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015)

<https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015>

## Introduction

Quantification of lactic acid bacteria is an important factor in assessing the quality of starter cultures, probiotics and fermented milk products. Examination of lactic acid bacteria in these products can be done following different method principles, with plate count techniques being the most traditional and widely used. Newer techniques include flow cytometry, which is able to determine cells as active and/or total units. Advantages of the use of flow cytometry include low variation, differentiation between active and total cells, and possibility of high analysis throughout. Furthermore, the quantification and use of the fraction of active cells per total cells is a key feature and an important flow cytometry tool to evaluate the fitness of a given cell population. This is of special relevance for certain applications such as optimization of production process and stability assessment during shelf-life.

The International Organization for Standardization (ISO) and the International Dairy Federation (IDF) draw attention to the fact that compliance with this document may involve the use of patents concerning the staining of protocol C as described in this document.

Neither ISO nor IDF take position concerning the evidence, validity and scope of these patent rights.

The holder of these patent rights has ensured ISO and IDF that he/she is willing to negotiate licences either free of charge or under reasonable and non-discriminatory terms and conditions with applicants throughout the patented territory. In this respect, the statement of the holder of these patent rights is registered with ISO. Information may be obtained from:

Chr. Hansen A/S  
Boege Alle 10-12  
2970 Hoersholm  
Denmark

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights other than those identified above. Neither ISO nor IDF shall be held responsible for identifying any or all such patent rights.

ISO 19344:2015  
patent rights: iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015

ISO ([www.iso.org/patents](http://www.iso.org/patents)) and IEC (<http://patents.iec.ch>) maintain online databases of patents relevant to their standards. Users are encouraged to consult the databases for the most up-to-date information concerning patents.

# Milk and milk products — Starter cultures, probiotics and fermented products — Quantification of lactic acid bacteria by flow cytometry

**WARNING** — The use of this International Standard may involve hazardous materials and operations. This International Standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish safety and health practices and to determine the applicability of regulatory limitations prior to use.

## 1 Scope

This International Standard specifies a standardized method for the quantification of active and/or total lactic acid bacteria and probiotic strains in starter cultures used in dairy products by means of flow cytometry. The method is also applicable to probiotics used in dairy products and to fermented milk products such as yogurts containing primarily lactic acid bacteria.

This International Standard does not apply to taxonomical differentiation of bacteria. Due to its non-specificity, the method may quantify other bacteria than those within the scope of this International Standard, when present in the sample. This may lead to overestimation of the counts.

The minimum bacterial cell concentration in the sample before applying this standardized method depends on the dilution rates used in the individual protocols. Typically  $10^6$  cells per gram or ml are considered within the minimum range.

[ISO 19344:2015](https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015)

[https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-](https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015)

## 2 Normative references

[0a2c35ea8367/iso-19344-2015](https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015)

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6887-1:1999, *Microbiology of food and animal feeding stuffs — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination — Part 1: General rules for the preparation of the initial suspension and decimal dilutions*

ISO 6887-5:2010, *Microbiology of food and animal feeding stuffs — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination — Part 5: Specific rules for the preparation of milk and milk products*

ISO 7218, *Microbiology of food and animal feeding stuffs — General requirements and guidance for microbiological examinations*

ISO 7889 | IDF 117, *Yogurt — Enumeration of characteristic microorganisms — Colony-count technique at 37 °C*

ISO 15214, *Microbiology of food and animal feeding stuffs — Horizontal method for the enumeration of mesophilic lactic acid bacteria — Colony-count technique at 30 °C*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1 lactic acid bacteria

gram-positive, non-motile, non-spore forming, catalase-negative, nitrate-reductase-negative and cytochrome oxidase-negative bacterium that does not liquefy gelatine or produce indole

Note 1 to entry: Lactic acid bacteria have a fermentative metabolism which is mainly saccharolytic. Lactic acid is the major end product from carbohydrate utilization.

EXAMPLE Lactic acid bacteria of importance for the dairy industry are: *Streptococcus thermophilus*, *Lactococcus lactis*, *Pediococcus*, *Enterococcus*, *Leuconostoc* and *Lactobacillus*.

### 3.2 probiotic strains

probiotic strains are live microorganisms which, when administered in adequate amounts, are intended to confer a health benefit to the host

EXAMPLE Probiotic strains of importance are: *Bifidobacterium animalis*, *Lactobacillus casei*, *Lactobacillus paracasei*, *Lactobacillus rhamnosus*, *Lactobacillus acidophilus*, *Lactobacillus reuteri*, *Lactobacillus plantarum* and *Propionibacterium freudenreichii*.

Note 1 to entry: See Reference [1].

### 3.3 active fluorescent units AFU

events counted in a gate specific for scatter/fluorescence characteristics of presumed live cells, i.e. cells stained for the specific activity indicator used in the protocol

### 3.4 non-active fluorescent units n-AFU

events counted in a gate specific for scatter/fluorescence characteristics of presumed dead cells, i.e. cells damaged to an extent that they do not stain for the specific activity indicator used in the protocol

### 3.5 total fluorescent units TFU

sum of AFU and n-AFU

### 3.6 % active fluorescent units % AFU

percentage ratio of AFU to TFU

## 4 Principle

4.1 A test portion or sample is prepared, and diluted if necessary.

4.2 Initial suspensions, and/or dilutions if needed, are stained according to one of the following three protocols, differing on the target of fluorescent cell staining, in order to discriminate active and total fluorescent units:

- a) dual staining targeting nucleic acid with the non-permeant red-fluorescent dye propidium iodide (PI) and intracellular enzyme activity based on cleavage of 5(6)-carboxyfluorescein diacetate (cFDA) mixed isomers to green-fluorescent carboxyfluorescein by intracellular esterases;



- b) dual nucleic acid staining with PI and a cell-permeant green fluorescent dye, i.e. SYTO®<sup>1)</sup> 24 green fluorescent cell-permeant nucleic acid stain;
- c) single staining with the membrane-potential-sensitive cyanine dye 3,3'-diethyloxycarbocyanine iodide (DiOC<sub>2</sub>). Wavelength of emitted light changes with metabolic activation of cells.

The choice of the staining protocol depends on the user's preferences or possibilities.

**4.3** The stained samples are analysed by means of a flow cytometer using a combination of light scattering (LS) and detection of emitted fluorescent light. As cells pass into the flow cytometer, each cell is counted and the fluorescence is recorded.

**4.4** Gating is conducted to separate cells from noise and to differentiate AFUs and n-AFUs.

**4.5** Calculation of the concentration in the original sample is a multiplication of AFUs (or TFUs) per volume of analysed sample and the dilution factors employed in the sample preparation.

## 5 Diluents and reagents

### 5.1 General

Unless otherwise specified, use only reagents of recognized analytical grade, and distilled or deionized water or water of equivalent purity, according to ISO 7218.

Prepare the initial suspension (common for all protocols) with the diluent as specified in 5.2.

The composition and the preparation of all the reagents used in each of the three staining protocols (A, B and C) are specified in 5.3. An overview of the diluents and reagents per protocol is given in Table 1.

### 5.2 Peptone-salt solution

The composition and preparation of the peptone-salt solution is according to ISO 6887-5:2010, 5.2.1.

**NOTE** For the preparation of the initial suspension, and dilutions if needed, other diluents for general use mentioned in ISO 6887-5:2010 can be used if they can be shown to lead to the same results.

### 5.3 Diluents and reagents for staining protocols

**WARNING — Propidium iodide is a potential mutagen. Proper actions for deactivation should be taken in case of spilling. Preparation and application of the dye solution shall be carried out in a fume cupboard, using protective equipment and following good laboratory practices.**

**Table 1 — Reagents used per protocol**

Reagent	Protocol A	Protocol B	Protocol C
Diluent	PBS (5.3.1.1)	PBS (5.3.2.1)	MRS or M17 broth (5.3.3.1 or 5.3.3.2)
Dye solution	cFDA (5.3.1.2) PI (5.3.1.3)	PI (5.3.2.2) SYTO® 24 (5.3.2.3)	Glucose solution (5.3.3.3.1) DiOC <sub>2</sub> (5.3.3.3.2) Buffer solution (5.3.3.3.3)

1) SYTO® 24 green fluorescent cell-permeant nucleic acid stain is supplied by Life Technologies. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

### 5.3.1 Protocol A

#### 5.3.1.1 Phosphate-buffered saline (PBS)

##### 5.3.1.1.1 Composition

- 9 g sodium chloride (NaCl)
- 795 mg sodium hydrogenphosphate heptahydrate ( $\text{Na}_2\text{HPO}_4 \cdot 7\text{H}_2\text{O}$ )
- 144 mg potassium dihydrogen phosphate ( $\text{KH}_2\text{PO}_4$ )

##### 5.3.1.1.2 Preparation

Dissolve the components (see [5.3.1.1.1](#)) in water. Add water to a final volume of 1 000 ml. Adjust the pH with HCl to  $7,4 \pm 0,05$ , if necessary. Distribute the solution into aliquots and sterilize in an autoclave set at  $121 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$  (liquid cycle) for 15 min. The diluent can be stored at cooling temperature ( $3 \text{ }^\circ\text{C} \pm 2 \text{ }^\circ\text{C}$ ) for up to 6 months.

#### 5.3.1.2 5(6)-Carboxyfluorescein diacetate (cFDA) mixed isomers solution

##### 5.3.1.2.1 Composition

- 230 mg 5(6)-cFDA mixed isomers
- 100 ml dimethyl sulfoxide (DMSO)

iteh STANDARD PREVIEW  
(standards.iteh.ai)

##### 5.3.1.2.2 Preparation

A 5 mmol/l solution is prepared by dissolving cFDA in DMSO at the amounts specified in [5.3.1.2.1](#). The solution can be stored at  $-18 \text{ }^\circ\text{C} \pm 2 \text{ }^\circ\text{C}$ , protected from light, for up to 6 months.

#### 5.3.1.3 Propidium iodide (PI)

##### 5.3.1.3.1 Composition

- 100 mg PI
- 100 ml ultrapure water

##### 5.3.1.3.2 Preparation

Dissolve the PI in ultrapure water to a final concentration of 1,0 mg/ml, corresponding to approximately 1,5 mmol/l. This can be stored at  $3 \text{ }^\circ\text{C} \pm 2 \text{ }^\circ\text{C}$ , protected from light, for up to 6 months.

NOTE The concentration of the PI solution used is 0,1 % and the final concentration is 0,002 %. This is below the potential toxicity level.

### 5.3.2 Protocol B

#### 5.3.2.1 Phosphate-buffered saline (PBS)

See [5.3.1.1](#).

### 5.3.2.2 Propidium iodide (PI)

See [5.3.1.3](#) for the preparation of the PI solution. The PI solution shall be further diluted to 0,2 mmol/l with water prior to use.

NOTE The concentration of the PI solution used is 0,01 % and the final concentration is 0,000 1 %. This is below the potential toxicity level.

### 5.3.2.3 SYTO® 24 green fluorescent cell-permeant nucleic acid stain

The stain is a 5 mmol/l solution in DMSO. Store at -20 °C, protected from light, for up to 12 months. The solution shall be diluted to 0,1 mmol/l with water before use.

## 5.3.3 Protocol C

### 5.3.3.1 MRS broth

The composition and the preparation are specified in ISO 15214 except for no addition of agar.

### 5.3.3.2 M17 broth

The composition and the preparation are specified in ISO 7889 except for no addition of agar.

### 5.3.3.3 Stain mixture

The stain mixture consists of 210 µl 50 % glucose solution ([5.3.3.3.1](#)), 210 µl 1,5 mmol/l DiOC<sub>2</sub> ([5.3.3.3.2](#)) and 50 ml buffer solution ([5.3.3.3.3](#)). The stain mixture is prepared the same day as it is used.

#### 5.3.3.3.1 Glucose solution

ISO 19344:2015

<https://standards.iteh.ai/catalog/standards/sist/bdf65b20-dc04-471f-bb5e-0a2c35ea8367/iso-19344-2015>

##### 5.3.3.3.1.1 Composition

- 50 g D(+)-glucose monohydrate
- 50 g water

##### 5.3.3.3.1.2 Preparation

A 50 % glucose solution is prepared by dissolving the glucose in the water. This is aided by warming the solution to below the boiling point. Avoid evaporation. The solution is autoclaved at 121 °C ± 1 °C for 15 min and can be stored unopened at 3 °C ± 2 °C for up to 3 months.

#### 5.3.3.3.2 3,3'-diethyloxycarbocyanine iodide (DiOC<sub>2</sub>)

##### 5.3.3.3.2.1 Composition

- 69 mg 3,3'-DiOC<sub>2</sub>, ≥ 98 %
- 100 ml dimethyl sulfoxide (DMSO)

##### 5.3.3.3.2.2 Preparation

The DiOC<sub>2</sub> staining is prepared as a 1,5 mmol/l solution by weighing DiOC<sub>2</sub> into DMSO at the amounts specified in [5.3.3.3.2.1](#). Dispense in, e.g., 1 ml tubes. Keep dark, as DiOC<sub>2</sub> is unstable in light, at 5 °C ± 3 °C for up to 12 months.

### 5.3.3.3.3 Buffer solution

#### 5.3.3.3.3.1 Composition

- 7,6 g sodium chloride (NaCl; 130 mmol/l)
- 0,5 g sodium dihydrogenphosphate dihydrate ( $\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$ ; 3 mmol/l)
- 1,24 g sodium hydrogenphosphate dihydrate ( $\text{Na}_2\text{HPO}_4 \cdot 2\text{H}_2\text{O}$ ; 7 mmol/l)
- 1 000 ml water

#### 5.3.3.3.3.2 Preparation

Weigh and dissolve the three salts in the water at the amounts specified in 5.3.3.3.3.1. Stirring is applied until the salts are dissolved. Adjust pH to  $6,5 \pm 0,05$  with 2,5 mol/l HCl. The solution shall then be filtered through a 0,22  $\mu\text{m}$  filter. The mixture can be kept at  $3 \text{ }^\circ\text{C} \pm 2 \text{ }^\circ\text{C}$  for up to one week. For longer periods, up to 6 months, storage at  $-20 \text{ }^\circ\text{C}$  is recommended.

## 6 Apparatus

Usual laboratory equipment and, in particular, the equipment required for the preparation of test samples and dilutions specified in ISO 6887-5, as well as the following, shall be used.

- 6.1 **Water bath**, capable of operating at  $21 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$ .
- 6.2 **Analytical balance**, capable of weighing to the nearest 1 mg, with readability to 0,1 mg.
- 6.3 **pH-meter**, with temperature compensation, accurate to  $\pm 0,1$  pH unit.
- 6.4 **Incubator, heating block or equivalent**, capable of operating at the temperatures specified in [Table 2](#).

Table 2 — Incubation temperatures required per protocol

Protocol	Incubation temperatures
A	$30 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$ and $37 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$
B	$37 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$
C	$30 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$ and $37 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$

6.5 **Flasks, bottles and test tubes**, of sufficient capacity to contain the required volumes and leave adequate head-space for mixing. The capacity depends on the staining protocol and on the flow cytometry equipment.

6.6 **Pipettes**, sterile, calibrated for bacteriological use, accurate to within 2 % of the volume being pipetted.

6.7 **Vortex mixer**.

6.8 **Filter**, sterile, with membrane filters of a pore size 0,22  $\mu\text{m}$  and 25  $\mu\text{m}$ .

6.9 **Flow cytometer**, instrument capable of detecting and counting particles or cells when passing individually in a directed flow through a beam of excitation light. The instrument must be equipped with