

SLOVENSKI STANDARD SIST EN IEC 60746-4:2019

01-julij-2019

Izražanje lastnosti elektrokemičnih analizatorjev - 4. del: Raztopljeni kisik v vodi merjen z membranskimi amperometričenimi celicami (IEC 60746-4:2018)

Expression of performance of electrochemical analyzers - Part 4: Dissolved oxygen in water measured by membrane covered amperometric sensors (IEC 60746-4:2018)

Angabe zum Betriebsverhalten von elektrochemischen Analysatoren - Teil 4: Gelöster Sauerstoff in Wasser mit Hilfe membranbedeckter amperometrischer Messzellen (IEC 60746-4:2018)

(standards.iteh.ai)

Expression des qualités de fonctionnement des analyseurs électrochimiques - Partie 4: Oxygène dissous dans l'eau mesure par des capteurs ampérométriques recouverts d'une membrane (IEC 60746-4:2018) la log standards sixt d'7291867-1596-4366-8217-

Ta slovenski standard je istoveten z: EN IEC 60746-4:2019

ICS:

19.080 Električno in elektronsko Electrical and electronic

preskušanje testing

71.040.10 Kemijski laboratoriji. Chemical laboratories.

Laboratorijska oprema Laboratory equipment

SIST EN IEC 60746-4:2019 en,fr,de

SIST EN IEC 60746-4:2019

iTeh STANDARD PREVIEW (standards.iteh.ai)

EUROPEAN STANDARD

EN IEC 60746-4

NORME EUROPÉENNE ...

EUROPÄISCHE NORM

March 2019

ICS 71.040.40; 19.080

English Version

Expression of performance of electrochemical analyzers - Part 4: Dissolved oxygen in water measured by membrane-covered amperometric sensors (IEC 60746-4:2018)

Expression des qualités de fonctionnement des analyseurs électrochimiques - Partie 4: Oxygène dissous dans l'eau mesuré par des capteurs ampérométriques recouverts d'une membrane (IEC 60746-4:2018)

Angabe zum Betriebsverhalten von elektrochemischen Analysatoren - Teil 4: Gelöster Sauerstoff in Wasser mit Hilfe membranbedeckter amperometrischer Messzellen (IEC 60746-4:2018)

This European Standard was approved by CENELEC on 2019-01-17. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member. III and III a

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions. It avoidable standards sixty described a version of the centre of t

519507d77a37/sist-en-iec-60746-4-2019

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 60746-4:2019 (E)

European foreword

The text of document 65B/1128/FDIS, future edition 2 of IEC 60746-4, prepared by SC 65B "Measurement and control devices" of IEC/TC 65 "Industrial-process measurement, control and automation" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60746-4:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2019-10-17 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the document have to be withdrawn

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

iTeh STANDARD PREVIEW

(Stendorsement hotice i)

SIST EN IEC 60746-4:2019

The text of the International Standard IEC 60746-4:2018 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

ISO 9001 NOTE Harmonized as EN ISO 9001

ISO 5814:2012 NOTE Harmonized as EN ISO 5814:2012 (not modified)

EN IEC 60746-4:2019 (E)

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Publication Year Title EN/HD Year
IEC 60746-1 2003 Expression of performance of electrochemical EN 60746-1 2003

(standards.iteh.ai)

SIST EN IEC 60746-4:2019

iTeh STANDARD PREVIEW (standards.iteh.ai)



IEC 60746-4

Edition 2.0 2018-12

INTERNATIONAL **STANDARD**

NORME INTERNATIONALE

Expression of performance of electrochemical analyzers -V Part 4: Dissolved oxygen in water measured by membrane-covered amperometric sensors

SIST EN IEC 60746-4:2019
Expression des qualités de fonctionnement des analyseurs électrochimiques – Partie 4: Oxygène dissous dans l'éau mesuré par des capteurs ampérométriques recouverts d'une membrane

INTERNATIONAL **ELECTROTECHNICAL** COMMISSION

COMMISSION **ELECTROTECHNIQUE INTERNATIONALE**

ISBN 978-2-8322-6274-0 ICS 19.080; 71.040.40

Warning! Make sure that you obtained this publication from an authorized distributor. Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FC	DREWO	RD	.4	
1	Scop	e	.6	
2	Norm	ative references	.6	
3	Term	s and definitions	.6	
	3.1	Oxygen sensor properties	.7	
	3.2	Electronics		
	3.3	Measurement units and solubility of oxygen		
	3.4	Test media		
4	Influence quantities for membrane covered amperometric sensors			
	4.1	Temperature	.9	
	4.2	Pressure		
	4.3	Dissolved substances	.9	
	4.4	Flow	.9	
5	Proce	edure for specification	.9	
	5.1	Zero and span drift	.9	
	5.2	Additional specifications for the sensor unit		
	5.2.1	Electrode and sensor materials		
	5.2.2	Dimensions of the sensor	10	
	5.2.3	Dimensions of the sensor	10	
	5.2.4			
	5.2.5			
	5.2.6	Zero currentSIST.EN.IEC 60746-4:2019	10	
	5.2.7	Sensors/sensitivityeh.ai/catalog/standards/sist/d7291867-f59b-4566-82f7-	10	
	5.2.8	Stabilization time 19507d77a37/sist-en-iec-60746-4-2019	10	
	5.2.9	Oxygen consumption	10	
	5.2.1	0 Flow rate	10	
	5.2.1	1 Method and extent of sensor regeneration	10	
6		mmended standard values and ranges of influence quantities affecting the rmance of electronic units	10	
7	Verifi	cation of values	11	
	7.1	General	11	
	7.1.1	General aspects of verification of values		
	7.1.2	·		
	7.1.3			
	7.2	Simulator for testing electronic units		
	7.3	Calibration solutions		
	7.4	Testing procedures for complete analyzer (sensor unit connected to electronic unit)		
	7.4.1	Intrinsic uncertainty		
	7.4.2	•		
	7.4.3			
	7.4.4	Interference uncertainty (whole analyzer)	12	
	7.4.5			
	7.4.6	·		
	7.4.7	Delay times T_{10} and 90 % rise or fall times T_{90}	13	
	7.4.8	Temperature compensation	13	
	7.4.9	Operating uncertainty of the whole analyzer	14	

7.4.10 Determination of the sensor unit residual signal	14
7.4.11 Oxygen consumption	14
Annex A (informative) Supplementary general information on amperometric oxygen sensors	15
A.1 Sensors' performance characteristics	15
A.2 Precautions	16
A.3 Sensor calibration techniques	16
Annex B (informative) Technique for the preparation of batch calibration standards by the saturation approach [10]	18
Annex C (informative) Calibration solutions for low levels of oxygen in water measurement	19
C.1 System development [10]	19
C.2 Description and operation of the system	19
C.3 Further developments	19
Bibliography	25
Figure C.1 – Laboratory rig to produce water with a low level of dissolved oxygen	21
Figure C.2 – Complete system for laboratory testing dissolved oxygen monitor	22
Figure C.3 – Dimensions of block A	23
Figure C.4 – Dimensions of block B ANDARD PREVIEW	24
(standards.iteh.ai)	

SIST EN IEC 60746-4:2019

https://standards.iteh.ai/catalog/standards/sist/d7291867-f59b-4566-82f7-519507d77a37/sist-en-iec-60746-4-2019

IEC 60746-4:2018 © IEC 2018

INTERNATIONAL ELECTROTECHNICAL COMMISSION

-4 -

EXPRESSION OF PERFORMANCE OF ELECTROCHEMICAL ANALYZERS –

Part 4: Dissolved oxygen in water measured by membrane-covered amperometric sensors

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.

 519507d77a37/sist-en-iec-60746-4-2019
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60746-4 has been prepared by subcommittee 65B: Measurement and control devices, of IEC technical committee 65: Industrial-process measurement, control and automation.

This second edition cancels and replaces the first edition published in 1992. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) terms and definitions have been revised to meet the requirements of ISO/IEC Directives Part 2:2016.
- b) ISO 5814:2012 is cited as reference for solubility tables of dissolved oxygen in water with variable salt content at different pressure and temperature.

IEC 60746-4:2018 © IEC 2018

- 5 -

The text of this International Standard is based on the following documents:

FDIS	Report on voting
65B/1128/FDIS	65B/1138/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 60746 series, published under the general title *Expression of performance of electrochemical analyzers*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW (standards.iteh.ai)

EXPRESSION OF PERFORMANCE OF ELECTROCHEMICAL ANALYZERS –

Part 4: Dissolved oxygen in water measured by membrane-covered amperometric sensors

1 Scope

This part of IEC 60746 is intended:

- to specify terminology, definitions and requirements for statements by manufacturers for analyzers, sensor units and electronic units used for the determination of dissolved oxygen partial pressure or concentration;
- to establish performance tests for such analyzers, sensor units and electronic units;
- to provide basic documents to support the applications of quality assurance standards [1]1.

This document applies to analyzers using membrane covered amperometric sensors. It applies to analyzers suitable for use in water containing liquids, ultrapure waters, fresh or potable water, sea water or other aqueous solutions, industrial or municipal waste water from water bodies (e.g. lakes, rivers, estuaries), as well as for industrial process streams and process liquids. Whilst in principle amperometric oxygen-analyzers are applicable in gaseous phases, the expression of performance in the gas phase is outside the scope of this document.

SIST EN IEC 60746-4:2019

This document is applicable to analyzers specified for permanent installation in any location (indoors or outdoors) using membrane-covered amperometric sensors.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60746-1:2003, Expression of performance of electrochemical analyzers – Part 1: General

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60746-1 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at http://www.iso.org/obp

¹ Numbers in square brackets refer to the Bibliography.