



Designation: D 5366 – 96

## Standard Test Method for Determining the Dynamic Performance of a Wind Vane<sup>1</sup>

This standard is issued under the fixed designation D 5366; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### 1. Scope

1.1 This test method covers the determination of the starting threshold, delay distance, and overshoot ratio of a wind vane from direct measurements in a wind tunnel. This test method is applicable only to wind vanes having measurable overshoot.

1.2 This test method provides for determination of the performance of a system consisting of a wind vane and its associated position-to-output transducer in wind tunnel flow. Use of values determined by this test method to describe performance in atmospheric flow of a wind direction measuring system incorporating the vane must be done with an understanding of the differences between the two systems and the two environments.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

### 2. Terminology

#### 2.1 Definitions:

2.1.1 *delay distance* ( $D$ )—the distance the air flows past a wind vane during the time it takes the vane to return to 50 % of the initial displacement.

2.1.2 *overshoot* ( $\theta_n$ )—the amplitude of a deflection of a wind vane as it oscillates about  $\theta_B$  after release from an initial displacement.

2.1.3 *overshoot ratio* ( $\Omega$ )—the ratio of two successive overshoots, as expressed by the equation:

$$\Omega = \theta_{(n+1)}/\theta_n \quad (1)$$

where  $\theta_n$  and  $\theta_{(n+1)}$  are the  $n$  and  $n + 1$  overshoots, respectively. In practice, since deflections after the first (to the side opposite the release point are normally small, the initial release point (that is, the  $n = 0$  deflection) and the first deflection after release ( $n = 1$ ) are used in determining the overshoot ratio.

2.1.4 *starting threshold* ( $U_o$ )—the lowest speed at which the vane can be observed or measured moving from a 10° offset in a wind tunnel.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D-22 on Sampling and Analysis of Atmospheres and is the direct responsibility of Subcommittee D22.11 on Meteorology.

Current edition approved Oct. 10, 1996. Published December 1996. Originally published as D 5366 – 93. Last previous edition D 5366 – 95.

#### 2.2 Symbols:

$D$	(m)	delay distance
$U_o$	(m/s)	starting threshold
$\Omega$	(none)	overshoot ratio
$\eta$	(none)	damping ratio
$\lambda_d$	(m)	damped natural wavelength
$\theta_n$	(degrees)	overshoot; maximum angular excursion
$\theta_o$	(degrees)	reference direction
$\theta_B$	(degrees)	vane equilibrium position
$\theta_B - \theta_o$	(degrees)	dynamic vane bias

#### 2.3 Calculated or Estimated Values:

2.3.1 *damping ratio* ( $\eta$ )—calculated from the overshoot ratio (1,2).<sup>2</sup>

$$\eta = \frac{\ln(1/\Omega)}{(\pi^2 + [\ln(1/\Omega)]^2)^{0.5}} \quad (2)$$

2.3.2 *damped natural wavelength* ( $\lambda_d$ )—at sea level in the U.S. Standard Atmosphere, damped natural wavelength is related to delay distance and damping ratio by the empirical expression (1,2).<sup>2</sup>

$$\lambda_d = \frac{D(6.0 - 2.4\eta)}{(1 - \eta^2)^{0.5}} \quad (3)$$

### 3. Summary of Test Method

3.1 *Reference Direction* ( $\theta_o$ , degrees) is the indicated angular position of the vane when aligned along the centerline of the wind tunnel.

3.2 *Vane Equilibrium Position* ( $\theta_B$ , degrees) is the final resting position of the vane after motion in response to an initial displacement. Ideally,  $\theta_B = \theta_o$ .

3.3 *Dynamic Vane Bias* ( $\theta_B - \theta_o$ , degrees) is the displacement of the vane from the wind tunnel centerline at 5 m/s. This measurement will identify wind vanes with unbalanced aerodynamic response because of damage (for example, bent tail) or poor design.

3.4 *Starting Threshold* ( $U_o$ , m/s) is determined by observing or measuring the lowest speed at which the vane, released from a 10° offset position in a wind tunnel, moves toward  $\theta_B$ . Movement must be distinguishable from vibration.

3.5 *Delay Distance* ( $D$ , m) may be determined at a number of wind speeds but shall include 5 m/s and 10 m/s. It is computed from the time required for the vane to reach 50 % of the initial displacement from 10° off  $\theta_B$ . This time in seconds

<sup>2</sup> The boldface numbers in parentheses refer to the list of references at the end of this standard.

is converted to delay distance by multiplying by the wind tunnel speed in metres per second. Tests shall include an equal number of displacements to each side of  $\theta_B$ .

3.6 *Overshoot Ratio* ( $\Omega$ ) may be determined at the same time as the delay distance. The maximum angular excursion on the opposite side of  $\theta_B$  from the initial  $10^\circ$  displacement from  $\theta_B$  is measured. This value is divided by the initial displacement to obtain  $\Omega$ .

#### 4. Significance and Use

4.1 This test method will provide a standard for comparison of wind vanes of different types. Specifications by regulatory agencies and industrial societies (3-5) have stipulated performance values. This test method provides an unambiguous method for measuring starting threshold, delay distance, and overshoot ratio.

#### 5. Apparatus

##### 5.1 Wind Tunnel (6):

5.1.1 *Size*—The wind tunnel shall be large enough so that the total projected area of supports, sensor apparatus, and the vane in its displaced position is less than 5 % of the cross-sectional area of its test section.

5.1.2 *Speed Range*—The wind tunnel shall have a speed control that will allow the flow rate to be varied from 0 to at least 10 m/s. The speed control shall maintain the flow rate within  $\pm 0.2$  m/s.

5.1.3 *Turbulence and Swirl*—Across the volume to be occupied by the vane, the flow profile shall vary by no more than 1 % about the mean speed and shall exhibit a turbulence of less than 1 %. **CAUTION:** Swirl in the wind tunnel may influence starting threshold measurements. Variations in the measurement of  $\theta_B$  at low speeds likely indicate the existence of swirl.

5.1.4 *Calibration*—The mean flow rate shall be verified at the mandatory speeds of 5 and 10 m/s by use of transfer standards that have been calibrated by the National Institute of Standards and Technology (formerly called the National Bureau of Standards)<sup>3</sup> or by a fundamental physical method.

5.1.4.1 Speeds below 2 m/s for threshold determination shall be verified by a sensitive anemometer or by some fundamental time and distance technique, such as measuring the transition time of smoke puffs, soap bubbles, or heat puffs between two points separated by a known distance.

5.1.5 *Environment*—The temperature and pressure of the environment within the wind tunnel test section shall be reported. Differences of greater than 3 % in the density of air within the test environment may result in poor inter-comparability of independent measurements of starting threshold, delay distance, and overshoot ratio since these values are density dependent.

##### 5.2 Measuring System:

5.2.1 *Direction*—The resolution of the wind vane position-to-output transducer limits the resolution of the measurements. The accuracy of the position-to-output conversion shall

be within  $\pm 0.1^\circ$ . **CAUTION:** Avoid potentiometer dead spots or crossover positions while performing these procedures.

5.2.2 *Time*—The resolution of time shall be consistent with the distance accuracy required. For this reason, the time resolution may be changed as the wind tunnel speed is changed. For example, for a distance constant measurement to 0.1 m, one must have a time resolution of 0.05 s at 2 m/s and 0.01 s at 10 m/s. If time accuracy is based on commercial electrical power frequency, it will be at least an order of magnitude better than the resolution suggested above.

5.3 *Signal Conditioning*—Care shall be taken to avoid electronic circuits in signal conditioning and recording devices that adversely affect the apparent vane performance. **CAUTION:** Time constants in signal conditioning and recording devices shall be less than 0.01 s.

5.4 *Recording Techniques*—The measuring or recording system shall represent the  $10^\circ$  displacement on each side of  $\theta_B$  with a resolution of  $0.2^\circ$ . One simple technique is to use a fast-response recorder (flat to 40–60 Hz or better) with enough gain so that a vane can be oriented in the wind tunnel with  $\theta_B$  represented at mid-scale, and  $\pm 10^\circ$  of vane displacement traversing the full span of the recorder.

5.4.1 The recorder shall have a fast chart speed of 50 mm/s or more. An alternative is to use an FM tape recorder to record the signal. When played back at lower speed, a proportionately slower analog strip chart recorder yielding an equivalent 50-mm/s chart speed is acceptable. Oscilloscopes with memory and hard copy capability may also be used.

5.4.2 Digital recording and data reduction systems are satisfactory if the sampling rate is at least 100 per second.

#### 6. Sampling

6.1 *Starting Threshold*—Ten consecutive tests at the same speed meeting the test method requirement, five in each direction from  $\theta_B$ , are required for a valid starting threshold measurement.

6.2 *Delay Distance and Overshoot Ratio*—The arithmetic mean of ten tests, five in each direction from  $\theta_B$ , is required for a valid measurement at each speed. The results of measurements at two or more speeds shall be averaged to a single value for delay distance and a single value for overshoot ratio.

#### 7. Procedure

##### 7.1 Dynamic Vane Bias:

7.1.1 Set vane at tunnel centerline with no flow in the wind tunnel.

7.1.2 Adjust the wind tunnel to give a flow of 5 m/s.

7.1.3 Measure the equilibrium vane position ( $\theta_B$ ) relative to the tunnel centerline ( $\theta_o$ ). The angular difference,  $\theta_B - \theta_o$  is the dynamic vane bias.

7.1.4 A dynamic vane bias greater than  $1^\circ$  indicates poor design or a problem with the vane. Appropriate corrections shall be made before continuing.

##### 7.2 Starting Threshold:

7.2.1 Provide a mechanical method for holding and releasing the vane at  $10 \pm 1^\circ$  from  $\theta_B$ . With no flow in the wind tunnel, verify that the vane moves by no more than  $\pm 0.5^\circ$  when the release mechanism is activated.

7.2.2 Adjust the wind tunnel to a speed expected to be lower

<sup>3</sup> Available from National Institute of Standards and Technology (NIST), Gaithersburg, MD 20899.