

---

**Živalske in rastlinske maščobe in olja - Določanje relativne sestave olj in derivatov s kapilarno plinsko kromatografijo (metoda fingerprint)**

Animal and vegetable fats and oils - Relative composition of oils and derivatives by capillary gas chromatography (fingerprint method)

Tierische und pflanzliche Fette und Öle - Relative Zusammensetzung von Ölen und Derivaten mittels Kapillargaschromatographie (Fingerprint-Verfahren)

Corps gras d'origines animale et végétale - Composition relative des corps gras et produits dérivés par chromatographie gazeuse capillaire (méthode fingerprint)

<https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-4bcb967aa04/kSIST-TS-FprCEN-ISO-TS-22115-2020>

**Ta slovenski standard je istoveten z: FprCEN ISO/TS 22115**

---

**ICS:**

67.200.10	Rastlinske in živalske maščobe in olja	Animal and vegetable fats and oils
-----------	--	------------------------------------

**kSIST-TS FprCEN ISO/TS 22115:2020**    **en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[KSIST-TS FprCEN ISO/TS 22115:2020](https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebec967aa0/ksist-ts-fprcen-iso-ts-22115-2020)

<https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebec967aa0/ksist-ts-fprcen-iso-ts-22115-2020>

TECHNICAL  
SPECIFICATION

ISO/TS  
22115

First edition

---

---

**Animal and vegetable fats and  
oils — Relative composition of oils  
and derivatives by capillary gas  
chromatography (fingerprint method)**

*Corps gras d'origines animale et végétale — Composition relative  
des corps gras et produits dérivés par chromatographie gazeuse  
capillaire (méthode fingerprint)*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[KSIST-TS FprCEN ISO/TS 22115:2020](https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebec967aa0/ksist-ts-fprcen-iso-ts-22115-2020)

[https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-  
ebec967aa0/ksist-ts-fprcen-iso-ts-22115-2020](https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebec967aa0/ksist-ts-fprcen-iso-ts-22115-2020)

**PROOF / ÉPREUVE**

---

---



Reference number  
ISO/TS 22115:2020(E)

© ISO 2020

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[kSIST-TS FprCEN ISO/TS 22115:2020](https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebecbc967aa0/ksist-ts-fprcen-iso-ts-22115-2020)  
<https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebecbc967aa0/ksist-ts-fprcen-iso-ts-22115-2020>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>1</b>
<b>5 Reagents</b> .....	<b>1</b>
<b>6 Apparatus</b> .....	<b>2</b>
<b>7 Sample</b> .....	<b>3</b>
7.1 Sampling.....	3
7.2 Preparation of test sample.....	3
<b>8 Procedure</b> .....	<b>3</b>
8.1 Preparation of the internal standard, Tridecanoylglycerol (5.2.5), c = 20 mg/ml.....	3
8.2 Preparation of the individual standard solutions for determination of response factors.....	3
8.3 Silylation of the standards.....	4
8.4 Preparation of the sample.....	4
8.5 Sample solution.....	4
8.6 Silylation of the sample.....	4
8.7 Gas chromatography.....	4
8.8 Peak identification and integration.....	5
<b>9 Result of the determination</b> .....	<b>5</b>
9.1 Calculation of the response factor.....	5
9.2 Quantitative determination.....	6
<b>10 Precision of the method</b> .....	<b>7</b>
10.1 Interlaboratory test.....	7
10.2 Repeatability.....	7
10.3 Reproducibility.....	7
<b>11 Test report</b> .....	<b>7</b>
<b>Annex A (informative) Typical chromatograms</b> .....	<b>8</b>
<b>Annex B (informative) Results of an interlaboratory test</b> .....	<b>18</b>
<b>Bibliography</b> .....	<b>22</b>

## ISO/TS 22115:2020(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 307, *Oilseeds, vegetable and animal fats and oils and their by-products — Methods of sampling and analysis*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# Animal and vegetable fats and oils — Relative composition of oils and derivatives by capillary gas chromatography (fingerprint method)

## 1 Scope

This document specifies a method for the semi-quantitative analysis of oils, fats and oil/fat-related samples (deodistillates).

It is applicable to the screening of oils, fats and oil/fat-related samples to obtain main (e.g. triglycerides) and minor (e.g. sterols, sterol esters, tocopherols, wax esters, fatty alcohols, glycerol) component information in one single analysis. For a truly quantitative analysis of pre-identified compound classes, specific methods are more appropriate.

The method can also be used as a useful qualitative screening tool for the relative comparison of sample compositions.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 661, *Animal and vegetable fats and oils, — Preparation of test sample*

<https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebecbc967aa0/ksist-ts-fprcen-iso-ts-22115-2020>

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

## 4 Principle

The hydroxylated compounds are transformed into silyl derivatives. This operation has no effect on the apolar (non-hydroxylated) compounds also present in the sample. The sample prepared is analysed by gas chromatography (GC) on a high-temperature capillary column with a low film thickness, with an on-column injector and flame-ionization detector.

For quantitative purposes, the compounds are quantified in the presence of an internal standard (1,2,3-Tridecanoylglycerol) and the response factors are determined from a reference standard from each class.

## 5 Reagents

**WARNING — Attention is drawn to the regulations which specify the handling of hazardous substances. Technical, organizational and personal safety measures shall be followed.**

**ISO/TS 22115:2020(E)**

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and distilled or demineralized water or water of equivalent purity.

**5.1 Silylation reagent mixture:** Methylimidazole (CAS 616-47-7) and MSHFBA (CAS 53296-64-3).

Prepare a mixture of silylation reagents: 1 ml of MSHFBA + 50 µl of Methylimidazole.

NOTE This mixture cannot be stored more than one week due to moisture absorption. Indeed, silyl derivatives are moisture sensitive.

**5.2 Reference substances<sup>1)</sup>.**

**5.2.1 Oleic acid** (CAS 112-80-1).

**5.2.2 1-Monoolein** (CAS 111-03-5).

**5.2.3 1,3-Diolein** (CAS 2465-32-9).

**5.2.4 Triolein** (CAS 122-32-7).

**5.2.5 Tridecanoylglycerol** (CAS 621-71-6).

**5.2.6 Eicosanol** (CAS 629-96-9).

**5.2.7 α-Tocopherol** (CAS 10191-41-0), **γ-tocopherol** (CAS 54-28-4), **δ tocopherol** (CAS 119-13-1).

**5.2.8 Cholesterol** (CAS 57-88-5).

**STANDARD PREVIEW**  
(standard.it) <https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-eb0bc967590/ksist-ts-fprcen-iso-ts-22115-2020>

**5.2.9 Cholesterol palmitate** (CAS 601-34-3).

**5.2.10 Squalene** (CAS 111-02-4).

**5.2.11 Plant sterol mix<sup>2)</sup>** (CAS 474-67-9, 474-62-4, 83-48-7, 83-46-5).

**5.3 Isooctane**, trace organic analysis grade, purity 99 % min.

**5.4 Chloroform**, chromatographic quality.

## 6 Apparatus

Usual laboratory equipment and, in particular, the following.

**6.1 Volumetric flask**, 100 ml and 10 ml.

**6.2 Oven**, at 103 °C ± 1 °C

**6.3 Conical glass sample vials**, 10 ml capacity.

1) Suitable suppliers are Sigma-Aldrich ([www.sigmaaldrich.com](http://www.sigmaaldrich.com)) or Larodan ([www.larodan.com](http://www.larodan.com)). This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this supplier.

2) A suitable supplier is Larodan ([www.larodan.com](http://www.larodan.com)). This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this supplier.



**6.4 Gas chromatograph for capillary columns**, equipped with an on-column injector or equivalent device, a temperature-programmable oven and a flame ionization detector (FID).

**6.5 Fused silica capillary column**, capable of being programmed up to 400 °C (“high temperature” type) for which the following characteristics are advised: 100 % dimethylpolysiloxane or 95 % dimethyl/ 5 % diphenyl polysiloxane stationary phase, length 30 m, internal diameter 0,32 mm or 0,25 mm, film thickness 0,1 µm. Other columns of similar polarity and selectivity may also be used.

**6.6 Microsyringe**, 5 µl to 10 µl capacity, suitable for on-column injection in gas chromatography.

**6.7 Analytical balance**, reading accuracy 0,001 g.

**6.8 Ultrasonic bath.**

**6.9 Nitrogen evaporator.**

## 7 Sample

### 7.1 Sampling

Sampling is not part of the method specified in this document. A recommended sampling method is given in ISO 5555<sup>[1]</sup>.

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

### 7.2 Preparation of test sample

Prepare the test sample in accordance with ISO 661.

## 8 Procedure

### 8.1 Preparation of the internal standard, Tridecanoylglycerol (5.2.5), c = 20 mg/ml

Weigh 200 mg Tridecanoylglycerol (5.2.5) into a 10 ml volumetric flask, fill it up with isooctane (5.3).

### 8.2 Preparation of the individual standard solutions for determination of response factors

A 5 mg/ml standard solution in isooctane (5.3) is prepared for each compound (see Table 1) to determine the response factor for each component group.

**Table 1 — List of the standards for the calculation of the response factors**

Class determined	Standard
Alcohols	Eicosanol (5.2.6)
Free fatty acids	Oleic acid (5.2.1)
Hydrocarbons	Squalene (5.2.10)
Monoglycerides	Monoolein (5.2.2)
Tocopherols	α-Tocopherol (5.2.7)
Sterols	Cholesterol (5.2.8)
Diglycerides	Diolein (5.2.3)

**Table 1** (continued)

Class determined	Standard
Sterol esters	Cholesterol palmitate (5.2.9)
Triglycerides	Triolein (5.2.4)

### 8.3 Silylation of the standards

Introduce 100 µl of the internal standard solution, Tridecanoylglycerol (5.2.5),  $c = 20$  mg/ml, and 200 µl of the individual standard solution (each component separately) into a silylation tube (6.3) and evaporate the solvent to dryness under nitrogen (6.9). Introduce 100 µl of the silylation mixture (5.1). Keep at 103 °C for 15 min (6.2), then take up into 5 ml isooctane (5.3).

It is recommended when preparing the following solution to check the method precision (optional).

In a 100 ml volumetric flask (6.1), exactly weigh about 50 mg of each standard (5.2), including 50 mg of the internal standard solution, Tridecanoylglycerol (5.2.5), and dilute to 100 ml with chloroform (5.4).

This solution can be stored at ambient temperature for six months.

To keep the solution homogenous over time, the solution has to be placed into an ultrasonic bath for 10 min before use.

Introduce 0,5 ml of the solution into a silylation tube (6.3) and dry under nitrogen (6.9). Introduce 100 µl of the silylation mixture (5.1). Keep at 103 °C for 15 min (6.2), then take up into 10 ml isooctane (5.3).

### 8.4 Preparation of the sample (standards.iteh.ai)

If the sample is solid at room temperature, completely melt the sample in a 103 °C oven (6.2) and homogenize it.

<https://standards.iteh.ai/catalog/standards/sist/8731f238-1e37-4cce-9b31-ebecbc967aa0/ksist-ts-fprcen-iso-ts-22115-2020>

### 8.5 Sample solution

In a 10 ml volumetric flask (6.1), exactly weigh about 100 mg of sample. Introduce exactly 1 ml of internal standard solution (see 8.1) and dilute to 10 ml with isooctane (5.3).

If the sample is not totally soluble in isooctane, chloroform or toluene may be used instead of isooctane.

### 8.6 Silylation of the sample

Introduce 0,5 ml of the sample solution (see 8.5) into a silylation tube (6.3) and evaporate the solvent to dryness under nitrogen (6.9). Introduce 100 µl of the silylation mixture (5.1). Keep at 103 °C for 15 min (6.2), then take up into 10 ml isooctane (5.3).

### 8.7 Gas chromatography

Install the column in the gas chromatograph and check the working conditions by injecting the solvent, isooctane (5.3). The baseline should be straight with a small positive drift. If the drift is high, proceed to condition the column. For a negative drift, check the connections of the column.

If the column is being used for the first time, it is necessary to condition the column by heating it in the oven using a temperature gradient up to 370 °C (depending on the oven temperature chosen for the analysis) in 4 h. Maintain the temperature for 2 h.

A 1 m long, 0,5 mm diameter, empty fused silica capillary column is recommended as a pre-column.

The conditions in Table 2 for the gas chromatograph have been found to give useful chromatograms.

Solvent vent mode is recommended if the inlet in use is capable of using it.

Optimize the temperature programme and the velocity of the carrier gas flow so that chromatograms similar to the ones presented in [Figures A.1](#) to [A.10](#) are obtained.

**Table 2 — Gas chromatographic conditions**

Function	Conditions
Column	HP1 (30 m long, 0,25 mm i.d., 0,10 µm film thickness)
Oven temperature	Initial temperature 60 °C, programmed at 20 °C/min to 170 °C, 10 °C/min to 360 °C (hold 10 min)
Carrier gas	Linear velocity 87,7 cm/sec (carrier gas used H <sub>2</sub> )
Detector temperature	370 °C
Injector	Cold on-column
Injection volume	1 µl of the solution prepared in <a href="#">8.6</a>

## 8.8 Peak identification and integration

Identify the peaks with the standard solution prepared in [8.2](#). Typical chromatograms are presented in [Annex A](#). The relative retention times of identified compounds or families of compounds are listed in [Table 3](#).

When the compound of interest is present in the form of a single peak, integration according to the conventional criteria is used. By family, the resolution level is not always obtained as there can be some minor compounds. It is then recommended to integrate the set of peaks together or to combine them in a sum.

**Table 3 — Indicative relative retention times of identified compounds or families of compounds**

Retention time relative to Tridecanoylglycerol (RRT)	Determination or class of compounds
< 0,53	Alcohols ( <i>silyl derivative</i> )
0,57 to 0,63	Free fatty acids ( <i>silyl derivative</i> )
0,87	Squalene
0,82 to 0,86	Monoglycerides ( <i>silyl derivative</i> )
0,90	Delta-tocopherol ( <i>silyl derivative</i> )
0,94	Gamma-tocopherol ( <i>silyl derivative</i> )
0,99	α-Tocopherol ( <i>silyl derivative</i> )
0,98 to 1,05	Sterols ( <i>silyl derivative</i> )
1,00	Tridecanoylglycerol ( <i>internal standard</i> )
1,30 to 1,36	Diglycerides ( <i>silyl derivative</i> )
1,52 to 1,57	Sterol esters
> 1,72	Triglycerides

## 9 Result of the determination

### 9.1 Calculation of the response factor

Calculate the response factor,  $F$ , for each family representative compound  $x$  as given in [Formula \(1\)](#):

$$F_x = \frac{A_{IS} \cdot m_x}{m_{IS} \cdot A_x} \quad (1)$$

where