

---

**Zobozdravstvo - Laboratorijska rezila - 2. del: Kovinska laboratorijska rezila  
(ISO/DIS 7787-2:2019)**

Dentistry - Laboratory cutters - Part 2: Carbide laboratory cutters (ISO/DIS 7787-2:2019)

Zahnheilkunde - Laborfräser - Teil 2: Laborfräser aus Hartmetall (ISO/DIS 7787-2:2019)

Médecine bucco-dentaire - Fraises techniques - Partie 2: Fraises en carbure (ISO/DIS 7787-2:2019)

**Ta slovenski standard je istoveten z: prEN ISO 7787-2**

**ICS:**

11.060.25      Zobotehnični instrumenti      Dental instruments

**oSIST prEN ISO 7787-2:2019****en**



# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 7787-2

ISO/TC 106/SC 4

Secretariat: DIN

Voting begins on:  
2019-06-14Voting terminates on:  
2019-09-06

---

## Dentistry — Laboratory cutters —

### Part 2: Carbide laboratory cutters

*Médecine bucco-dentaire — Fraises techniques —**Partie 2: Fraises en carbure*

ICS: 11.060.20

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 7787-2:2020](https://standards.iteh.ai/catalog/standards/sist/c38d97a1-e6bd-45ff-8dc0-c7a8bc742bec/sist-en-iso-7787-2-2020)

<https://standards.iteh.ai/catalog/standards/sist/c38d97a1-e6bd-45ff-8dc0-c7a8bc742bec/sist-en-iso-7787-2-2020>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

**ISO/CEN PARALLEL PROCESSING**



Reference number  
ISO/DIS 7787-2:2019(E)

© ISO 2019

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 7787-2:2020

<https://standards.iteh.ai/catalog/standards/sist/c38d97a1-e6bd-45ff-8dc0-c7a8bc742bec/sist-en-iso-7787-2-2020>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>v</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms, definitions and symbols</b>	<b>1</b>
3.1 Terms and definitions	1
3.2 Symbols	2
<b>4 Classification</b>	<b>2</b>
<b>5 Requirements</b>	<b>2</b>
5.1 Materials	2
5.1.1 Working part	2
5.1.2 Shank	2
5.2 Shapes	2
5.3 Dimensions and number of blades	7
5.3.1 Working part	7
5.3.2 Shank	7
5.4 Tothing	7
5.5 Run-out	8
<b>6 Sampling</b>	<b>8</b>
<b>7 Measurement and test methods</b>	<b>8</b>
7.1 Shapes	8
7.2 Dimensions and number of blades	8
7.3 Tothing	8
7.4 Run-out	8
7.5 Pass/fail evaluation	8
<b>8 Designation</b>	<b>8</b>
8.1 Designation code numbers	8
8.2 Designation of tothing and number of blades	8
<b>9 Marking on the instrument</b>	<b>9</b>
<b>10 Labelling on the package</b>	<b>9</b>
<b>11 Packaging</b>	<b>9</b>
<b>Annex A (informative) Designation of tothing and number of blades</b>	<b>10</b>

## ISO/DIS 7787-2:2019(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

SIST EN ISO 7787-2:2020

This second edition cancels and replaces the first edition (ISO 7787-2:2000), which has been technically revised. The main changes compared to the previous edition are as follows:

- normative references have been updated;
- definitions were added;
- classification was added as [Clause 4](#);
- requirement for designation of toothing and number of blades was moved from the main body (normative) to informative [annex A](#) (now optional; reference only).

A list of all parts in the ISO 7787- series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

This document is one of a series of International Standards relating to dental rotary instruments.

The various dimensional and other requirements specified for carbide laboratory cutters are those considered important to ensure the interchangeability of these instruments.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

SIST EN ISO 7787-2:2020

<https://standards.iteh.ai/catalog/standards/sist/c38d97a1-e6bd-45ff-8dc0-c7a8bc742bec/sist-en-iso-7787-2-2020>



# Dentistry — Laboratory cutters —

## Part 2: Carbide laboratory cutters

### 1 Scope

This document specifies dimensional and other requirements for the 11 most commonly used carbide cutters which are predominantly used in the dental laboratory.

NOTE These cutters are also used in podiatry.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1797, *Dentistry — Shanks for rotary and oscillating instruments*

ISO 1942, *Dentistry — Vocabulary*

ISO 2157, *Dentistry — Nominal diameters and designation code numbers for rotary instruments*

ISO 8325, *Dentistry — Test methods for rotary instruments*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

### 3 Terms, definitions and symbols

#### 3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

##### 3.1.1

##### **laboratory cutter**

rotary cutting instrument designed for use with dental materials in the dental laboratory

##### 3.1.2

##### **dental laboratory**

facility where dental technician procedures complementing dental clinical treatment are carried out

[SOURCE: ISO 1942:2010, 2.77]

## ISO/DIS 7787-2:2019(E)

### 3.2 Symbols

$d$	diameter of the working part, head diameter
$l$	length of the working part, head length
$\alpha$	angle of working part

## 4 Classification

Dental laboratory cutters are classified according to the material used for the working part as

- steel laboratory cutters;
- carbide laboratory cutters;
- other suitable material specified in ISO 1797.

NOTE Steel laboratory cutters are described in ISO 7787-1. Carbide laboratory cutters are described in this part of ISO 7787, ISO 7787-3 and ISO 7787-4.

## 5 Requirements

### 5.1 Materials

#### 5.1.1 Working part

The working part of the laboratory cutter shall be made of tungsten carbide. The selection of the type of tungsten carbide and the treatment given to it shall be left to the discretion of the manufacturer.

#### 5.1.2 Shank


The shank of the laboratory cutter shall be made of stainless steel or other suitable material. The selection of the type of material and the treatment given to it shall be left to the discretion of the manufacturer.

### 5.2 Shapes

The shapes of the working part shall be as specified in [Figures 1](#) to [11](#).

Variations of the shapes are permitted within the limited dimensions and the descriptions used in the subclause titles.

Testing shall be carried out in accordance with [7.1](#).



**Figure 1 — Ball (spherical; round)**