
Determination of theanine in tea and instant tea in solid form using high-performance liquid chromatography

Détermination de la théanine dans le thé et le thé instantané sous forme solide en utilisant la chromatographie en phase liquide à haute performance

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 19563:2017

<https://standards.iteh.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017>



iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 19563:2017

<https://standards.iteh.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	2
6 Apparatus	3
7 Sampling	3
8 Preparation of test samples	3
9 Procedure	3
9.1 General	3
9.2 Determination of dry matter content	4
9.3 Test portion	4
9.3.1 Sample preparations	4
9.3.2 Sample clean-up (using a polyamide column) — Optional	4
9.4 Determination	5
9.4.1 Adjustment of the apparatus	5
9.4.2 HPLC analysis	5
9.4.3 Identification	5
10 Calculation	6
11 Precision	6
11.1 Interlaboratory test	6
11.2 Repeatability	6
11.3 Reproducibility	6
12 Test report	7
Annex A (informative) Typical theanine calibration graph	8
Annex B (informative) Typical chromatograms	9
Annex C (informative) Results of interlaboratory tests	12
Bibliography	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 8, *Tea*.

[ISO 19563:2017](https://standards.iteh.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017)

<https://standards.iteh.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017>

Introduction

Theanine (N-ethyl- γ -L-glutamine) is a non-protein derived (free) amino acid. In the plant world, it is a unique amino acid found in *Camellia* genus and the mushroom *Boletus badius* and *Ilex guayusa*. The most common and the most prominent occurrence is in *Camellia sinensis* (L.) O. Kuntze species.

Theanine constitutes *circa* 0,1 % to 2 % w/w of the dry weight in tea leaves (*Camellia sinensis*) and is approximately 50 % of the total free amino acids. Separation of L- and D-theanine by HPLC is not achieved by this method; however, L-theanine is the major form (more than 95 %) in all teas^{[2][3]}. Theanine has been associated with the typical tea umami taste and is also being studied to evaluate potential beneficial effects to health.

The umami taste characteristic has been associated with theanine which is also being evaluated for its potential benefits to health. This is also creating interest in theanine as an important novel food and dietary supplement ingredient.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 19563:2017

<https://standards.iteh.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017>

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

ISO 19563:2017

<https://standards.iteh.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017>

Determination of theanine in tea and instant tea in solid form using high-performance liquid chromatography

1 Scope

This document specifies a high-performance liquid chromatographic (HPLC) method for the determination of theanine content in tea (*Camellia sinensis*). It is applicable to both tea and instant tea samples. Separation of L- and D-theanine is not possible using this method; however, the L-enantiomer is the major form in tea.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1572, *Tea — Preparation of ground sample of known dry matter content*

ISO 1573, *Tea — Determination of loss in mass at 103 °C*

ISO 1839, *Tea — Sampling*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 7513, *Instant tea in solid form — Determination of moisture content (loss in mass at 103 °C)*

ISO 7516, *Instant tea in solid form — Sampling*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Principle

A reversed-phase (RP)-C18 column is used which is capable of separating more polar compounds using water as an eluent and UV detection at 210 nm. Theanine is separated from the other amino acids; however, the potential co-elution of theanine with methionine is not relevant due to the very small concentration of methionine in tea. It is recommended that polyamide column extract clean-up is used to remove interfering polyphenols for column protection when separation of theanine from other constituents in the extract sample is difficult. Alternatively, the column should be washed after each tea sample using acetonitrile.

5 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified.

5.1 Water, conforming to grade 1 of ISO 3696.

5.2 Acetonitrile, HPLC grade (however, super grade is preferable).

5.3 Methanol, regular laboratory distilled grade.

5.4 Polyamide (Polyamide CC 6¹⁾, Particle size: 0,05 mm to 0,16 mm for column chromatography).

5.5 L-theanine reference standard, >99 %, anhydrous.

5.6 HPLC mobile phases.

WARNING — Wear gloves and eye protection and dispense reagents in a fume cupboard.

5.6.1 Mobile phase A, 100 % purified water (5.1).

5.6.2 Mobile phase B, 100 % acetonitrile (5.2).

5.7 Standard solution.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

5.7.1 Standard stock solution, corresponding to 1 mg/ml.

Weigh (0,050 0 ± 0,000 1) g of pure L-theanine (5.5) in a 50 ml volumetric flask dissolve with purified water using sonication to aid dissolution and make up to the volume with purified water.

5.7.2 Standard working solution.

The stock solution is diluted with purified water to prepare the standard solutions in the concentration range of 5 µg/ml to 200 µg/ml as detailed in Table 1.

Table 1 — Stock solution dilutions

Standard solution	Working solution (µg/ml)	Volume taken from the stock solution (ml)	Final volume (ml)
A	5	0,5	100
B	10	1,0	100
C	20	1,0	50
D	50	2,5	50
E	100	5,0	50
F	150	7,5	50
G	200	10,0	50

1) Polyamide CC 6 is the required/minimum specification required for the separation which is available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this product.

6 Apparatus

Usual laboratory apparatus and, in particular, the following:

- 6.1 Analytical balances**, capable of weighing to an accuracy of $\pm 0,000$ 1 g.
- 6.2 Magnetic stirrer**, capable of 500 r/min.
- 6.3 Filters**, membrane filter units of 0,45 μm pore size for filtration of mobile phase and diluted sample extracts. All membranes used should be checked using a simple calibration solution to ensure that theanine retention does not occur.
- 6.4 Centrifuge**, capable of 12 300 relative centrifugal force (R.C.F.) (typically 13 500 r/min).
- 6.5 Glass column**, 17 cm \times 2,2 cm in diameter, (for polyamide clean-up only).
- 6.6 One mark volumetric flasks**, of capacities 50 ml and 100 ml.
- 6.7 High performance liquid chromatograph**, equipped to perform binary gradient elution, with a thermostatically controlled column compartment and an ultraviolet detector set at 210 nm.
- 6.8 Chromatographic column for HPLC**, Phenomenex Aqua^{TM2)} 250 \times 4,6 in diameter or 2 mm in diameter or equivalent.

NOTE This method has been developed with the AquaTM which has the minimum specification required for this analysis.

[ISO 19563:2017](https://standards.itch.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017)

7 Sampling

<https://standards.itch.ai/catalog/standards/sist/a90b24fd-8e0b-4faa-81cd-7abb144e7ad4/iso-19563-2017>

The sampling methods that shall be used are detailed in ISO 1839 for tea or ISO 7516 for instant tea.

A representative sample(s) shall be sent to the laboratory. It shall not be damaged or changed during transport or storage.

8 Preparation of test samples

To ensure homogeneity, grind the sample of leaf tea in accordance with ISO 1572 and store samples in well-sealed containers protected from light.

Grinding of instant tea samples is only required with samples having a coarse granular structure (e.g. freeze-dried instant tea).

9 Procedure

9.1 General

If it is required to check whether the results are within the acceptable repeatability limit (see [11.2](#)), carry out two single determinations in accordance with [9.2](#) to [9.4](#) under repeatability conditions.

2) Phenomenex AquaTM is the required/minimum specification required for the separation which is available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this product.

9.2 Determination of dry matter content

Calculate the dry matter content from the moisture content (loss in mass at 103 °C) determined on a portion of the test sample (Clause 8) in accordance with ISO 1573 for tea or ISO 7513 for instant tea.

9.3 Test portion

9.3.1 Sample preparations

9.3.1.1 Leaf tea

9.3.1.1.1 Weigh ($1,00 \pm 0,01$) g of a finely ground sample into a 200 ml beaker and add 100 ml of boiling water.

9.3.1.1.2 Allow brewing for 5 min using a magnetic stirrer and then filter through a filter paper (Grade 4 or equivalent) into a 100 ml volumetric flask.

9.3.1.1.3 Allow the tea brew to cool down to ambient temperature, make up to volume of 100 ml with purified water and filter using a $0,45 \mu\text{m}$ membrane before injection. Alternatively, approximately 1 ml of the sample solution can be centrifuged at 13 000 r/min for 10 min prior to HPLC analysis.

9.3.1.2 Tea extract or instant tea

9.3.1.2.1 Weigh ($0,100 \pm 0,001$) g sample into a 50 ml volumetric flask, dissolve and make up with heated purified water (approximately 50 °C).

9.3.1.2.2 Sonicate for 2 min to 5 min for complete dissolution and filter using a $0,45 \mu\text{m}$ membrane before injection. Alternatively, approximately 1 ml of the sample solution can be centrifuged at 13 000 r/min for 10 min prior to HPLC analysis.

9.3.2 Sample clean-up (using a polyamide column) — Optional

Certain tea samples will not elute with clean chromatograms. When this occurs in the position where theanine elutes in the chromatogram, the interference will prevent an accurate determination of the peak area. On these occasions, it is necessary to carry out the clean-up procedures as described below in order to improve separation and to protect the column.

9.3.2.1 Suspend 1,2 kg polyamide in a mixture of 8 l of water and 1,5 l of methanol, mix well and allow settling overnight.

9.3.2.2 Remove supernatant and wash with 1,5 l methanol twice. Store in methanol and mix well before use.

9.3.2.3 Column chromatography:

9.3.2.3.1 Fill a glass column (17 cm \times 2,2 cm in diameter) equipped with a thin layer of glass wool and sea sand with swollen polyamide to a height of 8 cm. Condition with 250 ml of water.

9.3.2.3.2 Apply 50 ml of the tea solution to the column and immediately collect the eluate into a 100 ml volumetric flask.

9.3.2.3.3 Wash the column with purified water until the mark of 100 ml is reached. Mix well and filter using a 0,45 µm membrane before injection. Alternatively, approximately 1 ml of the sample solution can be centrifuged at 13 000 r/min for 10 min prior to HPLC analysis.

The polyphenols will be retained on the polyamide while the theanine elutes. This step corresponds to a twofold dilution.

9.4 Determination

9.4.1 Adjustment of the apparatus

Set up the chromatograph (6.7) in accordance with the manufacturer's instructions and adjust it as follows.

- Flow rate of the mobile phase (5.6.1 and 5.6.2): 1,0 ml/min for 4,6 mm columns and 0,25 ml/min for 2 mm columns.
- Binary gradient conditions: according to the elution programme as detailed in Table 2.
- Temperature of the column (6.8): ambient, optional 20 °C.
- UV detector set at 210 nm.
- Injection volume: 20 µl (optional 10 µl to 50 µl).

Table 2 — Elution programme

Time (min)	% A	% B	Comment
0	100	0	analysis
10	100	0	analysis
12	20	80	wash
20	20	80	wash
22	100	0	conditioning
40	100	0	conditioning

9.4.2 HPLC analysis

Once the flow rate of the mobile phase (5.6) and temperature [see 9.4.1 c)] are stable, condition the column with a blank gradient run (9.4.1). Then inject onto the column 20 µl of each of the standard working solutions (5.7.2), followed by an equal volume of the sample solution (9.3.1.1.3 or 9.3.1.2.2). Repeat the injection of the mixed working standard solutions (5.7.2) at regular intervals (typically after six test solutions). Collect data using the data collection/integration system of the HPLC (6.5) for the peak of theanine in the standard and test sample solutions.

After each batch of analysis, thoroughly clean the column by flushing the chromatographic system and column with 50 % acetonitrile, 50 % water mobile phase. Replace column sealing plugs if disconnected for storage.

9.4.3 Identification

Identify the theanine by comparing retention time from sample chromatograms with that obtained from the standard solutions under the same chromatographic conditions (see 9.4.1).

NOTE Typical chromatograms can be found in Annex B.