



Designation: E563-02<sup>ε1</sup> Designation: E 563 – 08

# Standard Practice for Preparation and Use of an Ice-Point Bath as a Reference Temperature <sup>1</sup>

This standard is issued under the fixed designation E 563; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

---

<sup>ε1</sup>NOTE—Footnote 1 was editorially corrected in March 2005.

---

## 1. Scope

1.1 This practice covers a method of preparing, maintaining, and using a temperature reference bath of a mixture of shaved ice and water, saturated with air at a pressure of 101 325 Pa (1 atm).

1.2 An industrial practice for relating values referenced to the ice point and to the water triple point on the ITS-90 is included.

1.3 Methods to promote uniformity of bath temperature by mechanical stirring or agitation are not described in detail.

1.4 Methods of approximating the ice point, as by thermostatically-controlled refrigeration, are not covered by this practice.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

D 1193 ~~Standard Specification for Reagent Water~~ Specification for Reagent Water

E 344 Terminology Relating to Thermometry and Hydrometry

E 1594 Guide for Expression of Temperature

## 3. Terminology

3.1 *Definitions*—Definitions given in Terminology E 344, unless otherwise defined herein, apply to terms as used in this practice.

3.2 Temperature relationships given in Guide E 1594, unless otherwise defined herein, apply to temperature values as used in this practice.

3.3 *Definitions of Terms Specific to This Standard:*

3.3.1 *ice-point bath, n*—physical system containing ice and water assembled to realize the ice point as a reference temperature, or to establish a constant temperature near 0 °C.

## 4. Summary of Practice

4.1 The ice-point bath described by this practice consists of an intimate mixture, without voids, of pure shaved ice or ice particles and distilled air-saturated water in a thermally insulating vessel open to the atmosphere.

4.2 The ice bath realization of the ice point physically approximates, with small uncertainty, a natural fixed-point temperature.

4.2.1 ~~An ice-point bath prepared using distilled water ice and air-saturated, chilled distilled water, typically has a temperature of  $0.000 \pm 0.002$  °C. Consequently, when the bath is used as a fixed-point temperature reference, the bath temperature is assumed to be 0 °C, with an uncertainty that depends on the care with which the bath was established and maintained.~~

4.2.2 ~~The ice-point bath is open to the atmosphere; the effect of barometric pressure on the ice point is  $-75$  nK/Pa ( $-7.6$  mK/atm). The solubility of air in the water is directly proportional to the atmospheric pressure.~~

4.3 ~~The ice-bath temperature can also be measured with an accurately calibrated thermometer or compared to a water triple point~~

---

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee E20 on Temperature Measurement and is the direct responsibility of Subcommittee E20.07 on Fundamentals in Thermometry.

Current edition approved October 10, 2002. Published May 2003. Originally approved in 1976. Discontinued February 1996 and reinstated in 1997 as E563-97.

Current edition approved Nov. 1, 2008. Published December 2008. Originally approved in 1976. Discontinued February 1996 and reinstated in 1997 as E 563 – 97. Last previous edition approved in 2002 as E 563–02<sup>ε1</sup>.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

cell and the bath temperature is reported as the measured temperature with an uncertainty that is attributed to the measurement, not to the ice point.

4.2.1 An ice-point bath prepared by rigorous application of this practice, using distilled-water ice and air-saturated, chilled distilled water, typically has a temperature of  $0.000 \pm 0.002 \text{ }^\circ\text{C}$  at a barometric pressure of 101,325 Pa (1 standard atmosphere). See 8, Precision and Bias.

4.2.2 The ice-point bath is open to the atmosphere. The solubility of air in water, which affects phase change, is directly proportional to the atmospheric pressure. The effect of barometric pressure on the ice point is  $-75 \text{ nK/Pa}$  ( $-7.6 \text{ mK/atm}$ ). Accordingly, the change in ice-point temperature resulting from an increase in elevation above sea level is  $-0.86 \text{ mK}$  per 1000 m increase in altitude ( $-0.26 \text{ mK}$  per 1000 ft increase in altitude). See Table 1.

4.3 The ice-bath temperature can also be measured with an accurately calibrated thermometer or compared to a water triple point cell and the bath temperature reported as the measured temperature with an uncertainty that is attributed to the measurement, not to the ice point.

## 5. Significance and Use

5.1 This practice is adequate for use with other ASTM standards that specify the ice point as a reference. It is also intended to be adequate for most other ice-point reference purposes.

5.2 The ice point is a common practical industrial reference point of thermometry. The ice point is relatively simple to realize and provides a readily available natural fixed-point reference temperature.

5.3 *Use in Resistance Thermometry :*

5.3.1 The ice point was a defining fixed point on practical temperature scales prior to 1960.

5.3.2 The ITS-90 defines  $W(T_{90}) = R(T_{90})/R(273.16 \text{ K})$ , the measured resistance ratio of a Standard Platinum Resistance Thermometer (SPRT), in reference to the water triple point, not the ice point **(1)**.<sup>3</sup> In many instances, where the water triple point is not available, or when the accuracy obtainable with the water triple point is not required, reference to a properly established and maintained ice-point reference is used. For industrial-quality resistance thermometers, the resistance value is determined for  $0 \text{ }^\circ\text{C}$ , and an uncertainty is assigned that is appropriate for the quality of the ice-point realization is assigned.

5.4 *Use in Thermoelectric Thermometry :*

5.4.1 In thermoelectric thermometry, the ice point is ordinarily used as the reference temperature **(2)**.

5.4.2 Adequate thermoelectric reference requires that thermocouple junctions be well-coupled thermally to the bath, electrically isolated from each other and from the bath, and adequately immersed to avoid perturbing the reference-junction temperatures by radiation and longitudinal conduction of heat along the thermoelements **(3 and 4)**.

5.5 *Use in Liquid-in-Glass Thermometry :*

5.5.1 In liquid-in-glass thermometry, the ice point is ordinarily used as the reference temperature **(6)**.

5.5.2 The periodic recalibration of a liquid-in-glass thermometer at the ice point provides a reliable indication of the effect of gradual relaxation of residual mechanical strains in the glass that have a significant effect on the volume of the bulb, and provides a means for the accurate adjustment of the remainder of the scale **(6)**.

## 6. Hazards

6.1 Excess water accumulating in any region, particularly around the reference location, can elevate the temperature in that vicinity above the ice point. Errors, usually somewhat less than  $4 \text{ }^\circ\text{C}$ , can occur from this cause in poorly maintained baths and with poorly positioned test objects **(3 and 4)**.

6.2 For a stirred bath, the temperature of the bath will depend on the heat gained by the bath, the amount of water and ice, and the vigor of stirring. The uniformity of temperature of the bath can be enhanced by slowly stirring or agitating the slush of ice and water either manually or by a powered stirring means so that all of the ice and water in the bath come into intimate contact.

6.3 Ice making machines operate below  $0 \text{ }^\circ\text{C}$ . Therefore, when excessively large ice particles are used to prepare the ice-point bath, the initial temperature of the bath can briefly be slightly below the ice point. Also, some of the water may freeze and bridge

<sup>3</sup> The boldface numbers in parentheses refer to the list of references at the end of this standard.

**TABLE 1 Change in Ice Point Temperature Corresponding to Elevation Above Sea Level**

Elevation	Change in ice point, mK
Sea Level	0
500 m (1640 ft)	-0.43
1000 m (3281 ft)	-0.86
1500 m (4921 ft)	-1.29
2000 m (6562 ft)	-1.72
2500 m (8202 ft)	-2.15
3000 m (9843 ft)	-2.58