



# SLOVENSKI STANDARD SIST EN ISO 22442-1:2021

01-februar-2021

Nadomešča:

SIST EN ISO 22442-1:2016

---

**Medicinski pripomočki, ki uporabljajo živalska tkiva in njihove derivate - 1. del:  
Uporaba obvladovanja tveganja (ISO 22442-1:2020)**

Medical devices utilizing animal tissues and their derivatives - Part 1: Application of risk management (ISO 22442-1:2020)

Tierische Gewebe und deren Derivate, die zur Herstellung von Medizinprodukten eingesetzt werden - Teil 1: Anwendung des Risikomanagements (ISO 22442-1:2020)

Dispositifs médicaux utilisant des tissus animaux et leurs dérivés - Partie 1: Application de la gestion des risques (ISO 22442-1:2020)

**Ta slovenski standard je istoveten z: EN ISO 22442-1:2020**

---

**ICS:**

11.040.99	Druga medicinska oprema	Other medical equipment
11.120.01	Farmacija na splošno	Pharmaceutics in general

**SIST EN ISO 22442-1:2021**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 22442-1:2021](https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 22442-1**

December 2020

ICS 11.100.20

Supersedes EN ISO 22442-1:2015

English Version

**Medical devices utilizing animal tissues and their  
derivatives - Part 1: Application of risk management (ISO  
22442-1:2020)**

Dispositifs médicaux utilisant des tissus animaux et  
leurs dérivés - Partie 1: Application de la gestion des  
risques (ISO 22442-1:2020)

Tierische Gewebe und deren Derivate, die zur  
Herstellung von Medizinprodukten eingesetzt werden -  
Teil 1: Anwendung des Risikomanagements (ISO  
22442-1:2020)

This European Standard was approved by CEN on 2 December 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	<b>Page</b>
<b>European foreword</b> .....	<b>3</b>
<b>Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC as amended by Commission Regulation (EU) No 722/2012</b> .....	<b>5</b>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 22442-1:2021](https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021)  
<https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021>

## European foreword

This document (EN ISO 22442-1:2020) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2021, and conflicting national standards shall be withdrawn at the latest by June 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 22442-1:2015.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZA, which is an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard "within the meaning of Annex ZA", the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

**NOTE** The way in which these references documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

**Table — Correlations between normative references and dated EN and ISO standards**

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 10993-1	EN ISO 10993-1:2020	ISO 10993-1:2018
ISO 14971	EN ISO 14971:2019	ISO 14971:2019
ISO 22442-2:2020	EN ISO 22442-2:2020	ISO 22442-2:2020
ISO 22442-3:2007	EN ISO 22442-3:2007	ISO 22442-3:2007

**EN ISO 22442-1:2020 (E)**

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

**Endorsement notice**

The text of ISO 22442-1:2020 has been approved by CEN as EN ISO 22442-1:2020 without any modification.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 22442-1:2021](https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021>

## Annex ZA (informative)

### Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC as amended by Commission Regulation (EU) No 722/2012

This European Standard has been prepared under a Commission's standardization request [M/023 concerning the development of European Standards related to medical devices] to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [O] L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining acceptable risk must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 This Annex ZA is based on normative references according to the table of references in the European foreword, replacing the references in the core text.

NOTE 4 When an Essential Requirement does not appear in Table ZA, it means that it is not addressed by this European Standard.

**Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC as amended by Commission Regulation (EU) No 722/2012**

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
7.1 first and second indents	4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	Covered for the choice of materials based on animal tissues and their derivatives with regards to toxicity and compatibility with biological tissues, cells and body fluids. Not covered for flammability. Not covered for manufacture.
7.2	4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	Covered for reducing risks to patients resulting from contaminants in materials used in medical devices based on animal tissues and their derivatives, providing any resulting risks are minimized. Not covered for packing, manufacture, transport or

## EN ISO 22442-1:2020 (E)

Essential Requirements of Directive 93/42/EEC	Clause(s)/sub-clause(s) of this EN	Remarks/Notes
		storage.
8.1 First sentence only	4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C	Covered for reducing risks of infection to patients resulting from using materials in medical devices based on animal tissues and their derivatives, providing any resulting risks are eliminated or reduced as far as possible. Not covered for manufacture.
8.2 first paragraph only	4.2.1.2 d)	Covered only for the veterinary control of animals from which tissues and their derivatives have been obtained.
Annex I of Commission Regulation (EU) No 722/2012	4.1, 4.2, 4.3, 4.4, 4.5, 4.6 and Annex C  SIST EN ISO 22442-1:2021 <a href="https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7222-4202-813c-1213a4c5a660/sist-en-iso-22442-1-2021">https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7222-4202-813c-1213a4c5a660/sist-en-iso-22442-1-2021</a>	The requirements detailed in Section 1 of Annex I of Reg. 722/2012/EC cover risk analysis and risk management of medical devices manufactured utilising non-viable tissues, or derivatives thereof, sourced from animals that are susceptible to TSEs, as defined in Art. 1.2. The Regulation is therefore specific to TSE risks.  Annex C is dedicated to TSE risk, but 4.1, 4.2, 4.3, 4.4, 4.5 and 4.6 are relevant to the risk analysis and risk management process required by the Regulation.

**WARNING 1** — Presumption of conformity stays valid only as long as a reference to this European Standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

**WARNING 2** — Other Union legislation may be applicable to the products falling within the scope of this standard.



INTERNATIONAL  
STANDARD

ISO  
22442-1

Third edition  
2020-09

---

---

**Medical devices utilizing animal  
tissues and their derivatives —**

**Part 1:  
Application of risk management**

*Dispositifs médicaux utilisant des tissus animaux et leurs dérivés —*

*Partie 1: Application de la gestion des risques*

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

[SIST EN ISO 22442-1:2021](https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021>



Reference number  
ISO 22442-1:2020(E)

© ISO 2020

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 22442-1:2021](https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Risk management process</b> .....	<b>3</b>
4.1 General.....	3
4.2 Risk analysis.....	4
4.2.1 Identification of qualitative and quantitative characteristics related to the safety of medical devices.....	4
4.2.2 Identification of hazards and hazardous situations.....	5
4.3 Risk evaluation.....	5
4.4 Risk control.....	5
4.4.1 General.....	5
4.4.2 Risk control for viruses and TSE agents.....	5
4.4.3 Risk control of other hazards.....	6
4.4.4 Residual risk evaluation.....	6
4.5 Evaluation of overall residual risk acceptability.....	7
4.5.1 General.....	7
4.5.2 Documentation.....	7
4.6 Production and post-production information system.....	7
<b>Annex A (informative) Guidance on the application of this document</b> .....	<b>8</b>
<b>Annex B (informative) Graphical representation of part of the risk management process for medical devices utilizing animal material</b> .....	<b>9</b>
<b>Annex C (normative) Special requirements for some animal materials considering the risk management for TSE agents</b> .....	<b>11</b>
<b>Annex D (informative) Information relevant to the management of TSE risk</b> .....	<b>17</b>
<b>Bibliography</b> .....	<b>24</b>

## ISO 22442-1:2020(E)

## Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)) or the IEC list of patent declarations received (see <http://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 194, *Biological and clinical evaluation of medical devices*, Subcommittee SC 1, *Tissue product safety*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 206, *Biological and clinical evaluation of medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 22442-1:2015), which has been technically revised.

The main changes compared to the previous edition are as follows:

- [4.4.2](#) has been updated;
- weblinks in [C.2](#), bullet point 1, [C.3.3](#) and [C.4.4](#) have been updated;
- the weblink in [D.3.3](#) has been updated;
- [C.10](#) has been added;
- the bibliography has been updated.

A list of all parts in the ISO 22442 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Certain medical devices utilize materials of animal origin.

Animal tissues and their derivatives are used in the design and manufacture of medical devices to provide performance characteristics that have been chosen for advantages over non-animal based materials. The range and quantities of materials of animal origin in medical devices vary. These materials can comprise a major part of the device (e.g. bovine/porcine heart valves, bone substitutes for use in dental or orthopaedic applications, haemostatic devices), can be a product coating or impregnation (e.g. collagen, gelatine, heparin), or can be used in the device manufacturing process (e.g. tallow derivatives such as oleates and stearates, foetal calf serum, enzymes, culture media).

ISO 14971 is a general standard which specifies a process for a manufacturer by identifying hazards and hazardous situations associated with medical devices, including *in vitro* medical devices, to estimate and evaluate the risks associated with those hazards, to control these risks and to monitor the effectiveness of the control throughout the life cycle. This document provides additional requirements and guidance for the evaluation of medical devices manufactured utilizing animal tissues or derivatives which are non-viable or rendered non-viable.

This document is intended to cover medical devices including active implantable medical devices such as implantable infusion pumps.

This document does not apply to *in vitro* diagnostic devices.

This document can only be used in combination with ISO 14971 and is not a “stand-alone” standard.

NOTE Compliance to this document is shown by fulfilling its specified requirements. The guidance given in the notes and the informative annexes is not normative and is not provided as a checklist for auditors.

[SIST EN ISO 22442-1:2021](https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/5a6ec7b3-7ad9-47e5-982a-1213a4c5a660/sist-en-iso-22442-1-2021>