

DRAFT INTERNATIONAL STANDARD

ISO/DIS 16063-17

ISO/TC 108/SC 3

Secretariat: DS

Voting begins on:
2015-08-24

Voting terminates on:
2015-11-24

Methods for the calibration of vibration and shock transducers —

Part 17: Primary calibration by centrifuge

*Méthodes pour l'étalonnage des transducteurs de vibrations et de chocs —
Partie 17: Étalonnage primaire par centrifugeur*

ICS: 17.160

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f42b80dc-38c7-4311-8d6b-93454d146a04/iso-16063-17-2016>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.



Reference number
ISO/DIS 16063-17:2015(E)

© ISO 2015

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f42b80dc-38c7-4311-8d6b-93454d146a04/iso-16063-17-2016>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland.

Contents

Foreword	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Requirements for apparatus and environmental conditions	4
5 Preferred values	5
6 Method 1 (with measurement of the radius of rotation)	6
7 Method 2 (without measurement of the radius of rotation)	7
Annex A (informative) Expression of uncertainty of measurement	8
A.1 Uncertainty analysis on Method 1	8
A.2 Uncertainty analysis on Method 2	9
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC ISO/TC 108, *Mechanical vibration, shock and condition monitoring*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.

ISO 16063-17 replaces ISO 5347-7:1993.

Methods for the calibration of vibration and shock transducers – Part 17: Primary calibration by centrifuge

1 Scope

ISO 16063 comprises a series of documents dealing with methods for the calibration of vibration and shock transducers.

This part of ISO 16063 lays down detailed specifications for the instrumentation and procedure to be used for primary calibration of accelerometers using centrifuge calibration.

This part of ISO 16063 applies to rectilinear accelerometers with zero-frequency response, mainly of the strain gauge or piezoresistive type, and to primary standard and working transducers.

It is applicable for a calibration range from 10 m/s² to 20 000 m/s² (higher accelerations possible) at 0 Hz.

The limits of uncertainty applicable are ± 1 % of reading.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2041, *Mechanical vibration, shock and condition monitoring – Vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 2041 apply.

4 Requirements for apparatus and environmental conditions

4.1 Calibration environment

The standard reference atmospheric conditions are: (23 ± 3) °C; and 75 % relative humidity max. The temperature, humidity and atmospheric pressure shall be measured and reported.

4.2 Balanced table or arm (rotational table)

The main component of the calibration apparatus consists of balanced table or arm which rotates about a vertical axis with uniform angular speed. For the calibration range from 10 m/s² to 100 m/s² the table/arm shall be level within $\pm 0,5^\circ$ of horizontal. For ranges higher than 100 m/s², levelling is allowed to within $\pm 2^\circ$. The calibration apparatus shall be placed on a sufficiently heavy base which is sufficiently isolated from the floor vibration.

The rotational frequency shall be uniform within $\pm 0,05$ % of the nominal value.

The transducers axis of sensitivity shall be aligned within $\pm 0,5^\circ$.

The radius of rotation to the centre of the transducer mass element shall be measured with an uncertainty less than $\pm 0,15$ %. If it is not possible, method 2 shall be obtained.

If the accelerometer is substituted by impedances not sensitive to acceleration, the hum and noise when the centrifuge is rotating at the calibration speeds shall be at least 0,5 % below reading.

The apparatus of the main component is illustrated by Figure 1.




Figure 1 – Illustration of the apparatus of the main component

4.3 Rotational frequency measuring instrumentation

The relative expanded measurement uncertainty ($k = 2$) contribution of the measurement of rotational frequency shall be 0,05 % or less.

4.4 Voltage-measuring instrumentation

The relative expanded measurement uncertainty ($k = 2$) contribution of the voltmeter shall be 0,01% or less.

5 Preferred values

Six acceleration values, in metres per second squared, equally covering the accelerometer range, shall be chosen from the following series: 10; 20; 50; 100; 200; 500 or their multiples of ten. The reference acceleration shall be 100 m/s² (second choice: 50 m/s²).

6 Method 1 (with measurement of the radius of rotation)

6.1 Test procedure

Rotate the table or arm at different frequencies determined by calculation from the standard levels using the following formula:

$$a = 4\pi^2 n^2 r \tag{1}$$

where

n is the rotational frequency, in hertz;

r is the radius of rotation to the centre of the accelerometer mass element, in metres.

Measure the transducer output for every level.

Determine the reference calibration factor at the reference acceleration. Then determine the sensitivity for the other calibration amplitudes. The results shall be given as a percentage deviation from the reference calibration factor.

6.2 Expression of results

The calibration factor, S , in volts per (metre per second squared) [V/(m/s²)], is given by the following formula:

$$S = \frac{V}{4\pi^2 n^2 r} \tag{2}$$

where

V is the transducer output, in volts;

n is the rotational frequency, in hertz;

r is the radius of rotation to the centre of the accelerometer mass element, in metres.

When the calibration results are reported, the total uncertainty of the calibration and the corresponding confidence level, calculated in accordance with Annex A, shall also be reported.

7 Method 2 (without measurement of the radius of rotation)

7.1 Test procedure

If the rotational radius cannot be measured with the specified uncertainty, the transducer can be rotated in two different positions, the radial distance between these to be measured with uncertainty maximum of $\pm 0,5 \%$.

Measure the two rotational frequencies giving the same pick-up output at the two positions.

7.2 Expression of results

The calibration factor, S , in volts per (metre per second squared), is given by the following formula:

$$S = \frac{V}{4\pi^2 n_2^2 \frac{\Delta r}{1 - (n_2/n_1)^2}} \quad (3)$$

where

V is the accelerometer output, in volts;

n_1 is the rotational frequency at the first accelerometer position, in hertz;

n_2 is the rotational frequency at the second accelerometer position, in hertz;

Δr is the distance between the two accelerometer positions, in metres.

When the calibration results are reported, the total uncertainty of the calibration and the corresponding confidence level, calculated in accordance with Annex A, shall also be reported.

Annex A (informative)

Expression of uncertainty of measurement in calibration

A.1 Uncertainty analysis on Method 1

Components of uncertainty that contribute to the uncertainty of the measurement on Method 1 are listed in Table A.1. The combined standard uncertainty u_c can be expressed by

$$u_c = \sqrt{\sum_{i=1}^{10} (u_i)^2}$$

where u_i is uncertainty component in Table A.1 respectively.

The expanded uncertainty U shall be determined by multiplying u_c by coverage factor k , where a value of $k = 2$.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f42b99dc-38c7-4311-8d6b-93454d146a04/iso-16063-17-2016>