



Standard Test Method for Vapor Pressure of Petroleum Products (Mini Method— Atmospheric)¹

This standard is issued under the fixed designation D 5482; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope *

1.1 This test method covers a procedure for the determination of total vapor pressure of petroleum products using automatic vapor pressure instruments. The test method is suitable for testing samples with boiling points above 0°C (32°F) that exert a vapor pressure between 7 and 110 kPa (1.0 and 16 psi) at 37.8°C (100°F) at a vapor-to-liquid ratio of 4:1. The test method is applicable to gasolines containing oxygenates. No account is made of dissolved water in the sample.

NOTE 1—Because the external atmospheric pressure does not influence the resultant vapor pressure, this vapor pressure is an absolute pressure at 37.8°C (100°F) in kPa (psi). This vapor pressure differs from the true vapor pressure of the sample due to some small vaporization of the sample and dissolved air into the air of the confined space.

1.1.1 Some gasoline-oxygenate blends may show a haze when cooled to 0 to 1 °C. If a haze is observed in 8.5, it shall be indicated in the reporting of results. The precision and bias statements for hazy samples have not been determined (see Note 6).

1.2 This test method is a modification of Test Method D 5191 (Mini Method) in which the test chamber is at atmospheric pressure prior to sample injection.

1.3 This test method covers the use of automated vapor pressure instruments that perform measurements on liquid sample sizes in the range from 1 to 10 mL.

1.4 This test method is suitable for the determination of the dry vapor pressure equivalent (DVPE) of gasoline and gasoline-oxygenate blends by means of a correlation equation (see 13.2). The calculated DVPE is considered equivalent to the result obtained on the same material when tested by Test Method D 4953.

1.5 The values stated in acceptable SI units are regarded as standard. The values given in parentheses are provided for information only.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the*

responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. (For specific hazard statements, see 7.2 through 7.7.)

2. Referenced Documents

2.1 ASTM Standards:

D 4057 Practice for Manual Sampling of Petroleum and Petroleum Products²

D 4953 Test Method for Vapor Pressure of Gasoline and Gasoline-Oxygenate Blends (Dry Method)²

D 5190 Test Method for Vapor Pressure of Petroleum Products (Automatic Method)²

D 5191 Test Method for Vapor Pressure of Petroleum Products (Mini Method)²

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *dry vapor pressure equivalent (DVPE)*—a value calculated by a correlation equation (see 13.2) from the total pressure.

3.1.1.1 *Discussion*—The DVPE is expected to be equivalent to the value obtained on the sample by Test Method D 4953.

3.1.2 *total pressure*—the observed pressure measured in the experiment that is the resultant pressure increase from the initial ambient atmospheric pressure.

4. Summary of Test Method

4.1 A known volume of chilled, air-saturated sample is introduced into a thermostatically controlled test chamber, the internal volume of which is five times that of the total test specimen introduced into the chamber. The test chamber is at atmospheric pressure prior to introduction of the sample. After introduction of the sample into the test chamber, the test specimen is allowed to reach thermal equilibrium at the test temperature, 37.8°C (100°F). The resulting rise in pressure in the chamber is measured using a pressure transducer sensor and indicator.

4.2 The measured total vapor pressure is converted to a DVPE by use of a correlation equation (see 13.2).

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum products and Lubricants and is the direct responsibility of Subcommittee D02.08.ODon RVP and V/L Ratio.

Current edition approved May 10, 2001. Published July 2001. Originally published as D 5482 – 93. Last previous edition D 5482 – 99.

² *Annual Book of ASTM Standards*, Vol 05.02.

*A Summary of Changes section appears at the end of this standard.

5. Significance and Use

5.1 Vapor pressure is an important physical property of volatile liquids.

5.2 Vapor pressure is critically important for both automotive and aviation gasolines, affecting starting, warm-up, and tendency to vapor lock with high operating temperatures or high altitudes. Maximum vapor pressure limits for gasoline are legally mandated in some areas as a measure of air pollution control.

6. Apparatus

6.1 *Vapor Pressure Apparatus*—The type of apparatus³ suitable for use in this test method employs a small volume test chamber incorporating a transducer for pressure measurements and associated equipment for thermostatically controlling the chamber temperature.

6.1.1 The test chamber shall be designed to contain between 2 and 50 mL of liquid and vapor and be capable of maintaining a vapor-liquid ratio between 3.95 and 1.00 and 4.05 and 1.00.

6.1.2 The pressure transducer shall have a minimum operational range from 0 to 172 kPa (0 to 25.0 psi) with a minimum resolution of 0.1 kPa (0.01 psi) and a minimum accuracy of ± 0.3 kPa (± 0.05 psi). The pressure measurement system shall include associated electronics and readout devices to display the resulting pressure reading.

6.1.3 A thermostatically controlled heater shall be used to maintain the test chamber at $37.8 \pm 0.1^\circ\text{C}$ ($100 \pm 0.2^\circ\text{F}$) for the duration of the test.

6.1.4 A platinum resistance thermometer shall be used for measuring the temperature of the test chamber. The minimum temperature range of the measuring device shall be from ambient to 75°C (167°F) with a resolution of 0.1°C (0.2°F) and accuracy of 0.1°C (0.2°F).

6.1.5 The vapor pressure apparatus shall have provisions for introduction of the test specimen into the test chamber and for the cleaning or purging of the chamber following the test.

6.2 *Syringe*, if required, gas tight, 1 to 20 mL capacity with a $\pm 1\%$, or better, accuracy and a $\pm 1\%$, or better, precision. The capacity of the syringe shall not exceed two times the volume of the test specimen being dispensed, and shall be chosen so as to provide maximum accuracy and resolution for the volume to be injected.

6.3 *Iced-Water Bath or Air Bath*, for chilling the samples and syringe to temperatures between 0 and 1°C (32 and 34°F).

6.4 *Pressure Measuring Device*, capable of measuring ambient and above ambient pressures with an accuracy of 0.20 kPa (0.03 psi) or better at the same elevation relative to sea level as the apparatus in the laboratory.

6.4.1 When a mercury manometer is not used as the pressure measuring device, the calibration of the pressure measuring device employed shall be periodically checked

(with traceability to a nationally recognized standard) to ensure that the device remains within the required accuracy specified in 6.4.

6.5 *Pressure Source*, clean, dry compressed gas or other suitable compressed air capable of providing pressure for calibration of the transducer and cleaning of the cell.

NOTE 2—A vacuum source is an alternate means for cleaning of the cell.

7. Reagents and Materials

7.1 *Purity of Reagents*—Use chemicals of at least 99 % purity for quality control checks (see Section 11). Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴ Lower purities can be used, provided it is first ascertained that the reagent is of sufficient purity to permit its use without lessening the accuracy of the determination.

7.2 *Cyclohexane*, (**Warning**—Cyclohexane is flammable and a health hazard).

7.3 *Cyclopentane*, (**Warning**—Cyclopentane is flammable and a health hazard).

7.4 *2,2-Dimethylbutane*, (**Warning**—2,2-dimethylbutane is flammable and a health hazard).

7.5 *2,3-Dimethylbutane*, (**Warning**—2,3-dimethylbutane is flammable and a health hazard).

7.6 *2-Methylpentane*, (**Warning**—2-methylpentane is flammable and a health hazard).

7.7 *Toluene*, (**Warning**—Toluene is flammable and a health hazard).

8. Sampling

8.1 General Requirements:

8.1.1 The extreme sensitivity of vapor pressure measurements to losses through evaporation and the resulting changes in composition is such as to require the utmost precaution and the most meticulous care in the handling of samples.

8.1.2 Obtain a sample and test specimen in accordance with 10.3 of Practice D 4057, except do not use 10.3.1.8 of Practice D 4057, Sampling by Water Displacement, for fuels containing oxygenates. Use a 1-L (1-qt) sized container filled between 70 to 80 % with sample.

8.1.3 Perform the vapor pressure determination on the first test specimen withdrawn from a sample container. Do not use the remaining sample in the container for a second vapor pressure determination. If a second determination is necessary, obtain a new sample.

8.1.4 Protect samples from excessive temperatures prior to testing. This can be accomplished by storage in an appropriate ice bath or refrigerator.

8.1.5 Do not test samples stored in leaky containers. Discard and obtain a new sample if leaks are detected.

³ The following instruments have been found satisfactory for use in this test procedure as determined by interlaboratory testing: Herzog Mini Reid Vapor Pressure Model MP970—available from Varlen Instruments, Inc., 2777 Washington Blvd., Bellwood, IL 60104 and ABB Model 4100—available from ABB Process Analytics, Lewisburg, WV.

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.