

SLOVENSKI STANDARD oSIST prEN ISO 13161:2019

01-marec-2019

Kakovost vode - Merjenje koncentracije aktivnosti polonija Po-210 v vodi z alfa spektrometrijo (ISO/DIS 13161:2019)

Water quality - Measurement of polonium 210 activity concentration in water by alpha spectrometry (ISO/DIS 13161:2019)

Wasserbeschaffenheit - Bestimmung der Aktivitätskonzentration von Polonium-210 in Wasser mittels Alphaspektrometrie (ISO/DIS 13161:2019)

Qualité de l'eau - Mesurage de l'activité du polonium 210 dans l'eau par spectrométrie alpha (ISO/DIS 13161:2019)

Ta slovenski standard je istoveten z: prEN ISO 13161 rev

ICS:

13.060.60	Preiskava fizikalnih lastnosti vode	Examination of physical properties of water
17.240	Merjenje sevanja	Radiation measurements

oSIST prEN ISO 13161:2019

en,fr,de

oSIST prEN ISO 13161:2019

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN ISO 13161:2021</u> https://standards.iteh.ai/catalog/standards/sist/fefe217f-8099-4ff5-aa12-4699b449cd38/sisten-iso-13161-2021

DRAFT INTERNATIONAL STANDARD ISO/DIS 13161

ISO/TC 147/SC 3

Voting begins on: **2019-02-06**

Secretariat: AFNOR

Voting terminates on: 2019-05-01

Water quality — Polonium 210 — Test method using alpha spectrometry

Qualité de l'eau — Polonium 210 — Méthode par spectrométrie alpha

ICS: 13.060.60; 17.240

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN ISO 13161:2021</u>

https://standards.iteh.ai/catalog/standards/sist/fefe217f-8099-4ff5-aa12-4699b449cd38/sisten-iso-13161-2021

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION. This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number ISO/DIS 13161:2019(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 13161:2021

https://standards.iteh.ai/catalog/standards/sist/fefe217f-8099-4ff5-aa12-4699b449cd38/sisten-iso-13161-2021



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Fax: +41 22 749 09 47 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword		iv
Introd	Introduction	
1	Scope	
2	Normative references	
3	Terms, definitions, symbols and abbreviated terms3.1Terms and definitions3.2Symbols and abbreviated terms	2 2 2
4	Principle 4.1 General 4.2 Treatment 4.2.1 Treatment for a deposition on a disc. 4.2.2 Treatment for a precipitation on a filter. 4.3 Principle of alpha spectrometry	3 3 3 3 4 4
5	Reagents and equipment 5.1 Reagents 5.2 Preparation material and treatment 5.3 Alpha spectrometry measuring equipment	4 4 5 5
6	Sampling and samples	5
7	Chemical treatment and deposit process 7.1 General 7.2 Chemical treatment 7.2.1 Autodeposition of polonium on a disc. 7.2.2 Microprecipitation on a filter	6
8 _{https}	Measurement by alpha spectrometry 12021 8.1 General 8.2 Quality control 8.3 Measurement	8 8 8 8
9	Expression of results9.1General9.2Total yield9.3Activity concentration of 210Po in the sample9.4Combined uncertainties9.5Decision threshold9.6Detection limit9.7Confidence limits	
10	Test report	
Annex	Annex A (informative) Cell deposit examples	
Annex	Annex B (informative) Spectrum examples	
Biblio	Bibliography	

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radiological methods*.

<u>SIST EN ISO 13161:2021</u>

This second edition cancels and replaces the first edition (ISO 13161:2011), which has been technically revised. The main changes compared to the previous edition are as follows:

Introduction

Radioactivity from several naturally-occurring and anthropogenic sources is present throughout the environment. Thus, water bodies (e.g. surface waters, ground waters, sea waters) can contain radionuclides of natural, human-made, or both origins:.

- Natural radionuclides, including 40K, 3H, 14C, and those originating from the thorium and uranium decay series, in particular 226Ra, 228Ra, 234U, 238U, and 210Pb, can be found in water for natural reasons (e.g. desorption from the soil and washoff by rain water) or can be released from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizer production and use);
- Human-made radionuclides, such as transuranium elements (americium, plutonium, neptunium, curium), 3H, 14C, 90Sr, and gamma emitting radionuclides can also be found in natural waters. Small quantities of these radionuclides are discharged from nuclear fuel cycle facilities into the environment as the result of authorized routine releases. Some of these radionuclides used for medical and industrial applications are also released into the environment after use. Anthropogenic radionuclides are also found in waters as a result of past fallout contaminations resulting from the explosion in the atmosphere of nuclear devices and accidents such as those that occurred in Chernobyl and Fukushima.

Radionuclide activity concentration in water bodies can vary according to local geological characteristics and climatic conditions and can be locally and temporally enhanced by releases from nuclear installation during planned, existing and emergency exposure situations.^[1] Drinking-water may thus contain radionuclides at activity concentrations which could present a risk to human health.

The radionuclides present in liquid effluents are usually controlled before being discharged into the environment^[2] and water bodies. Water Drinking water are monitored for their radioactivity content as recommended by the World Health Organization (WHO)^[3]so that proper actions can be taken to ensure that there is no adverse health effect to the public. Following these international recommendations, national regulation usually specify radionuclide authorized concentration limits for liquid effluent discharged to the environment and radionuclide guidance levels for water bodies and drinking waters for planned, existing and emergency exposure situations. Compliance with these limits can be assessed using measurement results with their associated uncertainties as specified by ISO/IEC Guide 98-3 and ISO 5667-20.^[4]

Depending on the exposure situation, there are different limits and guidance levels that would result in an action to reduce health risk. As an example, during a planned or existing situation, the WHO guidance level in drinking water is 0,1 Bq l-1 for polonium 210 activity concentration.

NOTE 1 The guidance level is the activity concentration with an intake of 2 l/d of drinking water for one year that results in an effective dose of 0,1 mSv/a for members of the public. This is an effective dose that represents a very low level of risk and which is not expected to give rise to any detectable adverse health effects.^[3]

In the event of a nuclear emergency, the WHO Codex Guideline Levels^[5] mentioned that the activity concentration might be greater.

NOTE 2 The Codex guidelines levels (GLs) apply to radionuclides contained in foods destined for human consumption and traded internationally, which have been contaminated following a nuclear or radiological emergency. These GLs apply to food after reconstitution or as prepared for consumption, i.e., not to dried or concentrated foods, and are based on an intervention exemption level of 1 mSv in a year for members of the public (infant and adult).^[5]

Thus, the test method can be adapted so that the characteristic limits, decision threshold, detection limit and uncertainties ensure that the radionuclide activity concentrations test results can be verified to be below the guidance levels required by a national authority for either planned/existing situations or for an emergency situation.^[6][\mathbb{Z}]

Usually, the test methods can be adjusted to measure the activity concentration of the radionuclide(s) in either wastewaters before storage or in liquid effluents before being discharged to the environment.

The test results will enable the plant/installation operator to verify that, before their discharge, wastewaters/liquid effluent radioactive activity concentrations do not exceed authorized limits.

The test method(s) described in this document may be used during planned, existing and emergency exposure situations as well as for wastewaters and liquid effluents with specific modifications that could increase the overall uncertainty, detection limit, and threshold.

The test method(s) may be used for water samples after proper sampling, sample handling, and test sample preparation (see the relevant part of the ISO 5667- series).

An International Standard on a test method of polonium 210 activity concentrations in water samples is justified for test laboratories carrying out these measurements, required sometimes by national authorities, as laboratories may have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

This document is one of a set of International Standards on test methods dealing with the measurement of the activity concentration of radionuclides in water samples.

Polonium 210 is a natural alpha-emitting radionuclide with a half-life of 138 d. It appears in the natural chain of uranium 238 (see Figure 1). It is a long-life decay product of radon 222 (in the frame of Figure 1) through lead 210 (see References^[8] to^[12]).



NOTE ²⁰⁶Pb is stable.

Figure 1 — Uranium 238 and its decay products

DRAFT INTERNATIONAL STANDARD

Water quality — Polonium 210 — Test method using alpha spectrometry

WARNING — Persons using this document should be familiar with normal laboratory practices. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices.

IMPORTANT — It is essential that tests conducted according to this test method be carried out by suitably trained staff.

1 Scope

This document specifies a method for the measurement of 210 Po in all types of waters by alpha spectrometry.

The method is applicable to test samples of supply/drinking water, rainwater, surface and ground water, marine water, as well as cooling water, industrial water, domestic, and industrial wastewater after proper sampling and handling, and test sample preparation. A filtration of the test sample is necessary.

The detection limit depends on the sample volume, the instrument used, the background count rate, the detection efficiency and the chemical yield. The method described in this document, using currently available alpha spectrometry apparatus, has a detection limit of approximately 5 mBq.l⁻¹, which is lower than the WHO criteria for safe consumption of drinking water (100 mBq·l⁻¹). This value can be achieved with a counting time of 24 hours for a sample volume of 500 ml.

The method described in this document is applicable in the event of an emergency situation.

The analysis of ²¹⁰Po adsorbed to suspended matter is not covered by this method. 9cd38/sist-

It is the user's responsibility to ensure the validity of this test method for the water samples tested.

If suspended material has to be removed or analysed, filtration at 0,45 μ m is recommended. The analysis of the insoluble fraction requires a mineralization step that is not covered by this document (see NF M 60-790-4^[13]). In this case, the measurement is made on the different phases obtained. The final activity is the sum of all the measured activity concentrations.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, Water for analytical laboratory use — Specification and test methods

ISO 5667-1, Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques

ISO 5667-3, Water quality — Sampling — Part 3: Preservation and handling of water samples

ISO 5667-10, Water quality — Sampling — Part 10: Guidance on sampling of waste waters

ISO 11929, Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application

ISO 80000-10, Quantities and units — Part 10: Atomic and nuclear physics

ISO/IEC 17025, General requirements for the competence of testing and calibration laboratories

ISO/IEC Guide 98-3, Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)

3 Terms, definitions, symbols and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 80000-10 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at http://www.electropedia.org/

ISO Online browsing platform: available at http://www.iso.org/obp

3.1.1

certified reference solution

solution of known concentration traceable to primary or secondary certified reference solution standards of radioactivity

3.1.2

tracer solution

usually a secondary standard or reference material, such as ²⁰⁸Po or ²⁰⁹Po, employed to determine the chemical yield of the analysis

3.1.3

quality control standard

radioactive source used to demonstrate that the measurement equipment employed performs within defined limits

Note 1 to entry: Quality control is usually carried out by the regular measurement of a suitable radioactive source in accordance with ISO 7870-1[14], ISO 7870-2,[15] and ISO 7870-4.[16]

3.2 Symbols and abbreviated terms

For the purposes of this document, the symbols and abbreviated terms defined in ISO 80000-10 and the following apply.

Α	activity of the tracer added	Bq
CA	activity concentration of ²¹⁰ Po	Bq l-1
<i>c</i> _A *	decision threshold	Bq l-1
<i>c</i> [#] _A	detection limit	Bq l-1
$C_{\rm A}^{\triangleleft}, C_{\rm A}^{\triangleright}$	lower and upper limits of the confidence interval	Bq l ⁻¹
R _c	chemical yield	1
R _T	total yield	1
r_0	background count rate in the ²¹⁰ Po region of interest	s-1

r _{0T}	background count rate in the tracer region of interest	s-1
rg	gross count rate of the sample in the ²¹⁰ Po region of interest	s ⁻¹
r_{T}	gross count rate in the tracer region of interest	s-1
t_0	background counting time	S
tg	sample counting time	S
U	expanded uncertainty calculated by $U = k \cdot \mathbb{Z}u(c_A)$ with $k = 1, 2 \dots$	Bq l-1
$u(c_A)$	standard uncertainty associated with the initial measurement result	Bq l-1
V	volume of the test sample aliquot	1
ε	counting efficiency	1

4 Principle

4.1 General

There are different techniques to measure ²¹⁰Po activity concentration in water: alpha spectrometry, liquid scintillation counting, and alpha proportional counting. This document describes the alpha spectrometry technique.

After sampling, the test sample undergoes a treatment which leads to an extremely thin deposit of the polonium on a metal disc or on a filter, for measurement by alpha spectrometry.

²¹⁰Po has a half-life of 138,376 d \pm 0,002 d (see Reference^[17]).

The sample shall be analysed as soon as possible to evaluate an activity concentration on the sampling date. If the time elapsed between sampling and measurement is long, the activity concentration measured requires correction and it is then necessary to know the 210 Pb and 210 Bi activity concentrations of the sample to adjust the 210 Po activity concentration on the sampling date.

4.2 Treatment

4.2.1 Treatment for a deposition on a disc

The main steps of the sample treatment are:

- filtration if necessary;
- acidification with concentrated hydrochloric acid (or nitric acid) and addition of a polonium tracer (²⁰⁸Po or ²⁰⁹Po) solution;

NOTE The polonium isotopes 208 Po (5,11 MeV alpha emission) or 209 Po (4,88 MeV alpha emission) can be used as tracers since interference with 210 Po (5,31 MeV alpha emission) is minimal for sources displaying good resolution (< 50 keV FWHM); 209 Po is preferred, but 208 Po is acceptable.

- addition of a reducing agent (e.g. ascorbic acid);
- spontaneous deposition of a thin layer on to a metal disc.

The activity concentration measurement as well as the determination of the total yield is carried out by alpha spectrometry.