
**Cardiovascular implants and
extracorporeal systems — Cardiac
valve repair devices**

*Implants cardiovasculaires et circuits extra-corporels — Dispositifs de
réparation de valves cardiaques*

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 5910:2018](https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018)

<https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 5910:2018](https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018)

<https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

| | Page |
|--|-----------|
| Foreword..... | v |
| Introduction..... | vi |
| 1 Scope..... | 1 |
| 2 Normative references..... | 1 |
| 3 Terms and definitions..... | 2 |
| 4 Abbreviations..... | 11 |
| 5 Fundamental requirements..... | 12 |
| 5.1 General..... | 12 |
| 5.2 Risk management..... | 12 |
| 6 Device description..... | 13 |
| 6.1 Intended use/Indication for use..... | 13 |
| 6.2 Design inputs..... | 13 |
| 6.2.1 General..... | 13 |
| 6.2.2 Operational specifications..... | 13 |
| 6.2.3 Functional, performance and safety requirements..... | 14 |
| 6.2.4 Usability..... | 15 |
| 6.2.5 Packaging, labelling, and sterilization..... | 16 |
| 6.3 Design outputs..... | 16 |
| 6.4 Design transfer (manufacturing verification/validation)..... | 16 |
| 7 Design verification testing and analysis / Design validation..... | 16 |
| 7.1 General requirements..... | 16 |
| 7.2 <i>In vitro</i> assessment..... | 16 |
| 7.2.1 General..... | 16 |
| 7.2.2 Test articles, sample selection, test conditions and reporting requirements..... | 17 |
| 7.2.3 Device material property assessment..... | 18 |
| 7.2.4 Functional performance assessment..... | 19 |
| 7.2.5 Device structural performance assessment..... | 19 |
| 7.2.6 Device corrosion assessment..... | 20 |
| 7.2.7 Delivery system design evaluation requirements..... | 20 |
| 7.2.8 Visibility..... | 20 |
| 7.2.9 Simulated use assessment..... | 20 |
| 7.2.10 Human factors / Usability assessment..... | 21 |
| 7.2.11 Device MRI safety..... | 21 |
| 7.2.12 Design specific testing..... | 21 |
| 7.3 Preclinical <i>in vivo</i> evaluation..... | 21 |
| 7.3.1 General..... | 21 |
| 7.3.2 Overall requirements..... | 21 |
| 7.3.3 Methods..... | 23 |
| 7.3.4 Test report..... | 24 |
| 7.4 Clinical Investigations..... | 24 |
| 7.4.1 General..... | 24 |
| 7.4.2 Study considerations..... | 25 |
| 7.4.3 Study endpoints..... | 28 |
| 7.4.4 Ethical considerations..... | 28 |
| 7.4.5 Distribution of subjects and investigators..... | 28 |
| 7.4.6 Statistical considerations including sample size and duration..... | 29 |
| 7.4.7 Patient selection criteria..... | 31 |
| 7.4.8 Clinical data requirements..... | 31 |
| 7.4.9 Clinical investigation analysis and reporting..... | 34 |
| 7.4.10 Post-market clinical follow-up..... | 34 |
| Annex A (informative) Rationale for the provisions of this document..... | 36 |

| | |
|---|------------|
| Annex B (informative) Examples of heart valve repair devices and delivery systems | 39 |
| Annex C (normative) Packaging | 45 |
| Annex D (normative) Product labels, instructions for use, and training | 46 |
| Annex E (normative) Sterilization | 49 |
| Annex F (informative) Heart valve repair system characteristics | 50 |
| Annex G (informative) Heart valve repair system hazards, associated failure modes and evaluation methods | 52 |
| Annex H (informative) <i>In vitro</i> test guidelines for paediatric devices | 63 |
| Annex I (informative) Examples and definitions of some physical and material properties of heart valve repair device components | 65 |
| Annex J (informative) Examples of standards applicable to testing of materials and components of heart valve repair devices | 79 |
| Annex K (informative) Considerations for device material properties undergoing alterations post implantation | 85 |
| Annex L (informative) Corrosion assessment | 86 |
| Annex M (informative) Guidelines for <i>in vitro</i> evaluation of functional performance of the repair | 89 |
| Annex N (informative) Durability testing | 95 |
| Annex O (informative) Fatigue assessment | 97 |
| Annex P (informative) Preclinical <i>in vivo</i> evaluation | 103 |
| Annex Q (normative) Adverse event classification during clinical investigation | 106 |
| Annex R (informative) Imaging protocol | 112 |
| Annex S (informative) Clinical investigation endpoints for valve repair devices: Suggestions for endpoints and their timing | 116 |
| Annex T (informative) Additional device design evaluation requirements | 120 |
| Annex U (informative) Guidelines for delivery system design evaluation | 122 |
| Bibliography | 124 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

<https://standards.iteh.ai>
ISO 5910:2018

<https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018>

Introduction

No heart valve repair device is ideal. Therefore, a group of engineers, scientists, and clinicians, experts well aware of the problems associated with heart valve repair devices and their development, has prepared this document. This document specifies types of tests, test methods, and requirements for test apparatus. It requires documentation of test methods and results. This document deals with those areas that will ensure adequate mitigation of device-associated risks for patients and other users of the device, facilitate quality assurance, aid the heart team in choosing a heart valve repair device, and ensure that the device will be provided in a convenient and usable form. This document emphasizes the need to specify and report types of *in vitro* testing, preclinical *in vivo* and clinical evaluations. It describes the labels and packaging of the device. Such a process involving *in vitro*, preclinical *in vivo* and clinical evaluations is intended to clarify the requirements prior to market release and to enable prompt identification and management of any subsequent problems.

With regard to *in vitro* testing and reporting, apart from basic material testing for mechanical, physical, chemical and biocompatibility characteristics, this document also covers important functional and durability characteristics of heart valve repair devices and their accessories. This document does not specify exact test methods for functional and durability testing but it offers guidelines for the test apparatus.

This document should be revised, updated, and amended as knowledge and techniques in heart valve repair device technology improve.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO 5910:2018](https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018)

<https://standards.iteh.ai/catalog/standards/iso/62b28e0f-cc05-4e4e-8691-fde6d8f03861/iso-5910-2018>

Cardiovascular implants and extracorporeal systems — Cardiac valve repair devices

1 Scope

1.1 This document applies to all heart valve repair systems that have an intended use to repair and/or improve the function of native human heart valves by acting either on the valve apparatus or on the adjacent anatomy (e.g. ventricle, coronary sinus).

1.2 This document outlines an approach for verifying/validating the design and manufacture of a heart valve repair system through risk management. The selection of appropriate verification/validation tests and methods are derived from the risk assessment. The tests include assessments of the physical, chemical, biological, and mechanical properties of components and materials of heart valve repair systems. The tests also include preclinical *in vivo* evaluation and clinical investigation of the finished heart valve repair system to assess the safety and effectiveness of the heart valve repair system.

NOTE For the purposes of this document, effectiveness endpoint includes clinical performance and benefits.

1.3 This document defines operational conditions and performance requirements for heart valve repair systems where adequate scientific and/or clinical evidence exists for their justification.

1.4 This document excludes Cardiac Resynchronization Therapy (CRT) devices, paravalvular leakage closure devices, systems that do not leave an implant in place (e.g. ablation, radio frequency annuloplasty), apical conduits and devices with components containing viable cells. This Standard also excludes materials not intended for repairing and/or improving the function of human heart valves according to its intended use (e.g. patch material and sutures used in general surgical practice).

NOTE A rationale for the provisions of this document is given in [Annex A](#).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-2, *Biological evaluation of medical devices — Part 2: Animal welfare requirements*

ISO 11135, *Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-2, *Sterilization of health care products — Radiation — Part 2: Establishing the sterilization dose*

ISO 11137-3, *Sterilization of health care products — Radiation — Part 3: Guidance on dosimetric aspects of development, validation and routine control*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 5910:2018(E)

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 13485, *Medical devices — Quality management systems — Requirements for regulatory purposes*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14160, *Sterilization of health care products — Liquid chemical sterilizing agents for single-use medical devices utilizing animal tissues and their derivatives — Requirements for characterization, development, validation and routine control of a sterilization process for medical devices*

ISO 14630, *Non-active surgical implants — General requirements*

ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 15223-2, *Medical devices — Symbols to be used with medical device labels, labelling, and information to be supplied — Part 2: Symbol development, selection and validation*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 17664, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices*

ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO/TS 17665-2, *Sterilization of health care products — Moist heat — Part 2: Guidance on the application of ISO 17665-1*

ISO/TS 17665-3, *Sterilization of health care products — Moist heat — Part 3: Guidance on the designation of a medical device to a product family and processing category for steam sterilization*

ISO 22442-1, *Medical devices utilizing animal tissues and their derivatives — Part 1: Application of risk management*

ISO 22442-2, *Medical devices utilizing animal tissues and their derivatives — Part 2: Controls on sourcing, collection and handling*

ISO 22442-3, *Medical devices utilizing animal tissues and their derivatives — Part 3: Validation of the elimination and/or inactivation of viruses and transmissible spongiform encephalopathy (TSE) agents*

ISO/TR 22442-4, *Medical devices utilizing animal tissues and their derivatives — Part 4: Principles for elimination and/or inactivation of transmissible spongiform encephalopathy (TSE) agents and validation assays for those processes*

IEC 62366-1, *Medical devices — Part 1: Application of usability engineering to medical devices*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

NOTE Additional definitions can be found in the informative annexes.

3.1

abnormal use

act or omission of an act by the operator or user of a medical device as a result of conduct which is beyond any means of risk control by the manufacturer

3.2

accessory

device-specific tool that is required to assist in the implantation and/or adjustment of the heart valve repair device, excluding the delivery system

3.3

active comparator

active control

intervention generally accepted or demonstrated to be safe and effective for the condition of interest that can be used as a basis of comparison of the safety and effectiveness of the heart valve repair device

Note 1 to entry: The active comparator is generally the standard of care for the condition.

3.4

actuarial analysis

statistical technique for calculating event rates over time

Note 1 to entry: Standard actuarial methods calculate the probability of freedom from events within pre-specified intervals of time. When the intervals approach zero width, the methods are called Kaplan-Meier methods.

3.5

adverse event

AE

untoward medical occurrence, unintended disease or injury, or untoward clinical signs (including abnormal laboratory findings) in subjects, users or other persons, whether or not related to the heart valve repair device implantation, adjustment or procedure

3.6

auxiliary device

device used during the procedure, not including accessories (e.g. sheath, guidewire) and delivery system

3.7

back pressure

differential pressure applied across the valve during the closed phase

3.8

body surface area

BSA

total surface area (m²) of the human body

Note 1 to entry: This can be calculated as the square root of product of the weight in kg times the height in cm divided by 3 600. See Reference [30].

3.9

cardiac index

cardiac output (3.10) (CO, l/min) divided by the *body surface area* (3.8) (BSA, m²), with units l/min/m²

3.10

cardiac output


CO

stroke volume multiplied by heart rate

3.11
closing volume

portion of the regurgitant volume that is associated with the dynamics of the valve closure during a single cycle

Note 1 to entry: See [Figure 1](#).



Key

- X time
- Y flowrate
- 1 forward flow volume
- 2 closing volume
- 3 leakage volume

Figure 1 — Schematic representation of flow waveform and regurgitant volumes for one cycle

3.12
coating

thin-film material that is applied to an element of a heart valve repair device to modify its properties

3.13
compliance

relationship between change in radius and change in pressure of a deformable tubular structure (e.g. valve annulus, aorta, conduit), defined in this document as:

$$C = 100 \% \times \frac{(r_2 - r_1) \times 100}{r_1 \times (p_2 - p_1)}$$

where

C is the compliance in units of % radial change/100 mmHg;

p_1 is the diastolic pressure, in mmHg;

p_2 is the systolic pressure, in mmHg;

r_1 is the inner radius at p_1 , in millimetres;

r_2 is the inner radius at p_2 , in millimetres.

Note 1 to entry: See Reference [Z].

3.14

component-joining material

material, such as a suture, adhesive or welding compound, used to assemble the components of a heart valve repair device, thereby becoming part of the implanted device

3.15 cycle

one complete sequence in the action of a heart valve under pulsatile flow conditions

3.16 cycle rate

number of complete cycles per unit of time, usually expressed as cycles per minute (cycles/min)

3.17

delivery system

system used to deliver, deploy, attach or adjust the device in the implant site

3.18

design validation

establishment by objective evidence that device specifications conform with user needs and *intended use(s)* (3.34)

3.19

design verification

establishment by objective evidence that the design output meets the design input requirements

3.20

device embolisation

dislodgement from the intended and documented original position to an unintended and nontherapeutic location

3.21

device failure

inability of a device to perform its intended function sufficient to cause a hazard

3.22

device migration

unintended movement or displacement of the device from its original position within the implant site, without embolisation

3.23
effective orifice area
EOA

orifice area that has been derived from flow and pressure or velocity data

Note 1 to entry: For *in vitro* testing, EOA is defined as:

$$EOA = \frac{q_{v_{RMS}}}{51,6 \times \sqrt{\frac{\Delta p}{\rho}}}$$

where

EOA is the effective orifice area (cm²);

$q_{v_{RMS}}$ is the root mean square forward flow (ml/s) during the positive differential pressure period;

Δp is the mean pressure difference (measured during the positive differential pressure period) (mmHg);

ρ is the density of the test fluid (g/cm³).

3.24
failure mode

mechanism of *device failure* ([3.21](#))

3.25
follow-up

continued assessment of subjects who have received the heart valve repair device

3.26
forward flow volume

volume of flow ejected through the heart valve in the forward direction during one cycle, not including any regurgitant flow through the valve

Note 1 to entry: See [Figure 1](#).

3.27
fracture

complete separation of any part of the *heart valve repair device* ([3.28](#)) that was previously intact

3.28
heart valve repair device

implant ([3.31](#)) intended to improve the function of native human heart valves by acting either on the valve apparatus or on the adjacent anatomy (e.g. ventricle, coronary sinus)

Note 1 to entry: See examples in [Annex B](#).

3.29
heart valve repair system

heart valve repair device, delivery system, other accessories as applicable, packaging, labelling, and instructions

3.30
imaging modality

method used to visualize and assess native anatomy and/or device position, geometry and/or function

3.31
implant

device placed surgically or non-surgically into the human body and intended to remain in place after the procedure

3.32**implant site**

location of heart valve repair device implantation or deployment

3.33**indication for use**

clinical condition of the patient population that the heart valve repair device is intended to treat or improve

3.34**intended use**

purpose of a heart valve repair device, in accordance with the specifications, instructions, and information provided by the manufacturer

3.35**Kaplan-Meier methods**

statistical approaches to calculating event rates over time when the actual dates of events for each person in the population are taken into account

3.36**leakage volume**

component of the regurgitant volume that is associated with leakage during closed phase of a valve in a single cycle

Note 1 to entry: See [Figure 1](#). The point of separation between the closing and leakage volumes is obtained according to a defined and stated criterion (the linear extrapolation shown in [Figure 1](#) is just an example).

3.37**linearized rate**

total number of events divided by the total time under evaluation

Note 1 to entry: Generally, the rate is expressed in terms of percent per patient year.

3.38**mean arterial pressure**

time-averaged arithmetic mean value of the arterial pressure during one cycle

3.39**mean pressure difference**

time-averaged arithmetic mean value of the pressure difference across a heart valve during the positive differential pressure period of the cycle

Note 1 to entry: See [Figure 2](#) for representative aortic and mitral flow and pressure waveforms. See [Figure 3](#) for representative pulmonary and tricuspid flow and pressure waveforms.

3.40**non-structural dysfunction**

abnormality extrinsic to the heart valve repair device that results in abnormal function of the device or causes clinical symptoms

3.41**pannus**

ingrowth of tissue onto the heart valve repair device which may interfere with normal functioning

3.42**pull-out**

situation in which the suture or anchoring device remains structurally intact but tears through the tissue in which it is implanted

3.43

reference device

heart valve substitute or heart valve repair device with known clinical history used for comparative preclinical and clinical evaluations

3.44

regurgitant fraction

regurgitant volume expressed as a percentage of the total ventricular stroke volume

3.45

regurgitant volume

volume of fluid that flows through a heart valve in the reverse direction during one cycle and is the sum of the closing volume and *leakage volume* (3.36)

Note 1 to entry: See [Figure 1](#).

3.46

repositioning

intentional change of implant position of a partially or fully deployed heart valve repair device

3.47

retrieval

removal of a partially or fully deployed heart valve repair device

3.48

risk

combination of the probability of occurrence of harm and the *severity* (3.53) of that harm

Note 1 to entry: See ISO 14971.

3.49

risk analysis

systematic use of available information to identify hazards and to estimate the associated *risks* (3.48)

Note 1 to entry: See ISO 14971.

3.50

risk assessment

overall process comprising a *risk analysis* (3.49) and a risk evaluation

Note 1 to entry: See ISO 14971.

3.51

root mean square forward flow

RMS forward flow

square root of the integral of the volume flow rate waveform squared during the positive differential pressure interval of the forward flow phase used to calculate EOA

Note 1 to entry: Defining the time interval for flow and pressure measurement as the positive pressure period of the forward flow interval for EOA computation provides repeatable and consistent results for comparison to the minimum device performance requirements.

Note 2 to entry: This is calculated using the following equation:

$$q_{vRMS} = \sqrt{\frac{\int_{t_1}^{t_2} q_v(t)^2 dt}{t_2 - t_1}}$$