

DRAFT INTERNATIONAL STANDARD

ISO/DIS 21033

ISO/TC 34/SC 11

Secretariat: BSI

Voting begins on:
2015-02-03

Voting terminates on:
2015-05-03

Animal and vegetable fats and oils — Determination of trace elements by inductively coupled plasma optical emission spectroscopy (ICP-OES)

Corps gras d'origines animale et végétale — Détermination des éléments traces dans les corps gras par spectrométrie d'émission optique à plasma induit par haute fréquence (ICP-OES)

ICS: 67.200.10

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/69080757-970a-4bfb-a7d1-a37c99199587/iso-21033-2016>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.



Reference number
ISO/DIS 21033:2014(E)

© ISO 2014

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/69080757-970a-4bfb-a7d1-a37c99199587/iso-21033-2016>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	1
6 Apparatus	2
7 Sampling	2
8 Preparation of test sample	2
9 Procedure	2
9.1 General	2
9.2 Preparation of standards	4
9.2.1 Blank	4
9.2.2 Standards	4
9.3 Standardization	4
9.4 Preparation of standards	4
10 Calculation	4
11 Precision	5
11.1 Interlaboratory test	5
11.2 Repeatability	5
11.3 Reproducibility	5
12 Test report	5
Annex A (informative) Results of interlaboratory studies	6
Annex B (informative) Results of interlaboratory studies	10
Bibliography	12

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21033 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*.

This second/third/... edition cancels and replaces the first/second/... edition (ISO/TS 21033:2011), [clause(s) / subclause(s) / table(s) / figure(s) / annex(es)] of which [has / have] been technically revised.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sis/690075-970a-4bfb-a7d1-a37c99199587/iso-21033-2014>

Animal and vegetable fats and oils — Determination of trace elements by inductively coupled plasma optical emission spectroscopy (ICP-OES)

1 Scope

This International Standard specifies an inductively coupled plasma optical emission spectroscopic method (ICP-OES) for the determination of the trace element content in oils. Depending on the dilution solvent used, most types of vegetable oils can be analysed (crude, degummed, refined, bleached, deodorized and hardened oils) and nearly all types of lecithins and phosphatides.

This procedure is only suitable when the elements are present in a solubilized form. Fine particles, such as bleaching earth, metal particles and rust, can result in poor recovery of the trace elements present as nebulization and atomization problems affect the ICP-OES analysis.

NOTE The only suitable non-ashing direct method for samples containing fine particles is graphite furnace atomic absorption spectrometry.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 661, *Animal and vegetable fats and oils — Preparation of test sample*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

trace element

element present in very low concentrations

Note 1 to entry: Trace element content is expressed as a mass fraction in milligrams per kilogram.

4 Principle

Solvent-diluted vegetable oils are analysed for trace elements by direct aspiration. Liquid samples are nebulized and carried into the excitation source by a flowing gas. Atoms are quantified by measuring the specific emission lines produced by atoms decaying from high energy levels.

5 Reagents

WARNING — Attention is drawn to national regulations that specify the handling of hazardous substances, and users' obligations thereunder. Technical, organizational and personal safety measures shall be followed.

Unless stated otherwise, use reagents specified in ISO 6353-2^[4] and ISO 6353-3^[5] if listed there, if not then use reagents of recognized analytical grade.

5.1 1-Butanol, ISO 6353-3^[5]

5.2 **Kerosene.**

5.3 **Xylene**, ISO 6353-3.^[5]

5.4 **Standard elements**, present in solution as an organic soluble material¹⁾. Multi-element standards may be used.

5.5 **Base oil [Base 20 Oil or Base 75 Oil from Accu-Standard^{1), may be used to check the blank oil used and for the dilution of the standard solutions as needed.}**

5.6 **Blank oil**, typically, refined and bleached soya bean or other oil, which has been shown to be free of trace elements.

5.7 **Mixture of the blank oil (5.6) with 1-butanol (5.1)**, mass fraction $w_{(\text{blank oil})} = 50 \text{ g}/100 \text{ g}$.

6 Apparatus

Usual laboratory equipment, and, in particular, the following.

6.1 **Inductively coupled plasma optical emission spectrometer.**

6.2 **Analytical balance**, readability 0,000 1 g, weighing precision 0,001 g

6.3 **Oven**, capable of maintaining a temperature of $(60 \pm 2) \text{ }^\circ\text{C}$.

6.4 **Tilt table mixer.**

6.5 **Volumetric flasks**, capacity 100 ml.

7 Sampling

Sampling is not part of the method in this Technical Specification. A recommended sampling method is given in ISO 5555.^[1]

It is important that the laboratory receive a truly representative sample which has not been damaged or changed during transport or storage.

8 Preparation of test sample

Prepare the test samples in accordance with ISO 661, except that the samples should not be clarified.

9 Procedure

9.1 General

9.1.1 Melt solid samples at approximately 10 °C above their melting point and homogenize prior to dilution. Keep the diluted sample warm and monitor throughout the analysis to ensure it remains in solution. The maximum temperature for the analysis of hardened fats is 60 °C.

1) SPEX and Accu-Standard are examples of suitable products available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of these products.

9.1.2 All samples, standards, and blanks are diluted (equal volume fractions) with 1-butanol (kerosene or xylene) to reduce the viscosity of the oil for better nebulization. Some samples are more soluble in 1-butanol than others.

1-Butanol is preferred because it has better moisture tolerance and allows a higher flow rate with higher pressure than kerosene without extinguishing the torch. The increased moisture tolerance permits the analysis of crude oils and lecithins without phase separation. The higher flow rate provides for improved detection limits.

If kerosene or xylene is used, all instrumental operating conditions, e.g. pump flow rate, change from those set for 1-butanol. Therefore, the analysis shall be standardized and all analyses shall be run with all standards, blank, and samples dissolved in the same solvent.

9.1.3 The instrument is ignited and allowed to warm. Profile the instrument according to the manufacturer's instructions. Elements can be detected at the major emission lines (see [Table 1](#)). Additional emission lines and equipment set up instructions are given in EN 14538:2006[6] and ISO 10540-3:2012[7].

Table 1 — Major emission lines and limits of detection

Element	Current limits of detection	Major emission line(s)		
	mg/kg	nm		
Aluminium	— ^a	167,078	308,215	
Barium	— ^a	455,404		
Boron	— ^a	249,773		
Cadmium	— ^a	226,502	214,441	228,802
Calcium	0,05	315,887	393,366	317,933
Chrome	— ^a	267,716	284,325	283,563
Copper	0,05	324,754	327,396	
Iron	0,05	259,940	238,204	261,187
Lead	— ^a	220,353		
Magnesium	0,05	285,213	279,553	280,270
Manganese	— ^a	257,611	259,373	
Molybdenum	— ^a	281,615	202,095	
Nickel	0,05	231,604	221,648	341,476
Phosphorus	0,05	213,618	178,287	177,495
Silicon	0,1	251,611	288,158	
Silver	— ^a	328,068		
Sodium	0,1	588,995		
Sulfur	1	180,731		
Tin	— ^a	242,949		
Titanium	— ^a	334,941	323,452	336,121
Vanadium	— ^a	309,311	311,071	
Zinc	— ^a	202,613	213,856	

^a Not reported at the time of publication

9.1.4 Standardize the instrument as specified in [9.3](#) and scan all the samples in triplicate.

NOTE Calibration drift has been noted. It can result from carbon build-up on the injector tip.

9.2 Preparation of standards

9.2.1 Blank

Typically, refined and bleached soya bean or other oil, which has been shown to be free of trace elements, is used. Blank oil is diluted (1: 1) as described in the sampling procedure. Base 20 or Base 75 oil (5.5) is used as an absolute reference blank to determine that the blank oil is free of trace elements.

9.2.2 Standards

The standard is prepared from commercially available single element organic-based standards. Weigh accurately the standard and add enough blank oil to total 50,00 g. Add 50,00 g of solvent (1-butanol, kerosene or xylene) to achieve a (1 + 1) dilution.

One standard concentration works; however, up to four multi-level, multi-element standards provide a better calibration for linearity and accuracy. Levels should include 2,5 mg/kg, 5 mg/kg, and 10 mg/kg standards depending on the range of values expected.

Instruments that are capable of simultaneous detection of multiple elements can achieve improved precision and accuracy by the inclusion of an internal standard in the analysis. If an internal standard is used, it should be incorporated as part of the dilution step. Typically, the resultant dilution should contain 10 mg/kg yttrium or scandium. Thus, under the dilution sequence for sampling (see 9.1.1 and 9.4), the diluent should contain 20 mg/kg internal standard yielding 10 mg/kg yttrium or scandium in the 1:1 dilution to accomplish this.

9.3 Standardization

Run the blank oil standard and diluted Base 20 or Base 75 oils at the specified wavelength for the element(s) of interest.

Run the standard solutions (9.2.2) at the wavelength(s) chosen.

Blanks, samples, and standards are scanned in triplicate for trace element(s) and are averaged.

Standards and the blank are run every 10 samples or fewer and the instrument is re-standardized as needed. For accuracy, use a narrow range of standardization (0 mg/kg to 25 mg/kg of each element). Although the linearity is somewhat greater, test samples should be diluted to keep the trace element(s) content within the range of standardization.

9.4 Preparation of standards

Weigh $2,5 \pm 0,02$ g of sample into an auto-sampler tube and dilute with 2,5 g of 1-butanol (kerosene or xylene) delivered from an automatic pipette. Cap the tube and invert 40 times to 50 times on a mixing table.

Dilute 0,2 g lecithins (up to 100 % acetone insoluble) to 5,0 g with blank soybean oil (dilution factor 25) and then to 10 g with 1- butanol (5.6), dilution factor 50. Mix the samples on a tilt table mixer for 1 h. Then dilute 1 part with 10 parts of the mixture of the blank oil and 1-butanol (5.7). The total dilution factor is 500.

10 Calculation

Computation is a feature of most instrument programs. Area counts from known standards are inserted into a linear regression formula versus the sample concentration. From this relationship, concentrations may be determined from the area counts of the samples.

Simultaneous instruments can usually accommodate internal standard calculations. If an internal standard is used, the ratio of the counts from the standard divided by the area counts of the internal standard are inserted into the linear regression formula.

It is important to include the correct dilution factor.

11 Precision

11.1 Interlaboratory test

Details of interlaboratory tests on the precision of the method are summarized in [Annex A](#). The values derived from these interlaboratory tests may not be applicable to concentration ranges and matrices other than those given.

11.2 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time will not be greater than the repeatability limit r as shown in [Annex A](#).

11.3 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will not be greater than the reproducibility limit R as shown in [Annex A](#).

12 Test report

The test report shall contain at least the following information:

- a) all information necessary for the complete identification of the sample;
- b) the sampling method used, if known;
- c) the test method used, with reference to this International Standard (ISO 21033:2014);
- d) all operating details not specified in this Technical Specification, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- e) the test result(s) obtained;
- f) if repeatability has been checked, the final quoted result obtained.