



SLOVENSKI STANDARD

oSIST prEN 15787:2020

01-marec-2020

Krma: metode vzorčenja in analize - Izolacija in štetje prisotnih *Lactobacillus* spp

Animal feeding stuffs: Methods of sampling and analysis - Isolation and enumeration of *Lactobacillus* spp.

Futtermittel: Probenahme- und Untersuchungsverfahren - Trennung und Zählung von *Lactobacillus* spp.

Aliments des animaux - Méthodes d'échantillonnage et d'analyse - Isolement et dénombrement des souches de *Lactobacillus* spp.

iTeh STANDARD PREVIEW

(standards.iTeh.ai)

oSIST prEN 15787:2020

Ta slovenski standard je istoveten z: <http://standards.iTeh.ai/catalog/standards/sist/15787-2020/e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020> prEN 15787

ICS:

65.120

Krmila

Animal feeding stuffs

oSIST prEN 15787:2020

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN 15787:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN 15787

January 2020

ICS 65.120

Will supersede EN 15787:2009

English Version

Animal feeding stuffs: Methods of sampling and analysis - Isolation and enumeration of *Lactobacillus* spp.

Aliments des animaux - Méthodes d'échantillonnage et
d'analyse - Isolement et dénombrement des souches de
Lactobacillus spp.

Futtermittel: Probenahme- und
Untersuchungsverfahren - Trennung und Zählung von
Lactobacillus spp.

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 327.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3
Introduction	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions	5
4 Principle	6
5 Diluents and selective media.....	6
6 Apparatus and glassware	9
7 Sampling.....	10
8 Preparation of test sample	10
9 Procedure.....	10
9.1 Samples with a single or multi-component added microflora	10
9.2 Preparation of poured agar plates.....	10
9.3 Preparation of MRS agar for pour plate method	11
9.4 Preparation of the initial suspension and decimal dilutions.....	11
9.5 Inoculation and incubation of plates	12
9.6 Counting of colonies.....	13
9.7 Confirmation.....	13
10 Expression of results.....	14
11 Precision.....	14
11.1 General.....	14
11.2 Interlaboratory study.....	15
11.3 Repeatability.....	15
11.4 Reproducibility	15
11.5 Test report.....	15
Annex A (informative) Notes on procedure.....	16
Annex B (informative) Results of the interlaboratory study.....	17
Bibliography	18

European foreword

This document (prEN 15787:2019) has been prepared by Technical Committee CEN/TC 327 “Animal feedings stuffs: Methods of sampling and analysis”, the secretariat of which is held by NEN.

This document is currently submitted to the CEN Enquiry.

This document will supersede EN 15787:2009.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN 15787:2020](https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020)

<https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020>

prEN 15787:2019 (E)

Introduction

This methodology has been developed to enumerate lactobacilli used as feed additives to enable the European Commission to control proper labelling of animal feeding products (EU project SMT4-CT98-2235 – “Methods for the official control of probiotics (microorganisms) used as feed additives”) [1]. The described methodology was validated in an interlaboratory study [1]. The validation was performed with one strain of *Lactobacillus acidophilus* and one strain of *Lactobacillus rhamnosus*. It can be assumed that the method is suitable also for other lactobacilli strains used as feed additives.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN 15787:2020](https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020)

<https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020>

1 Scope

This document defines general rules for the enumeration of lactobacilli in feedingstuffs (additives, premixtures and compound feeds excluding mineral feeds) that contain lactobacilli as a single microorganism component or in a mixture with other microorganisms. Applying the method to compound feeds with critical amounts of copper demands a special procedure (see Annex A). The document is not applicable to mineral feeds, which are defined as complementary feeding stuffs composed mainly of minerals and containing at least 40 % crude ash (Regulation R767/2009) [3].

There are different categories of feed samples:

- a) Additives containing about 10^{10} colony forming units (CFU)/g;
- b) Premixtures containing about 10^{11} CFU/kg;
- c) Compound feeds, meal or pellets, which contain about 10^9 CFU/kg.

The detection limit is defined in EN ISO 7218.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 6498, *Animal feeding stuffs - Guidelines for sample preparation (ISO 6498)*

EN ISO 7218, *Microbiology of food and animal feeding stuffs - General requirements and guidance for microbiological examinations (ISO 7218)*

EN ISO 6887-1, *Microbiology of the food chain - Preparation of test samples, initial suspension and decimal dilutions for microbiological examination - Part 1: General rules for the preparation of the initial suspension and decimal dilutions (ISO 6887-1)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

lactobacilli

gram-positive, catalase negative, rod-shaped bacteria in chains

Note 1 to entry This description is based on their characteristics as used for this standard

Note 2 to entry Lactobacilli form colonies fitting the description of these species on the specified selective media after incubation of 48 h to 72 h at a temperature of 37 °C under anaerobic conditions [6] (see 9.7)

prEN 15787:2019 (E)

4 Principle

- a) Preparation of sterile and dry poured agar plates or preparation of sterile liquid selective medium tempered at 44 °C to 47 °C;
- b) Drawing a representative test sample under sterile conditions;
- c) Preparation of the initial suspension with a tempered diluent to obtain a homogeneous distribution of bacterial cells from the test portion;
- d) Preparation of further decimal dilutions of the initial suspension in order to reduce the number of microorganisms per unit volume to allow, after incubation, the counting of colonies;
- e) Inoculation of the prepared poured plates with an aliquot of the optimum dilutions and dispersion of the inoculum by using a sterile spreader or inoculation of blank petri dishes with an aliquot of the optimum dilutions and pouring of the molten agar medium into each Petri dish, mixing and solidification;
- f) Incubation of inverted plates at 37 °C ± 1 °C under anaerobic conditions for 48 h to 72 h;
- g) Counting of typical colonies, considering the specific properties of lactobacilli;
- h) Morphological verification of isolates within the *Lactobacillus* genus through the use of microscope analysis if necessary;
- i) Calculation of the colony count per gram or kilogram of feed sample.

5 Diluents and selective media

oSIST prEN 15787:2020
<https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020>

5.1 Diluents

5.1.1 Diluent for initial suspension

This diluent is used for the preparation of the initial suspension and may also be used for the preparation of further decimal dilutions.

Table 1 — Phosphate buffered saline supplemented with Tween[®] 80 (tPBS)

Sodium chloride	NaCl	8,00 g
Potassium chloride	KCl	0,20 g
Disodium hydrogen phosphate anhydrous	Na ₂ HPO ₄	1,15 g
Potassium dihydrogen phosphate anhydrous	KH ₂ PO ₄	0,20 g
Polyoxyethylensorbitanmonooleate (Tween [®] 80)		1 ml
Water, distilled or deionized		1 000 ml

¹ Tween[®] is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by CEN of this product.

Dissolve the components in water. If necessary, adjust to a final pH of $7,3 \pm 0,2$ at $25\text{ }^{\circ}\text{C}$ after sterilization. The solution is filled into appropriate containers (e.g. bottles or flasks, test tubes) and sterilized at $121\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ for 15 min. To avoid loss during autoclaving, screw cap bottles are recommended.

Temper to $(40 \pm 1)^{\circ}\text{C}$ in a water bath or incubator immediately before usage.

NOTE The use of commercially available PBS buffer tablets is acceptable. However, please take note that variations in composition and pH can occur between products from different manufacturers and could therefore give results different from the ones obtained with the medium as specified in this International Standard.

5.1.2 Diluents for serial dilutions

For serial dilutions, the diluent for initial suspension (5.1.1) or alternatively peptone salt solution (PSS) according to EN ISO 6887-1 can be used.

Table 2 — Peptone salt solution PSS according to EN ISO 6887-1

Enzymatic digest casein	1,0g
Sodium chloride (NaCl)	8,5g
Water, distilled or deionized	1 000 ml

Dissolve the components in the water in flasks, bottles or test tubes. Adjust the pH if necessary so that, after sterilization, it is $7,0 \pm 0,2$ at $25\text{ }^{\circ}\text{C}$. For decimal dilutions, prepare test tubes containing $9,0\text{ ml} \pm 0,1\text{ ml}$ after sterilization or use screw cap bottles to avoid weight loss during autoclaving.

Sterilize at $121\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ for 15 min. Bring the diluent to room temperature before use.

NOTE Commercially available, ready-to-use PSS tubes of 9 ml are suitable.

5.2 Enumeration Media

oSIST prEN 15787:2020

5.2.1 General

<https://standards.iteh.ai/catalog/standards/sist/216ae76e-5980-4bdd-87fd-a243d6f8df85/osist-pren-15787-2020>

Three different media are proposed:

- a) MRS medium;
- b) MRS supplemented with Triphenyl Tetrazolium Chloride (TTC);
- c) AMRSA: Acidified MRS agar.

For routine enumeration of lactobacilli the use of MRS agar will be sufficient assuming that the probiotic strain is present in far higher numbers than any other microorganism. The medium is designed to encourage the growth of the 'lactic acid bacteria' such as lactobacilli, enterococci and pediococci. Selection can be made by pH adjustment, as lactobacilli will tolerate a lower pH than enterococci (pH 5,0 to pH 6,5), with pediococci growing best in this range. When enterococci are expected to be present in similar concentrations as lactobacilli, acidified MRS agar (AMRSA) should be used. When lactobacilli in combination with pediococci are expected, MRS agar supplemented with TTC allows differentiation of colonies by different coloration after anaerobic incubation but can exert a negative influence on the colony amount.

NOTE An antifungal agent like nystatin (50 U/ml) can be added to MRS agar, AMRSA or MRS+TTC to inhibit moulds and yeasts.

prEN 15787:2019 (E)

5.2.2 Composition

5.2.2.1 MRS agar

Table 3 — Composition of the MRS agar

Glucose	20 g
Peptone	10 g
Meat extract	8 g
Yeast extract	4 g
Sodium acetate 3 H ₂ O	5 g
Dipotassium hydrogen phosphate	2 g
Triammonium citrate	2 g
Sorbitan mono-oleate (Tween 80)	1 ml
Magnesium sulphate 7 H ₂ O	0,2 g
Manganese sulphate 4 H ₂ O	0,05 g
Agar agar	10-15 g ^a
Water, distilled or deionized	1000 ml
pH 6,2 ± 0,2 at 25 °C	
a Depending on the gel strength of the agar.	

NOTE The use of commercially available, ready-to-use media is acceptable. However, please take note that variations in composition and pH can occur between products from different manufacturers and could therefore give results different from the ones obtained with the medium as specified in this document.

5.2.2.2 MRS agar supplemented with TTC (0,01 %)

Sterilize MRS agar (5.2.2.1) by autoclaving at 121 °C ± 1 °C for 15 min. Supplement with 1 ml of a filter sterilized 1 g/100 ml water solution of Triphenyl Tetrazolium Chloride (TTC) per 100 ml MRS agar.

5.2.2.3 AMRSA

Acidified MRS agar can be obtained by adjusting the pH of MRS agar (see 5.2.2.1) to 5,4 ± 0,1 with HCl prior to autoclaving.

5.2.3 Preparation

5.2.3.1 MRS agar

Dispense the agar medium into suitable containers (bottles or flasks with non-toxic metal screw-caps may be used). Dissolve all components described in 5.2.2.1 in water by boiling. If necessary adjust the pH so that after sterilization it is pH 6,2 ± 0,2 or 5,4 ± 0,1. Sterilize at 121 °C ± 1 °C for 15 min. Excessive heating shall be avoided. MRS agar supplemented with TTC.

Prepare 1 g Triphenyl Tetrazolium Chloride (TTC) in 100 ml water and filter sterilize. Add 1 ml per 100 ml MRS agar medium (5.2.2.1) which is tempered at 44 °C to 47 °C after autoclaving.

NOTE TTC is destroyed by autoclaving. Protect the solution from light, and discard it if a pink tinge develops.

5.2.3.2 AMRSA

Prepare the medium as described in 5.2.3.1. Adjust the pH of MRS agar with HCl to $5,4 \pm 0,1$ prior to autoclaving. Sterilize at $121 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$ for 15 min.

6 Apparatus and glassware

Usual microbiological laboratory equipment and, in particular, the following:

6.1 Equipment for dry sterilization (oven) and wet sterilization (autoclave)

According to EN ISO 7218.

6.2 Incubator

Capable of maintaining a temperature of $30 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$. Optionally also capable of maintaining a temperature of $40 \text{ }^\circ\text{C} \pm 1 \text{ }^\circ\text{C}$ and/or $44 \text{ }^\circ\text{C}$ to $47 \text{ }^\circ\text{C}$.

6.3 Blending equipment

The following apparatus may be used (EN ISO 7218):

- a rotary homogenizer (blender) with a notional variable speed of 3 000 rpm to 10 000 rpm, as well as aseptic glass or metals bowls equipped with covers; or
- a peristaltic blender with sterile bags (paddle homogenizer), possibly with the option to adjust blending speed and time; or
- a vibrational mixer with sterile bags; or
- any other homogenizing system with equivalent efficiency (e.g. a hand blender with aseptic beaker).

6.4 Mechanical stirrer

A mechanical stirrer e.g. Vortex Mixer (see EN ISO 7218), or equivalent.

6.5 Balance

Balances of the required range and accuracy according to EN ISO 7218 for the different products to be weighed.

6.6 Flasks or screw-cap bottles of appropriate capacities

6.7 Test tubes of appropriate capacities

6.8 Pipettes or Pipettor and sterile tips to dispense 0.1 ml to 1 ml

6.9 Sterile 5 ml graduated pipettes

For full outlet with wide (approx. 3 mm) tips (e.g. serological pipette; alternatively: 5 ml-graduated pipettes without tips).

6.10 pH meter

Having an accuracy of calibration of $\pm 0,1$ pH unit at $20 \text{ }^\circ\text{C}$ to $25 \text{ }^\circ\text{C}$.