
Pšenična moka - Fizikalne značilnosti testa - 1. del: Ugotavljanje vpijanja vode in reoloških lastnosti s farinografom (ISO/DIS 5530-1:2020)

Wheat flour - Physical characteristics of doughs - Part 1: Determination of water absorption and rheological properties using a farinograph (ISO/DIS 5530-1:2020)

Weizenmehl - Physikalische Eigenschaften von Teigen - Teil 1: Bestimmung der Wasserabsorption und der rheologischen Eigenschaften mittels Farinograph (ISO/DIS 5530-1:2020)

Farines de blé tendre - Caractéristiques physiques des pâtes - Partie 1: Détermination de l'absorption d'eau et des caractéristiques rhéologiques au moyen du farinographe (ISO/DIS 5530-1:2020)

Ta slovenski standard je istoveten z: prEN ISO 5530-1

ICS:

67.060	Žita, stročnice in proizvodi iz njih	Cereals, pulses and derived products
--------	--------------------------------------	--------------------------------------

oSIST prEN ISO 5530-1:2020

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 5530-1:2020](https://standards.iteh.ai/catalog/standards/sist/e2933f47-7168-42b3-9f52-65d814ec422e/osist-pren-iso-5530-1-2020)

<https://standards.iteh.ai/catalog/standards/sist/e2933f47-7168-42b3-9f52-65d814ec422e/osist-pren-iso-5530-1-2020>

DRAFT INTERNATIONAL STANDARD

ISO/DIS 5530-1

ISO/TC 34/SC 4

Secretariat: SAC

Voting begins on:
2020-05-13Voting terminates on:
2020-08-05

Wheat flour — Physical characteristics of doughs —

Part 1:

Determination of water absorption and rheological properties using a farinograph

*Farines de blé tendre — Caractéristiques physiques des pâtes —**Partie 1: Détermination de l'absorption d'eau et des caractéristiques rhéologiques au moyen du farinographe*

ICS: 67.060

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 5530-1:2020](https://standards.iteh.ai/catalog/standards/sist/e2933f47-7168-42b3-9f52-65d814ec422e/osist-pren-iso-5530-1-2020)<https://standards.iteh.ai/catalog/standards/sist/e2933f47-7168-42b3-9f52-65d814ec422e/osist-pren-iso-5530-1-2020>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 5530-1:2020(E)

© ISO 2020

iTeh STANDARD PREVIEW (standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/e2933f47-7168-42b3-9f52-65d814ec422e/osist-pren-iso-5530-1-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Reagent	3
6 Apparatus	3
7 Sampling	3
8 Procedure	4
8.1 Determination of the moisture content of the flour.....	4
8.2 Preparation of Farinograph.....	4
8.3 Test portion.....	4
8.3.1 Constant flour mass procedure.....	4
8.3.2 Constant dough mass procedure.....	8
8.4 Common rules of determination.....	8
9 Evaluation of the Farinogram and calculation of the derived rheological characteristics	9
9.1 General.....	9
9.2 Water absorption of flour.....	9
9.3 Characteristics relating to the consistency of dough.....	10
10 Precision	11
10.1 Interlaboratory tests.....	11
10.2 Repeatability.....	11
10.3 Reproducibility.....	12
11 Test report	12
Annex A (informative) Description of the Farinograph	13
Annex B (informative) Examples of Farinograms	19
Annex C (informative) Results of interlaboratory tests	24
Bibliography	31

ISO/DIS 5530-1:2020(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, Food products, Subcommittee SC 4, Cereals and pulses.

This fourth edition cancels and replaces the third edition (ISO 5530-1:2013), which has been technically revised.

The main changes compared to the previous edition are as follows:

- Wheat flour interlaboratory test 2015 described in this annex (see [Tables C.1](#) to C.6) in order to evaluate the repeatability and reproducibility of the test method specified in this part of ISO 5530.

A list of all parts in the ISO 5530 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Wheat flour — Physical characteristics of doughs —

Part 1:

Determination of water absorption and rheological properties using a farinograph

1 Scope

This part of ISO 5530 specifies a method, using a Farinograph, for the determination of the water absorption of flours and the mixing behaviour of doughs made from them by a constant flour mass procedure, or by a constant dough mass procedure.

The method is applicable to experimental and commercial flour from wheat (*Triticum aestivum* L.).

NOTE This part of ISO 5530 is based on ICC 115/1^[1] and AACC Method 54-21.2^[2].

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 712, *Cereals and cereal products — Determination of moisture content — Reference method*

<https://standards.iteh.ai/catalog/standards/sist/e2933f47-7168-42b3-9f52-65d814ec422e/osist-pren-iso-5530-1-2020>

3 Terms and definitions

For the purposes of this part of ISO 5530, the following terms and definitions apply.

3.1

consistency

resistance of a dough to being mixed in specific conditions (mixing conditions, temperature, hydration, ...). In this document it will refer to resistance of dough being mixed in a Farinograph under the conditions specified in the following methodology

Note 1 to entry: It is expressed in Farinograph arbitrary units (see 3.2).

3.2

Farinograph unit

FU

arbitrary unit for consistency on the Farinogram

Note 1 to entry: For the mathematical expression of Farinograph units, see 6.1.

Note 2 to entry: It is also possible to define “Farinograph unit (FU)” as a torque expressed in Nm, measured in the axis of the mixer.

3.3

maximum consistency

consistency measured at the end of dough development time

Note 1 to entry: For the mathematical expression of maximum consistency, see 9.2.

Note 2 to entry: It is expressed in Farinograph units (FU).

ISO/DIS 5530-1:2020(E)

Note 3 to entry: See [3.7](#).

3.4 water absorption of flour (WA)

volume of water required to produce a dough with a maximum consistency of 500 FU, under the specified operating conditions

Note 1 to entry: Water absorption is expressed in millilitres per 100 g of flour at 14 % (mass fraction) moisture content to an accuracy of 0,1 ml.

3.5 dough development time (DDT, former wordings also use 'peak time')

time from the beginning of the addition of water to the point on the curve immediately before the first sign of the decrease of maximum consistency

Note 1 to entry: In those cases where two maxima are observed, use the second maximum to measure the dough development time.

Note 2 to entry: See [Figure 1](#) and 9.3 .

Note 3 to entry: It is expressed in minutes to the nearest 0,1 min.

3.6 Stability (S)

difference in time between the point where the top part of the curve intercepts, for the first time, the line of 500 FU and the last point where leaves this line

Note 1 to entry: This value measures the tolerance of the flour to mixing.

Note 2 to entry: When the maximum consistency deviates from the (500 ± 20) FU line, the line of this consistency should be used to read the interceptions.

Note 3 to entry: The stability is expressed in minutes, to an accuracy of 0,5 min.

3.7 degree of softening (DS)

Two definitions: First one the difference between the height of the centre of the curve at the point where it begins to decline (DDT) and the height of the centre of the curve 12 min after that point (ICC method). Secondly the difference between the height of the centre of the curve at DDT and the height of the centre of the curve 10 min after beginning of the measurement

Note 1 to entry: It is expressed in Farinograph units (FU).

Note 2 to entry: In the case where two peaks appear, the second peak is considered.

Note 3 to entry: The degree of softening should be expressed to the nearest 5 FU.

Note 4 to entry: If another time is used to carry out this method, this has to be detailed in the report along with information on the reference standard applied. The definite time is usually 12 min.

3.8 mixing tolerance index (MTI)

difference from the top of the curve at peak (DDT) to the top of the curve measured at 5 min after peak is reached

Note 1 to entry: It is expressed in Farinograph units (FU).

3.9 Farinograph quality number (FQN)

length, along the time axis, between the point of the addition of water and the point where the height of the centre of the curve has decreased by 30 FU, compared to the height of the centre of the curve at DDT

Note 1 to entry: It is expressed in millimetres to an accuracy of 1 mm.

4 Principle

Measuring and recording, by means of a Farinograph, the consistency of a dough as it is formed from flour and water, as it is developed, and as it changes with time.

NOTE The maximum consistency of the dough is adjusted to a fixed value by adapting the quantity of water added. The correct water addition, which is called the water absorption, is used to obtain a complete mixing curve, the various features of which are a guide to the rheological properties (strength) of the dough.

5 Reagent

Use only distilled or demineralized water or water complying with grade 3, ISO 3696.

6 Apparatus

The usual laboratory apparatus and, in particular, the following:

6.1 Farinograph¹⁾ (see [Annex A](#)), with the following operating characteristics:

- slow blade rotational frequency: $(63 \pm 2) \text{ min}^{-1}$ (rev/min); the ratio of the rotational frequencies of the mixing blades shall be $1,50 \pm 0,01$;
- torque per Farinograph unit:
 - for a 300 g mixer: $(9,8 \pm 0,2) \text{ mN}\cdot\text{m}/\text{FU}$ [$(100 \pm 2) \text{ gf}\cdot\text{cm}/\text{FU}$];
 - for a 50 g mixer: $(1,96 \pm 0,04) \text{ mN}\cdot\text{m}/\text{FU}$ [$(20 \pm 0,4) \text{ gf}\cdot\text{cm}/\text{FU}$];
 - chart speed: $(1,00 \pm 0,03) \text{ cm}/\text{min}$, for the chart recording devices;
 - for the electronic devices chart speed is not applicable, but time is measured.

6.1.1 Water dosing

6.1.2 Thermostat, with circulating water for constant temperature $(30^\circ \pm 0,2)^\circ\text{C}$.

6.2 Balance, capable of weighing to the nearest $\pm 0,1 \text{ g}$.

6.3 Spatula, thin, made of non metal material.

7 Sampling

Sampling is not part of the method specified in this part of ISO 5530. A recommended sampling method is given in ISO 24333^[3].

It is important that the laboratory receive a sample which is truly representative and which has not been damaged or changed during transport and storage.

1) This part of ISO 5530 has been drawn up on the basis of the Brabender Farinograph, which is an example of a suitable product available commercially. This information is given for the convenience of users of this part of ISO 5530 and does not constitute an endorsement by ISO of this product. Other equipment may be used if it can be shown to give comparable results.

ISO/DIS 5530-1:2020(E)

8 Procedure

8.1 Determination of the moisture content of the flour

Determine the moisture content of the flour using the method specified in ISO 712.

8.2 Preparation of Farinograph

NOTE See details of electronic Farinograph characteristics and procedure in [A.4](#).

8.2.1 Turn on the thermostat of the Farinograph ([6.1.2](#)) and circulate the water, until the required temperature is reached, prior to using the instrument. Before and during use, check the temperatures of the thermostat and of the mixing bowl, the latter in the hole provided for this purpose. The temperature of the mixing bowl shall be $(30 \pm 0,2)$ °C.

The laboratory temperature should be between 20 °C and 30 °C.

8.2.2 For the mechanical devices uncouple the mixer from the driving shaft and adjust the position of the counterweight(s) so as to obtain zero deflection of the pointer with the motor running at the specified rotational frequency (see [6.1](#)). Switch off the motor and then couple the mixer. For the electronic devices the zero adjustment is programmed to be done automatically at the start of each measurement.

8.2.3 For the mechanical devices lubricate the mixer with a drop of water between the back-plate and each of the blades. Check that the deflection of the pointer is within the range (0 ± 5) FU with the mixing blades rotating at the specified rotational frequency in the empty, clean bowl. If the deflection exceeds 5 FU, clean the mixer more thoroughly or eliminate other causes of friction. For the electronic controlled devices lubrication of the blades is done with silicon fat.

8.2.4 For the mechanical devices adjust the arm of the pen so as to obtain identical readings from the pointer and the recording pen.

8.2.5 For the mechanical devices adjust the damper so that, with the motor running, the time required for the pointer to go from 1 000 FU to 100 FU is $(1,0 \pm 0,2)$ s. This should result in a bandwidth of approximately 60 FU to 90 FU.

8.2.6 Fill the burette ([6.1.1](#)) with water at 30 °C. The time to flow from 0 ml to 225 ml or from 0 ml to 37,5 ml, respectively, shall be not more than 20 s. For the electronic Farinograph's the time for the water flow by means of the dosing system is the same.

8.3 Test portion

NOTE If necessary, bring the flour to a temperature of between 25 °C and 30 °C.

8.3.1 Constant flour mass procedure

Weigh ([6.2](#)), to the nearest 0,1 g, the equivalent of 300 g (for a 300 g mixer) or 50 g (for a 50 g mixer) of flour having a moisture content of 14 % mass fraction. Let this mass, in grams, be m ; see [Table 1](#) for m as a function of moisture content.

Place the test portion in the mixer. Cover the mixer, and keep it covered until the end of mixing except, for the shortest possible time, when water has to be added and the dough has to be scraped down. Temperature of the measurement is defined in note [8.2.1](#).

Table 1 — Mass of flour, in grams, equivalent to 300 g and 50 g at a moisture content of 14 % mass fraction

Moisture content % mass fraction	Mass, <i>m</i> , of flour equivalent to	
	300 g	50 g
9,0	283,5	47,3
9,1	283,8	47,3
9,2	284,1	47,4
9,3	284,5	47,4
9,4	284,8	47,5
9,5	285,1	47,5
9,6	285,4	47,6
9,7	285,7	47,6
9,8	286,0	47,7
9,9	286,3	47,7
10,0	286,7	47,8
10,1	287,0	47,8
10,2	287,3	47,9
10,3	287,6	47,9
10,4	287,9	48,0
10,5	288,3	48,0
10,6	288,6	48,1
10,7	288,9	48,2
10,8	289,2	48,2
10,9	289,6	48,3
11,0	289,9	48,3
11,1	290,2	48,4
11,2	290,5	48,4
11,3	290,9	48,5
11,4	291,2	48,5
11,5	291,5	48,6
11,6	291,9	48,6
11,7	292,2	48,7
11,8	292,5	48,8
11,9	292,8	48,8
12,0	293,2	48,9
12,1	293,5	48,9

NOTE The values in this table are calculated using the following formulae:

a) for the mass, in grams, equivalent to 300 g at 14 % mass fraction moisture content

$$m = \frac{25800}{100 - H}$$

b) for the mass, in grams, equivalent to 50 g at 14 % mass fraction moisture content:

$$m = \frac{4300}{100 - H}$$

where *H* is the moisture content of the sample, as a percentage by mass.

Calculation example: Mass of flour to be added, e.g. having 13 % moisture:

$$m(13\%) = 300\text{g} * (100\% - 14\%) / (100\% - 13\%) = 296,55\text{g}$$

Table 1 (continued)

Moisture content	Mass, <i>m</i> , of flour equivalent to	
12,2	293,8	49,0
12,3	294,2	49,0
12,4	294,5	49,1
12,5	294,9	49,1
12,6	295,2	49,2
12,7	295,5	49,3
12,8	295,9	49,3
12,9	296,2	49,4
13,0	296,6	49,4
13,1	296,9	49,5
13,2	297,2	49,5
13,3	297,6	49,6
13,4	297,9	49,7
13,5	298,3	49,7
13,6	298,6	49,8
13,7	299,0	49,8
13,8	299,3	49,9
13,9	299,7	49,9
14,0	300,0	50,0
14,1	300,3	50,1
14,2	300,7	50,1
14,3	301,1	50,2
14,4	301,4	50,2
14,5	301,8	50,3
14,6	302,1	50,4
14,7	302,5	50,4
14,8	302,8	50,5
14,9	303,2	50,5
15,0	303,5	50,6
15,1	303,9	50,6
15,2	304,2	50,7
15,3	304,6	50,8
15,4	305,0	50,8
15,5	305,3	50,9

NOTE The values in this table are calculated using the following formulae:

a) for the mass, in grams, equivalent to 300 g at 14 % mass fraction moisture content

$$m = \frac{25800}{100 - H}$$

b) for the mass, in grams, equivalent to 50 g at 14 % mass fraction moisture content:

$$m = \frac{4300}{100 - H}$$

where *H* is the moisture content of the sample, as a percentage by mass.

Calculation example: Mass of flour to be added, e.g. having 13 % moisture:

$$m(13\%) = 300\text{g} * (100\% - 14\%) / (100\% - 13\%) = 296,55\text{g}$$