



SLOVENSKI STANDARD

oSIST prEN 17462:2020

01-februar-2020

Krma: metode vzorčenja in analize - Določevanje radionuklidnega joda-131, cezija-134 in cezija-137 v krmi

Animal feeding stuffs: Methods of sampling and analysis - Determination of the radionuclides Iodine-131, Caesium-134 and Caesium-137 in feed

Futtermittel: Probenahme- und Untersuchungsverfahren - Bestimmung der Radionuklide Jod-131, Cäsium-134 und Cäsium-137 in Futtermittel

Aliments des animaux : Méthodes d'échantillonnage et d'analyse - Détermination des radionucléides iode 131, césium 134 et césium 137 dans les aliments composés pour animaux

<https://standards.iteh.ai/catalog/standards/sist/6584f4f6-db8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021>

Ta slovenski standard je istoveten z: prEN 17462

ICS:

65.120

Krmila

Animal feeding stuffs

oSIST prEN 17462:2020

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[kSIST FprEN 17462:2021](#)

<https://standards.iteh.ai/catalog/standards/sist/6584f4f6-db8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN 17462

December 2019

ICS 65.120

English Version

Animal feeding stuffs: Methods of sampling and analysis - Determination of the radionuclides Iodine-131, Caesium- 134 and Caesium-137 in feed

Aliments des animaux : Méthodes d'échantillonnage et
d'analyse - Détermination des radionucléides iode 131,
césium 134 et césium 137 dans les aliments composés
pour animaux

Futtermittel: Probenahme- und
Untersuchungsverfahren - Bestimmung der
Radionuklide Jod-131, Cäsium-134 und Cäsium-137 in
Futtermittel

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 327.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Symbols and abbreviations	8
4.1 Symbols	8
4.2 Abbreviations	11
5 Principle	11
6 Safety precautions	11
7 Equipment	12
7.1 General	12
7.2 Equipment for test portion preparation	12
7.3 Gamma-ray spectrometry equipment	12
7.4 Containers to be used	12
8 Procedure	13
8.1 Calibration	13
8.2 Test portion preparation	16
8.3 Spectrum recording	16
8.4 Spectrum analysis	17
8.5 Quality assurance	17
9 Expression of results	18
9.1 Massic activity calculation	18
9.2 Characteristic limits	24
9.3 Precision	25
9.4 Test report	26
Annex A (informative) List of possible interfering gamma rays from naturally occurring radionuclides as well as from radionuclides that could be present in environmental samples immediately after a nuclear accident or incident	27
Annex B (informative) Example of uncertainty budget in gamma-ray spectrometry using an HPGe detector	28
Annex C (informative) Results of the collaborative trial	29
Bibliography	34

European foreword

This document (prEN 17462:2019) has been prepared by Technical Committee CEN/TC 327 “Animal feeding stuffs: Methods of sampling and analysis”, the secretariat of which is held by NEN.

This document is currently submitted to the CEN Enquiry.

This document has been prepared under a standardization request given to CEN by the European Commission and the European Free Trade Association.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[kSIST FprEN 17462:2021](https://standards.iteh.ai/catalog/standards/sist/6584f4f6-db8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021)

<https://standards.iteh.ai/catalog/standards/sist/6584f4f6-db8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021>

Introduction

This document describes a method for ^{131}I , ^{134}Cs and ^{137}Cs massic activity determination (Bq/kg) in animal feed. It was initiated by Directorate General for Health and Food Safety (DG SANTE) of the European Commission following the accident in the Fukushima Daiichi nuclear power plant in March 2011. The event highlighted the need for standardized measurements of the three most common radioactive contaminants following such type of nuclear accident.

The most commonly used method for identification and quantification of these radionuclides in animal feed samples is high-resolution gamma-ray spectrometry. As this is a secondary measurement method based on analysis of photopeaks of the emitted gamma rays care should be taken to use appropriate energy and efficiency calibrations for the detector and test portion used. This method of massic activity determination is described in the present document.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[kSIST FprEN 17462:2021](https://standards.iteh.ai/catalog/standards/sist/6584f4f6-db8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021)

<https://standards.iteh.ai/catalog/standards/sist/6584f4f6-db8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021>

1 Scope

This document describes a method for determination of the massic activity (Bq/kg) of ^{131}I , ^{134}Cs and ^{137}Cs in animal feed in monitoring laboratories.

General guidance on the preparation of feed samples and the measurement of the three radionuclides ^{131}I , ^{134}Cs and ^{137}Cs by high resolution gamma-ray spectrometry is provided. The current document aims to be complementary to existing standards. More information on sample preparation, moisture content determination and gamma-ray spectrometry can be found in specific standards referred to in this document. For example, generic advice on the equipment selection, detectors and quality assurance for gamma-ray spectrometry can be found in ISO 20042.

The method was fully statistically tested and evaluated in a collaborative trial comprising five animal feeding stuff samples for the radionuclides ^{131}I , ^{134}Cs and ^{137}Cs . Details on the successfully tested working range for each of the examined radionuclides are described in Annex C.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the cited edition applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 6497, *Animal feeding stuffs - Sampling (ISO 6497)*

ISO 11929-1, *Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation - Fundamentals and application - Part 1: Elementary applications*

ISO 20042:2019, *Measurement of radioactivity - Gamma-ray emitting radionuclides - Generic test method using gamma-ray spectrometry*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN ISO 6497 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 background

spectrum recorded by the gamma-ray detector when no sample is measured; the spectral data, including full energy peaks, in such a spectrum is resulting from radioactive decay occurring in the environment surrounding the detector (including the cosmic ray interactions) or in the detector

prEN 17462:2019 (E)**3.2****background continuum**

events in the spectrum that form a smooth curve onto which the photopeaks are superimposed

Note 1 to entry: The continuum may arise from gamma rays scattered inside the test sample or any surrounding materials, from cosmic radiation or from radionuclides in the surrounding materials.

[SOURCE: ISO 20042:2019, 3.1]

3.3**blank sample**

sample, liquid or solid, with very low to no activity for radiation of the same type and region of interest, with a mass and a composition as close as possible to those of the test sample

[SOURCE: ISO 19581:2017, 3.1, see [1]]

3.4**coincidence summing****true coincidence summing (TCS)****cascade-summing**

simultaneous detection of two or more gamma-rays in the spectrometry system, due to the emission of a cascade of gamma-rays in the decay of a single nucleus in the test sample

[SOURCE: ISO 20042:2019, 3.24]

3.5**dead time**

time during spectrum acquisition (real time) during which pulses are not recorded or processed

Note 1 to entry: Dead time is given by real time minus live time.
<https://standards.iteh.ai/catalog/standards/sist/6584446-db8b-4076-9830-0ded59815316/ksis-prEN-17462-2021>

Note 2 to entry: The time is given in seconds.

[SOURCE: ISO 20042:2019, 3.5]

3.6**detection efficiency transfer****efficiency transfer**

calculation that enables the user to establish the value of the detection efficiency for a given gamma-ray peak in the spectrum of the test portion, when only the detection efficiency from an experimental calibration with a reference source that may have a different composition, density and/or geometry compared to the test portion is known

3.7**full energy peak (FEP)**

see photopeak [3.12]

3.8**high energy resolution**

relative term which in gamma-ray spectrometry refers to the energy resolution obtained with a Ge(Li) or an HPGe detector

3.9**laboratory sample**

sample as prepared (from the lot) for sending to the laboratory and intended for inspection or testing

[SOURCE: EN ISO 6498:2012, 2.1.2 [9]]

3.10**live time**

time during which pulses are processed during an acquisition (real) time

Note 1 to entry: The time is given in seconds.

[SOURCE: ISO 20042:2019, 3.12]

3.11**photopeak****full energy peak (FEP)**

peak observed above the background continuum in a gamma-ray spectrum due to events that deposit the full energy of the photon in the detector material, usually approximately Gaussian in shape

[SOURCE: ISO 20042:2019, 3.17]

3.12**real time**

time taken to acquire a spectrum

Note 1 to entry: The time is given in seconds.

[SOURCE: ISO 20042:2019, 3.10]

3.13**sample holder**

device that is specially designed to enable the placement of a given sample container in a well-defined position on top of a specific detector

3.14**spectrometry system**

complete assembly of the sensor and associated pulse-processing electronics that converts the gamma rays detected by the sensor into a pulse-height spectrum

[SOURCE: ISO 20042:2019, 3.22]

3.15**test portion**

quantity of material drawn from the test sample (or from the laboratory sample if both are the same)

[SOURCE: EN ISO 6498:2012, 2.1.4 [9]]

iTeh STANDARD PREVIEW
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/65844f6-d8b-4076-9830-6ded3981331b/ksist-fpren-17462-2021>

prEN 17462:2019 (E)

3.16

test sample

subsample or sample prepared from the laboratory sample and from which test portions will be taken

[SOURCE: EN ISO 6498:2012, 2.1.3 [9]]

4 Symbols and abbreviations**4.1 Symbols**

For the purposes of this document, the following symbols apply.

Symbol	Name of quantity	Unit
A	activity of a reference radionuclide emitting photons of energy E in the calibration source, at the time of calibration	Bq
a_m	massic activity at energy E of a radionuclide in the sample	Bq/kg
$a_{m,605}; a_{m,796}$	massic activity of ^{134}Cs obtained using the gamma-ray of energy i , which is either 604,72 keV or 795,86 keV	Bq/kg
a'_m	massic activity of ^{134}Cs based on the weighted mean calculation including the two major gamma rays of this radionuclide	Bq/kg
\tilde{a}	true value of massic activity at energy E of a radionuclide in the sample	Bq/kg
a_m^*	decision threshold kSIST FprEN 17462:2021	Bq/kg
$a_m^\#$	detection limit 6ded3981331b/ksist-fpren-17462-2021	Bq/kg
a	annum (year), the tropical year which is approximately equal to 365,2422 d	year
d	day (1 day = 86 400 s)	day
ε_E	detection efficiency at energy E for the specific measurement geometry and detector used	-
$f_{\text{att}(E)}$	factor to correct for gamma-ray attenuation within the test portion (self-attenuation)	-
f_d	factor to correct for decay between the reference time and the start of the measurement and during the measurement NOTE 1 The latter is important for short-lived radionuclides like ^{131}I .	-
f_{d1}	is the factor to correct for decay between the reference time and the start of the measurement	-
f_{d2}	is the factor to correct for decay during the measurement	-
f_E	composite correction factor for the gamma ray with energy E considering all necessary corrections as shown in Formula (5)	-

Symbol	Name of quantity	Unit
f_g	factor to correct for geometry differences	-
$f_{tcs,E}$	factor to correct for true coincidence summing effects NOTE 2 In this case, this is only applicable to ^{134}Cs .	-
λ	decay constant of a radionuclide	s^{-1}
m	quantity of the test portion	kg
m_c	corrected quantity of the test portion	kg
m_f	fresh mass of the test portion	kg
m_d	dry mass of the moisture content determination portion	kg
m_w	fresh mass of the moisture content determination portion	kg
N	number of gamma rays used for the calculation of massic activity for ^{134}Cs	-
$n_{N,E}$	number of counts in the net area of the photopeak at energy E , in the test portion spectrum	-
$n_{b,E}$	number of counts in the net area of the photopeak at energy E , in the background spectrum	-
P_E	probability (per 100 decays) of the emission of a gamma ray with energy E by a radionuclide NOTE 3 The probability can be expressed in percentage (%) or in absolute values.	-
$r_{N,E}$	net count rate in the full energy peak at energy E	s^{-1}
t_b	background spectrum live time	s
t_g	test portion spectrum live time	s
t_i	time elapsed between the reference time and the start of the measurement NOTE 4 It will have a negative value when the measurement was started before the reference time and positive value when the measurement was started after the reference time.	s
t_r	test portion spectrum real time	s
t_s	calibration spectrum live time	s
$t_{1/2}$	half-life of a radionuclide	s
$u_{rel}(A)$	relative uncertainty of the activity of a reference radionuclide emitting photons of energy E in the calibration source, at the time of calibration	-

prEN 17462:2019 (E)

Symbol	Name of quantity	Unit
$u(a_m)$	standard uncertainty of the massic activity	Bq/kg
$U(a_m)$	expanded uncertainty with coverage factor k calculated as $U = k \times u$	Bq/kg
$u(a_{m,605});$ $u(a_{m,796})$	standard uncertainty of the massic activity calculated for two most intense gamma rays (604,72 keV and 795,86 keV) of ^{134}Cs	Bq/kg
$\tilde{u}(\tilde{a}_m)$	standard uncertainty of a_m as a function of its true value	Bq/kg
$u_{rel}(\varepsilon_E)$	relative uncertainty of the detection efficiency at energy E for the specific measurement geometry and detector used	-
$u_{rel}(f_E)$	relative uncertainty of the composite correction factor for the gamma ray with energy E considering all necessary corrections as shown in Formula (5)	-
$u_{rel}(f_d)$	relative uncertainty of the factor to correct for decay between the reference time and the start of the measurement and during the measurement	-
$u_{rel}(f_{tcs,E})$	relative uncertainty of the factor to correct for true coincidence summing effects	-
$u_{rel}(m)$	relative uncertainty of the quantity of the test portion	-
$u(n_{N,E})$	uncertainty of the net number of counts in the photopeak at energy E in the test portion spectrum	-
$u(n_{b,E})$	uncertainty of the net number of counts in the photopeak at energy E in the background spectrum	-
$u_{rel}(P_E)$	relative uncertainty of the probability (per 100 decays) of the emission of a gamma ray with energy E by a radionuclide	-
$u(r_{N,E})$	uncertainty of the net count rate	-
$u_{rel}(r_{N,E})$	relative uncertainty of the net count rate	-
u_{rand}	random uncertainty component	-
u_{sys}	systematic uncertainty component	-
u_{tot}	total uncertainty calculated based on random and systematic components	-
$u(w)$	total standard uncertainty for coverage factor w	-
$u_{rel}(w)$	relative value of total standard uncertainty for coverage factor w	-