



SLOVENSKI STANDARD SIST EN ISO 19980:2021

01-september-2021

Nadomešča:
SIST EN ISO 19980:2012

Oftalmični instrumenti - Topografi roženice (ISO 19980:2021)

Ophthalmic instruments - Corneal topographers (ISO 19980:2021)

Ophthalmische Instrumente - Hornhauttopographen (ISO 19980:2021)

Instruments ophtalmiques - Topographes de la cornée (ISO 19980:2021)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Ta slovenski standard je istoveten z: EN ISO 19980:2021

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31cc73f2f7/sist-en-iso-19980-2021>

ICS:

11.040.70 Oftalmološka oprema Ophthalmic equipment

SIST EN ISO 19980:2021 **en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 19980:2021

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 19980

June 2021

ICS 11.040.70

Supersedes EN ISO 19980:2012

English Version

Ophthalmic instruments - Corneal topographers (ISO 19980:2021)

Instruments ophtalmiques - Topographes de la cornée
(ISO 19980:2021)

Ophthalmische Instrumente - Hornhauttopographen
(ISO 19980:2021)

This European Standard was approved by CEN on 6 June 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 19980:2021
<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>

European foreword

This document (EN ISO 19980:2021) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2021, and conflicting national standards shall be withdrawn at the latest by December 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 19980:2012.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of ISO 19980:2021 has been approved by CEN as EN ISO 19980:2021 without any modification.

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 19980:2021

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>

INTERNATIONAL
STANDARD

ISO
19980

Third edition
2021-06

**Ophthalmic instruments — Corneal
topographers**

Instruments ophtalmiques — Topographes de la cornée

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 19980:2021](https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021)

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>



Reference number
ISO 19980:2021(E)

© ISO 2021

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 19980:2021

<https://standards.iteh.ai/catalog/standards/sist/51a8cc33-2c67-4022-905a-b31ec73fc2f7/sist-en-iso-19980-2021>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	9
4.1 Area measured.....	9
4.2 Measurement sample density.....	9
4.3 Measurement and report of performance.....	9
4.4 Colour presentation of results.....	9
5 Test methods and test devices	9
5.1 Tests.....	9
5.1.1 Accuracy test.....	9
5.1.2 Repeatability test.....	9
5.2 Test surfaces.....	9
5.2.1 Reflection-based and Placido ring corneal topographers.....	9
5.2.2 Luminous surface corneal topographer.....	10
5.2.3 Optical-sectioning corneal topographer.....	10
5.2.4 Specification of test surfaces.....	10
5.2.5 Verification of test surfaces.....	10
5.2.6 Type testing of surfaces.....	10
5.3 Data collection — Test surfaces.....	11
5.4 Analysis of the data.....	11
5.4.1 General.....	11
5.4.2 Structure of the accuracy data set.....	12
5.4.3 Analysis of the paired data sets.....	12
5.4.4 Report of accuracy performance.....	13
6 Accompanying documents	13
7 Marking	13
Annex A (informative) Test surfaces for corneal topographers (CTs)	14
Annex B (informative) Standardized displays for corneal topographers (CTs)	16
Annex C (normative) Calculation of area-weighting values	19
Annex D (normative) Test methods for measuring human corneas	21
Bibliography	22

ISO 19980:2021(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 19980:2012), which has been technically revised. The main changes compared to the previous edition are as follows:

- a) normative references were updated;
- b) [5.2.6](#) regarding requirements for test surfaces and requirement for testing of accuracy was changed;
- c) in [5.4.3](#), formulae for data analysis have been updated;
- d) Table 4 was deleted;
- e) document editorially revised.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Ophthalmic instruments — Corneal topographers

1 Scope

This document specifies minimum requirements for instruments and systems that fall into the class of corneal topographers (CTs). It also specifies tests and procedures to verify that a system or instrument complies with this document and thus qualifies as a CT according to this document. It also specifies tests and procedures that allow the verification of capabilities of systems that are beyond the minimum requirements for CTs.

This document defines terms that are specific to the characterization of the corneal shape so that they may be standardized throughout the field of vision care.

This document is applicable to instruments, systems and methods that are intended to measure the surface shape of the cornea of the human eye.

NOTE The measurements can be of the curvature of the surface in local areas, three-dimensional topographical measurements of the surface or other more global parameters used to characterize the surface.

This document is not applicable to ophthalmic instruments classified as ophthalmometers.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005 + A1:2012 + A2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

corneal apex

location on the corneal surface where the mean of the local principal curvature is greatest

Note 1 to entry: See Figure 1.

3.2

corneal eccentricity

e_c

eccentricity, e , of the conic section that best fits the *corneal meridian* (3.3) of interest

Note 1 to entry: If the meridian is not specified, the corneal eccentricity is that of the flattest corneal meridian (see [Table 1](#) and [Annex A](#)).

ISO 19980:2021(E)

3.3 corneal meridian

θ

curve created by the intersection of the corneal surface and a plane that contains the corneal topographer axis

Note 1 to entry: A meridian is identified by the angle θ , that the plane creating it makes to the horizontal (see ISO 8429).

Note 2 to entry: The value of θ , for a full meridian, ranges from 0° to 180°.

3.4 corneal shape factor

E

value that specifies the type of conic section that best fits a *corneal meridian* (3.3), given by Formula (1):

$$E = 1 - p \quad (1)$$

where

p is the value that specifies a conic section such as a circle, ellipse, hyperbola, or parabola

value p is given by Formula (2):

$$p = \pm \frac{a^2}{b^2} \quad (2)$$

where

a and b are the semi-diameters of the axes of the conic section;
 + indicates a circle or ellipse;
 - indicates a hyperbola

a conic section is specified by Formula (3):

$$\frac{z^2}{b^2} \pm \frac{x^2}{a^2} = 1 \quad (3)$$

value E also is the square of the *eccentricity* (3.9) of the conic section, given by Formula (4):

$$E = e^2 \quad (4)$$

Note 1 to entry: Unless otherwise specified, E refers to the meridian with least curvature (flattest meridian). See Table 1 and Annex A.

Note 2 to entry: Although the magnitude of E is equal to the square of the eccentricity and so is always positive, the sign of E is a convention to signify whether an ellipse takes a prolate or oblate orientation.

Note 3 to entry: The negative value of E is defined by ISO 10110-12 as the conic constant designated by the symbol K . The negative value of E has also been called asphericity and given the symbol Q .

Note 4 to entry:

Note 5 to entry: