

---

**Medicinska električna oprema - 2-69. del: Posebne zahteve za osnovno varnost in bistvene lastnosti naprav za koncentriranje kisika (ISO/DIS 80601-2-69)**

Medical electrical equipment - Part 2-69: Particular requirements for basic safety and essential performance of oxygen concentrator equipment (ISO/DIS 80601-2-69)

Medizinische elektrische Geräte - Teil 2-69: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale für Sauerstoff-Konzentratoren (ISO/DIS 80601-2-69)

Appareils électromédicaux - Partie 2-69: Exigences particulières pour la sécurité de base et les performances essentielles des dispositifs concentrateurs d'oxygène (ISO/DIS 80601-2-69)

**Ta slovenski standard je istoveten z: prEN ISO 80601-2-69**

---

**ICS:**

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

**oSIST prEN ISO 80601-2-69:2019**                      **en**



# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 80601-2-69

ISO/TC 121/SC 3

Secretariat: ANSI

Voting begins on:  
2019-07-25Voting terminates on:  
2019-10-17

### Medical electrical equipment —

Part 2-69:

### Particular requirements for basic safety and essential performance of oxygen concentrator equipment

*Appareils électromédicaux —**Partie 2-69: Exigences particulières pour la sécurité de base et les performances essentielles des dispositifs concentrateurs d'oxygène*

ICS: 11.040.10

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

SIST EN ISO 80601-2-69:2021

<https://standards.iteh.ai/catalog/standards/sist/e156c41c-d98f-4f2d-be6f-0cd26d8bb9fe/sist-en-iso-80601-2-69-2021>

Member bodies are requested to consult relevant national interests in IEC/SC 62D before casting their ballot to the e-Balloting application.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

### ISO/CEN PARALLEL PROCESSING



Reference number  
ISO/DIS 80601-2-69:2019(E)

© ISO 2019

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 80601-2-69:2021

<https://standards.iteh.ai/catalog/standards/sist/e156c41c-d98f-4f2d-be6f-0cd26d8bb9fe/sist-en-iso-80601-2-69-2021>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

## CONTENTS

1		
2	CONTENTS .....	3
3	FOREWORD .....	5
4	INTRODUCTION .....	7
5	201.1 Scope, object and related standards .....	9
6	201.2 Normative references .....	11
7	201.3 Terms and definitions .....	13
8	201.4 General requirements .....	13
9	201.5 General requirements for testing of <i>ME equipment</i> .....	14
10	201.6 Classification of <i>ME equipment</i> and <i>ME systems</i> .....	14
11	201.7 <i>ME equipment</i> identification, marking and documents .....	14
12	201.8 Protection against electrical hazards from <i>ME equipment</i> .....	21
13	201.9 Protection against <i>mechanical hazards</i> of <i>ME equipment</i> and <i>ME systems</i> .....	21
14	201.10 Protection against unwanted and excessive radiation hazards .....	23
15	201.11 Protection against excessive temperatures and other hazards .....	23
16	201.12 Accuracy of controls and instruments and protection against hazardous	
17	outputs .....	27
18	201.13 Hazardous situations and fault conditions for <i>ME equipment</i> .....	33
19	201.14 Programmable electrical medical systems (PEMS) .....	33
20	201.15 Construction of <i>ME equipment</i> .....	33
21	201.16 <i>ME systems</i> .....	33
22	201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i> .....	34
23	201.101 Outlet connector .....	34
24	201.102 Requirements for parts and accessories .....	34
25	201.103 Functional connection .....	36
26	201.104 Indication of duration of operation .....	36
27	201.105 Integrated <i>conserving equipment</i> function .....	36
28	202 Electromagnetic disturbances – Requirements and tests .....	36
29	206 Usability .....	37
30	211 Requirements for medical electrical equipment and medical electrical systems	
31	used in the home healthcare environment .....	38
32	Annex C (informative) Guide to marking and labelling requirements for <i>ME equipment</i>	
33	and <i>ME systems</i> .....	39
34	Annex D (informative) Symbols on marking .....	43
35	Annex AA (informative) Particular guidance and rationale .....	45

## ISO/DIS 80601-2-69:2019(E)

– 4 –

ISO DIS 80601-2-69 © ISO 2019

36	Annex BB (informative) Reference to the <i>essential principles</i> .....	52
37	Annex CC (informative) Reference to the general safety and performance requirements.....	56
38	Annex DD (informative) Terminology — alphabetized index of defined terms.....	60
39	Bibliography.....	63
40		
41	Figure 201.101 – Standard resistance.....	23
42	Figure 201.102 — Oxygen delivery performance, typical test setup .....	31
43		
44	Table 201.101 — Distributed <i>essential performance</i> requirements.....	14
45	Table 201.C.101 — Marking on the outside of an <i>oxygen concentrator</i> , its parts or	
46	<i>accessories</i> .....	39
47	Table 201.C.102 — <i>Accompanying documents</i> , general .....	40
48	Table 201.C.103 — Instructions for use.....	40
49	Table 201.C.104 — Technical description.....	42
50	Table 201.D.1.101 — Additional symbols on marking.....	43
51	Table 201.D.2.101 — Additional safety signs .....	44
52	Table BB.1 — Correspondence between this document and the <i>essential principles</i> .....	52
53	Table CC.1 — Correspondence between this document and the general safety and	
54	performance requirements.....	56
55		

## INTERNATIONAL ORGANIZATION for STANDARDISATION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-69: Particular requirements for the basic safety and essential performance of oxygen concentrator equipment**

## FOREWORD

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. [www.iso.org/directives](http://www.iso.org/directives)

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. [www.iso.org/patents](http://www.iso.org/patents)

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

ISO 80601-2-69 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

This second edition of ISO 80601-2-69 cancels and replaces the first edition of ISO 80601-2-69:2014. This edition of ISO 80601-2-69 constitutes a technical revision of

89 ISO 80601-2-69:2014 and includes an alignment with the fourth edition of IEC 60601-1-2, the  
90 third edition of IEC 60601-1-6, including its Amendment 1, the second edition of IEC 60601-1-8,  
91 including its Amendment 1, and the second edition of IEC 60601-1-11.

92 The most significant changes are the following modifications:

- 93 – changes to the low oxygen concentration *alarm condition*;
- 94 – changes to the gas outlet connector;
- 95 – changes to the test method for the filter for the delivered gas; and
- 96 – reformatted to provide a unique identifier for each requirement.

97

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

SIST EN ISO 80601-2-69:2021

<https://standards.iteh.ai/catalog/standards/sist/e156c41c-d98f-4f2d-be6f-0cd26d8bb9fe/sist-en-iso-80601-2-69-2021>



## INTRODUCTION

Oxygen supplementation can be part of management of *patients* with chronic, acute-on-chronic or acute respiratory disorders. The amount of supplemental oxygen depends on the individual *patient's* needs under various conditions. The managing healthcare team typically prescribes the endpoint of treatment, for example a target value for oxygen saturation. The amount of supplemental oxygen can be controlled by the flowrate.

The goal of long-term oxygen therapy is to keep the oxygen saturation above 90 % in *patients* that require supplemental oxygen. The flowrate should be adjusted for rest, exertion, and sleep to meet the individual *patient's* needs under these various conditions. Ideally, the resting flowrate is adjusted to maintain  $\text{SpO}_2 > 90 \%$  as indicated by pulse oximetry.

Supplemental oxygen is supplied by various sources: *medical gas pipeline systems, oxygen concentrators*, compressed gas cylinders and liquid oxygen reservoirs. This document covers the particular requirements for *basic safety* and *essential performance* of *oxygen concentrators*. *Oxygen concentrators* produce oxygen-enriched air from room air for delivery to a *patient* requiring oxygen therapy. The most common *oxygen concentrator* uses molecular sieve beds to filter and concentrate oxygen molecules from the ambient air, generating oxygen concentrations of typically 82 % to 96 %. The main component of this type of *oxygen concentrator* is the molecular sieve, which adsorbs nitrogen from air to produce a product gas, which is a mixture of typically up to 95 % oxygen and 5 % of other gases. The periodic adsorbing and purging of nitrogen is referred to as the pressure swing adsorption process.

Long-term oxygen therapy has been demonstrated in randomized, controlled clinical trials to prolong survival in *patients* with chronic respiratory disease and documented hypoxemia. Typical sources of therapeutic long-term oxygen therapy include gaseous oxygen from cylinders or from liquid oxygen and oxygen from an *oxygen concentrator*.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications and terms defined in Clause 3 of the general standard, in this document or as noted: italic type; and*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

In referring to the structure of this document, the term

- “clause” means one of the three numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.); and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

136 In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any  
137 combination of the conditions is true.

138 The verbal forms used in this document conform to usage described in ISO/IEC Directives, Part 2.  
139 For the purposes of this document, the auxiliary verb:

140 – “shall” means that conformance with a requirement or a test is mandatory for conformance  
141 with this document;

142 – “should” means that conformance with a requirement or a test is recommended but is not  
143 mandatory for conformance with this document;

144 – “may” is used to describe a permission (e.g., a permissible way to achieve conformance with a  
145 requirement or test);

146 – “can” is used to describe a possibility or capability; and

147 – “must” is used express an external constraint.

148 Annex C contains a guide to the marking and labelling requirements in this document.

149 Annex D contains a summary of the symbols referenced in this document.

150 An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title  
151 indicates that there is guidance or rationale related to that item in Annex AA.

152 This document is a particular standard in the IEC 60601 and IEC/ISO 80601 series of standards.

153 SIST EN ISO 80601-2-69:2021  
<https://standards.iteh.ai/catalog/standards/sist/e156c41c-d98f-4f2d-be6f-0cd26d8bb9fe/sist-en-iso-80601-2-69-2021>

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-69: Particular requirements for the basic safety and essential performance of oxygen concentrator equipment

#### 201.1 \* Scope, object and related standards

IEC 60601-1:2005+A1:2012<sup>1</sup>, Clause 1 applies, except as follows:

##### 201.1.1 Scope

IEC 60601-1:2005+Amendment 1:2012, 1.1 is replaced by:

This document specifies requirements for the *basic safety* and *essential performance* of an *oxygen concentrator* in combination with its *accessories*, hereafter referred to as *ME equipment*, intended to increase the oxygen concentration of gas intended to be delivered to a single *patient*. Such *oxygen concentrators* are typically intended for use in the *home healthcare environment*, including *transit-operable* use by a single *patient* in various environments including any private and public transportation as well as in commercial aircraft.

NOTE 1 Such an *oxygen concentrator* can also be used in professional healthcare facilities.

This document is applicable to a *transit-operable* and non-*transit-operable oxygen concentrator*. This document is applicable to an *oxygen concentrator* integrated into or used with other medical devices, *ME equipment* or *ME systems*.

EXAMPLE 1 An *oxygen concentrator* with integrated oxygen conserving equipment<sup>[1]</sup> <sup>2</sup> function or humidifier function.

EXAMPLE 2 An *oxygen concentrator* used with a flowmeter stand.

EXAMPLE 3 An *oxygen concentrator* as part of an anaesthetic system for use in areas with limited logistical supplies of electricity and anaesthetic gases<sup>[2]</sup>.

EXAMPLE 4 An *oxygen concentrator* with an integrated liquid reservoir function or gas cylinder filling system function.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to an *oxygen concentrator*, where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *oxygen concentrator*.

This document does not specify the requirements for *oxygen concentrators* for use with a *medical gas pipeline system*.

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

<sup>1</sup> The general standard is IEC 60601-1:2005 +AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

<sup>2</sup> Figures in square brackets refer to the Bibliography.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE 2 See also 4.2 of the General Standard.

This document is a particular standard in the IEC 60601 and IEC/ISO 80601 series of standards.

### 201.1.2 Object

IEC 60601-1:2005, 1.2 is replaced by:

The object of this document is to establish particular *basic safety* and *essential performance* requirements for an *oxygen concentrator* [as defined in 201.3.202] and its *accessories*.

NOTE 1 *Accessories* are included because the combination of the *oxygen concentrator* and the *accessories* needs to be adequately safe. *Accessories* can have a significant impact on the *basic safety* or *essential performance* of an *oxygen concentrator*.

NOTE 2 This document has been prepared to address the relevant *essential principles of safety and performance* of ISO 16142-1:2016 as indicated in Annex BB.

NOTE 3 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745 as indicated in Annex CC.

### 201.1.3 Collateral standards

IEC 60601-1:2005+AMD 1:2012, 1.3 applies with the following addition:

IEC 60601-1-2:2014, IEC 60601-1-6:2010+AMD1:2013+AMD2:— and IEC 60601-1-11:2015 apply as modified in Clauses 202, 206 and 211 respectively. IEC 60601-1-3<sup>[3]</sup> does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

### 201.1.4 Particular standards

#### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular *ME equipment* under consideration, and may add other *basic safety* and *essential performance* requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x”, where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3<sup>[3]</sup> collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this document.

"Addition" means that the text of this document is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3 [3], etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

## 201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the Bibliography.

IEC 60601-1:2005+AMD 1:2012, Clause 2 applies, except as follows:

*Replacement:*

ISO 7010:2011+AMD1:2012+AMD2:2012+AMD3:2012+AMD4:2013+AMD5:2014  
+AMD6:2014+AMD7:2016, *Graphical symbols -- Safety colours and safety signs -- Registered safety signs*

IEC 60601-1-2:2014, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic disturbances — Requirements and tests*

## ISO/DIS 80601-2-69:2019(E)

– 12 –

ISO DIS 80601-2-69 © ISO 2019

- 260 IEC 60601-1-6:2010+AMD1:2013, *Medical electrical equipment – Part 1-6: General requirements*  
 261 *for basic safety and essential performance – Collateral standard: Usability*
- 262 IEC 60601-1-8:2006+AMD1:2012, *Medical electrical equipment - Part 1-8: General requirements*  
 263 *for basic safety and essential performance - Collateral Standard: General requirements, tests and*  
 264 *guidance for alarm systems in medical electrical equipment and medical electrical systems*
- 265 IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic*  
 266 *safety and essential performance – Collateral Standard: Requirements for medical electrical*  
 267 *equipment and medical electrical systems used in the home healthcare environment*
- 268 IEC 61672:2013, *Electroacoustics - Sound level meters - Part 1: Specifications*
- 269 *Addition:*
- 270 ISO 3744:2010, *Acoustics -- Determination of sound power levels and sound energy levels of noise*  
 271 *sources using sound pressure -- Engineering methods for an essentially free field over a reflecting*  
 272 *plane*
- 273 ISO 7396-1:2016, *Medical gas pipeline systems -- Part 1: Pipeline systems for compressed medical*  
 274 *gases and vacuum*
- 275 ISO 9000:2015, *Quality management systems -- Fundamentals and vocabulary*
- 276 ISO 14937:2009, *Sterilization of health care products -- General requirements for characterization*  
 277 *of a sterilizing agent and the development, validation and routine control of a sterilization process*  
 278 *for medical devices*
- 279 ISO 15223-1:2016, *Medical devices — Symbols to be used with medical device labels, labelling and*  
 280 *information to be supplied — Part 1: General requirements*
- 281 ISO 16142-1:2016, *Medical devices -- Recognized essential principles of safety and performance of*  
 282 *medical devices -- Part 1: General essential principles and additional specific essential principles for*  
 283 *all non-IVD medical devices and guidance on the selection of standards*
- 284 ISO 17664:2017, *Processing of health care products -- Information to be provided by the medical*  
 285 *device manufacturer for the processing of medical devices*
- 286 ISO 18562-1:2017, *Biocompatibility evaluation of breathing gas pathways in healthcare*  
 287 *applications -- Part 1: Evaluation and testing within a risk management process*
- 288 ISO 19223-1:2019, *Lung ventilators and related equipment -- Vocabulary and semantics*
- 289 ISO 80601-2-74:2017, *Medical electrical equipment — Part 2-74: Particular requirements for basic*  
 290 *safety and essential performance of respiratory humidifying equipment*
- 291 ISO 80601-2-67:—<sup>3</sup>, *Medical Electrical Equipment — Part 2-67: Particular requirements for basic*  
 292 *safety and essential performance of oxygen conserving equipment*
- 293 IEC 60601-1:2005+AMD1:2012, *Medical electrical equipment – Part 1: General requirements for*  
 294 *basic safety and essential performance*

<sup>3</sup> To be published. Stage at time of publication ISO/DIS 80601-2-67:2019.

- 295 IEC 62366-1:2015, *Medical devices – Application of usability engineering to medical devices*
- 296 EN 13544-2:2002+AMD1:2009, *Respiratory therapy equipment - Part 2: Tubing and connectors*
- 297 EN 15986:2011, *Symbol for use in the labelling of medical devices - Requirements for labelling of*
- 298 *medical devices containing phthalates*

### 299 **201.3 Terms and definitions**

300 For the purposes of this document, the terms and definitions given in ISO 3744:2010,  
 301 ISO 7396-1:2016, ISO 9000:2015, ISO 16142-1:2016, ISO 17664:2017, ISO 18562-1:2017,  
 302 ISO 19223:2019, ISO 80601-2-67:—, ISO 80601-2-74:2017, IEC 60601-1:2005+AMD 1:2012,  
 303 IEC 60601-1-2:2014, IEC 60601-1-8:2006+AMD 1:2012, IEC 60601-1-11:2015,  
 304 IEC 62366-1:2015 and the following apply.

305 NOTE An index of defined terms is found in Annex DD.

306 *Addition:*

#### 307 **201.3.201**

##### 308 ***flow-direction-sensitive component***

309 component or *accessory* through which gas flow has to be in one direction only for proper  
 310 functioning or *patient* safety

311 [ISO 4135:2001 <sup>[4]</sup>, definition 3.1.7, modified — Added 'or *accessory*' and replaced 'must' with 'has  
 312 to']

#### 313 **201.3.202**

##### 314 ***oxygen concentrator***

315 *ME equipment*, which by selective removal of constituents of ambient air, increases the  
 316 concentration of oxygen in the output gas

### 317 **201.4 General requirements**

318 IEC 60601-1:2005+AMD 1:2012, Clause 4 applies, except as follows:

#### 319 **201.4.3 Essential performance**

320 IEC 60601-1:2005+AMD 1:2012, 4.3 applies, except as follows:

321 *Additional subclause:*

##### 322 **201.4.3.101 \* Additional requirements for essential performance**

323 Additional *essential performance* requirements are found in the subclauses listed in  
 324 Table 201.101.