

SLOVENSKI STANDARD
oSIST prEN ISO 20785-1:2019
01-april-2019

Dozimetrija za merjenje izpostavljenosti kozmičnemu sevanju v civilnem letalskem prometu - 1. del: Konceptualna osnova za meritve (ISO/DIS 20785-1:2019)

Dosimetry for exposures to cosmic radiation in civilian aircraft - Part 1: Conceptual basis for measurements (ISO/DIS 20785-1:2019)

Dosimetrie zu Expositionen durch kosmische Strahlung in Flugzeugen der zivilen Luftfahrt - Teil 1: Konzeptionelle Grundlage für Messungen (ISO/DIS 20785-1:2019)

Dosimétrie pour l'exposition au rayonnement cosmique à bord d'un avion civil - Partie 1: Fondement théorique des mesures (ISO/DIS 20785-1:2019)

Ta slovenski standard je istoveten z: prEN ISO 20785-1

ICS:

13.280	Varstvo pred sevanjem	Radiation protection
49.020	Letala in vesoljska vozila na splošno	Aircraft and space vehicles in general

oSIST prEN ISO 20785-1:2019

en,fr,de

DRAFT INTERNATIONAL STANDARD

ISO/DIS 20785-1

ISO/TC 85/SC 2

Secretariat: AFNOR

Voting begins on:
2019-01-30Voting terminates on:
2019-04-24

Dosimetry for exposures to cosmic radiation in civilian aircraft —

Part 1: Conceptual basis for measurements

Dosimétrie pour l'exposition au rayonnement cosmique à bord d'un avion civil —

Partie 1: Fondement théorique des mesurages

ICS: 49.020; 13.280

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 20785-1:2020

<https://standards.iteh.ai/catalog/standards/sist/009b4242-34cc-444f-9a85-b4f5c09b1c45/sist-en-iso-20785-1-2020>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 20785-1:2019(E)

© ISO 2019

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 20785-1:2020

<https://standards.iteh.ai/catalog/standards/sist/009b4242-34cc-444f-9a85-b4f5c09b1c45/sist-en-iso-20785-1-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
3.1 General terms.....	2
3.2 Quantities and units.....	3
3.3 Atmospheric radiation field.....	8
4 General considerations	10
4.1 The cosmic radiation field in the atmosphere.....	10
4.2 General calibration considerations for the dosimetry of cosmic radiation fields in aircraft.....	11
4.2.1 Approach.....	11
4.2.2 Considerations concerning the measurement.....	12
4.2.3 Considerations concerning the radiation field.....	12
4.2.4 Considerations concerning calibration.....	13
4.2.5 Simulated aircraft fields.....	13
4.3 Conversion coefficients.....	14
5 Dosimetric devices	14
5.1 Introduction.....	14
5.2 Active devices.....	14
5.2.1 Devices to determine all field components.....	14
5.2.2 Devices for low LET/non-neutron.....	16
5.2.3 Devices for high-LET/neutron component.....	16
5.3 Passive devices.....	17
5.3.1 General considerations.....	17
5.3.2 Etched track detectors.....	18
5.3.3 Fission foil detectors.....	18
5.3.4 Superheated emulsion neutron detectors (bubble) detectors.....	18
5.3.5 Thermoluminescent detectors.....	19
5.3.6 Photoluminescent detectors.....	19
Annex A (informative) Representative particle fluence rate energy distributions for the cosmic radiation field at flight altitudes for solar minimum and maximum conditions and for minimum and maximum vertical cut-off rigidity	20
Bibliography	26

ISO/DIS 20785-1:2019(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, Nuclear energy nuclear technologies and radiological protection, Subcommittee SC 2, Radiation protection.

This second edition cancels and replaces the first edition (ISO 20785-1:2012), which has been technically revised. The main changes are:

A list of all the parts in the ISO 20785- series can be found on the ISO website.

Introduction

Aircraft crews are exposed to elevated levels of cosmic radiation of galactic and solar origin and secondary radiation produced in the atmosphere, the aircraft structure and its contents. Following recommendations of the International Commission on Radiological Protection in Publication 60,^[1] confirmed by Publication 103,^[2] the European Union (EU) introduced a revised Basic Safety Standards Directive^[3] which included exposure to natural sources of ionizing radiation, including cosmic radiation, as occupational exposure. The Directive requires account to be taken of the exposure of aircraft crews liable to receive more than 1 mSv per year. It then identifies the following four protection measures: (i) to assess the exposure of the crew concerned; (ii) to take into account the assessed exposure when organizing working schedules with a view to reducing the doses of highly exposed crews; (iii) to inform the workers concerned of the health risks their work involves; and (iv) to apply the same special protection during pregnancy to female crews in respect of the "child to be born" as to other female workers. The EU Council Directive has already been incorporated into laws and regulations of EU Member States and is being included in the aviation safety standards and procedures of the Joint Aviation Authorities and the European Air Safety Agency. Other countries such as Canada and Japan have issued advisories to their airline industries to manage aircraft crew exposure.

For regulatory and legislative purposes, the radiation protection quantities of interest are the equivalent dose (to the foetus) and the effective dose. The cosmic radiation exposure of the body is essentially uniform and the maternal abdomen provides no effective shielding to the foetus. As a result, the magnitude of equivalent dose to the foetus can be put equal to that of the effective dose received by the mother. Doses on board aircraft are generally predictable, and events comparable to unplanned exposure in other radiological workplaces cannot normally occur (with the rare exceptions of extremely intense and energetic solar particle events). Personal dosimeters for routine use are not considered necessary. The preferred approach for the assessment of doses of aircraft crews, where necessary, is to calculate directly the effective dose per unit time, as a function of geographic location, altitude and solar cycle phase, and to combine these values with flight and staff roster information to obtain estimates of effective doses for individuals. This approach is supported by guidance from the European Commission and the ICRP in Publication 75.^[4]

The role of calculations in this procedure is unique in routine radiation protection and it is widely accepted that the calculated doses should be validated by measurement. The effective dose is not directly measurable. The operational quantity of interest is ambient dose equivalent, $H^*(10)$. In order to validate the assessed doses obtained in terms of effective dose, calculations can be made of ambient dose equivalent rates or route doses in terms of ambient dose equivalent, and values of this quantity determined by measurements traceable to national standards. The validation of calculations of ambient dose equivalent for a particular calculation method may be taken as a validation of the calculation of the effective dose by the same computer code, but this step in the process may need to be confirmed. The alternative is to establish *a priori* that the operational quantity ambient dose equivalent is a good estimator of effective dose and equivalent dose to the foetus for the radiation fields being considered, in the same way that the use of the operational quantity personal dose equivalent is justified for the estimation of effective dose for radiation workers.

The radiation field in aircraft at altitude is complex, with many types of ionizing radiation present, with energies ranging up to many GeV. The determination of ambient dose equivalent for such a complex radiation field is difficult. In many cases, the methods used for the determination of ambient dose equivalent in aircraft are similar to those used at high-energy accelerators in research laboratories. Therefore, it is possible to recommend dosimetric methods and methods for the calibration of dosimetric devices, as well as the techniques for maintaining the traceability of dosimetric measurements to national standards. Dosimetric measurements made to evaluate ambient dose equivalent should be performed using accurate and reliable methods that ensure the quality of readings provided to workers and regulatory authorities. This document gives a conceptual basis for the characterization of the response of instruments for the determination of ambient dose equivalent in aircraft.

Requirements for the determination and recording of the cosmic radiation exposure of aircraft crews have been introduced into the national legislation of EU Member States and other countries.

ISO/DIS 20785-1:2019(E)

Harmonization of methods used for determining ambient dose equivalent and for calibrating instruments is desirable to ensure the compatibility of measurements performed with such instruments.

This document is intended for the use of primary and secondary calibration laboratories for ionizing radiation, by radiation protection personnel employed by governmental agencies, and by industrial corporations concerned with the determination of ambient dose equivalent for aircraft crews.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 20785-1:2020

<https://standards.iteh.ai/catalog/standards/sist/009b4242-34cc-444f-9a85-b4f5c09b1c45/sist-en-iso-20785-1-2020>

Dosimetry for exposures to cosmic radiation in civilian aircraft —

Part 1: Conceptual basis for measurements

1 Scope

This document describes the conceptual basis for the determination of ambient dose equivalent for the evaluation of exposure to cosmic radiation in civilian aircraft and for the calibration of instruments used for that purpose.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC Guide 98-1, *Uncertainty of measurement — Part 1: Introduction to the expression of uncertainty in measurement*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 4037-1, *X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy — Part 1: Radiation characteristics and production methods*

ISO 6980-1, *Nuclear energy — Reference beta-particle radiation — Part 1: Methods of production*

ISO 8529-1:2001, *Reference neutron radiations — Part 1: Characteristics and methods of production*

ISO 12789-1, *Reference radiation fields — Simulated workplace neutron fields — Part 1: Characteristics and methods of production*

ISO 12789-2, *Reference radiation fields — Simulated workplace neutron fields — Part 2: Calibration fundamentals related to the basic quantities*

ISO 20785-2, *Dosimetry for exposures to cosmic radiation in civilian aircraft — Part 2: Characterization of instrument response*

ISO 20785-3, *Dosimetry for exposures to cosmic radiation in civilian aircraft — Part 3: Measurements at aviation altitudes*

ISO 20785-4, *Dosimetry for exposures to cosmic radiation in civilian aircraft — Part 4: Validation of codes*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO/DIS 20785-1:2019(E)

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 General terms

3.1.1

calibration

operation that, under specified conditions, establishes a relation between the conventional quantity, H_0 , and the indication, G

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: : Calibration should not be confused with adjustment of a measuring system, often mistakenly called "self-calibration", or with verification of calibration.

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

3.1.2

calibration coefficient

N_{coeff}

quotient of the conventional quantity value to be measured and the corrected indication of the instrument

Note 1 to entry: The calibration coefficient is equivalent to the calibration factor multiplied by the instrument constant.

Note 2 to entry: The reciprocal of the calibration coefficient, N_{coeff} , is the response.

Note 3 to entry: For the calibration of some instruments, e.g. ionization chambers, the instrument constant and the calibration factor are not identified separately but are applied together as the calibration coefficient.

Note 4 to entry: It is necessary, in order to avoid confusion, to state the quantity to be measured, for example: the calibration coefficient with respect to fluence, N_Φ , the calibration coefficient with respect to kerma, N_K , the calibration coefficient with respect to absorbed dose, N_D .

3.1.3

indication

G

quantity value provided by a measuring instrument or a measuring system

Note 1 to entry: An indication can be presented in visual or acoustic form or can be transferred to another device. An indication is often given by the position of a pointer on the display for analogue outputs, a displayed or printed number for digital outputs, a code pattern for code outputs, or an assigned quantity value for material measures.

Note 2 to entry: An indication and a corresponding value of the quantity being measured are not necessarily values of quantities of the same kind.

3.1.4

reference conditions

conditions of use prescribed for testing the performance of a detector assembly or for comparison of results of measurements

Note 1 to entry: The reference conditions represent the values of the set of influence quantities for which the calibration result is valid without any correction.

Note 2 to entry: The value of the measurand can be chosen freely in agreement with the properties of the detector assembly to be calibrated. The quantity to be measured is not an influence quantity but can influence the calibration result and the response.

3.1.5**response** R

quotient of the indication, G , or the corrected indication, G_{corr} , and the conventional quantity value to be measured

Note 1 to entry: To avoid confusion, it is necessary to specify which of the quotients, given in the definition of the response (to G or to G_{corr}) is applied. Furthermore, it is necessary, in order to avoid confusion, to state the quantity to be measured, for example: the response with respect to fluence, R_{Φ} , the response with respect to kerma, R_K , the response with respect to absorbed dose, R_D .

Note 2 to entry: : The reciprocal of the response under the specified conditions is equal to the calibration coefficient N_{coeff} .

Note 3 to entry: The value of the response may vary with the magnitude of the quantity to be measured. In such cases the detector assembly's response is said to be non-constant.

Note 4 to entry: The response usually varies with the energy and direction distribution of the incident radiation. It is, therefore, useful to consider the response as a function, $R(E, \Omega)$, of the radiation energy, E , and of the direction, Ω of the incident monodirectional radiation. $R(E)$ describes the "energy dependence" and $R(\Omega)$ the "angle dependence" of response; for the latter, Ω may be expressed by the angle, α , between the reference direction of the detector assembly and the direction of an external monodirectional field.

3.2 Quantities and units**3.2.1****particle fluence
fluence** Φ

number, dN , at a given point in space, of particles incident on a small spherical domain, divided by the cross-sectional area, da , of that domain:

$$\Phi = \frac{dN}{da}$$

Note 1 to entry: The unit of the fluence is m^{-2} ; a frequently used unit is cm^{-2} .

Note 2 to entry: The energy distribution of the particle fluence, Φ_E , is the quotient, $d\Phi$, by dE , where $d\Phi$ is the fluence of particles of energy between E and $E+dE$. There is an analogous definition for the direction distribution, Φ_{Ω} , of the particle fluence. The complete representation of the double differential particle fluence can be written (with arguments) $\Phi_{E, \Omega}(E, \Omega)$, where the subscripts characterize the variables (quantities) for differentiation and where the symbols in the brackets describe the values of the variables. The values in the brackets are needed for special function values, e.g. the energy distribution of the particle fluence at energy $E = E_0$ is written as $\Phi_E(E_0)$. If no special values are indicated, the brackets may be omitted.

3.2.2**particle fluence rate
fluence rate** $\dot{\Phi}$

rate of the particle fluence expressed as

$$\dot{\Phi} = \frac{d\Phi}{dt} = \frac{d^2N}{da \cdot dt}$$

where $d\Phi$ is the increment of the particle fluence during an infinitesimal time interval with duration dt .

Note 1 to entry: The unit of the fluence rate is $\text{m}^{-2} \text{s}^{-1}$, a frequently used unit is $\text{cm}^{-2} \text{s}^{-1}$.

ISO/DIS 20785-1:2019(E)

3.2.3 energy imparted

ε
for ionizing radiation in the matter within a given three-dimensional domain,

$$\varepsilon = \sum \varepsilon_i$$

where

ε_i is the energy deposited in a single interaction, i , and given by $\varepsilon_i = \varepsilon_{in} - \varepsilon_{out} + Q$, where

ε_{in} is the energy of the incident ionizing particle, excluding rest energy;

ε_{out} is the sum of the energies of all ionizing particles leaving the interaction, excluding rest energy, and

Q is the change in the rest energies of the nucleus and of all particles involved in the interaction.

Note 1 to entry: Energy imparted is a stochastic quantity.

Note 2 to entry: The unit of the energy imparted is J.

3.2.4 mean energy imparted

$\bar{\varepsilon}$
mean energy imparted to the matter in a given domain, expressed as

$$\bar{\varepsilon} = R_{in} - R_{out} + \sum Q$$

where

R_{in} is the radiant energy of all those charged and uncharged ionizing particles that enter the domain;

R_{out} is the radiant energy of all those charged and uncharged ionizing particles that leave the domain, and

$\sum Q$ is the sum of all changes of the rest energy of nuclei and elementary particles that occur in that domain.

Note 1 to entry: This quantity has the meaning of expected value of the energy imparted.

Note 2 to entry: The unit of the mean energy imparted is J.

3.2.5 specific energy imparted specific energy

z
for any ionizing radiation,

$$z = \frac{\varepsilon}{m}$$

where

ε is the energy imparted to the irradiated matter;

m is the mass of the irradiated matter.

Note 1 to entry: Specific energy imparted is a stochastic quantity.

Note 2 to entry: In the limit of a small domain, the mean specific energy imparted is equal to the absorbed dose.