
Molekularne diagnostične preiskave in vitro - Specifikacije za predpreiskovalne procese pri aspiraciji s tanko iglo (FNA) - 2. del: Izolirani proteini

Molecular in vitro diagnostic examinations - Specifications for pre-examination processes for Fine Needle Aspirates (FNAs) - Part 2: Isolated proteins

Molekularanalytische in vitro diagnostische Verfahren - Spezifikationen für präanalytische Prozesse für Feinnadelaspirate - Teil 2: Isolierte Proteine

Analyses moléculaires de diagnostic in vitro - Spécifications pour les processus préanalytiques pour les ponctions à l'aiguille fine - Partie 2 : Protéines extradites

<https://standards.iteh.ai/catalog/standards/sist/56874e00-98d3-487b-b2ba-0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021>

<https://standards.iteh.ai/catalog/standards/sist/56874e00-98d3-487b-b2ba-0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021>

Ta slovenski standard je istoveten z: FprCEN/TS 17688-2

ICS:

11.100.10	Diagnostični preskusni sistemi in vitro	In vitro diagnostic test systems
-----------	---	----------------------------------

kSIST-TS FprCEN/TS 17688-2:2021 **en,fr,de**

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

[kSIST-TS FprCEN/TS 17688-2:2021](https://standards.iteh.ai/catalog/standards/sist/56874e00-98d3-487b-b2ba-0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/56874e00-98d3-487b-b2ba-0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021>

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

FINAL DRAFT
FprCEN/TS 17688-2

July 2021

ICS 11.100.10

English Version

**Molecular in vitro diagnostic examinations - Specifications
for pre-examination processes for Fine Needle Aspirates
(FNAs) - Part 2: Isolated proteins**

Analyses moléculaires de diagnostic in vitro -
Spécifications pour les processus préanalytiques pour
les ponctions à l'aiguille fine - Partie 2 : Protéines
extradites

Molekularanalytische in-vitro-diagnostische Verfahren
- Spezifikationen für präanalytische Prozesse für
Feinnadelaspirate - Teil 2: Isolierte Proteine

This draft Technical Specification is submitted to CEN members for Vote. It has been drawn up by the Technical Committee CEN/TC 140.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a Technical Specification. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a Technical Specification.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 General considerations	11
5 Outside the laboratory	12
5.1 Specimen collection.....	12
5.1.1 General.....	12
5.1.2 Information about the specimen donor/patient	12
5.1.3 Information about the specimen.....	13
5.1.4 Selection of the primary FNA collection device.....	13
5.1.5 FNA specimen collection and processing from the donor/patient	13
5.2 Specimen storage and transport	15
6 Inside the laboratory	15
6.1 Specimen reception.....	15
6.2 Specimen/sample storage after transport and reception.....	15
6.2.1 General.....	15
6.2.2 Storage of FNA specimen/samples using collection devices with stabilizer	16
6.2.3 Storage of FNA specimen/samples using collection devices without stabilizers	16
6.3 Specimen/sample processing for cytological examination prior to protein isolation	16
6.3.1 General.....	16
6.3.2 Handling of cell suspension.....	17
6.3.3 Preparation of paraffin-embedded cell blocks.....	18
6.3.4 Preparation of cell suspension slides	19
6.4 Evaluation of the pathology of the specimen or sample(s).....	19
6.5 Processed sample storage, transport and reception.....	20
6.5.1 General.....	20
6.5.2 Storage and transport of cell suspension	20
6.5.3 Storage and transport of paraffin-embedded cell blocks	20
6.5.4 Storage and transport of cell suspension slides	21
6.6 Isolation of protein.....	21
6.6.1 General.....	21
6.6.2 Using a commercial protein isolation kit intended for diagnostic use.....	21
6.6.3 Using the laboratory's own protein isolation procedure	22
6.6.4 Isolation of proteins from specific sample types.....	22
6.7 Quantity and quality assessment of isolated proteins.....	23
6.8 Storage of isolated proteins	23
Bibliography.....	25

European foreword

This document (FprCEN/TS 17688-2:2021) has been prepared by Technical Committee CEN/TC 140 “In vitro diagnostic medical devices”, the secretariat of which is held by DIN.

This document is currently submitted to the Vote on TS.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[kSIST-TS FprCEN/TS 17688-2:2021](https://standards.iteh.ai/catalog/standards/sist/56874e00-98d3-487b-b2ba-0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/56874e00-98d3-487b-b2ba-0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021>

Introduction

Molecular *in vitro* diagnostics has enabled significant progress in medicine. Further progress is expected by new technologies analysing profiles of nucleic acids, proteins, and metabolites in human tissues and body fluids. However, the profiles of these molecules can change drastically during the pre-examination process, including the specimen collection, transport, storage and processing.

Examination of proteins is commonly used in clinical practice. This includes e.g. prognostic and predictive biomarker examinations. This is a fast growing field in molecular diagnostics.

Fine needle aspiration is a non-surgical procedure that uses a thin, hollow-bore needle and syringe to collect a specimen from patients for cytopathological and molecular investigation. As a minimally-invasive technique, fine needle aspirates (FNAs) are commonly used to diagnose and monitor for example a range of cancer types e.g. breast, lung and thyroid cancer, and other diseases, such as inflammatory diseases. FNAs also provide the opportunity to sample metastatic sites (e.g. lymph nodes) and otherwise non-resectable tissues.

Besides cytological assessment, molecular biological analysis of FNAs is expected to become increasingly used for cancer and other disease diagnostics, including companion diagnostics.

One of the challenges facing molecular analysis of FNA samples is their small size and diversity in composition (cells, blood, body fluid). The low cellular content of FNAs means that the yield of isolated proteins is typically towards the lower end of detection for molecular examination. Therefore, the protein isolation procedure should provide a sufficient amount of protein as required by the specific examination.

Protein profiles, protein integrities, and protein-protein interactions in FNAs can change drastically during and after collection (due to, e.g. gene induction, gene down regulation, protein degradation and modification). Protein species amounts can change differently in different donors'/patients' FNAs.

Therefore, standardization of the entire process from specimen collection to protein examination is needed to minimize protein degradation and protein profile changes during and after FNA collection. This document describes special measures which need to be taken to obtain good quality FNA specimens/samples and isolated protein therefrom for molecular examination.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

1 Scope

This document gives guidelines on the handling, documentation, storage and processing of fine needle aspirates (FNAs) intended for protein examination during the pre-examination phase before a molecular examination is performed.

This document is applicable to molecular *in vitro* diagnostic examinations including laboratory developed tests performed by medical laboratories and molecular pathology laboratories that examine proteins isolated from FNAs. It is also intended to be used by laboratory customers, *in vitro* diagnostics developers and manufacturers, biobanks, institutions and commercial organisations performing biomedical research, and regulatory authorities.

Different dedicated measures are taken for collecting, stabilizing, transporting and storing of core needle biopsies (FNA Biopsy or FNA B) and are not covered in this document, but in EN ISO 20184-2, *Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for frozen tissue — Part 2: Isolated proteins* and EN ISO 20166-2, *Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for formalin fixed and paraffin-embedded (FFPE) tissue — Part 2: Isolated proteins*.

This document is not applicable for protein examination by immunohistochemistry.

NOTE International, national or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 15189, *Medical laboratories - Requirements for quality and competence (ISO 15189)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN ISO 15189 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

aliquot

portion of a larger amount of homogenous material, assumed to be taken with negligible sampling error

Note 1 to entry: The term is usually applied to fluids. Tissues are heterogeneous and therefore cannot be aliquoted.

Note 2 to entry: The definition is derived from the Compendium of Chemical Terminology Gold Book. International Union of Pure and Applied Chemistry. Version 2.3.3., 2014; the PAC, 1990, 62, 1193 (Nomenclature for sampling in analytical chemistry (Recommendations 1990)) p. 1206; and the PAC 1990, 62, 2167 (Glossary of atmospheric chemistry terms (Recommendations 1990)) p. 2173.

3.10

examination
analytical test

set of operations with the objective of determining the value or characteristics of a property

Note 1 to entry: Processes that start with the *in situ detection* using *antibodies*, nucleic acid probes or dyes and include all kinds of parameter testing or chemical manipulation for quantitative or qualitative examination.

[SOURCE: ISO 15189:2012, 3.7, modified — Notes to entry 1 to 3 have been removed. Note 1 to entry has been added and “analytical test” has been added as a preferred term.]

3.11

examination manufacturer
analytical test manufacturer

entity that manufactures and/or produces a specific analytical test

3.12

examination performance
analytical test performance
analytical performance

accuracy, precision, and sensitivity of a test to measure the analyte of interest

Note 1 to entry: Other test performance characteristics such as robustness, repeatability can apply as well.

[SOURCE: EN ISO 20184-1:2018, 3.4]

3.13

fixative

solution used to preserve or harden FNA specimens for microscopic and molecular examination

3.14

formalin

saturated aqueous formaldehyde solution which at 100 % contains 37 % formaldehyde by mass (corresponding to 40 % by volume)

[SOURCE: EN ISO 20166-1:2018, 3.11]

3.15

fine needle aspirate
FNA

specimen withdrawn by a non-operative procedure that uses a thin, hollow-bore needle

3.16

FNA primary collection device

thin, hollow-bore needle or syringe used for collecting the FNA specimen from the donor/patient

3.17

FNA secondary collection device

suitable container into which the specimen is transferred from the FNA primary collection device

FprCEN/TS 17688-2:2021 (E)

3.18

homogeneous

uniform in structure and composition

[SOURCE: EN ISO 20166-1:2018, 3.31]

3.19

laboratory developed procedure

modified commercially available *in vitro* diagnostic device or fully in house developed procedure

3.20

nonconformity

nonfulfillment of a requirement

[SOURCE: ISO 9000:2005, 3.6.9]

3.21

paraffin embedding

process in which a sample is placed in paraffin to generate a hard surrounding matrix so that thin microscopic sections can be cut

3.22

pre-examination process**pre-analytical workflow****pre-analytical phase****pre-examination phase**

process that starts, in chronological order, from the clinician's request and includes the examination request, preparation and identification of the patient, collection of the primary sample(s), transportation to and within the analytical laboratory, isolation of analytes, and ends when the analytical examination begins

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Note 1 to entry: The pre-examination phase includes preparative processes that influence the outcome of the intended examination.

[SOURCE: ISO 15189:2012, 3.15, modified — An additional term was added and more detail was included.]

3.23

primary sample**specimen**

discrete portion of a body fluid, breath, hair or tissue taken for examination, study or analysis of one or more quantities or properties assumed to apply for the whole

[SOURCE: ISO 15189:2012, 3.16, modified — The term and definition is used here without the original Notes.]

3.24

proficiency test

evaluation of participant performance against pre-established criteria by means of inter-laboratory comparisons

[SOURCE: EN ISO/IEC 17043:2010, 3.7, modified — Term and definition are used here without the original Notes.]

3.25**protein**

type of biological macromolecules composed of one or more chains with a defined sequence of amino acids connected through peptide bonds

3.26**protein profile**

amounts of the individual protein molecules that are present in a sample and that can be measured in the absence of any losses, inhibition and interference

3.27**protein species**

amounts of a chemically clearly-defined protein corresponding to one spot on a high-performance two-dimensional gel electrophoresis pattern

[SOURCE: [16]]

3.28**post translational modifications****PTM**

chemical alterations to a primary protein structure, often crucial for conferring biological activity on a protein

[SOURCE: [17]]

3.29**ribonucleic acid****RNA**

polymer of ribonucleotides occurring in a double-stranded or single-stranded form

[SOURCE: EN ISO 22174:2005, 3.1.3]

3.30**room temperature**

temperature in the range of 18 °C to 25 °C

Note 1 to entry: Local or national regulations can have different definitions.

[SOURCE: EN ISO 20166-1:2018, 3.22]

3.31**sample**

one or more parts taken from a specimen

[SOURCE: ISO 15189:2012, 3.24, modified — Example has been removed.]

3.32**stability**

ability of a sample material, when stored under specified conditions, to maintain a stated property value within specified limits for a specified period of time

[SOURCE: ISO Guide 30:2015, 2.1.15, modified — The words “reference material” were replaced by “sample material”.]

iTeh STANDARD PREVIEW
(standards.iteh.ai)

kSIST-TS FprCEN/TS 17688-2:2021

Leontides occurring in a double-stranded or single-stranded form

0442a7d9d616/ksist-ts-fprcen-ts-17688-2-2021